



**One Touch Select Plus Flex and Ultra Plus Flex User Manual - FCC  
Wording Memorandum**

Document ID: 3130961

Rev: 1

The purpose of this memo is to confirm the exact wording which will be added to the US manual for the One Touch Select Plus Flex and Ultra Plus Flex Blood Glucose Meters.

Item	Text
User Manual Statements to be added before product launch	<p>Changes or modifications not expressly approved by the party responsible for compliance could void the user's authority to operate the equipment.</p> <p>NOTE: This equipment has been tested and found to comply with the limits for a Class B digital device, pursuant to part 15 of the FCC Rules. These limits are designed to provide reasonable protection against harmful interference in a residential installation. This equipment generates, uses and can radiate radio frequency energy and, if not installed and used in accordance with the instructions, may cause harmful interference to radio communications. However, there is no guarantee that interference will not occur in a particular installation. If this equipment does cause harmful interference to radio or television reception, which can be determined by turning the equipment off and on, the user is encouraged to try to correct the interference by one or more of the following measures:</p> <ul style="list-style-type: none"><li>—Reorient or relocate the receiving antenna.</li><li>—Increase the separation between the equipment and receiver.</li><li>—Connect the equipment into an outlet on a circuit different from that to which the receiver is connected.</li><li>—Consult the dealer or an experienced radio/ TV technician for help.</li></ul> <p>This device complies with part 15 of the FCC Rules. Operation is subject to the following two conditions: (1) This device may not cause harmful interference, and (2) this device must accept any interference received, including interference that may cause undesired operation</p>

Sign:

Date:

27 APR 2015

Title:

GPL

Sign:

Date:

27 Apr 2015

Title:

RA Specialist