

RF System Check Test Report for CD1880:SN 1159

Hardware Setup

Probe | Calibration DateEF3DV3 - SN4041 | March 11, 2024

DAE | Calibration DateDAE4 Sn1798 | May 22, 2024

Software Version1.2.6.2199

Communication Systems

Band NameCD1880

Channel | Frequency [MHz]50 | 1880.0

Communication Systems' NameCW

Grid Settings

Extent X [mm]20.0Step X [mm]5.0

Extent Y [mm]90.0Step Y [mm]5.0

Distance [mm]15.0

Results

Emax [V/m]85.0

Drift [dB]0.01

