



Declaration of Product Identity

To whom it may concern:

We, LifeScan Europe GmbH,
Gubelstrasse 34
6300 Zug, Switzerland

Declare that our products:

Product name: Blood glucose meter with BLE 5.0

Model number: OneTouch Ultra Plus Flex

And

Product name: Blood glucose meter with BLE 5.0

Model number: OneTouch Verio Flex

have the same technical construction including software design, RF module, circuit design and PCB layout. The mentioned difference(s) between the models will not affect its electromagnetic emissions, radio-frequency output signal characteristics and output power level.

The differences between the models are the colour of the plastic housing (white for Verio Flex and blue for Ultra Plus Flex), keypad colour to match the plastic housings and the strip port connector contact positions. The Ultra Plus Flex will be the model on sale in Hong Kong.

Yours Faithfully,

A handwritten signature in blue ink, appearing to read "Oyinkan Donaldson", written in a fluid, cursive style.

Name: Oyinkan Donaldson

Title: VP Regulatory Affairs

Tel: +44 1463 382322

Email address: odonalds@lifescan.com

Date of Issue: 21 Dec 2022