


REVISION					
REV	DESCRIPTION	ECO NUMBER	DATE	DRFT	APRV
A	INITIAL PRODUCTION RELEASE	20-00686	9/4/2020	TY	KV
B	CORRECT INDUCTANCE TYPO	21-00751	1/12/2021	NV	KL

- NOTES:
1. THIS DRAWING IS INTENDED AS A MEANS OF COMMUNICATING LOCATIONS, SIZES, AND TOLERANCES OF MECHANICALLY RELEVANT FEATURES AND COMPONENTS. SEE LAYOUT FILES FOR COMPLETE MANUFACTURING INFORMATION.
 2. SEE PROVIDED DXF FILE FOR COMPLETE 2D GEOMETRY DEFINITION.
 3. PCB SHALL BE UL 94V-0 COMPLIANT AND MARKED AS SUCH.
 4. COIL INDUCTANCE, AS MEASURED BETWEEN TP1 AND TP2 WITH A CALIBRATED LCR METER, SHALL BE 247.4 uH ±2% AT 100 kHz.
 5. CTQ ARE CRITICAL INSPECTION DIMENSIONS (1).

IMS P/N	REV	DESCRIPTION
200-331-001	A	PCBA, PROGRAMMER CABLE HEAD

PROPERTY AND CONFIDENTIAL
The information contained in this drawing is the sole property of Inspire. Any reproduction in part or as a whole without the express written permission of Inspire is strictly prohibited.



Inspire Medical Systems, Inc.
Golden Valley, MN USA

TITLE: PCBA, PROGRAMMER CABLE HEAD

X.X ±.1
X.XX ±.01
X.XXX ±.005
ANGLES ±2.0°
All dimensions in inches unless otherwise noted

DOCUMENT CREATED BY:
T. YOCH

DATE
12/2/2019

APPROVED
K. LARSON

DATE
12/2/2019

SIZE
B

DWG NO.
200-331-000

REV.
B

SHEET
1 OF 1