

AIRSENSE 11 - SIMILARITIES AND DIFFERENCES for PACIFIC project

Title: AIRSENSE 11 - SIMILARITIES AND DIFFERENCES

Project/Product PACIFIC

ResMed Confidential Proprietary Information, not to be reproduced or made available to third parties without prior consent from ResMed and not to be used in any unauthorised way. © ResMed Ltd 2021

1. Purpose

This document discusses the similarities and differences between the AirSense 11 variants applicable to the following device models:

Model Name / Product Code	Marketing Name	Region of Sale
39485	AirSense 11 AutoSet	USA
39486	AirSense 11 CPAP	USA
39487	AirSense 11 Elite	USA
39488	AirSense 11 AutoSet	Canada
39489	AirSense 11 CPAP	Canada
39490	AirSense 11 Elite	Canada

2. Equivalence

All the device models mentioned in <u>Section 1</u> have the following <u>same</u> components/parts:

- Hardware
 - o Cellular
 - o Bluetooth
 - o Flow Generator
- Software
 - o Cellular
 - o Bluetooth
 - Flow Generator
 - Device Colour (paint)

3. Differences

Different models are derived for <u>marketing purposes</u>. The difference between Elite, CPAP and AutoSet models are the therapy modes and customer/region centric features supported by the device. These settings are configured at manufacturing via a configuration file and is independent of the cellular, Bluetooth and FG software.

In summary, the differences between the device models mentioned in Section 1 are:

- Therapy modes
- Data settings
- UI language
- Name plaque
- Data connection parameters.