

# PCTEST Engineering Laboratory, Inc.

6660-B Dobbin Road • Columbia, MD 21045 • U.S.A.

TEL (410) 290-6652 • FAX (410) 290-6654

<http://www.pctestlab.com>

## CERTIFICATE OF COMPLIANCE

### FCC Part 95 Certification

Medtronic, Inc.  
7000 Central Ave.  
Minneapolis, MN 55432-3576  
U.S.A.

Attention: Mr. Len Twetan,  
Senior Product Development Manager

Dates of Tests: Jan 26-27, 2006  
Test Report S/N: 0601090010  
Test Site: PCTEST Lab, MD

FCC ID

LF5MICSW

APPLICANT

Medtronic, Inc.

Classification:	Licensed Non-Broadcast Transmitter
FCC Rule Part(s):	Part 95
Procedure(s):	FCC 95, RSS 243, EN 301 839-1
EUT Type:	Medtronic Cure Link Monitor
Trade Name(s):	Medtronic
Model(s):	2490C, 2490W, 2490R
Tx/Rx Frequency Range:	402.15 – 404.85 MHz
Max. RF Output Power:	23 µW
Frequency Tolerance:	16 ppm
Emission Designator:	135KF1D
Channel Capacity:	10

This equipment has been shown to be capable of compliance with the applicable technical standards as indicated in the measurement report and was tested in accordance with the measurement procedures specified in §2.947.

I attest to the accuracy of data. All measurements reported herein were performed by me or were made under my supervision and are correct to the best of my knowledge and belief. I assume full responsibility for the completeness of these measurements and vouch for the qualifications of all persons taking them.

*PCTEST certifies that no party to this application has been denied the FCC benefits pursuant to Section 5301 of the Anti-Drug Abuse Act of 1988, 21 U.S.C. 862.*



## TABLE OF CONTENTS

SCOPE	1
INTRODUCTION	2
INSERTS	3
DESCRIPTION OF TESTS	4
RADIATED MEASUREMENTS	5
PLOTS OF EMISSIONS	6
LIST OF TEST EQUIPMENT	7
SAMPLE CALCULATIONS	9
CONCLUSION	10
ATTACHMENT A: TEST PLOTS	11
ATTACHMENT B: TEST SETUP PHOTOGRAPHS	12

PCTEST™ PT. 95 REPORT	FCC Certification			Reviewed By: Quality Manager
Test Report S/N: 0601090010	Test Dates: JAN. 26-27, 2006	EUT Type: Medtronic Cure Link Monitor	FCC ID: LF5MICSW	Page i of i



# MEASUREMENT REPORT

## 1.1 Scope

*Measurement and determination of electromagnetic emissions (EME) of radio frequency devices including intentional and/or unintentional radiators for compliance with the technical rules and regulations of the Federal Communications Commission.*

## General Information

**Applicant Name:** Medtronic, Inc.  
**Address:** 7000 Central Ave.  
Minneapolis, MN 55432-3576  
**Attention:** Mr. Len Twetan,  
Senior Product Development Manager

- **FCC ID:** LF5MICSW
- **Model(s):** 2490C, 2490W, 2490R
- **Quantity:** Quantity production is planned
- **Tx/Rx Freq. Range:** 402.15 – 404.85 MHz
- **Equipment Class:** Licensed Non-Broadcast Transmitter
- **Equipment Type:** Medtronic Cure Link Monitor
- **Emission Designator:** 135KF1D
- **Modulation:** F1D
- **Frequency Tolerance:** 16 ppm
- **Max. Power:** 23  $\mu$ W
- **FCC Rule Part(s):** §§§§ 95
- **Dates of Tests:** Jan. 26 – 27, 2006
- **Place of Tests:** PCTEST Lab, Columbia, MD U.S.A.
- **Test Report S/N:** 0601090010

PCTEST™ PT. 95 REPORT		FCC Certification		Reviewed By: Quality Manager
Test Report S/N: 0601090010	Test Dates: JAN. 26-27, 2006	EUT Type: Medtronic Cure Link Monitor	FCC ID: LF5MICSW	Page 1 of 14

## 2.1 INTRODUCTION



Figure 1. Map of the Greater Baltimore and Metropolitan Washington, D.C. area.

### Measurement Procedure

These measurement tests were conducted at **PCTEST Engineering Laboratory, Inc.** facility in New Concept Business Park, Guilford Industrial Park, Columbia, Maryland. The site address is 6660-B Dobbin Road, Columbia, MD 21045. The test site is one of the highest points in the Columbia area with an elevation of 390 feet above mean sea level. The site coordinates are 39° 11'15" N latitude and 76° 49'38" W longitude. The facility is 1.5 miles North of the FCC laboratory, and the ambient signal and ambient signal strength are approximately equal to those of the FCC laboratory. There are no FM or TV transmitters within 15 miles of the site. The detailed description of the measurement facility was found to be in compliance with the requirements of § 2.948 according to ANSI C63.4 on October 19, 1992.

The radiated and spurious measurements were made outdoors at a 3-meter test range (see Figure2). The equipment under testing was placed on a wooden turntable, 3-meters from the receive antenna. The receive antenna height and turntable rotations was adjusted for the highest reading on the receive spectrum analyzer. A half-wave dipole was substituted in place of the EUT. This dipole antenna was driven by a signal generator and the level of the signal generator was adjusted to obtain the same receive spectrum analyzer reading. This level was recorded.

For readings above 1 GHZ, the above procedure would be repeated using horn antennas and the difference between the gain of the horn and an isotropic antenna are taken into consideration.

PCTEST™ PT. 95 REPORT		FCC Certification		Reviewed By: Quality Manager
Test Report S/N: 0601090010	Test Dates: JAN. 26-27, 2006	EUT Type: Medtronic Cure Link Monitor	FCC ID: LF5MICSW	Page 2 of 14

## 3.1 INSERTS

---

### Block Diagram(s) & Circuit Diagram(s)

The block diagram is shown in Attachment I, and the circuit diagram is shown in Attachment J.

N/A

### Operating Instructions

The instruction manual is shown in Attachment K.

N/A

PCTEST™ PT. 95 REPORT	FCC Certification			Reviewed By: Quality Manager
Test Report S/N: 0601090010	Test Dates: JAN. 26-27, 2006	EUT Type: Medtronic Cure Link Monitor	FCC ID: LF5MICSW	Page 3 of 14

## 4.1 DESCRIPTION OF TESTS

---

### 4.2 Purpose

The tests conducted for this investigation was to determine the effect of minor adjustments in the matching parameters for the antenna. In this case the FCC requires measurements to determine the radiated fundamental and spurious levels. The results are used to determine if it's appropriate to file a Class II change request with the FCC or treat the changes as Class I not requiring a filing with the FCC.

### 4.3 Radiated Spurious and Harmonic Emissions

Radiated spurious and harmonic emissions above 1 GHz are measured on a 3-meter outdoor site. The EUT is placed on the turntable connected to a dummy load in normal operation using the intended power source. A receiving antenna located 3 meters from the turntable receives any signal radiated from the transmitter and its operating accessories. The antenna is varied from 1 to 4 meters and the polarization is varied (horizontal and vertical) to determine the worst-case emission level. To obtain actual radiated signal strength, a signal generator is adjusted in output until a reading identical to that obtained with the actual transmitter is obtained at the receiver. Signal level is read directly from the generator and recorded on the attached table.

PCTEST™ PT. 95 REPORT	<i>FCC Certification</i>			Reviewed By: Quality Manager
Test Report S/N: 0601090010	Test Dates: JAN. 26-27, 2006	EUT Type: Medtronic Cure Link Monitor	FCC ID: LF5MICSW	Page 4 of 14

## 5.1 TEST DATA

### 5.2 Radiated Measurements

#### Field Strength of SPURIOUS Radiation -Low

Frequency (MHz)	Level (dBm)	AFC L (dB)	POL (H/V)	F/S (dB $\mu$ V/m)	F/S ( $\mu$ V/m)	Power (W)	Margin (dB)
402.15	-50.100 (ref)		H			23 uW	
804.3	-99.200	31.45	H	39.25	91.72759	1.54E-09	6.75
1206.5	-124.000	32.1	H	15.1	5.688529	5.92E-12	38.9
1608.6	-124.600	34.5	H	16.9	6.99842	8.96E-12	37.1
2010.8	-135.000	36	H	8	2.511886	1.15E-12	46

#### Field Strength of SPURIOUS Radiation -Medium

Frequency (MHz)	Level (dBm)	AFC L (dB)	POL (H/V)	F/S (dB $\mu$ V/m)	F/S ( $\mu$ V/m)	Power (W)	Margin (dB)
403.35	-50.000 (ref)		H			22 uW	
806.7	-99.900	31.5	H	38.6	85.1138	1.33E-09	7.4
1210.1	-124.200	32.2	H	15	5.623413	5.78E-12	39
1613.4	-122.400	34.6	H	19.2	9.120108	1.52E-11	34.8
2016.8	-135.000	36.1	H	8.1	2.540973	1.18E-12	45.9

#### Field Strength of SPURIOUS Radiation -High

Frequency (MHz)	Level (dBm)	AFCL (dB)	POL (H/V)	F/S (dB $\mu$ V/m)	F/S ( $\mu$ V/m)	Power (W)	Margin (dB)
404.85	-50.200(ref)		H			23 uW	
809.70	-98.600	31.6	H	40	100	1.83E-09	-6
1214.60	-123.100	32.3	H	16.2	6.456542	7.63E-12	-37.8
1619.40	-123.600	34.7	H	18.1	8.035261	1.18E-11	-35.9
2024.30	-135.000	36.2	H	8.2	2.570396	1.21E-12	-45.8

PCTEST™ PT. 95 REPORT		FCC Certification			Reviewed By: Quality Manager
Test Report S/N: 0601090010	Test Dates: JAN. 26-27, 2006	EUT Type: Medtronic Cure Link Monitor	FCC ID: LF5MICSW	Page 5 of 14	

## 6.1 POWER LINE EMISSIONS

---

SEE ATTACHMENT A

PCTEST™ PT. 95 REPORT	FCC Certification			Reviewed By: Quality Manager
Test Report S/N: 0601090010	Test Dates: JAN. 26-27, 2006	EUT Type: Medtronic Cure Link Monitor	FCC ID: LF5MICSW	Page 6 of 14



## 7.1 TEST EQUIPMENT

8.2 Type	Model	Cal. Due Date	S/N
Microwave Spectrum Analyzer 3638A08713	HP 8566B (100Hz-22GHz)		08/15/06
Microwave Spectrum Analyzer 2542A11898	HP 8566B (100Hz-22GHz)		04/17/06
Spectrum Analyzer/Tracking Gen. 3144A02458	HP 8591A (100Hz-1.8GHz)		08/10/06
Signal Generator*	HP 8640B (500Hz-1GHz)	06/03/06	2232A19558
Signal Generator*	HP 8640B (500Hz-1GHz)	06/03/06	1851A09816
Signal Generator*	Rohde & Schwarz (0.1-1000MHz)	09/11/06	894215/012
Ailtech/Eaton Receiver 0792-03271	NM 37/57A-SL (30-1000MHz)		04/12/06
Ailtech/Eaton Receiver 0805-03334	NM 37/57A (30-1000MHz)		03/11/06
Ailtech/Eaton Receiver 0608-03241	NM 17/27A (0.1-32MHz)		09/17/06
Quasi-Peak Adapter Ailtech/Eaton Adapter 0194-04082	HP 85650A CCA-7 CISPR/ANSI QP Adapter	08/15/06 03/11/06	2043A00301
RG58 Coax Test Cable	No. 167		n/a
Harmonic/Flicker Test System 3531A00115	HP 6841A (IEC 555-2/3)		
Broadband Amplifier (2) 1937A03348	HP 8447D		1145A00470,
Broadband Amplifier Transient Limiter 2820A00300	HP 8447F HP 11947A (9kHz-200MHz)		2443A03784
Horn Antenna 9704-5182	EMCO Model 3115 (1-18GHz)		
Horn Antenna 9205-3874	EMCO Model 3115 (1-18GHz)		
Horn Antenna 9203-2178	EMCO Model 3116 (18-40GHz)		
Biconical Antenna (4) Design 1295, 1332, 0355	Eaton 94455/Eaton 94455-1/Singer		94455-1/Compliance
Log-Spiral Antenna (3) 1103, 1104	Ailtech/Eaton 93490-1		0608,
Roberts Dipoles	Compliance Design (1 set)		
Ailtech Dipoles 33448-111	DM-105A (1 set)		
EMCO LISN 3816/2			1079
EMCO LISN 3816/2			1077
EMCO LISN 3725/2			2009
Microwave Preamplifier 40dB Gain 3123A00181	HP 83017A (0.5-26.5GHz)		
Microwave Cables Ailtech/Eaton Receiver 0792-03271	MicroCoax (1.0-26.5GHz) NM37/57A-SL		
Spectrum Analyzer	HP 8594A		3051A00187
Spectrum Analyzer (2) 3108A02053	HP 8591A		3034A01395,
Modulation Analyzer	HP 8901A		2432A03467
NTSC Pattern Generator 0377433	Leader 408		

PCTEST™ PT. 95 REPORT	FCC Certification			Reviewed By: Quality Manager
Test Report S/N: 0601090010	Test Dates: JAN. 26-27, 2006	EUT Type: Medtronic Cure Link Monitor	FCC ID: LF5MICSW	Page 7 of 14

Noise Figure Meter	HP 8970B	3106A02189
Noise Figure Meter	Ailtech 7510	
TE31700		
Noise Generator	Ailtech 7010	1473
Microwave Survey Meter	Holaday Model 1501 (2.450GHz)	80931
Digital Thermometer	Extech Instruments 421305	
426966		
Attenuator	HP 8495A (0-70dB) DC-4GHz	
Bi-Directional Coax Coupler	Narda 3020A (50-1000MHz)	
Shielded Screen Room	RF Lindgren Model 26-2/2-0	
6710 (PCT270)		
Shielded Semi-Anechoic Chamber	Ray Proof Model S81	R2437
(PCT278)		
Enviromental Chamber	Associated Systems Model 1025 (Temperature/Humidity)	
PCT285		

\* Calibration traceable to the National Institute of Standards and Technology (NIST).

PCTEST™ PT. 95 REPORT		FCC Certification		Reviewed By: Quality Manager
Test Report S/N: 0601090010	Test Dates: JAN. 26-27, 2006	EUT Type: Medtronic Cure Link Monitor	FCC ID: LF5MICSW	Page 8 of 14

## 8.1 SAMPLE CALCULATIONS

---

The power for channel 05 (403.35 MHz) was measured at a 3 meter distance on our OATS. The turn table and antenna mast were adjusted to obtain the highest reading on a receiver spectrum analyzer with RBW and VBW set at 3MHz each. A dipole antenna driven by a signal generator was substituted in place of the EUT and adjusted to match the -50.1 dBm receiver spectrum analyzer reading. (The dipole antenna used during the substitution has a gain of 2.14 dBi at 403.35 MHz.) The power at the antenna terminals of the substituted dipole was -18.87 dBm. This value was corrected by adding the 2.14 dBi gain of the substituted dipole to yield - 16.73 dBm for EIRP.

PCTEST™ PT. 95 REPORT		FCC Certification		Reviewed By: Quality Manager
Test Report S/N: 0601090010	Test Dates: JAN. 26-27, 2006	EUT Type: Medtronic Cure Link Monitor	FCC ID: LF5MICSW	Page 9 of 14


## 9.1 CONCLUSION

---


The data collected shows that the modified **Medtronic Cure Link Monitor FCC ID: LF5MICSW** continues to comply with all the requirements of Parts 2 and 95 of the FCC rules, RSS 243, and EN 301 829-2.

<b>PCTEST™</b> PT. 95 REPORT	<i>FCC Certification</i>			<b>Reviewed By:</b> Quality Manager
<b>Test Report S/N:</b> 0601090010	<b>Test Dates:</b> JAN. 26-27, 2006	<b>EUT Type:</b> Medtronic Cure Link Monitor	<b>FCC ID:</b> LF5MICSW	Page 10 of 14

## Attachment A



No.	Freq. [MHz]	Quasi-Pk [dBuV]	Limit [dBuV]	Margin [dB]	Average [dBuV]	Limit [dBuV]	Margin [dB]
1	.150	39.25	65.98	-26.73	27.95	55.66	-27.71
2	.474	32.99	56.44	-23.45	22.91	46.27	-23.36
3	.461	32.30	56.67	-24.29	22.87	46.48	-23.61
4	.474	32.22	56.44	-24.22	22.86	46.57	-23.71
5	.467	32.10	56.56	-24.38	22.93	46.11	-23.18
6	.447	31.33	56.94	-25.61	22.00	46.76	-23.00
7	.439	31.15	57.00	-25.83	22.07	46.89	-24.82
8	.278	37.49	60.87	-23.38	23.36	50.50	-27.22
9	.284	35.75	60.70	-24.95	23.27	50.11	-26.84
10	.331	33.16	59.43	-26.27	23.15	49.42	-26.27



No.	Freq. [MHz]	Quasi-Pk [dBuV]	Limit [dBuV]	Margin [dB]	Average [dBuV]	Limit [dBuV]	Margin [dB]
1	.150	42.33	65.98	-23.65	28.48	56.00	-27.52
2	.150	40.13	65.98	-25.75	28.35	56.00	-27.65
3	.458	32.61	56.87	-24.26	22.94	46.41	-23.47
4	.462	33.40	56.66	-23.26	22.97	46.49	-23.52
5	.442	31.47	57.03	-25.56	22.96	47.87	-24.11
6	.421	31.15	57.43	-26.28	22.95	47.24	-24.39
7	.402	33.15	56.31	-23.16	22.87	46.88	-23.22
8	.315	33.10	59.83	-26.73	23.11	49.77	-26.66
9	.475	32.98	56.43	-23.45	22.94	46.28	-23.34
10	.210	39.46	63.21	-23.75	24.17	52.84	-28.67

PCTEST™ PT. 95 REPORT	FCC Certification			Reviewed By: Quality Manager
Test Report S/N: 0601090010	Test Dates: JAN. 26-27, 2006	EUT Type: Medtronic Cure Link Monitor	FCC ID: LFSMICSW	Page 11 of 14

## Attachement B



PCTEST™ PT. 95 REPORT	FCC Certification			Reviewed By: Quality Manager
Test Report S/N: 0601090010	Test Dates: JAN. 26-27, 2006	EUT Type: Medtronic Cure Link Monitor	FCC ID: LF5MICSW	Page 12 of 14