

1/15/2025

Federal Communications Commission
7435 Oakland Mills Road
Columbia, MD 21046

To whom it may concern:

This letter is to inform you that Medtronic Minimed Inc. ("Medtronic") requests dismissal for FCC ID OH2BLEGST1 for the following reason. We have discovered that our production line has had trouble during surveillance testing, where test devices were showing higher tolerance than expected. Due to these issues, we withheld public release of the product pending root cause analysis and no units with the existing FCC ID have been sold. The production line has since made modifications to fix the issues seen during the initial registration. In order to verify the fixes, we plan to characterize several batches of units to confirm consistency prior to obtaining a recertification.

Sincerely,

A handwritten signature in black ink, appearing to read "J. Tabalujan", is positioned below the word "Sincerely,".

Jonathan Tabalujan
Pr. Electrical Engineer – RF
Medtronic MiniMed, Inc.