



In Collaboration with
s p e a g
CALIBRATION LABORATORY



中国认可
国际互认
校准
CALIBRATION
CNAS L0570

Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62304633-2512 Fax: +86-10-62304633-2504
E-mail: cttl@chinattl.com Http://www.chinattl.cn

Client

ICAS

Certificate No: Z20-60372

CALIBRATION CERTIFICATE

Object

EX3DV4 - SN : 7475

Calibration Procedure(s)

FF-Z11-004-02

Calibration Procedures for Dosimetric E-field Probes

Calibration date:

October 29, 2020

This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements(SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature(22±3)°C and humidity<70%.

Calibration Equipment used (M&TE critical for calibration)

Primary Standards	ID #	Cal Date(Calibrated by, Certificate No.)	Scheduled Calibration
Power Meter NRP2	101919	16-Jun-20(CTTL, No.J20X04344)	Jun-21
Power sensor NRP-Z91	101547	16-Jun-20(CTTL, No.J20X04344)	Jun-21
Power sensor NRP-Z91	101548	16-Jun-20(CTTL, No.J20X04344)	Jun-21
Reference 10dBAttenuator	18N50W-10dB	10-Feb-20(CTTL, No.J20X00525)	Feb-22
Reference 20dBAttenuator	18N50W-20dB	10-Feb-20(CTTL, No.J20X00526)	Feb-22
Reference Probe EX3DV4	SN 7307	29-May-20(SPEAG, No.EX3-7307_May20)	May-21
DAE4	SN 1556	4-Feb-20(SPEAG, No.DAE4-1556_Feb20)	Feb-21
Secondary Standards	ID #	Cal Date(Calibrated by, Certificate No.)	Scheduled Calibration
SignalGenerator MG3700A	6201052605	23-Jun-20(CTTL, No.J20X04343)	Jun-21
Network Analyzer E5071C	MY46110673	10-Feb-20(CTTL, No.J20X00515)	Feb-21

	Name	Function	Signature
Calibrated by:	Yu Zongying	SAR Test Engineer	
Reviewed by:	Lin Hao	SAR Test Engineer	
Approved by:	Qi Dianyuan	SAR Project Leader	

Issued: October 31, 2020

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.



In Collaboration with

s p e a g
CALIBRATION LABORATORY

Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62304633-2512 Fax: +86-10-62304633-2504
E-mail: ctl@chinattl.com [Http://www.chinattl.cn](http://www.chinattl.cn)

Glossary:

TSL	tissue simulating liquid
NORM _{x,y,z}	sensitivity in free space
ConvF	sensitivity in TSL / NORM _{x,y,z}
DCP	diode compression point
CF	crest factor (1/duty_cycle) of the RF signal
A,B,C,D	modulation dependent linearization parameters
Polarization Φ	Φ rotation around probe axis
Polarization θ	θ rotation around an axis that is in the plane normal to probe axis (at measurement center), i $\theta=0$ is normal to probe axis
Connector Angle	information used in DASY system to align probe sensor X to the robot coordinate system

Calibration is Performed According to the Following Standards:

- IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013
- IEC 62209-1, "Measurement procedure for the assessment of Specific Absorption Rate (SAR) from hand-held and body-mounted devices used next to the ear (frequency range of 300 MHz to 6 GHz)", July 2016
- IEC 62209-2, "Procedure to determine the Specific Absorption Rate (SAR) for wireless communication devices used in close proximity to the human body (frequency range of 30 MHz to 6 GHz)", March 2010
- KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

Methods Applied and Interpretation of Parameters:

- NORM_{x,y,z}**: Assessed for E-field polarization $\theta=0$ ($f \leq 900\text{MHz}$ in TEM-cell; $f > 1800\text{MHz}$: waveguide). NORM_{x,y,z} are only intermediate values, i.e., the uncertainties of NORM_{x,y,z} does not effect the E^2 -field uncertainty inside TSL (see below ConvF).
- NORM(f)_{x,y,z}** = NORM_{x,y,z} * frequency_response (see Frequency Response Chart). This linearization is implemented in DASY4 software versions later than 4.2. The uncertainty of the frequency response is included in the stated uncertainty of ConvF.
- DCP_{x,y,z}**: DCP are numerical linearization parameters assessed based on the data of power sweep (no uncertainty required). DCP does not depend on frequency nor media.
- PAR**: PAR is the Peak to Average Ratio that is not calibrated but determined based on the signal characteristics.
- A_{x,y,z}; B_{x,y,z}; C_{x,y,z}; VR_{x,y,z}**: A,B,C are numerical linearization parameters assessed based on the data of power sweep for specific modulation signal. The parameters do not depend on frequency nor media. VR is the maximum calibration range expressed in RMS voltage across the diode.
- ConvF and Boundary Effect Parameters**: Assessed in flat phantom using E-field (or Temperature Transfer Standard for $f \leq 800\text{MHz}$) and inside waveguide using analytical field distributions based on power measurements for $f > 800\text{MHz}$. The same setups are used for assessment of the parameters applied for boundary compensation (alpha, depth) of which typical uncertainty valued are given. These parameters are used in DASY4 software to improve probe accuracy close to the boundary. The sensitivity in TSL corresponds to NORM_{x,y,z} * ConvF whereby the uncertainty corresponds to that given for ConvF. A frequency dependent ConvF is used in DASY version 4.4 and higher which allows extending the validity from $\pm 50\text{MHz}$ to $\pm 100\text{MHz}$.
- Spherical isotropy (3D deviation from isotropy)**: in a field of low gradients realized using a flat phantom exposed by a patch antenna.
- Sensor Offset**: The sensor offset corresponds to the offset of virtual measurement center from the probe tip (on probe axis). No tolerance required.
- Connector Angle**: The angle is assessed using the information gained by determining the NORM_x (no uncertainty required).



DASY/EASY – Parameters of Probe: EX3DV4 – SN:7475

Basic Calibration Parameters

	Sensor X	Sensor Y	Sensor Z	Unc (k=2)
Norm($\mu V/(V/m)^2$) ^A	0.53	0.51	0.64	±10.0%
DCP(mV) ^B	100.3	101.6	100.6	

Calibration Results for Modulation Response

UID	Communication System Name		A dB	B dB $\sqrt{\mu V}$	C	D dB	VR mV	Max Dev.	Max Unc ^E (k=2)
0	CW	X	0.0	0.0	1.0	0.00	179.6	±2.0%	±4.7%
		Y	0.0	0.0	1.0		174.9		
		Z	0.0	0.0	1.0		196.7		
10352-AAA	Pulse Waveform (200Hz, 10%)	X	2.38	64.73	9.72	10.00	60	±3.2%	±9.6%
		Y	2.47	65.22	9.93		60		
		Z	3.00	67.28	10.85		60		
10353-AAA	Pulse Waveform (200Hz, 20%)	X	2.23	65.79	9.32	6.99	80	±2.1%	±9.6%
		Y	2.59	67.42	9.99		80		
		Z	2.6	67.48	10.14		80		
10354-AAA	Pulse Waveform (200Hz, 40%)	X	1.38	64.53	7.94	3.98	95	±1.4%	±9.6%
		Y	2.55	69.38	9.82		95		
		Z	1.85	66.92	9.03		95		
10355-AAA	Pulse Waveform (200Hz, 60%)	X	0.61	61.69	5.92	2.22	120	±1.4%	±9.6%
		Y	1.14	66.12	7.83		120		
		Z	0.7	62.8	6.55		120		
10387-AAA	QPSK Waveform, 1 MHz	X	1.53	64.78	13.66	1.00	150	±2.5%	±9.6%
		Y	1.66	67.59	15.34		150		
		Z	1.56	65.31	14.14		150		
10388-AAA	QPSK Waveform, 10 MHz	X	2.1	66.89	14.67	0.00	150	±1.4%	±9.6%
		Y	2.27	69.16	16.26		150		
		Z	2.15	67.45	15.19		150		
10396-AAA	64-QAM Waveform, 100 kHz	X	2.88	70.52	19.16	3.01	150	±1.0%	±9.6%
		Y	2.68	70.37	19.62		150		
		Z	2.79	69.94	19.26		150		
10414-AAA	WLAN CCDF, 64-QAM, 40MHz	X	4.84	65.57	15.35	0.00	150	±3.6%	±9.6%
		Y	4.78	65.89	15.69		150		
		Z	4.93	65.94	15.69		150		

Note: For details on UID parameters see Appendix

The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor $k=2$, which for a normal distribution Corresponds to a coverage probability of approximately 95%.

^A The uncertainties of Norm X, Y, Z do not affect the E²-field uncertainty inside TSL (see Page 5 and Page 6).

^B Numerical linearization parameter: uncertainty not required.

^E Uncertainty is determined using the max. deviation from linear response applying rectangular distribution and is expressed for the square of the field value.



In Collaboration with

s p e a g

CALIBRATION LABORATORY

Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China

Tel: +86-10-62304633-2512

Fax: +86-10-62304633-2504

E-mail: ctl@chinattl.com

[Http://www.chinattl.cn](http://www.chinattl.cn)

DASY/EASY – Parameters of Probe: EX3DV4 – SN: 7475

Sensor Model Parameters

	C1 fF	C2 fF	α V ⁻¹	T1 ms.V ⁻²	T2 ms.V ⁻¹	T3 ms	T4 V ⁻²	T5 V ⁻¹	T6
X	44.85	338.96	36.06	17.73	0.00	5.03	0.52	0.32	1.02
Y	38.62	288.37	35.59	15.60	0.00	5.03	0.00	0.33	1.02
Z	44.41	339.88	36.98	18.00	0.00	5.04	0.00	0.41	1.02

Other Probe Parameters

Sensor Arrangement	Triangular
Connector Angle (°)	134.7
Mechanical Surface Detection Mode	enabled
Optical Surface Detection Mode	disable
Probe Overall Length	337mm
Probe Body Diameter	10mm
Tip Length	9mm
Tip Diameter	2.5mm
Probe Tip to Sensor X Calibration Point	1mm
Probe Tip to Sensor Y Calibration Point	1mm
Probe Tip to Sensor Z Calibration Point	1mm
Recommended Measurement Distance from Surface	1.4mm



In Collaboration with

s p e a g**CALIBRATION LABORATORY**

Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China

Tel: +86-10-62304633-2512

Fax: +86-10-62304633-2504

E-mail: cttl@chinattl.com

[Http://www.chinattl.cn](http://www.chinattl.cn)

DASY/EASY – Parameters of Probe: EX3DV4 – SN:7475

Calibration Parameter Determined in Head Tissue Simulating Media

f [MHz] ^C	Relative Permittivity ^F	Conductivity (S/m) ^F	ConvF X	ConvF Y	ConvF Z	Alpha ^G	Depth ^G (mm)	Unct. (k=2)
750	41.9	0.89	10.16	10.16	10.16	0.40	0.80	±12.1%
835	41.5	0.90	9.79	9.79	9.79	0.19	1.23	±12.1%
900	41.5	0.97	9.74	9.74	9.74	0.20	1.21	±12.1%
1450	40.5	1.20	8.68	8.68	8.68	0.17	1.01	±12.1%
1750	40.1	1.37	8.45	8.45	8.45	0.22	1.07	±12.1%
1900	40.0	1.40	8.07	8.07	8.07	0.23	1.11	±12.1%
2100	39.8	1.49	8.04	8.04	8.04	0.21	1.15	±12.1%
2300	39.5	1.67	7.90	7.90	7.90	0.61	0.71	±12.1%
2450	39.2	1.80	7.65	7.65	7.65	0.60	0.71	±12.1%
2600	39.0	1.96	7.45	7.45	7.45	0.66	0.70	±12.1%
3300	38.2	2.71	7.15	7.15	7.15	0.47	0.94	±13.3%
3500	37.9	2.91	6.84	6.84	6.84	0.47	0.97	±13.3%
3700	37.7	3.12	6.51	6.51	6.51	0.46	1.02	±13.3%
3900	37.5	3.32	6.57	6.57	6.57	0.40	1.15	±13.3%
4100	37.2	3.53	6.52	6.52	6.52	0.40	1.20	±13.3%
4200	37.1	3.63	6.40	6.40	6.40	0.35	1.33	±13.3%
4400	36.9	3.84	6.30	6.30	6.30	0.40	1.25	±13.3%
4600	36.7	4.04	6.18	6.18	6.18	0.45	1.25	±13.3%
4800	36.4	4.25	6.12	6.12	6.12	0.45	1.30	±13.3%
4950	36.3	4.40	5.80	5.80	5.80	0.50	1.25	±13.3%
5200	36.0	4.66	5.53	5.53	5.53	0.50	1.25	±13.3%
5300	35.9	4.76	5.29	5.29	5.29	0.45	1.30	±13.3%
5500	35.6	4.96	4.95	4.95	4.95	0.55	1.20	±13.3%
5600	35.5	5.07	4.74	4.74	4.74	0.55	1.31	±13.3%
5800	35.3	5.27	4.75	4.75	4.75	0.55	1.25	±13.3%

^C Frequency validity above 300 MHz of ±100MHz only applies for DASY v4.4 and higher (Page 2), else it is restricted to ±50MHz. The uncertainty is the RSS of ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below 300 MHz is ± 10, 25, 40, 50 and 70 MHz for ConvF assessments at 30, 64, 128, 150 and 220 MHz respectively. Above 5 GHz frequency validity can be extended to ± 110 MHz.

^F At frequency below 3 GHz, the validity of tissue parameters (ϵ and σ) can be relaxed to ±10% if liquid compensation formula is applied to measured SAR values. At frequencies above 3 GHz, the validity of tissue parameters (ϵ and σ) is restricted to ±5%. The uncertainty is the RSS of the ConvF uncertainty for indicated target tissue parameters.

^G Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than ± 1% for frequencies below 3 GHz and below ± 2% for the frequencies between 3-6 GHz at any distance larger than half the probe tip diameter from the boundary.

DASY/EASY – Parameters of Probe: EX3DV4 – SN:7475

Calibration Parameter Determined in Body Tissue Simulating Media

f [MHz] ^C	Relative Permittivity ^F	Conductivity (S/m) ^F	ConvF X	ConvF Y	ConvF Z	Alpha ^G	Depth ^G (mm)	Unct. (k=2)
750	55.5	0.96	10.17	10.17	10.17	0.40	0.82	±12.1%
835	55.2	0.97	9.83	9.83	9.83	0.20	1.32	±12.1%
900	55.0	1.05	9.81	9.81	9.81	0.26	1.12	±12.1%
1450	54.0	1.30	8.70	8.70	8.70	0.18	1.06	±12.1%
1750	53.4	1.49	8.22	8.22	8.22	0.27	1.11	±12.1%
1900	53.3	1.52	8.08	8.08	8.08	0.28	0.92	±12.1%
2100	53.2	1.62	7.89	7.89	7.89	0.24	1.25	±12.1%
2300	52.9	1.81	7.71	7.71	7.71	0.54	0.83	±12.1%
2450	52.7	1.95	7.59	7.59	7.59	0.63	0.75	±12.1%
2600	52.5	2.16	7.31	7.31	7.31	0.54	0.82	±12.1%
3300	51.6	3.08	6.64	6.64	6.64	0.43	1.14	±13.3%
3500	51.3	3.31	6.38	6.38	6.38	0.40	1.25	±13.3%
3700	51.0	3.55	6.32	6.32	6.32	0.40	1.25	±13.3%
3900	51.2	3.78	6.26	6.26	6.26	0.40	1.40	±13.3%
4100	50.5	4.01	6.20	6.20	6.20	0.40	1.22	±13.3%
4200	50.4	4.13	6.15	6.15	6.15	0.45	1.15	±13.3%
4400	50.1	4.37	6.11	6.11	6.11	0.45	1.23	±13.3%
4600	49.8	4.60	5.90	5.90	5.90	0.45	1.25	±13.3%
4800	49.6	4.83	5.60	5.60	5.60	0.45	1.65	±13.3%
4950	49.4	5.01	5.40	5.40	5.40	0.45	1.70	±13.3%
5200	49.0	5.30	5.00	5.00	5.00	0.50	1.55	±13.3%
5300	48.9	5.42	4.71	4.71	4.71	0.55	1.45	±13.3%
5500	48.6	5.65	4.36	4.36	4.36	0.55	1.50	±13.3%
5600	48.5	5.77	4.22	4.22	4.22	0.60	1.40	±13.3%
5800	48.2	6.00	4.25	4.25	4.25	0.60	1.45	±13.3%

^C Frequency validity above 300 MHz of ±100MHz only applies for DASY v4.4 and higher (Page 2), else it is restricted to ±50MHz. The uncertainty is the RSS of ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below 300 MHz is ± 10, 25, 40, 50 and 70 MHz for ConvF assessments at 30, 64, 128, 150 and 220 MHz respectively. Above 5 GHz frequency validity can be extended to ± 110 MHz.

^F At frequency below 3 GHz, the validity of tissue parameters (ϵ and σ) can be relaxed to ±10% if liquid compensation formula is applied to measured SAR values. At frequencies above 3 GHz, the validity of tissue parameters (ϵ and σ) is restricted to ±5%. The uncertainty is the RSS of the ConvF uncertainty for indicated target tissue parameters.

^G Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than ± 1% for frequencies below 3 GHz and below ± 2% for the frequencies between 3-6 GHz at any distance larger than half the probe tip diameter from the boundary.




In Collaboration with

s p e a g
CALIBRATION LABORATORY

Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62304633-2512 Fax: +86-10-62304633-2504
E-mail: cttl@chinattl.com [Http://www.chinattl.cn](http://www.chinattl.cn)

Frequency Response of E-Field (TEM-Cell: ifi110 EXX, Waveguide: R22)



Uncertainty of Frequency Response of E-field: $\pm 7.4\%$ ($k=2$)