

**Radiated Emissions Test Configuration****Radiated Emissions Environmental Conditions**

<b>Power Supply (to controller)</b>	120V AC, 60 Hz
<b>Temperature</b>	13.5°C
<b>Relative Humidity</b>	62%
<b>Barometric Pressure</b>	987mb

**Radiated Emissions Measurement Uncertainties**

<b>Frequency</b>	± 200kHz
<b>Amplitude</b>	± 4.6dB

The uncertainties stated are calculated in accordance with the requirements of UKAS with a confidence level of 95%.

**Test Equipment Used**

<b>Equipment Type</b>	<b>Model Number</b>
Loop Antenna	EMCO 6502
Spectrum Analyser	HP8563E