

Rhein Tech Laboratories  
360 Herndon Parkway  
Suite 1400  
Herndon, VA 20170  
<http://www.rheintech.com>

Client: Ohmart/VEGA  
Model: VEGAPULS 68  
FCC ID: MOIPULS68  
Standards: FCC 15.209/IC RSS-210  
Report Number: 2004236

## **APPENDIX C: DUTY CYCLE ATTESTATION**

Please see the following page.



Ohmart/VEGA Corporation  
4241 Allendorf Drive  
Cincinnati, Ohio 45209-9961  
USA

(800) FOR LEVEL Toll-Free in USA & Canada  
(513) 272. 0131 Telephone  
(513) 272. 0133 FAX

<http://www.ohmartvega.com> URL  
[info@ohmartvega.com](mailto:info@ohmartvega.com) E-mail

Desmond Fraser  
Rhein Tech Laboratories, Inc.  
360 Herndon Parkway  
Suite 1400  
Herndon, VA 20170

January 5, 2005

Dear Mr. Fraser:

Per your request the following values apply to the VEGAPULS 62 and the VEGAPULS 68.

VEGAPULS 62: Rise time = 200pS, Pulse width = 600pS,  
Pulse repetition rate = 3.6MHz (=280nS)

VEGAPULS 68: Rise time = 200pS, Pulse width = 1.2nS,  
Pulse repetition rate = 1.8MHz (=560nS)

The values are measured at the 50% point of the amplitude.

Sincerely,

A handwritten signature in cursive script, appearing to read "Ken Geiger".

Ken Geiger  
Certification Project Coordinator