



4339 Nancy Place, Suite 100
Saint Paul, MN 55126

Tel: 651 217-2176
www.vivaquant.com

September 19, 2018

To Whom it may concern,

We are seeking a Class II permissive change for application of the RX-1-SQVZ Cat M cellular modem used in the Model RX-1 Rhythm Express Remote Cardiac Monitor. The applicable rule parts are 47 CFR part 27. The current FCC ID is 2AQ8D-RX-1-SQVZ.

The RX-1 uses this modularly approved device as described in the application literature for the modem and has not modified the hardware or software of the approved module. The RX-1 uses an antenna that meets the requirements specified by the manufacturer for the modem in agreement with its approval.

The RX-1 limits the output power through duty cycle of the modem to allow for use within 20 cm of the torso on the body.

Best Regards,

A handwritten signature in black ink, appearing to read 'Brian P. Brockway', with a large, stylized loop at the end.

Brian Brockway
President and CEO