

FCC Test Report

Product Name	Automatic Upper Arm Blood Pressure Monitor
Model No.	HL858DM
FCC ID.	2ABTAHNL85DM1

Applicant	Health & Life Co., Ltd.
Address	9F, No. 186, Jian Yi Road, Zhonghe District, New Taipei City 23553, Taiwan

Date of Receipt	May 24, 2022
Issued Date	Jun. 27, 2022
Report No.	2250736R-RFUSBLEV01-A
Report Version	V1.0



The test results relate only to the samples tested.

The test results shown in the test report are traceable to the national/international standard through the calibration report of the equipment and evaluated measurement uncertainty herein.

This report must not be used to claim product endorsement by TAF or any agency of the government.

The test report shall not be reproduced without the written approval of DEKRA Testing and Certification Co., Ltd.

Measurement uncertainties evaluated for each testing system and associated connections are given here to provide the system information for reference. Compliance determinations do not take into account measurement uncertainties for each testing system, but are based on the results of the compliance measurement.

Test Report

Issued Date: Jun. 27, 2022

Report No.: 2250736R-RFUSBLEV01-A




Product Name	Automatic Upper Arm Blood Pressure Monitor
Applicant	Health & Life Co., Ltd.
Address	9F, No. 186, Jian Yi Road, Zhonghe District, New Taipei City 23553, Taiwan
Manufacturer	Health & Life Co., Ltd.
Model No.	HL858DM
FCC ID.	2ABTAHNL85DM1
EUT Rated Voltage	AC 100-240V, 50-60Hz
EUT Test Voltage	AC 120V/60Hz
Trade Name	Health & Life
Applicable Standard	FCC CFR Title 47 Part 15 Subpart C ANSI C63.4: 2014, ANSI C63.10: 2013
Test Result	Complied

Documented By :



(Senior Project Specialist / Joanne Lin)

Tested By :



(Senior Engineer / Bill Lin)

Tested By :



(Senior Engineer / Alan Chen)

TABLE OF CONTENTS

Description	Page
1. GENERAL INFORMATION	5
1.1. EUT Description.....	5
1.2. Tested System Details.....	7
1.3. Configuration of Tested System	7
1.4. EUT Exercise Software	7
1.5. Test Facility	8
1.6. List of Test Equipment.....	9
1.7. Uncertainty	10
2. Conducted Emission	11
2.1. Test Setup	11
2.2. Limits.....	11
2.3. Test Procedure	12
2.4. Test Result of Conducted Emission	12
3. Peak Power Output	15
3.1. Test Setup	15
3.2. Limit	15
3.3. Test Procedure	15
3.4. Test Result of Peak Power Output	16
4. Radiated Emission	16
4.1. Test Setup	17
4.2. Limits.....	18
4.3. Test Procedure	19
4.4. Test Result of Radiated Emission.....	20
5. RF Antenna Conducted Test	25
5.1. Test Setup	25
5.2. Limits.....	25
5.3. Test Procedure	25
5.4. Test Result of RF Antenna Conducted Test.....	26
6. Band Edge	26
6.1. Test Setup	27
6.2. Limit	28
6.3. Test Procedure	28
6.4. Test Result of Band Edge	30
7. 6dB Bandwidth	32
7.1. Test Setup	32
7.2. Limits.....	32
7.3. Test Procedure	32
7.4. Test Result of 6dB Bandwidth.....	33
8. Power Density	34
8.1. Test Setup	34
8.2. Limits.....	34
8.3. Test Procedure	34
8.4. Test Result of Power Density	35
9. Duty Cycle	36
9.1. Test Setup	36
9.2. Test Procedure	36
9.3. Test Result of Duty Cycle.....	37
10. EMI Reduction Method During Compliance Testing.....	38

Appendix 1: EUT Test Photographs

Appendix 2: Product Photos-Please refer to the file: 2250736R-Product Photos

Revision History

Report No.	Version	Description	Issued Date
2250736R-RFUSBLEV01-A	V1.0	Initial issue of report.	Jun. 27, 2022

1. GENERAL INFORMATION

1.1. EUT Description

Product Name	Automatic Upper Arm Blood Pressure Monitor
Trade Name	Health & Life
Model No.	HL858DM
FCC ID.	2ABTAHNL85DM1
Frequency Range	2402-2480MHz
Channel Number	Bluetooth V4.2: 40CH
Type of Modulation	Bluetooth V4.2: GFSK(1Mbps)
Channel Separation	Bluetooth V4.2: 2MHz
Data Rate	Bluetooth V4.2: 1Mbps
Antenna Type	PCB Antenna
Channel Control	Auto
Antenna Gain	Refer to the table “Antenna List”
USB Cable	Shielded, 0.6m
Power Adapter (1)	MFR: Zhongshan Baolijin Electronic Co. Ltd., M/N: BLJ06L060100P-V Input: 100-240V, 50-60Hz, 0.2A Output: 6V=1A, 6W Cable Out: Non-shielded, 1.5m
Power Adapter (2)	MFR: Zhongshan Baolijin Electronic Co. Ltd., M/N: BLJ06L060100P-U Input: 100-240V, 50-60Hz, 0.2A Output: 6V=1A Cable Out: Non-shielded, 1.5m
Power Adapter (3)	MFR: Zhongshan Baolijin Electronic co. ltd., M/N: BLJ06L050120U-R Input: 100-240V, 50-60Hz, 0.2A Output: 5V=1.2A, 6W
Power Adapter (4)	MFR: FranMar, M/N: FRM06-S05-UU Input: 100-240V, 50-60Hz, 0.3-1.5A Output: 5V=1.2A

Antenna List

No.	Manufacturer	Part No.	Antenna Type	Peak Gain
1	LaBest Technology Inc.	LB-BLE-006	PCB Antenna	2.23dBi for 2.4GHz

Note: The antenna of EUT is conforming to FCC 15.203.

Center Frequency of Each Channel:

Channel	Frequency	Channel	Frequency	Channel	Frequency	Channel	Frequency
Channel 00:	2402 MHz	Channel 01:	2404 MHz	Channel 02:	2406 MHz	Channel 03:	2408 MHz
Channel 04:	2410 MHz	Channel 05:	2412 MHz	Channel 06:	2414 MHz	Channel 07:	2416 MHz
Channel 08:	2418 MHz	Channel 09:	2420 MHz	Channel 10:	2422 MHz	Channel 11:	2424 MHz
Channel 12:	2426 MHz	Channel 13:	2428 MHz	Channel 14:	2430 MHz	Channel 15:	2432 MHz
Channel 16:	2434 MHz	Channel 17:	2436 MHz	Channel 18:	2438 MHz	Channel 19:	2440 MHz
Channel 20:	2442 MHz	Channel 21:	2444 MHz	Channel 22:	2446 MHz	Channel 23:	2448 MHz
Channel 24:	2450 MHz	Channel 25:	2452 MHz	Channel 26:	2454 MHz	Channel 27:	2456 MHz
Channel 28:	2458 MHz	Channel 29:	2460 MHz	Channel 30:	2462 MHz	Channel 31:	2464 MHz
Channel 32:	2466 MHz	Channel 33:	2468 MHz	Channel 34:	2470 MHz	Channel 35:	2472 MHz
Channel 36:	2474 MHz	Channel 37:	2476 MHz	Channel 38:	2478 MHz	Channel 39:	2480 MHz

Note:

1. The EUT is an Automatic Upper Arm Blood Pressure Monitor with built-in Bluetooth V4.2 transceiver.
2. DEKRA has evaluated each test mode. Only the worst case is shown in the report.
3. Regarding to the operation frequency, the lowest, middle and highest frequency are selected to perform the test.
4. These tests were conducted on a sample for the purpose of demonstrating compliance of transmitter with Part 15 Subpart C Paragraph 15.247 for spread spectrum devices.

Test Mode	Mode 1: Transmit - 1Mbps
-----------	--------------------------


1.2. Tested System Details

The types for all equipment, plus descriptions of all cables used in the tested system (including inserted cards) are:

Product	Manufacturer	Model No.	Serial No.	Power Cord
1 Power Adapter	zhongshan baolijin electronic co., ltd.	BLJ06L060100P-V	N/A	N/A

Signal Cable Type	Signal cable Description
A Power Cable	Non-shielded, 1.5m

1.3. Configuration of Tested System



1.4. EUT Exercise Software

- (1) Setup the EUT as shown in Section 1.3.
- (2) Press the button control.
- (3) Configure the test mode, the test channel, and the data rate.
- (4) Verify that the EUT works properly.

1.5. Test Facility

Ambient conditions in the laboratory:

Performed Item	Items	Required	Actual
Conducted Emission	Temperature (°C)	10~40 °C	24.9 °C
	Humidity (%RH)	10~90 %	59.6 %
Radiated Emission	Temperature (°C)	10~40 °C	25.0 °C
	Humidity (%RH)	10~90 %	63.4 %
Conductive	Temperature (°C)	10~40 °C	22.0 °C
	Humidity (%RH)	10~90 %	55.0 %

USA : **FCC Registration Number: TW0033**

Canada : **CAB Identifier Number: TW3023 / Company Number: 26930**

Site Description : Accredited by TAF
Accredited Number: 3023

Test Laboratory : DEKRA Testing and Certification Co., Ltd
Address : No. 5-22, Ruishukeng Linkou District, New Taipei City, 24451, Taiwan
Performed Location : No. 26, Huaya 1st Rd., Guishan Dist., Taoyuan City 333411, Taiwan,
R.O.C.
Phone number : +886-3-275-7255
Fax number : +866-3-327-8031
Email address : info.tw@dekra.com
Website : <http://www.dekra.com.tw>

1.6. List of Test Equipment

For Conduction measurements / HY-SR01

	Equipment	Manufacturer	Model No.	Serial No.	Cal. Date	Due. Date
X	EMI Test Receiver	R&S	ESR7	101601	2021.06.19	2022.06.18
X	Two-Line V-Network	R&S	ENV216	101306	2022.05.23	2023.05.22
X	Coaxial Cable	SUHNER	RG400_BNC	RF001	2022.05.24	2023.05.23

Note:

1. All equipments are calibrated every one year.
2. The test instruments marked with "X" are used to measure the final test results.
3. Test Software version : E3 210616 dekra V9.

For Conducted measurements / HY-SR02

	Equipment	Manufacturer	Model No.	Serial No.	Cal. Date	Due. Date
X	Spectrum Analyzer	R&S	FSV30	103466	2021/12/27	2022/12/26
X	Peak Power Analyzer	KEYSIGHT	8990B	MY51000410	2021.08.04	2022.08.03
X	Wideband Power Sensor	KEYSIGHT	N1923A	MY56080003	2021.08.03	2022.08.02
X	Wideband Power Sensor	KEYSIGHT	N1923A	MY56080004	2021.08.03	2022.08.02

Note:

1. All equipments are calibrated every one year.
2. The test instruments marked with "X" are used to measure the final test results.
3. Test Software version : RF Conducted Test Tools R3 V3.0.1.19.

For Radiated measurements / HY-CB01

	Equipment	Manufacturer	Model No.	Serial No.	Cal. Date	Due. Date
X	Loop Antenna	AMETEK	HLA6121	49611	2022/03/18	2023/03/17
X	Bi-Log Antenna	SCHWARZBECK	VULB9168	9168-675	2021.08.10	2022.08.09
X	Horn Antenna	ETS-Lindgren	3117	00201259	2021.11.09	2022.11.08
X	Horn Antenna	Com-Power	AH-840	101101	2021.11.30	2022.11.29
X	Pre-Amplifier	SGH	SGH0301	20211007-7	2022/02/22	2023/02/21
X	Pre-Amplifier	EMCI	EMC051835SE	980312	2022/02/22	2023/02/21
X	Pre-Amplifier	EMCI	EMC05820SE	980362	2021/08/24	2022/08/23
X	Pre-Amplifier	EMCI	EMC184045SE	980369	2022/05/12	2023/05/11
	Coaxial Cable	EMCI	EMC102-KM-KM-600	1160314		
	Coaxial Cable	EMCI	EMC102-KM-KM-7000	170242		
X	Filter	MICRO TRONICS	BRM50702	G251	2021/09/16	2022/09/15
	Filter	MICRO TRONICS	BRM50716	G188	2021/09/16	2022/09/15
X	EMI Test Receiver	R&S	ESR	102792	2021/12/15	2022/12/14
X	Spectrum Analyzer	R&S	FSV3044	101113	2022/01/25	2023/02/24
X	Coaxial Cable	SUHNER	SUCOFLEX 106	25450/6	2022/03/22	2023/03/21
	Coaxial Cable	SGH	HA800	GD20110222-8		
	Coaxial Cable	SGH	SGH18	2021003-8		
	Coaxial Cable	EMCI	EMC106	151113		

Note:

1. All equipments are calibrated every one year.
2. The test instruments marked with "X" are used to measure the final test results.
3. Test Software version : E3 210616 dekra V9.

1.7. Uncertainty

Uncertainties have been calculated according to the DEKRA internal document.


The reported expanded uncertainties are based on a standard uncertainty multiplied by a coverage factor of $k=2$, providing a level of confidence of approximately 95%.

Measurement uncertainties evaluated for each testing system and associated connections are given here to provide the system information for reference. Compliance determinations do not take into account measurement uncertainties for each testing system, but are based on the results of the compliance measurement.

Test item	Uncertainty	
Conducted Emission	± 3.42 dB	
Peak Power Output	± 0.89 dB	
Radiated Emission	Under 1GHz ± 4.05 dB	Above 1GHz ± 3.73 dB
RF Antenna Conducted Test	± 2.06 dB	
Band Edge	Under 1GHz ± 4.05 dB	Above 1GHz ± 3.73 dB
6dB Bandwidth	± 1544.74 Hz	
Power Density	± 2.06 dB	
Duty Cycle	± 2.31 ms	

2. Conducted Emission

2.1. Test Setup



2.2. Limits

FCC Part 15 Subpart C Paragraph 15.207 (dBuV) Limit		
Frequency MHz	Limits	
	QP	AV
0.15 - 0.50	66-56	56-46
0.50-5.0	56	46
5.0 - 30	60	50

Remarks: In the above table, the tighter limit applies at the band edges.

2.3. Test Procedure

The EUT and Peripherals are connected to the main power through a line impedance stabilization network (L.I.S.N.). This provides a 50 ohm /50uH coupling impedance for the measuring equipment. The peripheral devices are also connected to the main power through a LISN that provides a 50ohm /50uH coupling impedance with 50ohm termination. (Please refer to the block diagram of the test setup and photographs.)


Both sides of A.C. line are checked for maximum conducted interference. In order to find the maximum emission, the relative positions of equipment and all the interface cables must be changed according to ANSI C63.4: 2014 on conducted measurement.

Conducted emissions were investigated over the frequency range from 0.15MHz to 30MHz using a receiver bandwidth of 9kHz.


The EUT was setup to ANSI C63.4, 2014; tested to DTS test procedure of FCC KDB-558074 for compliance to FCC 47CFR Subpart C requirements.

2.4. Test Result of Conducted Emission

Power Adapter: BLJ06L050120U-R




Power Adapter: BLJ06L060100P-U



Power Adapter: BLJ06L060100P-V


Site :HY-SR01
Condition :Line
Mode :BLE-1M_2440MHz_TX
test by :Pounder



No.	Frequency	Level	Limit	Over	Read	Factor	Remark
	MHz	dBuV/m	Line	Limit	Level	dB	
1	0.183	46.78	64.34	-17.64	37.02	9.68	QP
2	0.183	27.87	54.34	-26.47	18.19	9.68	Average
3	0.294	39.93	60.41	-20.48	30.26	9.67	QP
4	0.294	22.84	50.41	-27.57	13.17	9.67	Average
5	0.465	34.03	56.60	-22.57	24.35	9.68	QP
6	0.465	21.21	46.60	-25.39	11.53	9.68	Average
7	0.861	19.49	56.00	-36.51	9.78	9.71	QP
8	0.861	13.39	46.00	-32.61	3.68	9.71	Average
9	4.250	16.97	56.00	-39.03	6.39	10.58	QP
10	4.250	9.73	46.00	-36.27	-0.85	10.58	Average
11	16.612	24.62	60.00	-35.38	13.87	10.75	QP
12	16.612	17.87	50.00	-32.13	7.12	10.75	Average

Note:
1. Level = Read Level + Factor
2. Factor = LISN insertion loss + Cable loss
3. Over Limit = Level - Limit Line

Site :HY-SR01
Condition :Neutral
Mode :BLE-1M_2440MHz_TX
test by :Pounder




No.	Frequency	Level	Limit	Over	Read	Factor	Remark
	MHz	dBuV/m	Line	Limit	Level	dB	
1	0.183	46.69	64.33	-17.64	37.03	9.66	QP
2	0.183	27.90	54.33	-26.43	18.74	9.66	Average
3	0.309	39.17	60.00	-20.83	29.52	9.65	QP
4	0.309	21.11	50.00	-28.89	11.46	9.65	Average
5	0.615	26.36	56.00	-29.64	16.69	9.67	QP
6	0.615	12.15	46.00	-33.85	2.48	9.67	Average
7	2.084	11.27	56.00	-44.73	1.44	9.83	QP
8	2.084	6.25	46.00	-39.75	-3.58	9.83	Average
9	5.356	12.02	60.00	-47.98	1.19	10.83	QP
10	5.356	7.32	50.00	-42.68	-3.51	10.83	Average
11	11.464	17.90	60.00	-42.10	7.09	10.81	QP
12	11.464	11.90	50.00	-38.10	1.09	10.81	Average

Note:
1. Level = Read Level + Factor
2. Factor = LISN insertion loss + Cable loss
3. Over Limit = Level - Limit Line

Power Adapter: FRM06-S05-UU


Site :HY-SR01
Condition :Line
Mode :BLE-1M_2440MHz_TX
test by :Pounder



No.	Frequency	Level	Limit	Over	Read	Factor	Remark
	MHz	dBuV/m	Line	Limit	Level	dB	
1	0.162	41.10	65.34	-24.24	31.42	9.68	QP
2	0.162	23.46	55.34	-31.88	13.78	9.68	Average
3	0.408	25.20	57.68	-32.48	15.52	9.68	QP
4	0.408	17.58	47.68	-30.10	7.90	9.68	Average
5	0.758	23.98	56.00	-32.02	14.27	9.71	QP
6	0.758	17.66	46.00	-28.34	7.95	9.71	Average
7	2.395	17.41	56.00	-38.59	7.46	9.95	QP
8	2.395	10.91	46.00	-35.09	0.96	9.95	Average
9	7.161	19.38	60.00	-40.62	8.55	10.83	QP
10	7.161	12.74	50.00	-37.26	1.91	10.83	Average
11	17.264	9.61	60.00	-50.39	-1.12	10.73	QP
12	17.264	4.48	50.00	-45.52	-6.25	10.73	Average

Note:
1. Level = Read Level + Factor
2. Factor = LISN insertion loss + Cable loss
3. Over Limit = Level - Limit Line

Site :HY-SR01
Condition :Neutral
Mode :BLE-1M_2440MHz_TX
test by :Pounder




No.	Frequency	Level	Limit	Over	Read	Factor	Remark
	MHz	dBuV/m	Line	Limit	Level	dB	
1	0.160	41.43	65.47	-24.04	31.77	9.66	QP
2	0.160	23.79	55.47	-31.68	14.13	9.66	Average
3	0.353	25.15	58.89	-33.74	15.49	9.66	QP
4	0.353	18.17	48.89	-30.72	8.51	9.66	Average
5	0.874	21.31	56.00	-34.69	11.62	9.69	QP
6	0.874	14.66	46.00	-31.34	4.97	9.69	Average
7	2.613	18.03	56.00	-37.97	8.03	10.00	QP
8	2.613	11.20	46.00	-34.80	1.20	10.00	Average
9	6.942	21.85	60.00	-38.15	11.04	10.81	QP
10	6.942	15.68	50.00	-34.32	4.87	10.81	Average
11	13.683	16.94	60.00	-43.06	6.12	10.82	QP
12	13.683	11.59	50.00	-38.41	0.77	10.82	Average

Note:
1. Level = Read Level + Factor
2. Factor = LISN insertion loss + Cable loss
3. Over Limit = Level - Limit Line

3. Peak Power Output

3.1. Test Setup



3.2. Limit

The maximum peak power shall be less 1Watt.

3.3. Test Procedure

The EUT was tested according to C63.10:2013 for compliance to FCC 47CFR 15.247 requirements. The maximum peak conducted output power using C63.10:2013 Section 11.9.1.3 PKPM1 Peak power meter method.

3.4. Test Result of Peak Power Output


Product : Automatic Upper Arm Blood Pressure Monitor
Test Item : Peak Power Output
Test Mode : Mode 1: Transmit - 1Mbps
Test Date : 2022/06/01

Channel No.	Frequency (MHz)	Measurement (dBm)	Required Limit	Result
00	2402	-1.56	1 Watt= 30 dBm	Pass
19	2440	-2.33	1 Watt= 30 dBm	Pass
39	2480	-2.92	1 Watt= 30 dBm	Pass


4. Radiated Emission

4.1. Test Setup


Radiated Emission Under 30MHz



Radiated Emission Below 1GHz



Radiated Emission Above 1GHz



4.2. Limits

➤ General Radiated Emission Limits

Emissions radiated outside of the specified frequency bands, except for harmonics, shall be attenuated by at least 20dB below the level of the fundamental or to the general radiated emission limits in paragraph 15.209, whichever is the lesser attenuation.

FCC Part 15 Subpart C Paragraph 15.209 Limits		
Frequency MHz	Field strength (microvolts/meter)	Measurement distance (meter)
0.009-0.490	2400/F(kHz)	300
0.490-1.705	24000/F(kHz)	30
1.705-30	30	30
30-88	100	3
88-216	150	3
216-960	200	3
Above 960	500	3

- Remarks:
1. RF Voltage (dBuV) = 20 log RF Voltage (uV)
 2. In the Above Table, the tighter limit applies at the band edges.
 3. Distance refers to the distance in meters between the measuring instrument antenna and the closed point of any part of the device or system.

4.3. Test Procedure

The EUT was setup according to ANSI C63.10: 2013 and tested according to C63.10:2013 Section 11.12.1 for compliance to FCC 47CFR 15.247 requirements.

Measuring the frequency range below 1GHz, the EUT is placed on a turn table which is 0.8 meter above ground, when measuring the frequency range above 1GHz, the EUT is placed on a turn table which is 1.5 meter above ground.

The turn table is rotated 360 degrees to determine the position of the maximum emission level.

The EUT was positioned such that the distance from antenna to the EUT was 3 meters.

The antenna is scanned between 1 meter and 4 meters to find out the maximum emission level. This is repeated for both horizontal and vertical polarization of the antenna. In order to find the maximum emission, all of the interface cables were manipulated according to ANSI C63.10: 2013 on radiated measurement.

The resolution bandwidth below 30MHz setting on the field strength meter is 9kHz and 30MHz~1GHz is 120kHz and above 1GHz is 1MHz.

Radiated emission measurements below 30MHz are made using Loop Antenna and 30MHz~1GHz are made using broadband Bilog antenna and above 1GHz are made using Horn Antennas.

The measurement is divided into the Preliminary Measurement and the Final Measurement.

The suspected frequencies are searched for in Preliminary Measurement with the measurement antenna kept pointed at the source of the emission both in azimuth and elevation, with the polarization of the antenna oriented for maximum response. The antenna is pointed at an angle towards the source of the emission, and the EUT is rotated in both height and polarization to maximize the measured emission. The emission is kept within the illumination area of the 3 dB bandwidth of the antenna. The measurement frequency range from 9kHz - 10th Harmonic of fundamental was investigated.

RBW and VBW Parameter setting:

According to C63.10 Section 11.12.2.4 Peak measurement procedure.

RBW = as specified in Table 1.

$VBW \geq 3 \times RBW$.

Table 1 - RBW as a function of frequency

Frequency	RBW
9-150 kHz	200-300 Hz
0.15-30 MHz	9-10 kHz
30-1000 MHz	100-120 kHz
> 1000 MHz	1 MHz

According to C63.10 Section 11.12.2.5 Average measurement procedure.

RBW = 1MHz.

VBW = 10Hz, when duty cycle $\geq 98\%$

$VBW \geq 1/T$, when duty cycle $< 98\%$

(T refers to the minimum transmission duration over which the transmitter is on and is transmitting at its maximum power control level for the tested mode of operation.)


2.4GHz band	Duty Cycle (%)	T (ms)	1/T (Hz)	VBW (Hz)
BLE (1Mbps)	94.35	2.1700	461	500

Note: Duty Cycle Refer to Section 9.

4.4. Test Result of Radiated Emission

Power Adapter: BLJ06L060100P-V


Site :HY-CB01
Condition :3m ,HORIZONTAL
Mode :TX_ble1M_2402MHz
TEST BY :Johnny



No.	Frequency	Level	Limit	Over	Read	Factor	Remark
	MHz	dBuV/m	dBuV/m	dB	dBuV	dB	
1	4804.000	52.00	74.00	-22.00	55.08	-3.08	Peak
2	7206.000	52.54	74.00	-21.46	51.66	0.88	Peak

Note:
1. Level = Read Level + Factor
2. Factor = Antenna- Factor + Cable Loss - Preamp Factor
3. Over Limit = Level - Limit Line


Site :HY-CB01
Condition :3m ,VERTICAL
Mode :TX_ble1M_2402MHz
TEST BY :Johnny



No.	Frequency	Level	Limit	Over	Read	Factor	Remark
	MHz	dBuV/m	dBuV/m	dB	dBuV	dB	
1	4804.000	52.33	54.00	-1.67	55.41	-3.08	Average
2	4804.000	55.39	74.00	-18.61	58.47	-3.08	Peak
3	7206.000	49.16	54.00	-4.84	48.28	0.88	Average
4	7206.000	55.29	74.00	-18.71	54.41	0.88	Peak

Note:
1. Level = Read Level + Factor
2. Factor = Antenna- Factor + Cable Loss - Preamp Factor
3. Over Limit = Level - Limit Line


Site :HY-CB01
Condition :3m ,HORIZONTAL
Mode :TX_ble1M_2440MHz
TEST BY :Johnny



No.	Frequency	Level	Limit	Over	Read	Factor	Remark
	MHz	dBuV/m	dBuV/m	dB	dBuV	dB	
1	4880.000	50.82	74.00	-23.18	54.05	-3.23	Peak
2	7320.000	51.02	74.00	-22.98	50.13	0.89	Peak


Note:
1. Level = Read Level + Factor
2. Factor = Antenna- Factor + Cable Loss - Preamp Factor
3. Over Limit = Level - Limit Line

Site :HY-CB01
Condition :3m ,VERTICAL
Mode :TX_ble1M_2440MHz
TEST BY :Johnny




No.	Frequency	Level	Limit	Over	Read	Factor	Remark
	MHz	dBuV/m	dBuV/m	dB	dBuV	dB	
1	4880.000	52.26	54.00	-1.74	55.49	-3.23	Average
2	4880.000	55.37	74.00	-18.63	58.60	-3.23	Peak
3	7320.000	52.90	74.00	-21.10	52.01	0.89	Peak

Note:
1. Level = Read Level + Factor
2. Factor = Antenna- Factor + Cable Loss - Preamp Factor
3. Over Limit = Level - Limit Line



Power Adapter: BLJ06L050120U-R

Site :HY-CB01
Condition :3m ,HORIZONTAL
Mode :TX_ble1M_2440MHz
Model :BLJ06L050120U-R
TEST BY :Caster




No.	Frequency MHz	Level dBuV/m	Limit Line dBuV/m	Over Limit dB	Read Level dBuV	Factor dB	Remark
1	48.430	32.41	40.00	-7.59	56.35	-23.94	QP
2	296.750	35.20	46.00	-10.80	58.81	-23.61	QP
3	437.400	28.81	46.00	-17.19	48.50	-19.69	QP
4	594.540	35.42	46.00	-10.58	51.55	-16.13	QP
5	875.840	38.74	46.00	-7.26	51.47	-12.73	QP
6	955.380	29.85	46.00	-16.95	48.66	-11.61	QP

Note:

1. Level = Read Level + Factor
2. Factor = Antenna Factor + Cable Loss - Preamp Factor
3. Over Limit = Level - Limit Line
4. The emission under 30MHz was not included since the emission levels are very low against the limit.

Site :HY-CB01
Condition :3m ,VERTICAL
Mode :TX_ble1M_2440MHz
Model :BLJ06L050120U-R
TEST BY :Caster




No.	Frequency MHz	Level dBuV/m	Limit Line dBuV/m	Over Limit dB	Read Level dBuV	Factor dB	Remark
1	48.430	32.48	40.00	-7.52	56.42	-23.94	QP
2	296.750	38.54	46.00	-7.46	62.15	-23.61	QP
3	482.990	32.42	46.00	-13.58	51.34	-18.92	QP
4	644.010	34.78	46.00	-11.22	50.20	-15.42	QP
5	875.840	37.64	46.00	-8.36	50.37	-12.73	QP
6	937.920	31.38	46.00	-14.62	43.28	-11.90	QP

Note:

1. Level = Read Level + Factor
2. Factor = Antenna Factor + Cable Loss - Preamp Factor
3. Over Limit = Level - Limit Line
4. The emission under 30MHz was not included since the emission levels are very low against the limit.

Power Adapter: BLJ06L060100P-U

Site :HY-CB01
Condition :3m ,HORIZONTAL
Mode :TX_ble1M_2440MHz
Model :BLJ06L060100P-U
TEST BY :Caster




No.	Frequency MHz	Level dBuV/m	Limit Line dBuV/m	Over Limit dB	Read Level dBuV	Factor dB	Remark
1	48.430	33.35	40.00	-6.65	57.29	-23.94	QP
2	296.750	34.04	46.00	-11.96	57.65	-23.61	QP
3	371.440	34.23	46.00	-11.77	55.81	-21.58	QP
4	497.540	31.73	46.00	-14.27	50.17	-18.44	QP
5	644.010	33.76	46.00	-12.24	49.18	-15.42	QP
6	946.650	38.26	46.00	-7.74	50.01	-11.75	QP

Note:

1. Level = Read Level + Factor
2. Factor = Antenna Factor + Cable Loss - Preamp Factor
3. Over Limit = Level - Limit Line
4. The emission under 30MHz was not included since the emission levels are very low against the limit.

Site :HY-CB01
Condition :3m ,VERTICAL
Mode :TX_ble1M_2440MHz
Model :BLJ06L060100P-U
TEST BY :Caster




No.	Frequency MHz	Level dBuV/m	Limit Line dBuV/m	Over Limit dB	Read Level dBuV	Factor dB	Remark
1	48.430	33.88	40.00	-6.12	57.82	-23.94	QP
2	296.750	37.55	46.00	-8.45	61.16	-23.61	QP
3	482.990	32.48	46.00	-13.52	51.40	-18.92	QP
4	562.530	33.42	46.00	-12.58	50.78	-17.36	QP
5	668.260	34.09	46.00	-11.91	49.37	-15.28	QP
6	946.650	35.67	46.00	-10.33	47.42	-11.75	QP

Note:

1. Level = Read Level + Factor
2. Factor = Antenna Factor + Cable Loss - Preamp Factor
3. Over Limit = Level - Limit Line
4. The emission under 30MHz was not included since the emission levels are very low against the limit.

Power Adapter: BLJ06L060100P-V


Site :HY-CB01
Condition :3m ,HORIZONTAL
Mode :TX_bleIM_2440MHz
Model :BLJ06L060100P-V
TEST BY :Caster



No.	Frequency	Level	Limit	Over	Read	Factor	Remark
	MHz	dBuV/m	dBuV/m	dB	dBuV	dB	
1	48.430	32.23	40.00	-7.77	56.17	-23.94	QP
2	371.440	35.67	46.00	-10.33	57.25	-21.58	QP
3	514.030	31.37	46.00	-14.63	49.56	-18.19	QP
4	594.540	36.57	46.00	-9.43	52.70	-16.13	QP
5	816.670	29.18	46.00	-16.82	42.46	-13.28	QP
6	946.650	39.15	46.00	-6.85	50.90	-11.75	QP

Note:
1. Level = Read Level + Factor
2. Factor = Antenna Factor + Cable Loss - Preamp Factor
3. Over Limit = Level - Limit Line
4. The emission under 30MHz was not included since the emission levels are very low against the limit.

Site :HY-CB01
Condition :3m ,VERTICAL
Mode :TX_bleIM_2440MHz
Model :BLJ06L060100P-V
TEST BY :Caster




No.	Frequency	Level	Limit	Over	Read	Factor	Remark
	MHz	dBuV/m	dBuV/m	dB	dBuV	dB	
1	30.970	34.05	40.00	-5.95	59.51	-25.46	QP
2	288.990	24.00	46.00	-22.00	47.81	-23.81	QP
3	371.440	36.85	46.00	-9.15	58.43	-21.58	QP
4	519.850	33.03	46.00	-12.97	51.09	-18.06	QP
5	692.510	27.42	46.00	-18.58	42.30	-14.88	QP
6	891.360	27.77	46.00	-18.23	40.25	-12.48	QP

Note:
1. Level = Read Level + Factor
2. Factor = Antenna Factor + Cable Loss - Preamp Factor
3. Over Limit = Level - Limit Line
4. The emission under 30MHz was not included since the emission levels are very low against the limit.

Power Adapter: FRM06-S05-UU


Site :HY-CB01
Condition :3m ,HORIZONTAL
Mode :TX_bleIM_2440MHz
Model :FRM06-S05-UU
TEST BY :Caster



No.	Frequency	Level	Limit	Over	Read	Factor	Remark
	MHz	dBuV/m	dBuV/m	dB	dBuV	dB	
1	48.430	33.33	40.00	-6.67	57.27	-23.94	QP
2	296.750	34.08	46.00	-11.92	57.69	-23.61	QP
3	532.460	30.01	46.00	-15.99	47.92	-17.91	QP
4	644.010	32.35	46.00	-13.65	47.77	-15.42	QP
5	875.840	37.52	46.00	-8.48	50.25	-12.73	QP
6	938.890	31.74	46.00	-14.26	43.62	-11.88	QP

Note:
1. Level = Read Level + Factor
2. Factor = Antenna Factor + Cable Loss - Preamp Factor
3. Over Limit = Level - Limit Line
4. The emission under 30MHz was not included since the emission levels are very low against the limit.

Site :HY-CB01
Condition :3m ,VERTICAL
Mode :TX_bleIM_2440MHz
Model :FRM06-S05-UU
TEST BY :Caster




No.	Frequency	Level	Limit	Over	Read	Factor	Remark
	MHz	dBuV/m	dBuV/m	dB	dBuV	dB	
1	48.430	32.96	40.00	-7.04	56.90	-23.94	QP
2	296.750	36.87	46.00	-9.13	60.48	-23.61	QP
3	482.990	31.90	46.00	-14.10	50.82	-18.92	QP
4	668.260	34.46	46.00	-11.54	49.74	-15.28	QP
5	875.840	38.00	46.00	-8.00	50.73	-12.73	QP
6	955.380	27.67	46.00	-18.33	39.28	-11.61	QP

Note:
1. Level = Read Level + Factor
2. Factor = Antenna Factor + Cable Loss - Preamp Factor
3. Over Limit = Level - Limit Line
4. The emission under 30MHz was not included since the emission levels are very low against the limit.

5. RF Antenna Conducted Test

5.1. Test Setup



5.2. Limits

According to FCC Section 15.247(d). In any 100 kHz bandwidth outside the frequency band in which the spread spectrum or digitally modulated intentional radiator is operating, the radio frequency power that is produced by the intentional radiator shall be at least 20 dB below that in the 100 kHz bandwidth within the band that contains the highest level of the desired power, based on either an RF conducted or measurement, provided the transmitter demonstrates compliance with the peak conducted power limits. If the transmitter complies with the conducted power limits based on the use of RMS averaging over a time interval, the attenuation required under this paragraph shall be 30 dB instead of 20 dB.

5.3. Test Procedure


The EUT was tested according to C63.10:2013 Section 11.11 for compliance to FCC 47CFR 15.247 requirements.

Set RBW = 100 kHz, Set VBW > RBW, scan up through 10th harmonic.

5.4. Test Result of RF Antenna Conducted Test

Product : Automatic Upper Arm Blood Pressure Monitor
Test Item : RF Antenna Conducted Test
Test Mode : Mode 1: Transmit - 1Mbps
Test Date : 2022/06/01

Figure Channel 19:




Date: 1.JUN.2022 12:40:56


6. Band Edge

6.1. Test Setup

RF Conducted Measurement



RF Radiated Measurement:



6.2. Limit

According to FCC Section 15.247(d). In any 100 kHz bandwidth outside the frequency band in which the spread spectrum or digitally modulated intentional radiator is operating, the radio frequency power that is produced by the intentional radiator shall be at least 20 dB below that in the 100 kHz bandwidth within the band that contains the highest level of the desired power, based on either an RF conducted or measurement, provided the transmitter demonstrates compliance with the peak conducted power limits. If the transmitter complies with the conducted power limits based on the use of RMS averaging over a time interval, the attenuation required under this paragraph shall be 30 dB instead of 20 dB. Attenuation below the general limits specified in Section 15.209(a) is not required. In addition, radiated emissions which fall in the restricted bands, as defined in Section 15.205(a), must also comply with the radiated emission limits specified in Section 15.209(a) (see Section 15.205(c)).

6.3. Test Procedure

The EUT was setup according to ANSI C63.10, 2013 and tested according to C63.10:2013 Section 11.12.1 for compliance to FCC 47CFR 15.247 requirements.

The EUT is placed on a turn table which is 1.5 meter above ground. The turn table is rotated 360 degrees to determine the position of the maximum emission level. The EUT was positioned such that the distance from antenna to the EUT was 3 meters.

The antenna is scanned from 1 meter to 4 meters to find out the maximum emission level. This is repeated for both horizontal and vertical polarization of the antenna. In order to find the maximum emission, all of the interface cables were manipulated according to ANSI C63.10:2013 on radiated measurement.

RBW and VBW Parameter setting:

According to C63.10 Section 11.12.2.4 Peak measurement procedure.

RBW = as specified in Table 1.

$VBW \geq 3 \times RBW$.

Table 1 - RBW as a function of frequency

Frequency	RBW
9-150 kHz	200-300 Hz
0.15-30 MHz	9-10 kHz
30-1000 MHz	100-120 kHz
> 1000 MHz	1 MHz

According to C63.10 Section 11.12.2.5 Average measurement procedure.

RBW = 1MHz.

VBW = 10Hz, when duty cycle $\geq 98\%$


$VBW \geq 1/T$, when duty cycle $< 98\%$

(T refers to the minimum transmission duration over which the transmitter is on and is transmitting at its maximum power control level for the tested mode of operation.)


2.4GHz band	Duty Cycle (%)	T (ms)	1/T (Hz)	VBW (Hz)
BLE (1Mbps)	94.35	2.1700	461	500

Note: Duty Cycle Refer to Section 9.

6.4. Test Result of Band Edge




Site :HY-CB01
Condition :3m ,HORIZONTAL
Mode :TX_ble1M_2480MHz
TEST BY :Johnny



Note:

1. Level = Read Level + Factor
2. Factor = Antenna- Factor + Cable Loss - Preamp Factor
3. Over Limit = Level - Limit Line


Site :HY-CB01
Condition :3m ,HORIZONTAL
Mode :TX_ble1M_2480MHz
TEST BY :Johnny



Note:

1. Level = Read Level + Factor
2. Factor = Antenna- Factor + Cable Loss - Preamp Factor
3. Over Limit = Level - Limit Line


Site :HY-CB01
Condition :3m ,Vertical
Mode :TX_ble1M_2480MHz
TEST BY :Johnny



Note:

1. Level = Read Level + Factor
2. Factor = Antenna- Factor + Cable Loss - Preamp Factor
3. Over Limit = Level - Limit Line

Site :HY-CB01
Condition :3m ,Vertical
Mode :TX_ble1M_2480MHz
TEST BY :Johnny




Note:

1. Level = Read Level + Factor
2. Factor = Antenna- Factor + Cable Loss - Preamp Factor
3. Over Limit = Level - Limit Line

7. 6dB Bandwidth

7.1. Test Setup



7.2. Limits

The minimum bandwidth shall be at least 500 kHz.

7.3. Test Procedure


The EUT was setup according to ANSI C63.4, 2014; tested according to ANSI C63.10 Section 11.8 for compliance to FCC 47CFR 15.247 requirements.

7.4. Test Result of 6dB Bandwidth

Product : Automatic Upper Arm Blood Pressure Monitor
 Test Item : 6dB Bandwidth Data
 Test Mode : Mode 1: Transmit - 1Mbps
 Test Date : 2022/06/01

Channel No.	Frequency (MHz)	Measurement Level (kHz)	Required Limit (kHz)	Result
00	2402	684	>500	Pass
19	2440	704	>500	Pass
39	2480	704	>500	Pass


Figure Channel 19:



Date: 1.JUN.2022 12:39:49

8. Power Density

8.1. Test Setup



8.2. Limits

The transmitted power density averaged over any 1 second interval shall not be greater +8dBm in any 3kHz bandwidth.

8.3. Test Procedure

The EUT was setup according to ANSI C63.10, 2013; tested according to DTS test procedure of KDB 558074 for compliance to FCC 47CFR 15.247 requirements.


The maximum power spectral density using C63.10 Section 11.10.2 Method PKPSD (peak PSD)

8.4. Test Result of Power Density

Product : Automatic Upper Arm Blood Pressure Monitor
 Test Item : Power Density Data
 Test Mode : Mode 1: Transmit - 1Mbps
 Test Date : 2022/06/01

Channel No.	Frequency (MHz)	Measure Level (dBm)	Limit (dBm)	Result
00	2402	-20.08	$\leq 8\text{dBm}$	Pass
19	2440	-14.76	$\leq 8\text{dBm}$	Pass
39	2480	-15.38	$\leq 8\text{dBm}$	Pass


Figure Channel 19:



Date: 1.JUN.2022 12:40:08

9. Duty Cycle

9.1. Test Setup



9.2. Test Procedure

The EUT was setup according to ANSI C63.10 2013; tested according to ANSI C63.10 2013 for compliance to FCC 47CFR 15.247 requirements.

9.3. Test Result of Duty Cycle

Product : Automatic Upper Arm Blood Pressure Monitor
 Test Item : Duty Cycle


Formula:

Duty Cycle = $T_{on} / (T_{on} + T_{off})$

Duty Factor = $10 \log (1/\text{Duty Cycle})$

2.4GHz band	Ton (ms)	Ton + Toff (ms)	Duty Cycle (%)	Duty Factor (dB)
BLE (1Mbps)	2.1700	2.3000	94.35	0.25

BLE (1Mbps)



16:52:10 01.06.2022

10. EMI Reduction Method During Compliance Testing

No modification was made during testing.