

ERRATA

Change to the Specifications table on page 18 — values in **bold** have been updated.

• filter settings	ON: 0.5– 25 Hz, ± 0.2 dB; OFF: 0.05–100 Hz ± 0.2 dB
• gain settings	1, 2, 5, 10, 20 mm/mV $\pm$ 25 %

Addition to the Prepare the Telemetry Wand section on page 5, after the caution and before the Make External Cable Connections section:

This device complies with Part 15 of the Federal Communications Commission (FCC) rules. Operation is subject to the following two conditions:

1. This device may not cause harmful interference, and
2. This device must accept any interference received, including interference that may cause undesired operation.

CAUTION: Changes or modifications not expressly approved by Guidant could void the user's authority to operate the equipment.