

Manufacturer's Statement

FCC ID location justification

2021-04-23

FCC ID: 2AXDT-RFM003

To whom it may concern


We, the undersigned, hereby submit the justification for placing the FCC-ID in the associated User Manual for the Hearing Aids containing the module with the above-mentioned FCC-ID.

According to KDB Publication 784748 D01, when the device is so small, or for such use that it is impracticable to label with a font size that is four points or larger (and the device does not utilize electronic labeling), then the FCC-ID shall be placed in the User Manual.

Hearing Aids having the above-mentioned FCC-ID are so small and considering the dimensions of the equipment and the label area, it is impracticable to label FCC-ID with a font size required on the devices. Therefore, the FCC-ID label is placed in the User Manual of the end product. Instruction for module integration regarding the FCC-ID labelling is included in the integration manual.


If you have any questions, feel free to contact us at the address shown above.

Yours sincerely,



Stefan Menzl
Chief R&D Officer

stefan.menzl@wsa.com



Jan Knittel
Vice President Q QMS

jan.knittel@wsa.com