

akkreditiert durch die / *accredited by the*
Deutsche Akkreditierungsstelle GmbH

als Kalibrierlaboratorium im / *as calibration laboratory in the*

Deutsche
Akkreditierungsstelle
D-K-15195-01-00

Deutschen Kalibrierdienst
DKD

Kalibrierschein

Calibration certificate

Kalibrierzeichen

Calibration mark

606113

D-K-
15195-01-00

2021-08

Gegenstand <i>Object</i>	ZNB40 VECT.NETW.ANALYZER 2 PORT
Hersteller <i>Manufacturer</i>	ROHDE & SCHWARZ
Typ <i>Type</i>	ZNB40
Fabrikat/Serien-Nr. <i>Serial number</i>	101544
Auftraggeber <i>Customer</i>	RISE Research Institutes of Sweden AB
	Brinellgatan 4 504 62 Borås SE
Auftragsnummer <i>Order No.</i>	8800006610 10, X1004157
Anzahl der Seiten des Kalibrierscheines <i>Number of pages of the certificate</i>	3 Certificate 18 Outgoing Results 18 Incoming Results
Datum der Kalibrierung <i>Date of calibration</i>	2021-08-04

Dieser Kalibrierschein dokumentiert die Rückführung auf nationale Normale zur Darstellung der Einheiten in Übereinstimmung mit dem Internationalen Einheitensystem (SI).

Die DAkkS ist Unterzeichner der multilateralen Übereinkommen der European co-operation for Accreditation (EA) und der International Laboratory Accreditation Cooperation (ILAC) zur gegenseitigen Anerkennung der Kalibrierscheine.

Für die Einhaltung einer angemessenen Frist zur Wiederholung der Kalibrierung ist der Benutzer verantwortlich.

This calibration certificate documents the traceability to national standards, which realize the units of measurement according to the International System of Units (SI).

The DAkkS is signatory to the multilateral agreements of the European co-operation for Accreditation (EA) and of the International Laboratory Accreditation Cooperation (ILAC) for the mutual recognition of calibration certificates.

The user is obliged to have the object recalibrated at appropriate intervals.

Dieser Kalibrierschein darf nur vollständig und unverändert weiterverbreitet werden. Auszüge oder Änderungen bedürfen der Genehmigung des ausstellenden Kalibrierlaboratoriums. Kalibrierscheine sind bei Nennung des für die Freigabe Verantwortlichen in Klarschrift auch ohne Unterschrift gültig.

This calibration certificate may not be reproduced other than in full except with the permission of the issuing laboratory. Calibration certificates with the full name of the approval responsible person are valid without signature.

Datum der Ausstellung
Date of issue

2021-08-04

Freigabe des Kalibrierscheins durch
Approval of the calibration certificate by


Dr. Gerhard Rösel
Leiter des Kalibrierlaboratoriums
Head of the calibration laboratory


Philipp Bartsch
Bearbeiter
Person in charge

Object ZNB40 VECT.NETW.ANALYZER 2
Type ZNB40 **Serial No.** 101544
Date 2021-08-04 **Material No.** 1311.6010K72
Page 2 / 3

606113
D-K- 15195-01-00
2021-08



ROHDE & SCHWARZ

Place of Calibration

87700 Memmingen, Rohde-und-Schwarz-Str. 1

Calibration Procedure

The calibration of the object can be classified as a direct measurement. Frequency was compared using a GPS synchronized rubidium oscillator. RF power was measured with a power sensor. Linearity was measured with a spectrum analyzer.

The traceability is represented in the table Working Standards used.

Calibration Method

See first page of Outgoing Results

Statement of Compliance

Incoming: Defective

Outgoing: All measured values are within
the data sheet specifications.

Working Standards used

Item	Type	Serial Number	Calibration Certificate Number	Cal. Due
Spectrum Analyzer 20Hz...67GHz	FSU67	100056	576320 D-K-15195-01-00 2020-11	2022-05-31
Therm.Power Sensor DC-40GHz	NRP-Z55	100030	576360 D-K-15195-01-00 2020-11	2022-05-31
Therm.Power Sensor DC-40GHz	NRP-Z55	100031	576363 D-K-15195-01-00 2020-11	2022-05-31
Therm.Power Sensor DC-44GHz	NRP-Z55	140160	549276 D-K-15195-01-01 2020-04	2021-10-31
Therm.Power Sensor DC-40GHz	NRP-Z55	100271	576370 D-K-15195-01-00 2020-11	2022-05-31
Standardfrequency unit 1X1	SYSTEM2000	808	100029 D-K-15195-01-01 2020-06	2023-06-30
RMS Peak-Voltmeter DC-30 MHz	URE3	100183	598563 D-K-15195-01-00 2021-05	2022-11-30
Vect. Netw. Analyzer 4PORT	ZVA67	101347	892929 D-K-15195-01-01 2020-05	2021-11-30
2,92mm Fixed Match Cal. Kit	ZV-Z34	100123	606917 D-K-15195-01-00 2021-07	2022-07-31

Custom Data

Asset Number: BX50051

Object ZNB40 VECT.NETW.ANALYZER 2
Type ZNB40 **Serial No.** 101544
Date 2021-08-04 **Material No.** 1311.6010K72
Page 3 / 3

606113
D-K- 15195-01-00
2021-08



ROHDE & SCHWARZ

Measurement Uncertainty

The expanded measurement uncertainty corresponds to the measurement results from the standard measurement uncertainty multiplied by the coverage factor $k = 2$.

It was determined in accordance with EA-4/02 M:2013. The true value is located in the corresponding interval with a probability of 95 %.

Environmental Conditions

Ambient Temperature $(23 \pm 3) ^\circ\text{C}$ Relative Humidity 20%-70%

Ancillary Functional Measurements

In addition to the calibration results, the calibration certificate includes functional measurements that might have an influence on the measurement uncertainty of the calibration results. The functional measurement results are marked and are not intended to be used to support the further dissemination of metrological traceability. They are intended to verify the requirements on the measurement object according to manufacturer specifications and technical standards.

Comments on Measurement Results

The measurement results in the test report stated below have been tested for compliance with the given specifications and marked if necessary. The associated uncertainty of measurement has been taken into account. Measurement results that are not covered by the DAkkS accreditation are marked with ¹.

Ref.: ILAC G8:09/2019 'Guidelines on Decision Rules and Statements of Conformity'.

Outgoing Results

Designation:	Network Analyzer
Type:	ZNB40-2Port
Material No.:	1311.6010K72
Serial No.:	101544
Certificate No.:	606113 D-K-15195-01-00 2021-08
Referring to Test Documentation:	1311.6010.01-PB-10.03

Test Department:	3MES2
Name:	see certificate
Date:	2021-08-04



The following abbreviations may be used in this document

{a}	No measurement uncertainty stated because the errors always add together. So it is sure that a measurement result evaluated as "PASS" is pass.
{b}	The measurement uncertainty depends on the measurement result. The stated measurement uncertainty is valid for the close area around the specification. Measurement results outside the close area have a higher measurement uncertainty but are within the specification.
{c}	Functional test, therefore no measurement uncertainty is stated.
{d}	Typical value, refer to performance test.
{e}	The measurement uncertainty is taken into account when setting the measuring system.
{f}	Verification of specified requirements. Technical operation that consist of the determination of one or more characteristics to a specified procedure.
DL or DT	Data Limit for symmetrical tolerance limits
DLL	Datasheet Lower Limit
DUL	Datasheet Upper Limit
MU	Symmetrical Measurement Uncertainty
MLL or MLV	Measurement Uncertainty Lower Value
MUL or MUV	Measurement Uncertainty Upper Value
Nom.	Nominal Value
Dev.	Deviation
Act.	Actual Value
UGB	Uncertainty Guard Band: Measuring uncertainty violates the data (spec.) limit.
UGB1	A compliance statement may be possible where a confidence level of less than 95 % is acceptable.
UGB2	A non-compliance statement may be possible where a confidence level of less than 95 % is acceptable.
DU	Datasheet Uncertainty

Explanation of charts




Table of contents

Software used for measurement	4
1. Connector gauge	5
2. Static frequency accuracy	5
3. Output harmonics	5
3.1 Harmonics PORT1	5
3.2 Harmonics PORT2	6
4. Maximum output power	6
4.1 Maximum output power PORT1	6
4.2 Maximum output power PORT2	7
5. Accuracy of output power	7
5.1 Accuracy of output power PORT1	7
5.2 Accuracy of output power PORT2	7
6. Output level linearity	8
6.1 Output level linearity PORT1	8
6.2 Output level linearity PORT2	9
7. Input power measurement accuracy	11
7.1 Input power measurement accuracy PORT1	11
7.2 Input power measurement accuracy PORT2	11
8. Input power linearity (low level)	12
8.1 Input power linearity (low level) PORT1	12
8.2 Input power linearity (low level) PORT2	12
9. Input power linearity (high level)	12
9.1 Input power linearity PORT1 (high level)	12
9.2 Input power linearity PORT2 (high level)	14
10. Trace noise magnitude	16
10.1 Trace noise magnitude PORT1	16
10.2 Trace noise magnitude PORT2	16
11. Trace noise phase	17
11.1 Trace noise phase PORT1	17
11.2 Trace noise phase PORT2	17
12. Dynamic range	17
12.1 Dynamic range S21	17
12.2 Dynamic range S12	18
13. Rear connectors and signals	18
13.1 Reference output	18
14. BIAS - functional test	18

Software used for measurement			
Item	Type	Version	Remark
7010.9063.00_.G5LIM	Limit File	2019-12-04 12:15	Test Management Software G5
Suite	Setup	V12.16	
Test Program (7010.9063.00_)	Component	V04.96	

1. Connector gauge

Connector	Port	DDL	DUL	Actual	MU
2.92 mm (male)	1	-40 μm	0 μm	-2 μm	1 μm
2.92 mm (male)	2	-40 μm	0 μm	-4 μm	1 μm



2. Static frequency accuracy

Frequency	Max. Deviation	Actual Deviation	MU
1 GHz	+/- 500 Hz	-394 Hz	1 Hz

3. Output harmonics



3.1 Harmonics PORT1

Source power level = 0 dBm, measurement uncertainty: {b}



3.2 Harmonics PORT2



Source power level = 0 dBm, measurement uncertainty: {b}



4. Maximum output power



4.1 Maximum output power PORT1

Power setting: +15 dBm



4.2 Maximum output power PORT2



Power setting: +15 dBm



5. Accuracy of output power



5.1 Accuracy of output power PORT1

Source power = -10 dBm



5.2 Accuracy of output power PORT2









Source power = -10 dBm









6. Output level linearity

6.1 Output level linearity PORT1



Reference level = -10 dBm

















6.2 Output level linearity PORT2

Reference level = -10 dBm









7. Input power measurement accuracy



7.1 Input power measurement accuracy PORT1

Test level = -10 dBm



7.2 Input power measurement accuracy PORT2


Test level = -10 dBm



8. Input power linearity (low level)


8.1 Input power linearity (low level) PORT1

Referred to a nominal input power level of -10 dBm



8.2 Input power linearity (low level) PORT2



Referred to a nominal input power level of -10 dBm














9. Input power linearity (high level)

9.1 Input power linearity PORT1 (high level)

Referred to a nominal input power level of -10 dBm




















9.2 Input power linearity PORT2 (high level)

Referred to a nominal input power level of -10 dBm









10. Trace noise magnitude



10.1 Trace noise magnitude PORT1

Measurement: S11 total reflection RMS StdDev, CW mode, source power = 0 dBm, points = 201, BW = 1 kHz ($f < 100$ kHz), BW = 10 kHz ($f \geq 100$ kHz), MU: {c}



10.2 Trace noise magnitude PORT2



Measurement: S22 total reflection RMS StdDev, CW mode, source power = 0 dBm, points = 201, BW = 1 kHz ($f < 100$ kHz), BW = 10 kHz ($f \geq 100$ kHz), MU: {c}



11. Trace noise phase



11.1 Trace noise phase PORT1

Measurement: S11 total reflection RMS StdDev, CW mode, source power = 0 dBm, points = 201, BW = 1 kHz ($f < 100$ kHz), BW = 10 kHz ($f \geq 100$ kHz), MU: {c}



11.2 Trace noise phase PORT2



Measurement: S22 total reflection RMS StdDev, CW mode, source power = 0 dBm, points = 201, BW = 1 kHz ($f < 100$ kHz), BW = 10 kHz ($f \geq 100$ kHz), MU: {c}



12. Dynamic range



12.1 Dynamic range S21

Measurement: S21, source power: 20 dBm, all ports are terminated by SHORTs, MU: {a}



12.2 Dynamic range S12

Measurement: S12, source power: 20 dBm, all ports are terminated by SHORTs, MU: {a}



13. Rear connectors and signals

13.1 Reference output

DLL	DUL	Actual	MU
+5 dBm	+13 dBm	7.7 dBm	0.3 dB

14. BIAS - functional test

DC input voltage = 10 V

Test point		Max. deviation	Actual	MU
Input (BNC)	1	+/-0.1 V	0.00 V	0.01 V
Output Port	1	+/-0.1 V	0.00 V	0.01 V
Input (BNC)	2	+/-0.1 V	0.00 V	0.01 V
Output Port	2	+/-0.1 V	0.00 V	0.01 V