



6600 Washington Avenue S.
Eden Prairie, Minnesota 55344
T: 1.800.328.8602

www.starkey.com

November 27, 2018

Federal Communications Commission
445 12th Street SW
Washington, DC 20554
Attention: Application Examiner

Reference:	Certification:	Part 15.247 Intentional Radiator
	Applicant:	Starkey Laboratories, Inc.
	Equipment:	Low Power Device (2400-2483.5 MHz)
	Model #:	Livio micro RIC 312 Hearing Aid
		FCC ID: EOA-24LIVIAM312

Dear Sir/Madam:

Enclosed please find an FCC EOA-24LIVIAM312 certification application. The device is a wireless hearing aid intended for use by the hearing impaired. The wireless features are used during the initial fitting of a hearing aid to program the hearing aid, and used during normal usage to communicate control information and digital audio to and from any properly configured Bluetooth Low Energy compliant accessory and between a binaural pair of Livio hearing aids.

Further details and supporting documents are attached.

Should you have any questions, please feel free to contact the undersigned.

Sincerely,

A handwritten signature in black ink that reads "William J. Mitchell". The signature is written in a cursive style with a large, stylized "W" and "M".

William J. Mitchell, PE
Principal Regulatory Affairs Engineer
952-918-5683
bill_mitchell@starkey.com