FDA Information FCC Information

## **TECHNICAL DATA**

Multiflex Tinnitus Technology Maximum Output = 87 dB SPL (typical) when measured in a 2cc coupler per ANSI S3.22 or IEC 60118-7.

## REGULATORY NOTICES

RIC 13

FCC ID: EOA-24HALO2R13 IC: 6903A-24HALO2R13

RIC 312

FCC ID: EOA-24HALO2R312 IC: 6903A-24HALO2R312

## **FCC Notice**

This device complies with part 15 of the FCC rules and with RSS-247. Operation is subject to the following two conditions: (1) This device may not cause harmful interference, and (2) this device must accept any interference received, including interference that may cause undesired operation of the device.

Note: The manufacturer is not responsible for any radio or TV interference caused by unauthorized modifications to this equipment. Such modifications could void the user's authority to operate the equipment.

Hereby, Starkey® Hearing Technologies declares that the products listed above are in compliance with the essential requirements and other relevant provisions of Directive 1999/5/EC. A copy of the Declaration of Conformity can be obtained from the addresses on the following page.

52