 <b>Medtronic</b> MEDTRONIC CONFIDENTIAL	Specification Number		Revision	Page
	DL01658		02	1 of 2
	Deliverable	Device Regulatory Label		
	Title	DEVICE REGULATORY LABEL – PAL CONTROLLER US, CAN		

## 1 Purpose

This record describes the implementation for a Device Regulatory Label PAL Controller US and Canada.

## 2 Scope

This document defines the minimum design specifications for the Device Regulatory Label PAL Controller US and Canada. All specifications are located in Section 3; all other sections are for reference purposes only.

## 3 Specifications

This section contains the artwork for the Device Regulatory Label.

3.1 Graphics: Per the supplied artwork file (zip file). All borders are represented by line-art.

3.2 Colors:

- a. Background, Black per the supplied color chip Infinity LTL number RX32001.
- b. Graphics, Reverse Process White, Blue PMS 3005c and Red PMS 1807c. See artwork for color location.

3.3 Label Materials and Dimensions: Per CMP02869.

3.4 Shelf Life: Upon receipt, this component shall have a minimum of 12 months of shelf life remaining determined by the Use-Before-Date.

3.5 Labeling: Label each package with part number, revision and the Use-Before-Date.

3.6 Packaging: Bulk package in polybags as to prevent damage and foreign material contamination.


◇ 3.7 Workmanship: Graphic content to be legible and free of ink voids and smudges.


### **Label Artwork – For reference only not to scale**

---

This document is the property of Medtronic, Inc. and must be accounted for. Information herein is confidential. Do not reproduce it, reveal it to unauthorized persons or send it outside Medtronic without proper authorization.

Form FRM001244, Text-based Physical Specification, Revision 1B

 <b>Medtronic</b>  MEDTRONIC CONFIDENTIAL	Specification Number		Revision	Page
	DL01658		02	2 of 2
	Deliverable	Device Regulatory Label		
	Title	DEVICE REGULATORY LABEL – PAL CONTROLLER US, CAN		



4 Abbreviations, Acronyms, and Definitions

Term	Definition
PMS	Pantone Matching System for color

5 Change History

Revision	Description of Change
01	Initial Release
02	Delete humidity symbol from artwork

Appendix A

None

This document is the property of Medtronic, Inc. and must be accounted for. Information herein is confidential. Do not reproduce it, reveal it to unauthorized persons or send it outside Medtronic without proper authorization.

Form FRM001244, Text-based Physical Specification, Revision 1B