

Appendix B

PICTURES OF THE EVALUATION SETUP

 Page 52 of 69
 Project # UESB-TRU5885-cordless phone-3916

 51 Spectrum Way
 Tel. (613) 820-2730

 Nepean, Ontario, K2R 1E6
 Fax (613) 820 4161

 e-mail: info@aprel.com
 © APREL 2002

 This report shall not be reproduced, except in full, without the express written approval of APREL Laboratories.



RIGHT HEAD, CHEEK/TOUCH POSITION



Page 53 of 69 51 Spectrum Way Nepean, Ontario, K2R 1E6 e-mail: info@aprel.com This report shall not be reproduced, except in full, without the express written approval of APREL Laboratories.



RIGHT HEAD, 15° TILT POSITION



Page 54 of 69 51 Spectrum Way Nepean, Ontario, K2R 1E6 e-mail: info@aprel.com Project # UESB-TRU5885-cordless phone-3916 Tel. (613) 820-2730 Fax (613) 820 4161 © APREL 2002



LEFT HEAD, CHEEK/TOUCH POSITION



Page 55 of 69 51 Spectrum Way Nepean, Ontario, K2R 1E6 e-mail: info@aprel.com This report shall not be reproduced, except in full, without the express written approval of APREL Laboratories.



RIGHT HEAD, 15° TILT POSITION



Page 56 of 69 51 Spectrum Way Nepean, Ontario, K2R 1E6 e-mail: info@aprel.com Project # UESB-TRU5885-cordless phone-3916 Tel. (613) 820-2730 Fax (613) 820 4161 © APREL 2002



BODY-WORN POSITION



Page 57 of 69 51 Spectrum Way Nepean, Ontario, K2R 1E6 e-mail: info@aprel.com This report shall not be Project # UESB-TRU5885-cordless phone-3916 Tel. (613) 820-2730 Fax (613) 820 4161 © APREL 2002



APPENDIX C: VALIDATION SCAN





tile SAI Values

Figure 5. Contour Plot of 1 gram Validation Scan

Figure 6. Contour Plot of 10 gram Validation Scan

Frequency: 5800 MHz Tissue Type: Brain **Conversion Factor: 2.4** Input Power to Dipole: 0.1 W (Normalized to 1W) Distance from Dipole to Tissue: 10 mm Tissue Deapth: 15 cm

Measured 1 Gram SAR (W/Kg)	Target 1 Gram SAR (W/Kg)	Delta (%)	Measured 10 Gram SAR (W/Kg)	Target 1 Gram S (W/Kg)
136.9	132.3	+3.5	39.9	38.4

Measured 10 Gram SAR (W/Kg)	Target 10 Gram SAR (W/Kg)	Delta (%)
39.9	38.4	+3.9

Project # UESB-TRU5885-cordless phone-3916 Tel. (613) 820-2730 Fax (613) 820 4161 © APREL 2002



Appendix D: UNCERTAINTY BUDGET

Calculated Uncertainties		
Type of Uncertainty	Specific to	Uncertainty
Power variation due to battery condition	DUI	0%
Extrapolation due to depth measurement	Setup	3.8%
Conductivity	Setup	1.0%
Permitivity	Setup	2.0%
Probe Calibration	Setup	7.0%
Probe Positioning	Setup	1.0%
Probe Isotropicity	Setup	1.5%
Other Setup Uncertainty (Ambient)	Setup	3.0%
	18.0%	Expanded Uncertainty K=2



Appendix E: Probe Calibration

NCL CALIBRATION LABORATORIES

Calibration File No.: C-P-0258

CERTIFICATE OF CALIBRATION

It is certified that the equipment identified below has been calibrated in the **NCL CALIBRATION LABORATORIES** by qualified personnel following recognized procedures and using transfer standards traceable to NRC/NIST.

Equipment: Miniature Isotropic RF Probe 5.8 GHz

Manufacturer: APREL Laboratories Model No.: E-010 Serial No.: 163

Calibration Procedure: SSI/DRB-TP-D01-032 Project No: Probe Cal Internal

> Calibrated: May 8th 2002 Recalibration required: may 7th 2003 Released on: May 8th 2002

Released By:



Page 60 of 69 51 Spectrum Way Nepean, Ontario, K2R 1E6 e-mail: info@aprel.com Project # UESB-TRU5885-cordless phone-3916 Tel. (613) 820-2730 Fax (613) 820 4161 © APREL 2002



INTRODUCTION

This Calibration Report reproduces the results of the calibration performed in line with the SSI/DRB-TP-D01-032 E-Field Probe Calibration Procedure. The results contained within this report are for APREL E-Field Probe E-010 163.

REFERENCES

SSI/DRB-TP-D01-032 E-Field Probe Calibration Procedure IEEE P-1528 *DRAFT* "Recommended Practice for Determining the Peak Spatial-Average Specific Absorption Rate (SAR) in the Human Body Due to Wireless Communications Devices: Experimental Techniques" SSI-TP-014 Tissue Calibration Procedure

Conditions

Probe 163 was a new probe taken from stock prior to calibration.

Ambient Temperature of the Laboratory:	23 °C +/- 0.5 °C
Temperature of the Tissue:	22 °C +/- 0.5 °C



CALIBRATION RESULTS SUMMARY

Probe Type:	E-Field Probe E-010
Serial Number:	163
Frequency:	5.8 GHz
Sensor Offset:	2.4 mm
Sensor Length:	2.5 mm
Tip Enclosure:	Glass*
Tip Enclosure: Tip Diameter:	Glass* 7 mm
Tip Enclosure: Tip Diameter: Tip Length:	Glass* 7 mm 40 mm

*Resistive to recommended tissue recipes per IEEE-P1528

SENSITIVITY IN AIR

Channel X:	0.58 ìV/(V/m) ²
Channel Y:	0.58 iV/(V/m) ²
Channel Z:	0.58 ìV/(V/m) ²

Diode Compression Point:

76 mV



SENSITIVITY IN HEAD TISSUE

Frequency:		5.8 GHz	
Epsilon:	35.3(+/-10%)	Sigma:	5.27 S/m(+/-10%)
ConvF			
Channel X:	2.4		
Channel Y:	2.4		
Channel Z:	2.4		

Tissue sensitivity values were calculated using a load impedance of 5 M Ω .

Boundary Effect:

Uncertainty resulting from the boundary effect is less than 2% for the distance between the tip of the probe and the tissue boundary, when less than 2.6mm.

Spatial Resolution:

The measured probe tip diameter is 7 mm (+/- 0.01 mm) and therefore meets the requirements of SSI/DRB-TP-D01-032 for spatial resolution.



RECEIVING PATTERN 5.8 GHZ (AIR)





ISOTROPY ERROR 5.8 GHZ (AIR)



51 Spectrum Way Nepean, Ontario, K2R 1E6 e-mail: info@aprel.com

Page 65 of 69

Project # UESB-TRU5885-cordless phone-3916 Tel. (613) 820-2730 Fax (613) 820 4161 © APREL 2002





DYNAMIC RANGE





Video Bandwidth



Probe Frequency Characteristics

Video Bandwidth at 500 Hz	1 dB
Video Bandwidth at 1.02 KHz	2: 3 dB



CONVERSION FACTOR UNCERTAINTY ASSESSMENT

Frequency:		5.8 GHz	
Epsilon:	35.3 (+/-10%)	Sigma:	5.27 S/m (+/-10%)
ConvF			
Channel X:	2.4	7%(K=2)	
Channel Y:	2.4	7%(K=2)	
Channel Z:	2.4	7%(K=2)	

To minimize the uncertainty calculation all tissue sensitivity values were calculated using a load impedance of 5 M $_{\Omega}$.

Boundary Effect:

For a distance of 2.6mm the evaluated uncertainty (increase in the probe sensitivity) is less than 2%.



TEST EQUIPMENT

The test equipment used during Probe Calibration, manufacturer, model number and, current calibration status are listed in the following table:

Instrument	Manufacturer	Model Number	Asset Number Serial Number	Calibration Due
UniPhantom	APRFI		API -085	N/A
Universal Frame	APREL		APL-114	N/A
<1GHz TEM Cell			APL-1GHZ- TEM	Jan 2003
>1GHz-2GHz Tem Cell			APL-2GHZ- TEM	Jan 2003
ALIDX-500	APREL/IDX		-	March 2003
RF Amplifier	APREL		301467	October 2003
Signal Generator	HP		301463	November 2002
Power Meter	R&S		301451	September 2002
Power Sensor	R&S		301461	September 2002
Directional Coupler	HP		100251	October 2002
VNA	Anritsu		Z0107643 <i>TEMP</i>	August 2002
Slotted Line	HP		100195	N/A
Slotted Line Probe	APREL		APL-SLP- 001	December 2002
APREL D-835- S1			301463	March 2003
D-900-S1	APREL		301472	March 2003
D-1900-S1	APREL		301459	March 2003
Measuring Amplifier	B&K		100675	Feb 2003
Signal Generator	B&K		100677	Feb 2003