

## **Attestation Letter**

April 13, 2021

Reference: FCC ID: 2AC7Z-ESPWROOM32DC

To whom it may concern:

This letter is confirmation of our request to apply for a CI2PC for the Wi-Fi module supplied by ESPRESSIF SYSTEMS (SHANGHAI) CO., LTD, model ESP32-WROOM-32DC, approved under FCC ID: 2AC7Z-ESPWROOM32DC.

This evaluation concerns the use of an existing Wi-Fi module of ESPRESSIF SYSTEMS (SHANGHAI) CO., LTD in Neteera technologies Ltd. host "Respiratory and cardiac rate monitor".

This letter is provided to state Neteera technologies Ltd. has not performed any modification and not changed any characteristics of radio module, including increase output power, frequency range and antenna. This Class 2 Permissive Change addresses the exposure distance when the radio is housed in the Neteera technologies Ltd. host "Respiratory and cardiac rate monitor". Originally, the minimum exposure distance was 20 cm. The exposure distance now is 5 cm as the device is considered portable. The file "RF Env evaluation 41599 Wi-Fi" attached to this application.

Accompanying this letter is a letter from ESPRESSIF SYSTEMS (SHANGHAI) CO., LTD confirming their permission for us to carry out this change.

Sincerely yours,

Mr. Shahar Yaron VP product

Neteera technologies Ltd.