

Medtronic

Simplera™, Simplera Sync™
MMT-5100J, MMT-5120
RF ID: 033686

Radio Compliance Information

United States

FCC ID: OH2BLEGST1

This device complies with part 15 of the FCC Rules. Operation is subject to two conditions:

- (1) This device may not cause harmful interference, and
- (2) this device must accept any interference received, including interference that may cause undesired operation.

IMPORTANT: Do not change or modify the internal RF transmitter or antenna unless expressly approved by Medtronic Diabetes. Doing so could interfere with your authority to operate the device.

Australia and New Zealand



This product conforms to Australia and New Zealand Radio Requirements.

Singapore

**Complies with
IMDA Standards
DA103452**

This product complies with IMDA Standards.

该产品符合 IMDA 标准

Japan



Ⓡ 211-211210

Canada

IC:3408B-BLEGST1

This device contains license-exempt transmitter(s)/receiver(s) that comply with Innovation, Science and Economic Development Canada's license-exempt RSS(s). Operation is subject to the following two conditions:

- (1) this device may not cause interference, and
- (2) this device must accept any interference, including interference that may cause undesired operation of the device.

L'émetteur/récepteur exempt de licence contenu dans le présent appareil est conforme aux CNR d'Innovation, Sciences et Développement économique Canada applicables aux appareils radio exempts de licence. L'exploitation est autorisée aux deux conditions suivantes:

- 1) L'appareil ne doit pas produire de brouillage;
- 2) L'appareil doit accepter tout brouillage radioélectrique subi, même si le brouillage est susceptible d'en compromettre le fonctionnement.

United Kingdom

**UK
CA**

Hereby, Medtronic declares that this Radio Equipment is in compliance with the relevant statutory requirements.

The full text of the declaration of conformity is available at the following internet address:

www.MedtronicCompliance.eu

This product operates in the 2.4–2.4835 GHz band with a radiated output power of less than 1.10 mW e.i.r.p.



M056624C005_1

European Union/ Conformité
Européenne



(European Conformity)

EN Hereby, Medtronic declares
that this Radio Equipment
is in compliance with
Directive 2014/53/EU.

The full text of the EU
declaration of conformity is
available at the following
internet address:

www.MedtronicCompliance.eu

This product operates in the
2.4–2.4835 GHz band with a
radiated output power of less
than 1.10 mW e.i.r.p.

Medtronic



Medtronic MiniMed
18000 Devonshire Street
Northridge, CA 91325
USA
1 800 646 4633
+1 818 576 5555
www.medtronicdiabetes.com



EC REP
Medtronic B.V.
Earl Bakkenstraat 10
6422 PJ Heerlen
The Netherlands