

Appendix - Calibration Certificate for Test Equipment

For Above 40GHz:

	Description & Manufacturer	Model No.	Serial No.	Cal. Date	Cal. Due
01	Spectrum Analyzer Keysight	N9030A	MY55330160	Feb. 07, 2020	Feb. 06, 2021
02	Spectrum Analyzer Keysight	N9030B	MY57140953	Jun. 28, 2019	Jun. 27, 2020
03	*Horn Antenna (33~55GHz) OML	M22RH	110215-1	Oct. 17, 2017	Oct. 16, 2020
04	*Horn Antenna (50~75GHz) OML	M15HWD	110215-1	Oct. 17, 2017	Oct. 16, 2020
05	*Horn Antenna(75~110GHz) OML	M10RH	110215-1	Oct. 17, 2017	Oct. 16, 2020
06	N9029AV15-DC9 - 50-75 GHz VDI Standard Downconverter with 9VDC supply Keysight	SA Extension WR15	SAX 381	CoC	CoC
07	N9029AV10-DC9 - 75-110 GHz VDI Standard Downconverter with 9VDC supply Keysight	SA Extension WR10	SAX 378	CoC	CoC
08	*Millimeter-Wave Signal Generator Frequency Extension Module (50~75 GHz) Keysight	E8257DV15	SGX 050	CoC	CoC
09	*Millimeter-Wave Signal Generator Frequency Extension Module (75~110 GHz) Keysight	E8257DV10	SGX 051	CoC	CoC
10	*Power meter Keysight	E4417A	MY55276004	Oct. 17, 2019	Oct. 16, 2022
11	*Waveguide Power Sensor Keysight	V8486A	MY55170003	Oct. 17, 2017	Oct. 16, 2020
12	*Waveguide Power Sensor Keysight	W8486A	MY55230006	Oct. 17, 2017	Oct. 16, 2020
13	Antenna Tower & Turn Table CT	NA	NA	NA	NA

工服 NO. 20-02-BCC-060-01L

財團法人台灣電子檢驗中心



收件日期: Feb.06,2020

校正報告

Receipt Date

CALIBRATION REPORT

發行日期: Feb.11,2020

ELECTRONICS TESTING CENTER, TAIWAN

Report Issue Date

Page 1 of 16

顧客名稱 香港商立德國際商品試驗有限公司桃園分公司

Customer

顧客地址 新竹縣芎林鄉上山村文德路206巷49號

Address

供校儀器 ITEM CALIBRATED

儀器名稱: PXA Signal Analyzer

Nomenclature

製造商: KEYSIGHT

Manufacturer

型別: N9030A

Model No.

識別號碼: MY55330160

ID. No.

上述儀器經本實驗室校正，結果如內文。未經本實驗室書面許可，不得部份複製本報告，完整複製則不在此限。

The above instruments were calibrated by the laboratory and please refer to the content for the calibration results. This report may not be reproduced in part without the written permission of the laboratory, except for full reproduction.

校正資料: ☒ 僅量測 ☐ 調整

Calibration Information Calibration Only Adjusted

實際環境: 溫度: 22 °C 相對濕度: 48 %

Actual Environments Temperature Relative Humidity

環境管制條件: 溫度: (23 ± 2) °C, 相對濕度: (50 ± 10) %

Environmental Conditions

校正日期: Feb.07,2020

Calibration Date

建議再校日期: Feb.06,2021 註: 建議再校日期為應顧客要求列入。

Recommended Recalibration Date Note: The recommended recalibration date is agreed by the customer.

校正地點: 財團法人台灣電子檢驗中心校正實驗室

Laboratory Location

- 實驗室名稱地址: ☒ 1. 校正實驗室 33383 桃園市龜山區文明路29巷8號 TEL:+886-3-3280026
Laboratory Name and Address 2. 新竹校正實驗室 30075 新竹市科學園區園區二路47號205室 TEL:+886-3-5798806
3. 台中校正實驗室 40766 台中市西屯區福中二街8號2樓之2 TEL:+886-4-23584899
4. 台南校正實驗室 70248 台南市南區新和二路5號 TEL:+886-6-2925787#50,51

財團法人台灣電子檢驗中心特此證明報告內記載之受校儀器已與下方標準做過比較校正，用以校正之標準器可追溯至中華民國國家度量衡標準實驗室，美國標準及技術研究院，或其它國家之度量衡國家標準。本中心的校正服務均符合ISO/IEC 17025 之規定。

ETC hereby certifies that the equipment noted herein has been compared with the below listed standards. The Standards used to perform this calibration are traceable to NML/ROC,NIST/USA or other countries. The calibration services from ETC are capable of performing services in compliance with the requirements of ISO/IEC 17025.

財團法人台灣電子檢驗中心
ELECTRONICS TESTING CENTER,
TAIWAN



實驗室主管
Laboratory Head



報告簽署人
Signature



ELECTRONICS TESTING
CENTER, TAIWAN

CALIBRATION REPORT

Page 2 of 16

使用校正依據 CALIBRATION PROCEDURE USED

1. 「頻譜(信號)分析儀校正程序書」, B00-CD-142, 5th Edition。
2. 「測試接收機/信號(頻譜)分析儀校正程序書」, B00-CD-376, 5th Edition。

使用標準器及附配件 STANDARD AND ACCESSORIES USED

儀器名稱【廠牌/型號】 Nomenclature【Mfg./Model No.】	【識別號碼】 【ID. No.】	校正單位(認可編號) Cal. Source(ACRED Code)	報告號碼 Cal. Report No.	校正日期 Date Cal.	有效日期 Due Date
RF Step Attenuator 【R/S RSG】 【13050122-001】		R&S(DAKS D-K-15195-01-01)	D-K-15195-01-01 2018-03	2018/03/08	2021/09/07
EPM Series Power Meter 【HP E4419B】 【13050609-001】		NML(TAF N0688)	U190100A	2019/07/24	2021/01/23
Swept Signal Generator 【AGILENT 83650B】 【13051703-001】		ETC(TAF 0025)	19-06-BAC-512-16L	2019/07/12	2020/07/11
Signal Generator 【R&S/SMB100A】 【13051715-001】		R&S(DAKS D-K-15195-01-01)	532388	2019/11/26	2023/05/25
Power Sensor 【AGILENT 8482A】 【13053506-001】		KEYSIGHT(ANAB AC-1498)	1-11911589522-1	2019/11/14	2022/05/13
Power Sensor 【AGILENT 8487A】 【13053508-003】		KEYSIGHT(ANAB AC-1498)	1-12179698623-1	2020/01/15	2022/07/14
Universal Counter 【HP 53132A】 【13060804-001】		ETC(TAF 0025)	19-12-BAC-630-02L	2020/01/03	2020/07/02

校正報告

財團法人台灣電子檢驗中心

工 服NO. 20-02-BCC-060-01L

CALIBRATION REPORT

ELECTRONICS TESTING
CENTER, TAIWAN

Page 3 of 16

Test Description	Actual Value	Expanded Uncertainty
1. Freq. Readout Accuracy and Freq. Count Marker Accuracy:		
(a) Freq. Readout Accuracy		
(a.1) 1.5 GHz Center Freq.		
1 MHz SPAN	1.500000 GHz	1.2×10^{-6}
10 MHz SPAN	1.500000 GHz	1.2×10^{-5}
20 MHz SPAN	1.500000 GHz	1.2×10^{-5}
50 MHz SPAN	1.500000 GHz	1.2×10^{-5}
100 MHz SPAN	1.500000 GHz	1.2×10^{-4}
1 GHz SPAN	1.499 GHz	1.2×10^{-3}
(a.2) 4.0 GHz Center Freq.		
1 MHz SPAN	4.000000 GHz	1.2×10^{-6}
10 MHz SPAN	4.000000 GHz	1.2×10^{-5}
20 MHz SPAN	4.000000 GHz	1.2×10^{-5}
50 MHz SPAN	4.000000 GHz	1.2×10^{-5}
100 MHz SPAN	4.000000 GHz	1.2×10^{-4}
1 GHz SPAN	3.999 GHz	1.2×10^{-3}
(a.3) 9.0 GHz Center Freq.		
1 MHz SPAN	9.000000 GHz	1.2×10^{-5}
10 MHz SPAN	9.000000 GHz	1.2×10^{-5}
20 MHz SPAN	9.000000 GHz	1.2×10^{-5}
50 MHz SPAN	9.000000 GHz	1.2×10^{-4}
100 MHz SPAN	9.000000 GHz	1.2×10^{-3}
1 GHz SPAN	8.999 GHz	1.2×10^{-6}
(a.4) 16.0 GHz Center Freq.		
1 MHz SPAN	16.000000 GHz	1.2×10^{-5}
10 MHz SPAN	16.000000 GHz	1.2×10^{-5}
20 MHz SPAN	16.000000 GHz	1.2×10^{-4}
50 MHz SPAN	16.000000 GHz	1.2×10^{-3}
100 MHz SPAN	16.000000 GHz	1.2×10^{-6}
1 GHz SPAN	15.999 GHz	1.2×10^{-5}

校正報告

財團法人台灣電子檢驗中心

工 服NO. 20-02-BCC-060-01L

CALIBRATION REPORT

ELECTRONICS TESTING
CENTER, TAIWAN

Page 4 of 16

Test Description	Actual Value	Expanded Uncertainty
1. Freq. Readout Accuracy and Freq. Count Marker Accuracy: (@ Continued):		
(a)Freq. Readout Accuracy		
(a.5)21.0 GHz Center Freq.		
1 MHz SPAN	21.000000 GHz	1.2×10^{-6}
10 MHz SPAN	21.000000 GHz	1.2×10^{-5}
20 MHz SPAN	21.000000 GHz	1.2×10^{-5}
50 MHz SPAN	21.000000 GHz	1.2×10^{-5}
100 MHz SPAN	21.000000 GHz	1.2×10^{-4}
1 GHz SPAN	20.999 GHz	1.2×10^{-3}
(a.6)29.0 GHz Center Freq.		
1 MHz SPAN	29.000000 GHz	1.2×10^{-6}
10 MHz SPAN	29.000000 GHz	1.2×10^{-5}
20 MHz SPAN	29.000000 GHz	1.2×10^{-5}
50 MHz SPAN	29.000000 GHz	1.2×10^{-5}
100 MHz SPAN	29.000000 GHz	1.2×10^{-4}
1 GHz SPAN	29.000 GHz	1.2×10^{-3}
(a.7)35.0 GHz Center Freq.		
1 MHz SPAN	35.000000 GHz	1.2×10^{-5}
10 MHz SPAN	35.000000 GHz	1.2×10^{-5}
20 MHz SPAN	35.000000 GHz	1.2×10^{-5}
50 MHz SPAN	35.000000 GHz	1.2×10^{-4}
100 MHz SPAN	35.000000 GHz	1.2×10^{-3}
1 GHz SPAN	34.999 GHz	1.2×10^{-6}
(b)Freq. Count Marker Accuracy:		
1.5 GHz Center Freq.	1.500000000 GHz	5.0×10^{-8}
4.0 GHz Center Freq.	4.000000000 GHz	5.0×10^{-8}
9.0 GHz Center Freq.	9.000000000 GHz	5.0×10^{-8}
16.0 GHz Center Freq.	16.00000000 GHz	5.0×10^{-8}
21.0 GHz Center Freq.	21.00000000 GHz	5.0×10^{-8}
29.0 GHz Center Freq.	29.00000000 GHz	5.0×10^{-8}
35.0 GHz Center Freq.	35.00000000 GHz	5.0×10^{-8}

校正報告

財團法人台灣電子檢驗中心

工 服NO. 20-02-BCC-060-01L

CALIBRATION REPORT

ELECTRONICS TESTING
CENTER, TAIWAN

Page 5 of 16

Test Description	Actual Value	Expanded Uncertainty
2. Freq. Span Accuracy:		
1 kHz Span	0.00 %	0.15 %
2 kHz Span	0.00 %	0.15 %
5 kHz Span	0.00 %	0.15 %
10 kHz Span	0.00 %	0.15 %
20 kHz Span	0.00 %	0.15 %
50 kHz Span	0.00 %	0.15 %
100 kHz Span	0.00 %	0.15 %
200 kHz Span	0.00 %	0.15 %
500 kHz Span	0.00 %	0.15 %
1 MHz Span	0.00 %	0.15 %
2 MHz Span	0.00 %	0.15 %
5 MHz Span	0.00 %	0.15 %
10 MHz Span	0.00 %	0.15 %
20 MHz Span	0.00 %	0.15 %
50 MHz Span	0.00 %	0.15 %
100 MHz Span	0.00 %	0.15 %
200 MHz Span	0.00 %	0.15 %
500 MHz Span	0.00 %	0.15 %
1 GHz Span	0.00 %	0.15 %
2 GHz Span	0.00 %	0.15 %

校正報告

財團法人台灣電子檢驗中心

工 服NO. 20-02-BCC-060-01L

CALIBRATION REPORT

ELECTRONICS TESTING
CENTER, TAIWAN

Page 6 of 16

Test Description	Actual Value	Expanded Uncertainty
3. Image, Multiple, and Out-of-Band Response:		
(a)2 GHz Center Freq.		
2021.4 MHz	-98.33 dBc	1.4 dB
2621.4 MHz	-98.26 dBc	0.96 dB
2321.4 MHz	-99.44 dBc	0.98 dB
2600.0 MHz	-98.96 dBc	1.1 dB
7910.7 MHz	-98.73 dBc	1.1 dB
9821.4 MHz	-98.19 dBc	1.2 dB
(b)4 GHz Center Freq.		
4021.4 MHz	-94.12 dBc	0.95 dB
4621.4 MHz	-95.40 dBc	0.90 dB
4321.4 MHz	-95.47 dBc	0.91 dB
4600.0 MHz	-95.44 dBc	0.97 dB
8310.7 MHz	-95.55 dBc	0.90 dB
8932.1 MHz	-95.71 dBc	1.2 dB
(c)9 GHz Center Freq.		
9021.4 MHz	-93.89 dBc	0.83 dB
9621.4 MHz	-94.86 dBc	1.2 dB
9321.4 MHz	-94.31 dBc	1.2 dB
9600.0 MHz	-95.14 dBc	1.4 dB
18310.7 MHz	-94.69 dBc	0.91 dB
18932.1 MHz	-94.37 dBc	0.92 dB
(d)15 GHz Center Freq.		
15021.400 MHz	-94.38 dBc	1.1 dB
15621.400 MHz	-95.97 dBc	1.2 dB
22655.350 MHz	-96.45 dBc	0.94 dB
23276.750 MHz	-95.17 dBc	1.4 dB
7344.650 MHz	-95.27 dBc	1.4 dB
7966.050 MHz	-94.05 dBc	1.1 dB
(e)20 GHz Center Freq.		
20021.400 MHz	-93.26 dBc	1.4 dB
20621.400 MHz	-92.46 dBc	1.0 dB
15543.725 MHz	-92.38 dBc	1.6 dB
25699.075 MHz	-92.78 dBc	1.1 dB
9844.650 MHz	-93.01 dBc	0.93 dB
10466.050 MHz	-93.03 dBc	1.4 dB

校正報告

財團法人台灣電子檢驗中心

工 服NO. 20-02-BCC-060-01L

CALIBRATION REPORT

ELECTRONICS TESTING
CENTER, TAIWAN

Page 7 of 16

Test Description	Actual Value	Expanded Uncertainty
3. Image, Multiple, and Out-of-Band Response(Continued):		
(f)29 GHz Center Freq.		
28378.600 MHz	-90.40 dBc	1.2 dB
28978.600 MHz	-90.42 dBc	1.3 dB
24450.925 MHz	-91.56 dBc	1.2 dB
28700.000 MHz	-89.50 dBc	1.2 dB
16455.350 MHz	-89.62 dBc	1.3 dB
35272.325 MHz	-89.77 dBc	1.3 dB
(g)35 GHz Center Freq.		
35021.400 MHz	-88.59 dBc	1.5 dB
35621.400 MHz	-87.58 dBc	1.4 dB
33093.725 MHz	-88.64 dBc	1.4 dB
35321.400 MHz	-88.49 dBc	1.3 dB
8744.538 MHz	-88.45 dBc	1.3 dB
15544.650 MHz	-88.55 dBc	1.3 dB
4. Frequency accuracy:	10.0000002 MHz	3.5×10^{-8}
5. IF bandwidth level accuracy Check:		
100 Hz	0.00 dB	0.32 dB
300 Hz	0.00 dB	0.32 dB
1 kHz	0.00 dB	0.32 dB
3 kHz	0.00 dB	0.32 dB
10 kHz	0.00 dB(Ref.)	----
30 kHz	0.00 dB	0.32 dB
100 kHz	0.00 dB	0.32 dB
300 kHz	0.00 dB	0.32 dB
1 MHz	0.00 dB	0.32 dB
3 MHz	0.01 dB	0.32 dB

校正報告

財團法人台灣電子檢驗中心

工 服NO. 20-02-BCC-060-01L

CALIBRATION REPORT

ELECTRONICS TESTING
CENTER, TAIWAN

Page 8 of 16

Test Description	Actual Value	Expanded Uncertainty
6. IF bandwidth (3 dB) Check:		
Bandwidth		
100 Hz	100.1 Hz	4.4 %
300 Hz	300.0 Hz	4.4 %
1 kHz	1.000 kHz	4.4 %
3 kHz	3.000 kHz	4.4 %
10 kHz	10.00 kHz	4.4 %
30 kHz	30.08 kHz	4.4 %
100 kHz	100.1 kHz	4.4 %
300 kHz	300.7 kHz	4.4 %
1 MHz	998.0 kHz	4.4 %
3 MHz	2.987 MHz	4.4 %
7. IF Bandwidths (3 dB)		
Shape factor Check:		
100 Hz	3.80	6.1 %
300 Hz	4.34	6.1 %
1 kHz	4.37	6.1 %
3 kHz	4.04	6.1 %
10 kHz	4.17	6.1 %
30 kHz	4.27	6.1 %
100 kHz	4.13	6.1 %
300 kHz	4.08	6.1 %
1 MHz	4.04	6.1 %
3 MHz	4.02	6.1 %

校正報告

財團法人台灣電子檢驗中心

工 服NO. 20-02-BCC-060-01L

CALIBRATION REPORT

ELECTRONICS TESTING
CENTER, TAIWAN

Page 9 of 16

Test Description	Actual Value	Expanded Uncertainty
8. Noise Display Check :		
9 kHz	-146.1 dBm	0.76 dB
100 kHz	-148.2 dBm	0.76 dB
999 kHz	-148.4 dBm	0.76 dB
10.99 MHz	-150.3 dBm	0.76 dB
19.99 MHz	-155.1 dBm	0.76 dB
49.99 MHz	-155.5 dBm	0.76 dB
99.99 MHz	-158.4 dBm	0.76 dB
199.9 MHz	-157.3 dBm	0.76 dB
499.9 MHz	-155.6 dBm	0.76 dB
999.9 MHz	-152.6 dBm	0.76 dB
1499 MHz	-152.1 dBm	0.88 dB
1999 MHz	-151.7 dBm	0.88 dB
2499 MHz	-150.9 dBm	0.88 dB
2999 MHz	-150.3 dBm	0.88 dB
3099 MHz	-148.9 dBm	0.88 dB
3499 MHz	-147.4 dBm	0.88 dB
3999 MHz	-148.2 dBm	0.88 dB
4499 MHz	-149.0 dBm	0.88 dB
4999 MHz	-148.4 dBm	0.88 dB
5499 MHz	-148.1 dBm	0.88 dB
5999 MHz	-147.3 dBm	0.88 dB
6499 MHz	-146.5 dBm	0.88 dB
6999 MHz	-146.9 dBm	0.88 dB
7999 MHz	-147.3 dBm	0.88 dB
8999 MHz	-146.8 dBm	0.88 dB
9999 MHz	-147.1 dBm	0.88 dB
10999 MHz	-144.6 dBm	0.88 dB
11999 MHz	-147.1 dBm	0.88 dB
12999 MHz	-146.0 dBm	0.88 dB
13999 MHz	-145.7 dBm	0.88 dB
14999 MHz	-144.8 dBm	0.88 dB
15999 MHz	-144.2 dBm	0.88 dB
16999 MHz	-143.7 dBm	0.88 dB

校正報告

財團法人台灣電子檢驗中心

工 服NO. 20-02-BCC-060-01L

CALIBRATION REPORT

ELECTRONICS TESTING
CENTER, TAIWAN

Page 10 of 16

Test Description	Actual Value	Expanded Uncertainty
8. Noise Display Check (@ Continued):		
17999 MHz	-144.1 dBm	0.88 dB
18999 MHz	-145.3 dBm	0.88 dB
19999 MHz	-146.1 dBm	0.88 dB
20999 MHz	-147.2 dBm	0.88 dB
21999 MHz	-146.9 dBm	0.88 dB
22999 MHz	-147.1 dBm	0.88 dB
23999 MHz	-146.9 dBm	0.88 dB
24999 MHz	-146.1 dBm	0.88 dB
25999 MHz	-145.4 dBm	0.88 dB
26999 MHz	-145.7 dBm	0.88 dB
27999 MHz	-145.8 dBm	0.88 dB
28999 MHz	-144.6 dBm	0.88 dB
29999 MHz	-145.3 dBm	0.88 dB
30999 MHz	-143.1 dBm	0.88 dB
31999 MHz	-146.3 dBm	0.88 dB
32999 MHz	-147.0 dBm	0.88 dB
33999 MHz	-146.5 dBm	0.88 dB
34999 MHz	-146.8 dBm	0.88 dB
35999 MHz	-145.3 dBm	0.88 dB
36999 MHz	-143.4 dBm	0.88 dB
37999 MHz	-142.7 dBm	0.88 dB
38999 MHz	-141.2 dBm	0.88 dB
39999 MHz	-139.2 dBm	0.88 dB

校正報告

財團法人台灣電子檢驗中心

工服NO. 20-02-BCC-060-01L

CALIBRATION REPORT

ELECTRONICS TESTING
CENTER, TAIWAN

Page 11 of 16

Test Description	Actual Value	Expanded Uncertainty
9. Frequency response Check:		
RF Attenuation 10 dB		
100 kHz	0.12 dB	0.66 dB
1 MHz	0.11 dB	0.66 dB
10 MHz	0.08 dB	0.66 dB
50 MHz	0.13 dB	0.66 dB
100 MHz	0.13 dB	0.66 dB
200 MHz	0.12 dB	0.66 dB
300 MHz	0.13 dB	0.66 dB
400 MHz	0.14 dB	0.66 dB
500 MHz	0.17 dB	0.66 dB
600 MHz	0.12 dB	0.66 dB
700 MHz	0.13 dB	0.66 dB
800 MHz	0.15 dB	0.66 dB
900 MHz	0.17 dB	0.66 dB
1000 MHz	0.19 dB	0.66 dB
1500 MHz	0.21 dB	0.66 dB
2000 MHz	0.22 dB	0.66 dB
2500 MHz	0.21 dB	0.66 dB
2990 MHz	0.23 dB	0.66 dB
3010 MHz	0.20 dB	0.70 dB
3500 MHz	0.20 dB	0.70 dB
4000 MHz	-0.22 dB	0.70 dB
4500 MHz	-0.49 dB	0.70 dB
5000 MHz	-0.65 dB	0.70 dB
5500 MHz	-0.41 dB	0.70 dB
6000 MHz	-0.31 dB	0.70 dB
6500 MHz	-0.28 dB	0.70 dB
6990 MHz	-0.23 dB	0.70 dB
7100 MHz	-0.22 dB	0.70 dB
8000 MHz	-0.24 dB	0.70 dB
9000 MHz	-0.32 dB	0.70 dB
10000 MHz	-0.39 dB	0.70 dB
11000 MHz	-0.42 dB	0.70 dB
12000 MHz	-0.51 dB	0.70 dB
13000 MHz	-0.60 dB	0.70 dB

校正報告

財團法人台灣電子檢驗中心

工服NO. 20-02-BCC-060-01L

CALIBRATION REPORT

ELECTRONICS TESTING
CENTER, TAIWAN

Page 12 of 16

Test Description	Actual Value	Expanded Uncertainty
9. Frequency response Check(@ Continued):		
RF Attenuation 10 dB		
13500 MHz	-0.71 dB	0.70 dB
14000 MHz	-0.58 dB	0.70 dB
15000 MHz	-0.43 dB	0.70 dB
16000 MHz	-0.51 dB	0.70 dB
17000 MHz	-0.55 dB	0.70 dB
18000 MHz	-0.53 dB	0.70 dB
19000 MHz	-0.57 dB	0.98 dB
20000 MHz	-0.56 dB	0.98 dB
21000 MHz	-0.52 dB	0.98 dB
22000 MHz	-0.41 dB	0.98 dB
23000 MHz	-0.52 dB	0.98 dB
24000 MHz	-0.59 dB	0.98 dB
25000 MHz	-0.63 dB	0.98 dB
26000 MHz	0.59 dB	0.98 dB
27000 MHz	-0.57 dB	1.3 dB
28000 MHz	-0.49 dB	1.3 dB
29000 MHz	-0.53 dB	1.3 dB
29900 MHz	-0.51 dB	1.3 dB
31000 MHz	-0.52 dB	1.3 dB
32000 MHz	-0.42 dB	1.3 dB
33000 MHz	-0.53 dB	1.3 dB
34000 MHz	-0.58 dB	1.3 dB
35000 MHz	-0.60 dB	1.3 dB
36000 MHz	-0.52 dB	1.3 dB
37000 MHz	-0.69 dB	1.3 dB
38000 MHz	-0.88 dB	1.3 dB
39000 MHz	-0.96 dB	1.3 dB
39999 MHz	-1.11 dB	1.3 dB
42990 MHz	-1.45 dB	1.3 dB
44000 MHz	-0.43 dB	1.3 dB
45000 MHz	-0.13 dB	1.3 dB
46000 MHz	0.14 dB	1.3 dB
47000 MHz	0.12 dB	1.3 dB
48000 MHz	0.10 dB	1.3 dB
49000 MHz	0.24 dB	1.3 dB
49999 MHz	0.00 dB	1.3 dB

校正報告

財團法人台灣電子檢驗中心

工 服NO. 20-02-BCC-060-01L

CALIBRATION REPORT

ELECTRONICS TESTING
CENTER, TAIWAN

Page 13 of 16

Test Description	Actual Value	Expanded Uncertainty
9. Frequency response Check(@ Continued):		
RF Attenuation 20 dB		
100 kHz	0.11 dB	0.64 dB
1 MHz	0.12 dB	0.64 dB
10 MHz	0.09 dB	0.64 dB
50 MHz	0.08 dB	0.64 dB
100 MHz	0.07 dB	0.64 dB
200 MHz	0.05 dB	0.64 dB
300 MHz	0.06 dB	0.64 dB
400 MHz	0.09 dB	0.64 dB
500 MHz	0.08 dB	0.64 dB
600 MHz	0.06 dB	0.64 dB
700 MHz	0.05 dB	0.64 dB
800 MHz	0.06 dB	0.64 dB
900 MHz	0.08 dB	0.64 dB
1000 MHz	0.11 dB	0.64 dB
1500 MHz	0.14 dB	0.64 dB
2000 MHz	0.16 dB	0.64 dB
2500 MHz	0.19 dB	0.64 dB
2990 MHz	0.22 dB	0.64 dB
Frequency response Check (@ Continued):		
RF Attenuation 40 dB		
100 kHz	0.11 dB	0.62 dB
1 MHz	0.12 dB	0.62 dB
10 MHz	0.09 dB	0.62 dB
50 MHz	0.08 dB	0.62 dB
100 MHz	0.07 dB	0.62 dB
200 MHz	0.08 dB	0.62 dB
300 MHz	0.09 dB	0.62 dB
400 MHz	0.11 dB	0.62 dB
500 MHz	0.13 dB	0.62 dB
600 MHz	0.11 dB	0.62 dB
700 MHz	0.09 dB	0.62 dB
800 MHz	0.08 dB	0.62 dB
900 MHz	0.09 dB	0.62 dB
1000 MHz	0.12 dB	0.62 dB
1500 MHz	0.14 dB	0.62 dB
2000 MHz	0.16 dB	0.62 dB
2500 MHz	0.19 dB	0.62 dB
2990 MHz	0.22 dB	0.64 dB

校正報告

財團法人台灣電子檢驗中心

工 服NO. 20-02-BCC-060-01L

CALIBRATION REPORT

ELECTRONICS TESTING
CENTER, TAIWAN

Page 14 of 16

Test Description	Actual Value	Expanded Uncertainty
10. Display linearity Check:		
RBW 300 Hz		
10 dB	9.99 dB	0.64 dB
12 dB	7.99 dB	0.64 dB
14 dB	5.98 dB	0.64 dB
16 dB	3.98 dB	0.64 dB
18 dB	1.97 dB	0.64 dB
20 dB	0.00 dB(Ref.)	----
22 dB	-2.00 dB	0.64 dB
24 dB	-4.00 dB	0.64 dB
26 dB	-6.00 dB	0.64 dB
28 dB	-8.01 dB	0.64 dB
30 dB	-10.00 dB	0.64 dB
32 dB	-12.01 dB	0.64 dB
34 dB	-14.01 dB	0.64 dB
36 dB	-16.02 dB	0.64 dB
38 dB	-18.03 dB	0.64 dB
40 dB	-20.00 dB	0.64 dB
42 dB	-21.99 dB	0.64 dB
44 dB	-23.98 dB	0.64 dB
46 dB	-25.98 dB	0.64 dB
48 dB	-27.99 dB	0.64 dB
50 dB	-30.00 dB	0.64 dB
52 dB	-31.99 dB	0.64 dB
54 dB	-33.98 dB	0.64 dB
56 dB	-35.99 dB	0.64 dB
58 dB	-38.00 dB	0.64 dB
60 dB	-39.97 dB	0.64 dB
65 dB	-44.98 dB	0.64 dB
70 dB	-49.99 dB	0.64 dB
75 dB	-54.99 dB	0.64 dB
80 dB	-59.98 dB	0.64 dB
85 dB	-64.98 dB	0.64 dB
90 dB	-69.97 dB	0.64 dB
95 dB	-74.96 dB	0.64 dB
100 dB	-79.94 dB	0.64 dB

校正報告

財團法人台灣電子檢驗中心

工 服NO. 20-02-BCC-060-01L

CALIBRATION REPORT

ELECTRONICS TESTING
CENTER, TAIWAN

Page 15 of 16

Test Description	Actual Value	Expanded Uncertainty
10. Display linearity Check		
RBW 300 kHz:		
10 dB	9.99 dB	0.64 dB
12 dB	7.99 dB	0.64 dB
14 dB	5.98 dB	0.64 dB
16 dB	3.97 dB	0.64 dB
18 dB	1.96 dB	0.64 dB
20 dB	0.00 dB(Ref.)	
22 dB	-2.00 dB	0.64 dB
24 dB	-4.01 dB	0.64 dB
26 dB	-6.02 dB	0.64 dB
28 dB	-8.03 dB	0.64 dB
30 dB	-10.01 dB	0.64 dB
32 dB	-12.01 dB	0.64 dB
34 dB	-14.00 dB	0.64 dB
36 dB	-16.01 dB	0.64 dB
38 dB	-18.02 dB	0.64 dB
40 dB	-20.00 dB	0.64 dB
42 dB	-21.99 dB	0.64 dB
44 dB	-23.98 dB	0.64 dB
46 dB	-25.98 dB	0.64 dB
48 dB	-27.99 dB	0.64 dB
50 dB	-29.99 dB	0.64 dB
52 dB	-31.98 dB	0.64 dB
54 dB	-33.98 dB	0.64 dB
56 dB	-35.99 dB	0.64 dB
58 dB	-38.01 dB	0.64 dB
60 dB	-40.01 dB	0.64 dB
65 dB	-44.98 dB	0.64 dB
70 dB	-49.99 dB	0.64 dB
75 dB	-55.01 dB	0.64 dB
80 dB	-60.02 dB	0.64 dB

校正報告

財團法人台灣電子檢驗中心

工 服NO. 20-02-BCC-060-01L

CALIBRATION REPORT

ELECTRONICS TESTING
CENTER, TAIWAN

Page 16 of 16

Test Description	Actual Value	Expanded Uncertainty
11. Attenuator accuracy Check:		
0 dB	-9.98 dB	0.38 dB
10 dB	0.00 dB(Ref.)	-----
20 dB	10.00 dB	0.38 dB
30 dB	20.01 dB	0.38 dB
40 dB	30.00 dB	0.38 dB
50 dB	40.00 dB	0.38 dB
60 dB	50.01 dB	0.38 dB
70 dB	60.04 dB	0.38 dB
12. Reference level		
switching accuracy		
Reference level Check:		
0 dBm	10.00 dB	0.36 dB
-10 dBm	0.00 dB(Ref.)	-----
-20 dBm	-10.00 dB	0.36 dB
-30 dBm	-20.00 dB	0.36 dB
-40 dBm	-30.01 dB	0.36 dB
-50 dBm	-40.01 dB	0.36 dB
-11 dBm	-1.00 dB	0.36 dB
-12 dBm	-2.00 dB	0.36 dB
-13 dBm	-3.00 dB	0.36 dB
-14 dBm	-4.00 dB	0.36 dB
-15 dBm	-5.00 dB	0.36 dB
-16 dBm	-6.00 dB	0.36 dB
-17 dBm	-7.00 dB	0.36 dB
-18 dBm	-8.00 dB	0.36 dB
-19 dBm	-9.00 dB	0.36 dB

說明:

1.本校正報告內的項次1、2、6、7相對擴充不確定度評估與評估表示係依據「ISO Guide 98-3 量測不確定度表示方式指引」，相對擴充不確定度 $U = ku_c$ 其中 u_c 為相對組合標準不確定度， $k = 2.0$ ，為信賴水準 95 %之涵蓋因子。

本校正報告內的項次3、4、5、8、9、10、11、12擴充不確定度與評估表示係依據「ISO Guide 98-3 量測不確定度表示方式指引」，擴充不確定度 $U = ku_c$ 其中 u_c 為組合標準不確定度， $k = 2.0$ ，為信賴水準 95 %之涵蓋因子。

2.參考ANSI C63.2及CISPR 16-1-1相關規範。

工服 NO. 19-06-BAC-534-04L

財團法人台灣電子檢驗中心



收件日期: Jun.25,2019

校正報告

Receipt Date

CALIBRATION REPORT

發行日期: Jul.01,2019

ELECTRONICS TESTING CENTER, TAIWAN

Report Issue Date

Page 1 of 16

顧客名稱 香港商立德國際商品試驗有限公司桃園分公司

Customer

顧客地址 桃園市龜山區文化里華亞二路19號

Address

供校儀器 ITEM CALIBRATED

儀器名稱: PXA Signal Analyzer

Nomenclature

製造商: KEYSIGHT

Manufacturer

型別: N9030B

Model No.

識別號碼: MY57140953

ID. No.

上述儀器經本實驗室校正，結果如內文。未經本實驗室書面許可，不得部份複製本報告，完整複製則不在此限。

The above instruments were calibrated by the laboratory and please refer to the content for the calibration results. This report may not be reproduced in part without the written permission of the laboratory, except for full reproduction.

校正資料: ☒ 僅量測 ☐ 調整

Calibration Information Calibration Only Adjusted

實際環境: 溫度: 23 °C 相對濕度: 47 %

Actual Environments Temperature Relative Humidity

環境管制條件: 溫度: (23 ± 2) °C ; 相對濕度: (50 ± 10) %

Environmental Conditions

校正日期: Jun.28,2019

Calibration Date

建議再校日期: Jun.27,2020

註: 建議再校日期為應客戶要求列入。

Recommended Recalibration Date

Note: The recommended recalibration date is agreed by the customer.

校正地點: 財團法人台灣電子檢驗中心校正實驗室

Laboratory Location

實驗室名稱地址: ☒ 1. 校正實驗室 33383 桃園市龜山區文明路29巷8號 TEL:+886-3-3280026

Laboratory Name and 2. 新竹校正實驗室 30075 新竹市科學園區區二路47號205室 TEL:+886-3-5798806

Address 3. 台中校正實驗室 40766 台中市西屯區福中二街8號2樓之2 TEL:+886-4-23584899

4. 台南校正實驗室 70248 台南市南區新和二路5號 TEL:+886-6-2925787#50,51

財團法人台灣電子檢驗中心特此證明報告內記載之受校儀器已與下方標準做過比較校正，用以校正之標準器可追溯至中華民國國家度量衡標準實驗室，美國標準及技術研究院，或其它國家之度量衡國家標準。本中心的校正服務均符合ISO/IEC 17025 之規定。

ETC hereby certifies that the equipment noted herein has been compared with the below listed standards. The Standards used to perform this calibration are traceable to NML/ROC,NIST/USA or other countries. The calibration services from ETC are capable of performing services in compliance with the requirements of ISO/IEC 17025.

財團法人台灣電子檢驗中心


實驗室主管

報告簽署人

ELECTRONICS TESTING CENTER,
TAIWAN

Laboratory Head

Signature



使用校正依據 CALIBRATION PROCEDURE USED

1. 「頻譜(信號)分析儀校正程序書」, B00-CD-142, 6th Edition。
2. 「測試接收機/頻譜分析儀校正程序書」, B00-CD-376, 4th Edition。

使用標準器及附配件 STANDARD AND ACCESSORIES

儀器名稱【廠牌/型號】 Nomenclature【Mfg. /Model No.】	【識別號碼】 【ID. No.】	校正單位(認可編號) Cal. Source(ACRED Code)	報告號碼 Cal. Report No.	校正日期 Date Cal.	有效日期 Due Date
RF Step Attenuator 【R/S RSG】 【13050122-001】		R&S(DAKKS D-K-15195-01-01)	D-K-15195-01-01 2018-03	2018/03/08	2021/09/07
EPM Series Power Meter 【HP E4419B】 【13050609-001】		NML(TAF N0688)	U180018A	2018/03/21	2019/09/20
Swept Signal Generator 【AGILENT 83650B】 【13051703-001】		ETC(TAF 0025)	18-07-BAC-245-03L	2018/07/16	2019/07/15
Power Sensor 【AGILENT 8482A】 【13053506-001】		KEYSIGHT(ANAB AC-1498)	1-9328170856-1	2017/10/16	2020/04/15
Power Sensor 【AGILENT 8487A】 【13053508-003】		KEYSIGHT(ANAB AC-1498)	1-9230656623-1	2017/09/14	2020/03/13
Universal Counter 【HP 53132A】 【13060804-001】		ETC(TAF 0025)	19-01-BAC-002-19L	2019/01/08	2019/07/07

校正報告

財團法人台灣電子檢驗中心

工 服NO. 19-06-BAC-534-04L

CALIBRATION REPORT

ELECTRONICS TESTING
CENTER, TAIWAN

Page 3 of 16

Test Description	Actual Value	Expanded Uncertainty
1. Freq. Readout Accuracy and Freq. Count Marker Accuracy:		
(a) Freq. Readout Accuracy		
(a.1) 1.5 GHz Center Freq.		
1 MHz SPAN	1.500000 GHz	1.4×10^{-6}
10 MHz SPAN	1.500000 GHz	1.4×10^{-5}
20 MHz SPAN	1.500000 GHz	2.1×10^{-5}
50 MHz SPAN	1.500000 GHz	2.7×10^{-5}
100 MHz SPAN	1.500000 GHz	2.5×10^{-4}
1 GHz SPAN	1.499 GHz	1.4×10^{-3}
(a.2) 4.0 GHz Center Freq.		
1 MHz SPAN	4.000000 GHz	1.3×10^{-6}
10 MHz SPAN	4.000000 GHz	1.3×10^{-5}
20 MHz SPAN	4.000000 GHz	1.5×10^{-5}
50 MHz SPAN	4.000000 GHz	1.5×10^{-5}
100 MHz SPAN	4.000000 GHz	1.3×10^{-4}
1 GHz SPAN	3.999 GHz	1.3×10^{-3}
(a.3) 9.0 GHz Center Freq.		
1 MHz SPAN	9.000000 GHz	1.3×10^{-6}
10 MHz SPAN	9.000000 GHz	1.3×10^{-5}
20 MHz SPAN	9.000000 GHz	1.3×10^{-5}
50 MHz SPAN	9.000000 GHz	1.3×10^{-5}
100 MHz SPAN	9.000000 GHz	1.3×10^{-4}
1 GHz SPAN	8.999 GHz	1.3×10^{-3}
(a.4) 16.0 GHz Center Freq.		
1 MHz SPAN	16.000000 GHz	1.3×10^{-6}
10 MHz SPAN	16.000000 GHz	1.3×10^{-5}
20 MHz SPAN	16.000000 GHz	1.3×10^{-5}
50 MHz SPAN	16.000000 GHz	1.3×10^{-5}
100 MHz SPAN	16.000000 GHz	1.3×10^{-4}
1 GHz SPAN	15.999 GHz	1.3×10^{-3}

校正報告

財團法人台灣電子檢驗中心

工 服NO. 19-06-BAC-534-04L

CALIBRATION REPORT

ELECTRONICS TESTING
CENTER, TAIWAN

Page 4 of 16

Test Description	Actual Value	Expanded Uncertainty
1. Freq. Readout Accuracy and Freq. Count Marker Accuracy: (@ Continued):		
(a)Freq. Readout Accuracy		
(a.5)21.0 GHz Center Freq.		
1 MHz SPAN	21.000000 GHz	1.3×10^{-6}
10 MHz SPAN	21.000000 GHz	1.3×10^{-5}
20 MHz SPAN	21.000000 GHz	1.3×10^{-5}
50 MHz SPAN	21.000000 GHz	1.3×10^{-5}
100 MHz SPAN	21.000000 GHz	1.3×10^{-4}
1 GHz SPAN	20.999 GHz	1.3×10^{-3}
(a.6)29.0 GHz Center Freq.		
1 MHz SPAN	29.000000 GHz	1.3×10^{-6}
10 MHz SPAN	29.000000 GHz	1.3×10^{-5}
20 MHz SPAN	29.000000 GHz	1.3×10^{-5}
50 MHz SPAN	29.000000 GHz	1.3×10^{-5}
100 MHz SPAN	29.000000 GHz	1.3×10^{-4}
1 GHz SPAN	28.999 GHz	1.3×10^{-3}
(a.7)35.0 GHz Center Freq.		
1 MHz SPAN	35.000000 GHz	1.3×10^{-6}
10 MHz SPAN	35.000000 GHz	1.3×10^{-5}
20 MHz SPAN	35.000000 GHz	1.3×10^{-5}
50 MHz SPAN	35.000000 GHz	1.3×10^{-5}
100 MHz SPAN	35.000000 GHz	1.3×10^{-4}
1 GHz SPAN	34.999 GHz	1.3×10^{-3}
(b)Freq. Count Marker Accuracy:		
1.5 GHz Center Freq.	1.500000000 GHz	1.3×10^{-7}
4.0 GHz Center Freq.	4.000000000 GHz	1.3×10^{-7}
9.0 GHz Center Freq.	9.000000000 GHz	1.3×10^{-7}
16.0 GHz Center Freq.	16.00000000 GHz	1.3×10^{-7}
21.0 GHz Center Freq.	21.00000000 GHz	1.3×10^{-7}
29.0 GHz Center Freq.	29.00000000 GHz	1.3×10^{-7}
35.0 GHz Center Freq.	35.00000000 GHz	1.3×10^{-7}

校正報告

財團法人台灣電子檢驗中心

工 服NO. 19-06-BAC-534-04L

CALIBRATION REPORT

ELECTRONICS TESTING
CENTER, TAIWAN

Page 5 of 16

Test Description	Actual Value	Expanded Uncertainty
2. Freq. Span Accuracy:		
1 kHz Span	0.00 %	1.3 %
2 kHz Span	0.00 %	1.3 %
5 kHz Span	0.00 %	1.3 %
10 kHz Span	0.00 %	1.3 %
20 kHz Span	0.00 %	1.3 %
50 kHz Span	0.00 %	1.3 %
100 kHz Span	0.00 %	1.3 %
200 kHz Span	0.00 %	1.3 %
500 kHz Span	0.00 %	1.3 %
1 MHz Span	0.00 %	1.3 %
2 MHz Span	0.00 %	1.3 %
5 MHz Span	0.00 %	6.3 %
10 MHz Span	0.00 %	6.3 %
20 MHz Span	0.00 %	6.3 %
50 MHz Span	0.00 %	6.3 %
100 MHz Span	0.00 %	6.3 %
200 MHz Span	0.00 %	6.3 %
500 MHz Span	0.00 %	6.3 %
1 GHz Span	0.00 %	6.3 %
2 GHz Span	0.00 %	6.3 %

校正報告

財團法人台灣電子檢驗中心

工 服NO. 19-06-BAC-534-04L

CALIBRATION REPORT

ELECTRONICS TESTING
CENTER, TAIWAN

Page 6 of 16

Test Description	Actual Value	Expanded Uncertainty
3. Image, Multiple, and Out-of-Band Response:		
(a)2 GHz Center Freq.		
2021.4 MHz	-97.73 dBc	2.1 dB
2621.4 MHz	-100.20 dBc	2.0 dB
2321.4 MHz	-98.46 dBc	2.1 dB
2600.0 MHz	-99.55 dBc	2.1 dB
7910.7 MHz	-99.72 dBc	2.1 dB
9821.4 MHz	-98.78 dBc	2.1 dB
(b)4 GHz Center Freq.		
4021.4 MHz	-92.57 dBc	2.1 dB
4621.4 MHz	-93.27 dBc	2.0 dB
4321.4 MHz	-93.33 dBc	2.1 dB
4600.0 MHz	-93.43 dBc	2.1 dB
8310.7 MHz	-93.34 dBc	2.0 dB
8932.1 MHz	-92.40 dBc	2.1 dB
(c)9 GHz Center Freq.		
9021.4 MHz	-95.16 dBc	2.0 dB
9621.4 MHz	-94.85 dBc	2.1 dB
9321.4 MHz	-94.46 dBc	2.2 dB
9600.0 MHz	-94.43 dBc	2.1 dB
18310.7 MHz	-94.29 dBc	2.1 dB
18932.1 MHz	-94.33 dBc	2.2 dB
(d)15 GHz Center Freq.		
15021.400 MHz	-93.13 dBc	2.1 dB
15621.400 MHz	-93.43 dBc	2.1 dB
22655.350 MHz	-93.45 dBc	2.1 dB
23276.750 MHz	-93.45 dBc	2.1 dB
7344.650 MHz	-93.50 dBc	2.1 dB
7966.050 MHz	-93.26 dBc	2.0 dB
(e)20 GHz Center Freq.		
20021.400 MHz	-88.54 dBc	2.1 dB
20621.400 MHz	-89.43 dBc	2.1 dB
15543.725 MHz	-90.47 dBc	2.1 dB
25699.075 MHz	-90.33 dBc	2.1 dB
9844.650 MHz	-90.33 dBc	2.1 dB
10466.050 MHz	-90.63 dBc	2.1 dB

校正報告

財團法人台灣電子檢驗中心

工 服NO. 19-06-BAC-534-04L

CALIBRATION REPORT

ELECTRONICS TESTING
CENTER, TAIWAN

Page 7 of 16

Test Description	Actual Value	Expanded Uncertainty
3. Image, Multiple, and Out-of-Band Response: (@Continued):		
(f)29 GHz Center Freq.		
28378.600 MHz	-90.08 dBc	2.0 dB
28978.600 MHz	-90.24 dBc	2.1 dB
24450.925 MHz	-90.24 dBc	2.0 dB
28700.000 MHz	-90.69 dBc	2.1 dB
16455.350 MHz	-90.51 dBc	2.1 dB
35272.325 MHz	-90.54 dBc	2.1 dB
(g)35 GHz Center Freq.		
35021.400 MHz	-85.60 dBc	2.1 dB
35621.400 MHz	-85.34 dBc	2.0 dB
33093.725 MHz	-85.46 dBc	2.2 dB
35321.400 MHz	-85.39 dBc	2.1 dB
8744.538 MHz	-85.43 dBc	2.1 dB
15544.650 MHz	-86.07 dBc	2.0 dB
4. Frequency accuracy:	10.0000000 MHz	1.2×10^{-7}
5. IF bandwidth level accuracy Check:		
100 Hz	0.00 dB	0.26 dB
300 Hz	0.00 dB	0.26 dB
1 kHz	0.00 dB	0.26 dB
3 kHz	0.01 dB	0.26 dB
10 kHz	0.00 dB(Ref.)	----
30 kHz	0.01 dB	0.26 dB
100 kHz	0.01 dB	0.26 dB
300 kHz	0.01 dB	0.26 dB
1 MHz	0.01 dB	0.26 dB
3 MHz	0.02 dB	0.26 dB

校正報告

財團法人台灣電子檢驗中心

工 服NO. 19-06-BAC-534-04L

CALIBRATION REPORT

ELECTRONICS TESTING
CENTER, TAIWAN

Page 8 of 16

Test Description	Actual Value	Expanded Uncertainty
6. IF bandwidth (3 dB) Check:		
Bandwidth		
100 Hz	100.0 Hz	4.4 %
300 Hz	300.5 Hz	4.4 %
1 kHz	1.002 kHz	4.4 %
3 kHz	3.011 kHz	4.4 %
10 kHz	10.02 kHz	4.4 %
30 kHz	30.08 kHz	4.4 %
100 kHz	100.7 kHz	4.4 %
300 kHz	301.7 kHz	4.4 %
1 MHz	1.000 MHz	4.4 %
3 MHz	2.973 MHz	4.4 %
7. IF Bandwidths (3 dB)		
Shape factor Check:		
100 Hz	3.94	6.1 %
300 Hz	3.96	6.1 %
1 kHz	4.00	6.1 %
3 kHz	4.01	6.1 %
10 kHz	4.08	6.1 %
30 kHz	4.02	6.1 %
100 kHz	3.99	6.1 %
300 kHz	3.97	6.1 %
1 MHz	4.03	6.1 %
3 MHz	4.07	6.1 %

校正報告

財團法人台灣電子檢驗中心

工 服NO. 19-06-BAC-534-04L

CALIBRATION REPORT

ELECTRONICS TESTING
CENTER, TAIWAN

Page 9 of 16

Test Description	Actual Value	Expanded Uncertainty
8. Noise Display Check :		
9 kHz	-149.3 dBm	0.80 dB
100 kHz	-150.2 dBm	0.80 dB
999 kHz	-151.3 dBm	0.80 dB
10.99 MHz	-151.2 dBm	0.80 dB
19.99 MHz	-151.1 dBm	0.80 dB
49.99 MHz	-151.2 dBm	0.80 dB
99.99 MHz	-151.9 dBm	0.80 dB
199.9 MHz	-152.7 dBm	0.80 dB
499.9 MHz	-152.8 dBm	0.80 dB
999.9 MHz	-151.7 dBm	0.80 dB
1499 MHz	-152.8 dBm	0.94 dB
1999 MHz	-151.7 dBm	0.94 dB
2499 MHz	-151.2 dBm	0.94 dB
2999 MHz	-150.9 dBm	0.94 dB
3099 MHz	-150.2 dBm	0.94 dB
3499 MHz	-150.8 dBm	0.94 dB
3999 MHz	-150.6 dBm	0.94 dB
4499 MHz	-150.4 dBm	0.94 dB
4999 MHz	-150.9 dBm	0.94 dB
5499 MHz	-150.6 dBm	0.94 dB
5999 MHz	-150.8 dBm	0.94 dB
6499 MHz	-150.5 dBm	0.94 dB
6999 MHz	-150.4 dBm	0.94 dB
7999 MHz	-150.7 dBm	0.94 dB
8999 MHz	-150.9 dBm	0.94 dB
9999 MHz	-150.8 dBm	0.94 dB
10999 MHz	-150.5 dBm	0.94 dB
11999 MHz	-150.9 dBm	0.94 dB
12999 MHz	-150.4 dBm	0.94 dB
13999 MHz	-150.9 dBm	0.94 dB
14999 MHz	-150.4 dBm	0.94 dB
15999 MHz	-150.7 dBm	0.94 dB
16999 MHz	-149.4 dBm	0.94 dB

校正報告

財團法人台灣電子檢驗中心

工 服NO. 19-06-BAC-534-04L

CALIBRATION REPORT

ELECTRONICS TESTING
CENTER, TAIWAN

Page 10 of 16

Test Description	Actual Value	Expanded Uncertainty
8. Noise Display Check (@ Continued):		
17999 MHz	-148.9 dBm	0.94 dB
18999 MHz	-147.1 dBm	0.94 dB
19999 MHz	-146.3 dBm	0.94 dB
20999 MHz	-145.6 dBm	0.94 dB
21999 MHz	-147.6 dBm	0.94 dB
22999 MHz	-148.1 dBm	0.94 dB
23999 MHz	-148.8 dBm	0.94 dB
24999 MHz	-148.2 dBm	0.94 dB
25999 MHz	-147.9 dBm	0.94 dB
26999 MHz	-147.3 dBm	0.94 dB
27999 MHz	-146.4 dBm	0.94 dB
28999 MHz	-145.9 dBm	0.94 dB
29999 MHz	-145.7 dBm	0.94 dB
30999 MHz	-146.0 dBm	0.94 dB
31999 MHz	-145.5 dBm	0.94 dB
32999 MHz	-145.1 dBm	0.94 dB
33999 MHz	-144.8 dBm	0.94 dB
34999 MHz	-143.7 dBm	0.94 dB
35999 MHz	-141.6 dBm	0.94 dB
36999 MHz	-140.5 dBm	0.94 dB
37999 MHz	-139.1 dBm	0.94 dB
38999 MHz	-137.5 dBm	0.94 dB
39999 MHz	-136.7 dBm	0.94 dB

校正報告

財團法人台灣電子檢驗中心

工 服NO. 19-06-BAC-534-04L

CALIBRATION REPORT

ELECTRONICS TESTING
CENTER, TAIWAN

Page 11 of 16

Test Description	Actual Value	Expanded Uncertainty
9. Frequency response Check:		
RF Attenuation 10 dB		
1 MHz	0.05 dB	0.62 dB
10 MHz	0.04 dB	0.62 dB
50 MHz	0.06 dB	0.62 dB
100 MHz	-0.01 dB	0.62 dB
200 MHz	-0.02 dB	0.62 dB
300 MHz	-0.01 dB	0.62 dB
400 MHz	-0.02 dB	0.62 dB
500 MHz	-0.02 dB	0.62 dB
600 MHz	-0.01 dB	0.62 dB
700 MHz	-0.01 dB	0.62 dB
800 MHz	-0.02 dB	0.62 dB
900 MHz	-0.02 dB	0.62 dB
1000 MHz	-0.02 dB	0.62 dB
1500 MHz	-0.03 dB	0.62 dB
2000 MHz	-0.02 dB	0.62 dB
2500 MHz	-0.02 dB	0.62 dB
2990 MHz	-0.01 dB	0.62 dB
3010 MHz	-0.04 dB	0.71 dB
3500 MHz	-0.06 dB	0.71 dB
4000 MHz	-0.09 dB	0.71 dB
4500 MHz	-0.12 dB	0.71 dB
5000 MHz	-0.18 dB	0.71 dB
5500 MHz	-0.19 dB	0.71 dB
6000 MHz	-0.23 dB	0.71 dB
6500 MHz	-0.24 dB	0.71 dB
6990 MHz	-0.29 dB	0.71 dB
7100 MHz	-0.32 dB	0.71 dB

校正報告

財團法人台灣電子檢驗中心

工 服NO. 19-06-BAC-534-04L

CALIBRATION REPORT

ELECTRONICS TESTING
CENTER, TAIWAN

Page 12 of 16

Test Description	Actual Value	Expanded Uncertainty
9. Frequency response Check (@ Continued):		
RF Attenuation 10 dB		
8000 MHz	-0.37 dB	0.71 dB
9000 MHz	-0.41 dB	0.71 dB
10000 MHz	-0.46 dB	0.71 dB
11000 MHz	-0.42 dB	0.71 dB
12000 MHz	-0.38 dB	0.71 dB
13000 MHz	-0.35 dB	0.71 dB
13500 MHz	-0.36 dB	0.71 dB
14000 MHz	-0.37 dB	0.71 dB
15000 MHz	-0.39 dB	0.71 dB
16000 MHz	-0.31 dB	0.71 dB
17000 MHz	-0.28 dB	0.71 dB
18000 MHz	-0.32 dB	0.71 dB
19000 MHz	-0.25 dB	0.79 dB
20000 MHz	0.08 dB	0.79 dB
21000 MHz	-0.62 dB	0.79 dB
22000 MHz	-0.54 dB	0.79 dB
23000 MHz	-0.45 dB	0.79 dB
24000 MHz	-0.51 dB	0.79 dB
25000 MHz	-0.63 dB	0.79 dB
26000 MHz	-0.62 dB	0.79 dB
27000 MHz	-0.58 dB	0.88 dB
28000 MHz	-0.61 dB	0.88 dB
29000 MHz	-0.62 dB	0.88 dB
29900 MHz	-0.66 dB	0.88 dB
31000 MHz	-0.68 dB	0.88 dB
32000 MHz	-0.71 dB	0.88 dB
33000 MHz	-0.72 dB	0.88 dB
34000 MHz	-0.74 dB	0.88 dB
35000 MHz	-0.73 dB	0.88 dB
36000 MHz	-0.67 dB	0.88 dB
37000 MHz	-0.61 dB	0.88 dB
38000 MHz	-0.59 dB	0.88 dB
39000 MHz	-0.46 dB	0.88 dB
39999 MHz	-0.49 dB	0.88 dB

校正報告

財團法人台灣電子檢驗中心

工 服NO. 19-06-BAC-534-04L

CALIBRATION REPORT

ELECTRONICS TESTING
CENTER, TAIWAN

Page 13 of 16

Test Description	Actual Value	Expanded Uncertainty
9. Frequency response Check (@ Continued):		
RF Attenuation 20 dB		
1 MHz	0.05 dB	0.62 dB
10 MHz	0.03 dB	0.62 dB
50 MHz	0.00 dB	0.62 dB
100 MHz	-0.04 dB	0.62 dB
200 MHz	-0.03 dB	0.62 dB
300 MHz	-0.04 dB	0.62 dB
400 MHz	-0.05 dB	0.62 dB
500 MHz	-0.06 dB	0.62 dB
600 MHz	-0.05 dB	0.62 dB
700 MHz	-0.05 dB	0.62 dB
800 MHz	-0.06 dB	0.62 dB
900 MHz	-0.05 dB	0.62 dB
1000 MHz	-0.06 dB	0.62 dB
1500 MHz	-0.05 dB	0.62 dB
2000 MHz	-0.04 dB	0.62 dB
2500 MHz	-0.04 dB	0.62 dB
2990 MHz	-0.05 dB	0.62 dB
Frequency response Check (@ Continued):		
RF Attenuation 40 dB		
1 MHz	0.05 dB	0.62 dB
10 MHz	0.03 dB	0.62 dB
50 MHz	0.01 dB	0.62 dB
100 MHz	-0.03 dB	0.62 dB
200 MHz	-0.02 dB	0.62 dB
300 MHz	-0.03 dB	0.62 dB
400 MHz	-0.04 dB	0.62 dB
500 MHz	-0.03 dB	0.62 dB
600 MHz	-0.02 dB	0.62 dB
700 MHz	-0.02 dB	0.62 dB
800 MHz	-0.01 dB	0.62 dB
900 MHz	-0.02 dB	0.62 dB
1000 MHz	-0.01 dB	0.62 dB
1500 MHz	-0.02 dB	0.62 dB
2000 MHz	-0.04 dB	0.62 dB
2500 MHz	-0.03 dB	0.62 dB
2990 MHz	-0.02 dB	0.62 dB

校正報告

財團法人台灣電子檢驗中心

工 服NO. 19-06-BAC-534-04L

CALIBRATION REPORT

ELECTRONICS TESTING
CENTER, TAIWAN

Page 14 of 16

=====

Test Description	Actual Value	Expanded Uncertainty
10. Display linearity Check:		
RBW 300 Hz		
10 dB	9.99 dB	0.62 dB
12 dB	7.98 dB	0.62 dB
14 dB	5.98 dB	0.62 dB
16 dB	3.98 dB	0.62 dB
18 dB	1.97 dB	0.62 dB
20 dB	0.00 dB(Ref.)	-----
22 dB	-2.00 dB	0.62 dB
24 dB	-4.00 dB	0.62 dB
26 dB	-6.01 dB	0.62 dB
28 dB	-8.02 dB	0.62 dB
30 dB	-10.00 dB	0.62 dB
32 dB	-12.00 dB	0.62 dB
34 dB	-14.01 dB	0.62 dB
36 dB	-16.02 dB	0.62 dB
38 dB	-18.02 dB	0.62 dB
40 dB	-19.99 dB	0.62 dB
42 dB	-22.00 dB	0.66 dB
44 dB	-23.99 dB	0.66 dB
46 dB	-25.98 dB	0.66 dB
48 dB	-27.98 dB	0.66 dB
50 dB	-29.99 dB	0.66 dB
52 dB	-31.98 dB	0.66 dB
54 dB	-33.98 dB	0.66 dB
56 dB	-35.99 dB	0.66 dB
58 dB	-37.99 dB	0.66 dB
60 dB	-39.96 dB	0.66 dB
65 dB	-44.98 dB	0.66 dB
70 dB	-49.98 dB	0.66 dB
75 dB	-54.98 dB	0.66 dB
80 dB	-59.97 dB	0.66 dB
85 dB	-64.98 dB	0.66 dB
90 dB	-69.98 dB	0.66 dB
95 dB	-74.97 dB	0.66 dB
100 dB	-79.96 dB	0.66 dB

校正報告

財團法人台灣電子檢驗中心

工 服NO. 19-06-BAC-534-04L

CALIBRATION REPORT

ELECTRONICS TESTING
CENTER, TAIWAN

Page 15 of 16

Test Description	Actual Value	Expanded Uncertainty
10. Display linearity Check		
RBW 300 kHz:		
10 dB	9.99 dB	0.62 dB
12 dB	7.99 dB	0.62 dB
14 dB	5.98 dB	0.62 dB
16 dB	3.98 dB	0.62 dB
18 dB	1.97 dB	0.62 dB
20 dB	0.00 dB(Ref.)	-----
22 dB	-2.00 dB	0.62 dB
24 dB	-4.01 dB	0.62 dB
26 dB	-6.02 dB	0.62 dB
28 dB	-8.02 dB	0.62 dB
30 dB	-10.01 dB	0.62 dB
32 dB	-12.00 dB	0.62 dB
34 dB	-14.01 dB	0.62 dB
36 dB	-16.02 dB	0.62 dB
38 dB	-18.02 dB	0.62 dB
40 dB	-20.00 dB	0.62 dB
42 dB	-21.99 dB	0.66 dB
44 dB	-24.00 dB	0.66 dB
46 dB	-25.99 dB	0.66 dB
48 dB	-27.99 dB	0.66 dB
50 dB	-29.99 dB	0.66 dB
52 dB	-31.98 dB	0.66 dB
54 dB	-33.97 dB	0.66 dB
56 dB	-35.98 dB	0.66 dB
58 dB	-37.98 dB	0.66 dB
60 dB	-39.97 dB	0.66 dB
65 dB	-44.98 dB	0.66 dB
70 dB	-49.97 dB	0.66 dB
75 dB	-54.97 dB	0.66 dB
80 dB	-59.96 dB	0.66 dB

校正報告

財團法人台灣電子檢驗中心

工 服NO. 19-06-BAC-534-04L

CALIBRATION REPORT

ELECTRONICS TESTING
CENTER, TAIWAN

Page 16 of 16

Test Description	Actual Value	Expanded Uncertainty
11. Attenuator accuracy Check:		
0 dB	-9.98 dB	0.38 dB
10 dB	0.00 dB(Ref.)	-----
20 dB	10.00 dB	0.38 dB
30 dB	20.00 dB	0.38 dB
40 dB	30.01 dB	0.38 dB
50 dB	40.01 dB	0.38 dB
60 dB	50.02 dB	0.38 dB
70 dB	60.05 dB	0.38 dB
12. Reference level		
switching accuracy		
Reference level Check:		
0 dBm	10.00 dB	0.34 dB
-10 dBm	0.00 dB(Ref.)	-----
-20 dBm	-10.00 dB	0.34 dB
-30 dBm	-20.00 dB	0.34 dB
-40 dBm	-30.00 dB	0.34 dB
-50 dBm	-40.00 dB	0.34 dB
-11 dBm	-1.00 dB	0.34 dB
-12 dBm	-2.00 dB	0.34 dB
-13 dBm	-3.00 dB	0.34 dB
-14 dBm	-4.00 dB	0.34 dB
-15 dBm	-5.00 dB	0.34 dB
-16 dBm	-6.00 dB	0.34 dB
-17 dBm	-7.00 dB	0.34 dB
-18 dBm	-8.00 dB	0.34 dB
-19 dBm	-9.00 dB	0.34 dB

說明:

1.本校正報告內的項次2、6、7相對擴充不確定度評估與評估表示係依據「ISO Guide 98-3 量測不確定度表示方式指引」，相對擴充不確定度 $U = ku_c$ 其中 u_c 為相對組合標準不確定度， $k=2.0$ ，為信賴水準 95 %之涵蓋因子。

本校正報告內的項次1、3、4、5、8、9、10、11、12擴充不確定度與評估表示係依據「ISO Guide 98-3 量測不確定度表示方式指引」，擴充不確定度 $U = ku_c$ 其中 u_c 為相對組合標準不確定度， $k=2.0$ ，為信賴水準 95 %之涵蓋因子。

2.參考ANSI C63.2及CISPR 16-1-1相關規範。



Calibration certificate

ISO 17025
ACCREDITED LABORATORY



Accreditation certificate No. № BY/112 02.5.0.0065 of 09.01.2015

Certificate number 95-17 Date when calibrated 10/17/2017 Page 1 of 2

Item

calibrated

Mixer M22HWD # 110215-1

Description of measurement standard / measuring instrument / identification

Customer

Bureau Veritas Group Consumer Products Services Division, Taiwan
Branch E-2, No.1, Lixing 1st Rd., East Dist., Hsinchu City 300, Taiwan,
R.O.C.

Name of the customer, address

**Method of
calibration**

GOST 20271.1, MK KL 8.2-16

Name of the method / identification

All measurements are traceable to the SI units which are realized by national measurement standards of NMI and state standards of Ukraine. This certificate shall not be reproduced, except in full. Any publication extracts from the calibration certificate requires written permission of the issuing calibration laboratory of microwave measuring equipment.

**Authorising
signature**



M. Svirid/ Technical manager
Name and position

Date of issue 10/17/2017

Calibration Certificate

Certificate number **95-17**

Page 2 of 2

Calibration is performed by using

1. Wattmeter M 568
2. Calibrator of power M1-11
3. Wattmeter M3-22A
4. Spectrum analyzer E4407B
5. Signal generator G4-161
6. Signal generator MG3694C
7. Voltmeter V7-34
8. Frequency meter RCH3-72
9. Frequency meter CH3-66
10. Diplexer DPL26

Calibration conditions

Temperature 22.2 °C
Humidity 45 %
Pressure 99.7 kPa

Calibration results are given in the Measuring report # 95-17.

#	Parameter	Specifications required	Specifications tested and measured
1	System Operating Frequency	33 – 55 GHz	Corresponds
2	LO Input	+12 – +17 dBm	Corresponds
3	IF Frequency Range	321 – 2400 MHz	Corresponds
4	Mixer Bias	+1.86 mA	Corresponds
5	System Waveguide Interface	WR-22	Corresponds
6	Conversion Loss	< 35 dB	Corresponds (Table 1)
7	System LO/IF Interface	SMA (f)	Corresponds
8	Typical RF Power to Avoid Compression	-20 dBm (10 µW)	Corresponds

Signature of the person who has performed calibration


M. Kasperovich/ Engineer
Name and function

**Calibration Laboratory of
Microwave Measuring Equipment**

Accreditation certificate

No. BY/112 02.5.0.0065

Address: 6, P. Brovki str., Minsk

220027, Belarus

Phone/Fax: +375 17 2938496

Technical Manager



MEASURING REPORT # 95-17

October 17, 2017

Customer:	Bureau Veritas Group Consumer Products Services Division, Taiwan Branch E-2, No.1, Lixing 1st Rd., East Dist., Hsinchu City 300, Taiwan, R.O.C.
Item calibrated:	Mixer M22HWD # 110215-1
Method of calibration:	GOST 20271.1, MK KL 8.2-16
Number of samples:	One
Delivery date of the sample:	09/18/2017
Date of calibration:	From 09/18/2017 to 10/17/2017

MEASURING CONDITIONS

Temperature: 22.2 °C	Humidity: 45 %	Pressure: 99.7 kPa
----------------------	----------------	--------------------

MEASURING EQUIPMENT

#	Measuring equipment	Serial number
1	Wattmeter M 568	164
2	Calibrator of power M1-11	841202
3	Wattmeter M3-22A	037410
4	Spectrum analyzer E4407B	MY45110807
5	Signal generator G4-161	3
6	Signal generator MG3694C	133805
7	Voltmeter V7-34	0067787
8	Frequency meter RCH3-72	931200
9	Frequency meter CH3-66	98051
10	Diplexer DPL26	01

MEASURING RESULTS

IF Frequency 321.4 MHz $\pm$ 5 MHz;

Mixer Bias +1.86 mA;

LO Input Power 14.5 to 16 dBm (2.9 to 7.1 GHz).

LO Insertion Loss of Diplexer 0.7 dB.

Table 1

Frequency, GHz	33	44	55
Input RF Power, dBm	-20.0	-20.0	-20.0
Measured Value, dBm	-54.9	-53.5	-53.5
Conversion Loss, dB	34.9	33.5	33.5
Expanded uncertainty, dB	2.1	2.1	2.2

Engineer

M. Kasperovich

Quality Manager

A. Kostrikin

This Measuring report issued in duplicate and sent to:

1. Bureau Veritas Group Consumer Products Services Division, Taiwan Branch E-2, No.1,
Lixing 1st Rd., East Dist., Hsinchu City 300, Taiwan, R.C.C.

2. Calibration Laboratory of Microwave Measuring Equipment

Duplication of Measuring report (complete or partial) must be authorized by the laboratory.

Calibration Laboratory of Microwave Measuring Equipment
of MWM lab



Calibration certificate

ISO 17025
ACCREDITED LABORATORY



Accreditation certificate No. № BY/112 02.5.0.0065 of 09.01.2015

Certificate number 96-17 Date when calibrated 10/17/2017 Page 1 of 2

Item

calibrated

Standard gain horn antenna M22RH

Description of measurement standard / measuring instrument / identification

Customer

Bureau Veritas Group Consumer Products Services Division, Taiwan
Branch E-2, No.1, Lixing 1st Rd., East Dist., Hsinchu City 300, Taiwan,
R.O.C.

Name of the customer, address

Method of
calibration

GOST 20271.1, MK KL 8.2-16

Name of the method / identification

All measurements are traceable to the SI units which are realized by national measurement standards of NMI and state standards of Ukraine. This certificate shall not be reproduced, except in full. Any publication extracts from the calibration certificate requires written permission of the issuing calibration laboratory of microwave measuring equipment.

Authorising
signature



M. Syrid/ Technical manager
Name and position

Date of issue 10/17/2017

Calibration Certificate

Certificate number 96-17

Page 2 of 2

Calibration is performed by using

1. Wattmeter M 568
2. Calibrator of power M1-11
3. Wattmeter M3-22A
4. Signal generator G4-161
5. Signal generator MG3694C
6. Voltmeter V7-34
7. Frequency meter RCH3-72
8. Frequency meter CH3-66
9. Horn antenna P6-11B
10. Horn antenna P6-80/3

Calibration conditions

Temperature 22.2 °C
Humidity 45 %
Pressure 99.7 kPa

Calibration results are given in the Measuring report # 96-17.

#	Parameter	Specifications required	Specifications tested and measured
1	Frequency range	33 – 55 GHz	Corresponds
2	Waveguide Interface	WR-22	Corresponds
3	Gain	23.9 dBi	Corresponds (Table 1)
4	Antenna Factor	38.7 dB/m	Corresponds (Table 1)

Signature of the person who has performed calibration



M. Kasperovich/ Engineer
Name and function

**Calibration Laboratory of
Microwave Measuring Equipment**

Accreditation certificate

No. BY/112 02.5.0.0065

Address: 6, P. Brovki str., Minsk
220027, Belarus

Phone/Fax: +375 17 2938496

Technical Manager



M. Svirid

October 17, 2017

MEASURING REPORT # 96-17

October 17, 2017

Customer:	Bureau Veritas Group Consumer Products Services Division, Taiwan Branch E-2, No.1, Lixing 1st Rd., East Dist., Hsinchu City 300, Taiwan, R.O.C.
Item calibrated:	Standard gain horn antenna M22RH
Method of calibration:	GOST 20271.1, MK KL 8.2-16
Number of samples:	One
Delivery date of the sample:	09/18/2017
Date of calibration:	From 09/18/2017 to 10/17/2017

MEASURING CONDITIONS

Temperature: 22.2 °C	Humidity: 45 %	Pressure: 99.7 kPa
----------------------	----------------	--------------------

MEASURING EQUIPMENT

#	Measuring equipment	Serial number
1	Wattmeter M 568	164
2	Calibrator of power M1-11	841202
3	Wattmeter M3-22A	037410
4	Signal generator G4-161	3
5	Signal generator MG3694C	133805
6	Voltmeter V7-34	0067787
7	Frequency meter RCH3-72	931200
8	Frequency meter CH3-66	98051
9	Horn antenna P6-11B	08051
10	Horn antenna P6-80/3	08112

MEASURING RESULTS

Distance between tested and generating antenna 0.9 m.

Table 1

Frequency, GHz	33	44	55
Input Power, mW	10.0	10.0	10.0
Power density of electromagnetic field, W/m ²	0.117	0.199	0.291
Maximum level of measured power, μ W	139	186	180
Gain, dB	22.6	24.0	24.2
Antenna Factor, dB/m	38.0	39.1	40.8
Expanded uncertainty, dB	2.5	2.6	2.5

Engineer

M. Kasperovich

Quality Manager

A. Kostrikin

This Measuring report issued in duplicate and sent to:

1. Bureau Veritas Group Consumer Products Services Division, Taiwan Branch E-2, No.1,
Lixing 1st Rd., East Dist., Hsinchu City 300, Taiwan, R.O.C.

2. Calibration Laboratory of Microwave Measuring Equipment

Duplication of Measuring report (complete or partial) must be authorized by the laboratory.



Calibration certificate

ISO 17025
ACCREDITED LABORATORY



Accreditation certificate No. № BY/112 02.5.0.0065 of 09.01.2015

Certificate number 97-17 Date when calibrated 10/17/2017 Page 1 of 2

Item

calibrated **Mixer M22HWD # 110215-1 + Standard gain horn antenna M22RH**

Description of measurement standard / measuring instrument / identification

Customer

Bureau Veritas Group Consumer Products Services Division, Taiwan
Branch E-2, No.1, Lixing 1st Rd., East Dist., Hsinchu City 300, Taiwan,
R.O.C.

Name of the customer, address

Method of calibration

GOST 20271.1, MK KL 8.2-16

Name of the method / identification

All measurements are traceable to the SI units which are realized by national measurement standards of NMI and state standards of Ukraine. This certificate shall not be reproduced, except in full. Any publication extracts from the calibration certificate requires written permission of the issuing calibration laboratory of microwave measuring equipment.

Authorising
signature



M. Svirid/ Technical manager
Name and position

Date of issue 10/17/2017

Calibration Certificate

Certificate number **97-17**

Page 2 of 2

Calibration is performed by using

1. Wattmeter M 568
2. Calibrator of power M1-11
3. Wattmeter M3-22A
4. Signal generator G4-161
5. Signal generator MG3694C
6. Voltmeter V7-34
7. Frequency meter RCH3-72
8. Frequency meter CH3-66
9. Horn antenna P6-11B
10. Horn antenna P6-80/3
11. Spectrum analyzer E4407B
12. Diplexer DPL26

Calibration conditions

Temperature 22.2 °C

Humidity 45 %

Pressure 99.7 kPa

Calibration results are given in the Measuring report # 97-17.

#	Parameter	Specifications required	Specifications tested and measured
1	Frequency range	33 – 55 GHz	Corresponds
2	Waveguide Interface	WR-22	Corresponds
3	LO Input	+12 – +17 dBm	Corresponds
4	IF Frequency Range	321 – 2400 MHz	Corresponds
5	Mixer Bias	+1.86 mA	Corresponds
6	Conversion Loss	< 35 dB	Corresponds (Table 1)
7	System LO/IF Interface	SMA (f)	Corresponds

Signature of the person who has performed calibration



M. Kasperovich/ Engineer
Name and function

**Calibration Laboratory of
Microwave Measuring Equipment**

Accreditation certificate

No. BY/112 02.5.0.0065

Address: 6, P. Brovki str., Minsk

220027, Belarus

Phone/Fax: +375 17 2938496

Technical Manager


M. Svirid
October 17, 2017

MEASURING REPORT # 97-17

October 17, 2017

Customer:	Bureau Veritas Group Consumer Products Services Division, Taiwan Branch E-2, No.1, Lixing 1st Rd., East Dist., Hsinchu City 300, Taiwan, R.O.C.
Item calibrated:	Mixer M22HWD # 110215-1 + standard gain horn antenna M22RH
Method of calibration:	GOST 20271.1, MK KL 8.2-16
Number of samples:	One
Delivery date of the sample:	09/18/2017
Date of calibration:	From 09/18/2017 to 10/17/2017

MEASURING CONDITIONS

Temperature: 22.2 °C	Humidity: 45 %	Pressure: 99.7 kPa
----------------------	----------------	--------------------

MEASURING EQUIPMENT

#	Measuring equipment	Serial number
1	Wattmeter M 568	164
2	Calibrator of power M1-11	841202
3	Wattmeter M3-22A	037410
4	Signal generator G4-161	3
5	Signal generator MG3694C	133805
6	Voltmeter V7-34	0067787
7	Frequency meter RCH3-72	931200
8	Frequency meter CH3-66	98051
9	Horn antenna P6-11B	08051
10	Horn antenna P6-80/3	08112
11	Spectrum analyzer E4407B	MY45110807
12	Diplexer DPL26	01

MEASURING RESULTS

IF Frequency 321.4 MHz $\pm$ 5 MHz;
 Mixer Bias +1.86 mA;
 LO Input Power 14.5 to 16 dBm (2.9 to 7.1 GHz).
 LO Insertion Loss of Diplexer 0.7 dB.

Distance between tested and generating antenna 0.9 m.

Table 1

Frequency, GHz	33	44	55
Input RF power, mW	0.71	0.51	0.57
Power density of electromagnetic field, W/m ²	0.0091	0.0068	0.0055
Measured level, dBm	-54.9	-53.6	-53.5
Power received by antenna, dBm	-20.1	-20.2	-19.9
Conversion Loss, dB	34.8	33.4	33.6
Expanded uncertainty, dB	3.0	2.9	2.8

Engineer

M. Kasperovich

Quality Manager

A. Kostrikin

This Measuring report issued in duplicate and sent to:

1. Bureau Veritas Group Consumer Products Services Division, Taiwan Branch E-2, No.1, Lixing 1st Rd., East Dist., Hsinchu City 300, Taiwan, R.O.C.
2. Calibration Laboratory of Microwave Measuring Equipment

Duplication of Measuring report (complete or partial) must be authorized by the laboratory.



Calibration certificate

ISO 17025
ACCREDITED LABORATORY



Accreditation certificate No. № BY/112 02.5.0.0065 of 09.01.2015

Certificate number 101-17 Date when calibrated 10/17/2017 Page 1 of 2

Item

calibrated Mixer M15HWD # 110215-1

Description of measurement standard / measuring instrument / identification

Customer

Bureau Veritas Group Consumer Products Services Division, Taiwan
Branch E-2, No.1, Lixing 1st Rd., East Dist., Hsinchu City 300, Taiwan,
R.O.C.

Name of the customer, address


**Method of
calibration**

GOST 20271.1, MK KL 8.2-16

Name of the method / identification

All measurements are traceable to the SI units which are realized by national measurement standards of NMI and state standards of Ukraine. This certificate shall not be reproduced, except in full. Any publication extracts from the calibration certificate requires written permission of the issuing calibration laboratory of microwave measuring equipment.

**Authorising
signature**



M. Svirid/ Technical manager
Name and position

Date of issue 10/17/2017

Calibration Certificate

Certificate number **101-17**

Page 2 of 2

Calibration is performed by using

1. Wattmeter M 568
2. Wattmeter M 546
3. Spectrum analyzer E4407B
4. Signal generator G4-186
5. Signal generator G4-161
6. Voltmeter V7-34
7. Frequency meter RCH3-72
8. Frequency meter CH3-66
9. Diplexer DPL26

Calibration conditions

Temperature 22.9 °C

Humidity 44.5 %

Pressure 98.4 kPa

Calibration results are given in the Measuring report # 101-17.

#	Parameter	Specifications required	Specifications tested and measured
1	System Operating Frequency	50 – 75 GHz	Corresponds
2	LO Input	+12 – +17 dBm	Corresponds
3	IF Frequency Range	321 – 2400 MHz	Corresponds
4	Mixer Bias	+4.05 mA	Corresponds
5	System Waveguide Interface	WR-15	Corresponds
6	Conversion Loss	< 37 dB	Corresponds (Table 1)
7	System LO/IF Interface	SMA (f)	Corresponds
8	Typical RF Power to Avoid Compression	-20 dBm (10 μ W)	Corresponds

Signature of the person who has performed calibration

M. Kasperovich/ Engineer

Name and function

**Calibration Laboratory of
Microwave Measuring Equipment**

Accreditation certificate

No. BY/112 02.5.0.0065

Address: 6, P. Brovki str., Minsk
220027, Belarus

Phone/Fax: +375 17 2938496

Technical Manager

 M. Svirid

October 17, 2017

MEASURING REPORT # 101-17

October 17, 2017

Customer:	Bureau Veritas Group Consumer Products Services Division, Taiwan Branch E-2, No.1, Lixing 1st Rd., East Dist., Hsinchu City 300, Taiwan, R.O.C.
Item calibrated:	Mixer M15HWD # 110215-1
Method of calibration:	GOST 20271.1, MK KL 8.2-16
Number of samples:	One
Delivery date of the sample:	09/18/2017
Date of calibration:	From 09/18/2017 to 10/17/2017

MEASURING CONDITIONS

Temperature: 22.9 °C	Humidity: 44.5 %	Pressure: 98.4 kPa
----------------------	------------------	--------------------

MEASURING EQUIPMENT

#	Measuring equipment	Serial number
1	Wattmeter M 568	164
2	Wattmeter M 546	163
3	Spectrum analyzer E4407B	MY45110807
4	Signal generator G4-186	5
5	Signal generator G4-161	3
6	Voltmeter V7-34	0067787
7	Frequency meter RCH3-72	931200
8	Frequency meter CH3-66	98051
9	Diplexer DPL26	01

MEASURING RESULTS

IF Frequency 321.4 MHz $\pm$ 5 MHz;

Mixer Bias +4.05 mA;

LO Input Power 14.5 to 16 dBm (2.9 to 7.1 GHz).

LO Insertion Loss of Diplexer 0.7 dB.


Table 1

Frequency, GHz	50	55	65	75
Input RF Power, dBm	-20.0	-20.0	-20.0	-20.0
Measured Value, dBm	-51.2	-52.6	-55.0	-57.4
Conversion Loss, dB	31.2	32.6	35.0	37.4
Expanded uncertainty, dB	2.0	2.1	2.2	2.5

Engineer

 M. Kasperovich

Quality Manager

 A. Kostrikin

This Measuring report issued in duplicate and sent to:

1. Bureau Veritas Group Consumer Products Services Division, Taiwan Branch E-2, No.1,
Lixing 1st Rd., East Dist., Hsinchu City 300, Taiwan, R.O.C.

2. Calibration Laboratory of Microwave Measuring Equipment

Duplication of Measuring report (complete or partial) must be authorized by the laboratory.



Calibration certificate

ISO 17025
ACCREDITED LABORATORY



Accreditation certificate No. № BY/112 02.5.0.0065 of 09.01.2015

Certificate number 102-17 Date when calibrated 10/17/2017 Page 1 of 2

Item

calibrated

Standard gain horn antenna M15RH

Description of measurement standard / measuring instrument / identification

Customer

Bureau Veritas Group Consumer Products Services Division, Taiwan
Branch E-2, No.1, Lixing 1st Rd., East Dist., Hsinchu City 300, Taiwan,
R.O.C.

Name of the customer, address

**Method of
calibration**

GOST 20271.1, MK KL 8.2-16

Name of the method / identification

All measurements are traceable to the SI units which are realized by national measurement standards of NMI and state standards of Ukraine. This certificate shall not be reproduced, except in full. Any publication extracts from the calibration certificate requires written permission of the issuing calibration laboratory of microwave measuring equipment.

**Authorising
signature**



M. Svirid/ Technical manager
Name and position

Date of issue 10/17/2017