

二维码及其变动信息：

(01) 16932825326007
(21) 2001612240600001

——

(01) 16932825326007
(21) 2001612240600100



TelerPM

Blood Pressure Monitor

Model: BB2284-AE01

Rating: Battery powered mode: 4 AAA size batteries, 6V ≐

AC adapter powered mode: 5V ≐ 1A (not included)

(Please use the AC adapter authorized by the manufacturer!)

Guangdong Transtek Medical Electronics Co., Ltd.

Zone A, No.105, Dongli Road, Torch Development

District, 528437 Zhongshan, Guangdong, China

FCC ID: OUI6B2284-AE01

IP CLASSIFICATION: IP21

MADE IN CHINA

2024-05



GS1 Datamatrix 二维码
(UDI), (不带括号)

此处的二维码只是示意图，供参考。二
维码生成的格式、尺寸位置，实际要
生成的内容见“二维码及变动信息”。
具体内容请以订单要求生成为准”

背贴上UDI中含有的SN流水号要与彩盒贴纸上UDI中含有的SN流水号一一对应

技术要求：

- 1、黏合不可露胶
- 2、保持印刷面板上的清洁
- 3、注意套印的准确性
- 4、表面处理不可爆开
- 5、须满足RoHS、Reach的环保要求
- 6、结构工艺以结构受控图为准
- 7、颜色参考：

广东乐心医疗电子股份有限公司

2024-06-12

受控文件

产品型号	BB2284-AE01	材质	50#哑银咪钠不干胶	零件名称	
产品名称	血压计	尺寸	44.6 x 24.9 mm		
对应结构图纸	-	印色	单黑+Pantone Reflex Blue C	2BMB1612-BB2284-AE01-GB-01-背贴-A0	
	比例	1:1	表面处理	零件图号	
	单位	mm	设计	BB2284-AE01-GB-01	
			审核	共 1 张	第 1 张
Transtek			批准	版 本	A/0
广东乐心医疗电子股份有限公司					