



Medtronic, Inc.

710 Medtronic Parkway
Minneapolis, MN-55432, USA

April 19, 2019

Federal Communications Commission
Authorization and Evaluation Division
7435 Oakland Mills Road
Columbia, Maryland 21046

www.medtronic.com
christiaan.masson@medtronic.com
tel 31 43 356 6540

Reference: Application for FCC Grant LF5BLEIMPLANT3, Medtronic Model LINQ II

To whom it may concern,

I, the undersigned, as the authorized signatory for Medtronic, Inc. hereby apply to seek original grant for Medtronic Model LINQ II.

The Medtronic LINQ II Insertable Cardiac Monitor (ICM) is a programmable device that continuously monitors a patient's ECG and other physiological parameters. The device records cardiac information in response to automatically detected arrhythmias and patient-initiated activation or markings. The device is designed to automatically record the occurrence of an episode of arrhythmia in a patient.

The Medtronic Model LINQ II implements Bluetooth Low Energy Telemetry in order to communicate with external devices, including but not limited to patient or clinician monitors, or smart phones or tablets with a Medtronic app installed. Authorization is sought under FCC rule part 15.247.

In case of any additional questions please feel free to contact me.

Sincerely,

A handwritten signature in blue ink, appearing to read "Masson", written over a horizontal line.

Christiaan Masson
Regulatory Affairs Program Manager
Cardiac Rhythm and Heart Failure

