

ANNEX B – RELEVANT PAGES FROM CALIBRATION REPORTS

DAE4 Sn:546

In Collaboration with TTL s p e e g CALIBRATION LABORATORY		In Collaboration with TTL s p e e g CALIBRATION LABORATORY									
Client: SRTC Certificate No: Z18-60400 CALIBRATION CERTIFICATE Object: DAE4 - SN: 546 Calibration Procedure(s): FF-Z11-002-01 Calibration Procedure for the Data Acquisition Electronics (DAE) Calibration date: October 15, 2018 This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements(SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate. All calibrations have been conducted in the closed laboratory facility: environment temperature23±2°C and humidity<70%. Calibration Equipment used (M&TE critical for calibration) <table border="1"> <thead> <tr> <th>Primary Standards</th> <th>ID #</th> <th>Cal Date(Calibrated by, Certificate No.)</th> <th>Scheduled Calibration</th> </tr> </thead> <tbody> <tr> <td>Process Calibrator 753</td> <td>1971018</td> <td>20-Jun-18 (CTTL, No.J180X02034)</td> <td>June-19</td> </tr> </tbody> </table> Calibrated by: Yu Zengying SAR Test Engineer Reviewed by: Lin Han SAR Test Engineer Approved by: Qi Dianyuan SAR Project Leader issued: October 17, 2018 This calibration certificate shall not be reproduced except in full without written approval of the laboratory.		Primary Standards	ID #	Cal Date(Calibrated by, Certificate No.)	Scheduled Calibration	Process Calibrator 753	1971018	20-Jun-18 (CTTL, No.J180X02034)	June-19	Glossary: DAE: data acquisition electronics Connector angle: information used in DASY system to align probe sensor X to the robot coordinate system. Methods Applied and Interpretation of Parameters: • DC Voltage Measurement: Calibration Factor assessed for use in DASY system by comparison with a calibrated instrument traceable to national standards. The figure given corresponds to the full scale range of the voltmeter in the respective range. • Connector angle: The angle of the connector is assessed measuring the angle mechanically by a tool inserted. Uncertainty is not required. • The report provide only calibration results for DAE, it does not contain other performance test results.	
Primary Standards	ID #	Cal Date(Calibrated by, Certificate No.)	Scheduled Calibration								
Process Calibrator 753	1971018	20-Jun-18 (CTTL, No.J180X02034)	June-19								
Certificate No: Z18-60400		Page 2 of 3									

In Collaboration with TTL s p e e g CALIBRATION LABORATORY																	
ADD: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China Tel: +86-10-62204633-2312 Fax: +86-10-62204633-2304 E-mail: cti@ttsi.com.cn Http://www.cti.com.cn DC Voltage Measurement A/D - Converter Resolution nominal 1LSB = 5.1uV Full range = -100...+200 mV High Range: 1LSB = 810V Full range = +1...+3mV Low Range: 1LSB = 810V Full range = +1...+3mV DASY measurement parameters: Auto Zero Time: 2 sec. Measuring time: 2 sec <table border="1"> <thead> <tr> <th>Calibration Factors</th> <th>X</th> <th>Y</th> <th>Z</th> </tr> </thead> <tbody> <tr> <td>High Range</td> <td>405.306 ± 0.12% (k=2)</td> <td>404.059 ± 0.13% (k=2)</td> <td>404.180 ± 0.13% (k=2)</td> </tr> <tr> <td>Low Range</td> <td>3.99893 ± 0.7% (k=2)</td> <td>3.99878 ± 0.7% (k=2)</td> <td>3.99821 ± 0.7% (k=2)</td> </tr> </tbody> </table> Connector Angle <table border="1"> <thead> <tr> <th>Connector Angle to be used in DASY system</th> <th></th> </tr> </thead> <tbody> <tr> <td></td> <td>238° ± 1°</td> </tr> </tbody> </table>		Calibration Factors	X	Y	Z	High Range	405.306 ± 0.12% (k=2)	404.059 ± 0.13% (k=2)	404.180 ± 0.13% (k=2)	Low Range	3.99893 ± 0.7% (k=2)	3.99878 ± 0.7% (k=2)	3.99821 ± 0.7% (k=2)	Connector Angle to be used in DASY system			238° ± 1°
Calibration Factors	X	Y	Z														
High Range	405.306 ± 0.12% (k=2)	404.059 ± 0.13% (k=2)	404.180 ± 0.13% (k=2)														
Low Range	3.99893 ± 0.7% (k=2)	3.99878 ± 0.7% (k=2)	3.99821 ± 0.7% (k=2)														
Connector Angle to be used in DASY system																	
	238° ± 1°																
Certificate No: Z: 8-60400	Page 1 of 3																

ES3DV3 Sn:3127

In Collaboration with
TTL Calibration Laboratory
Add: No.51 Xuyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-4230603-2312 Fax: +86-10-4230603-2304
E-mail: csl@ttl.com.cn <http://www.ttl.com.cn>

Client: **SRTC** Certificate No: **Z18-60398**

CALIBRATION CERTIFICATE

Object: **ES3DV3 - SN3127**

Calibration Procedure(s): **FF-Z11-004-01**
Calibration Procedures for Dosimetric E-field Probes

Calibration date: **November 02, 2018**

This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements(SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature(23±0.5)°C and humidity<70%.

Calibration Equipment used (M&TE critical for calibration)

Primary Standards	ID #	Cal Date(Calibrated by Certificate No.)	Scheduled Calibration
Power Meter NRP2	101919	20-Jun-18 (CTTL No.J18X05032)	Jun-19
Power sensor NRP-Z31	101547	20-Jun-18 (CTTL No.J18X05032)	Jun-19
Power sensor NRP-Z31	101548	20-Jun-18 (CTTL No.J18X05032)	Jun-19
Reference10dBAttenuator	18N50W-10dB	09-Feb-18(CTTL No.J18X01133)	Feb-20
Reference20dBAttenuator	18N50W-20dB	09-Feb-18(CTTL No.J18X01133)	Feb-20
Reference Probe EX3DV4	SN 3846	25-Jan-18(SPEAG No.EK3-3846_Jan18)	Jan-19
DAE4	SN 777	15-Dec-17(SPEAG No.DAE4-777_Dec17)	Dec-18

Secondary Standards	ID #	Cal Date(Calibrated by Certificate No.)	Scheduled Calibration
SignalGeneratorMG3700A	6201052605	21-Jun-18 (CTTL No.J18X05033)	Jun-19
Network Analyzer E5071C	MY46110673	14-Jan-18 (CTTL No.J18X00581)	Jan-19

Calibrated by: **Yu Zongying** SAR Test Engineer *[Signature]*

Reviewed by: **Lin Hao** SAR Test Engineer *[Signature]*

Approved by: **Qi Dianyan** SAR Project Leader *[Signature]*

Issued: November 04, 2018

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: Z18-60398

Page 1 of 11

In Collaboration with
TTL Calibration Laboratory
Add: No.51 Xuyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-4230603-2312 Fax: +86-10-4230603-2304
E-mail: csl@ttl.com.cn <http://www.ttl.com.cn>

Glossary:

TS: tissue simulating liquid
NORM_{x,y,z}: sensitivity in free space
ConF: sensitivity in TSL / NORM_{x,y,z}
DCP: diode compression point
OF: level factor (1/duty cycle) of the RF signal
A,E,C,D: modulation dependent linearization parameters
Polarization θ : θ rotation around probe axis
Polarization ϕ : ϕ rotation around an axis that is in the plane normal to probe axis (at measurement center), i
b=0 is normal to probe axis
Connector Angle: information used in DASY system to align probe sensor K to the root coordinate system

Calibration is Performed According to the Following Standards:
a) IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013
b) IEC 62209-1, "Measurement procedure for the assessment of Specific Absorption Rate (SAR) from hand-held and body mounted devices used next to the ear (frequency range of 300 MHz to 3 GHz)", July 2018
c) IEC 62209-2, "Procedure to determine the Specific Absorption Rate (SAR) for wireless communication devices used in close proximity to the human body (frequency range of 30 MHz to 6 GHz)", March 2018
d) KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

Methods Applied and Interpretation of Parameters:
• NORM_{x,y,z}: Assessed for E-field polarization $\theta=0$ (600MHz in TEM-cell; $F>1800$ MHz: waveguide) NORM_{x,y,z} are only intermediate values, i.e., the uncertainties of NORM_{x,y,z} does not effect the E^2 -field uncertainty inside TSL (see below ConF).
• NORM_(θ , ϕ , γ) = NORM_{x,y,z}: frequency response (see Frequency Response Chart). This linearity is implemented in DASY4 software versions later than 4.2. The uncertainty of the frequency response is included in the stated uncertainty of ConF.
• DCP_{x,y,z}: DCP are numerical linearization parameters assessed based on the data of power sweep (no uncertainty required). DCP does not depend on frequency nor media.
• PAR: PAR is the Peak to Average Ratio that is not calibrated but determined based on the signal characteristics.
• A_k, B_k, C_k, D_k, E_k, F_k, G_k, H_k, I_k, J_k, K_k, L_k, M_k, N_k, O_k, P_k, Q_k, R_k, S_k, T_k, U_k, V_k, W_k, X_k, Y_k, Z_k, A_k, B_k, C_k are numerical linearization parameters assessed based on the data of power sweep for specific modulation signal. The parameters do not depend on frequency nor media. VR is the maximum calibration range expressed in RMS voltage across the diode.
• ConF and Boundary Effect Parameters: Assessed in flat phantom using E-field (or Temperature Transfer Standard for 600MHz) and inside waveguide using analytical field distributions based on power measurements for $F>1800$ MHz. The semi-edges are used for assessment of the parameters applied for boundary compensation (alpha, depth) of which typical uncertainty valued are given. These parameters are used in DASY4 software to improve probe accuracy close to the boundary. The sensitivity in TSL corresponds to NORM_{x,y,z} ConF whereby the uncertainty corresponds to that given for ConF. A frequency dependent ConF is used in DASY version 4.4 and higher which allows extending the validity from 50MHz to 100MHz.
• Spherical isotropy (SD deviation from isotropy): in a field of low gradients realized using a fix phantom exposed by a patch antenna.
• Sensor Offset: The sensor offset corresponds to the offset of virtual measurement center from the probe to on probe axis. No tolerance required.
• Connector Angle: The angle is assessed using the information gained by determining the NORM_x (no uncertainty required).

Certificate No: Z18-60398

Page 2 of 11

Probe ES3DV3

SN: 3127

Calibrated: November 02, 2018

Calibrated for DASY/EASY Systems

(Note: non-compatible with DASY2 system!)

Certificate No: Z18-60398

Page 3 of 11



DASY/EASY – Parameters of Probe: ES3DV3 - SN: 3127

Basic Calibration Parameters

	Sensor X	Sensor Y	Sensor Z	Unc (k=2)
Norm($\mu V/(V/m)^2$) ^A	1.27	1.29	1.21	$\pm 10.0\%$
DCP(mV) ^B	103.3	104.4	105.0	

Modulation Calibration Parameters

UID	Communication System Name	A dB	B dB/μV	C	D dB	VR mV	Unc ^C (k=2)
0	CW	X 0.0	0.0	1.0	0.00	285.6	$\pm 2.2\%$
		Y 0.0	0.0	1.0		287.9	
		Z 0.0	0.0	1.0		282.9	

The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

^A The uncertainties of Norm X, Y, Z do not affect the E² field uncertainty inside TSL (see Page 5 and Page 6).
^B Numerical linearization parameter, uncertainty not required.
^C Uncertainty is determined using the max. deviation from linear response applying rectangular distribution and is expressed for the square of the field value.

Certificate No: Z18-60398

Page 4 of 11



DASY/EASY – Parameters of Probe: ES3DV3 - SN: 3127

Calibration Parameter Determined in Head Tissue Simulating Media

f [MHz] ^B	Relative Permittivity ^C	Conductivity (S/m) ^C	ConvF X	ConvF Y	ConvF Z	Alpha ^D	Depth ^D (mm)	Unc ^E (k=2)
750	41.9	0.89	6.34	6.34	6.34	0.40	1.35	$\pm 12.1\%$
835	41.5	0.90	6.18	6.18	6.18	0.35	1.58	$\pm 12.1\%$
1810	40.0	1.40	5.07	5.07	5.07	0.66	1.24	$\pm 12.1\%$
2000	40.0	1.40	4.98	4.98	4.98	0.70	1.20	$\pm 12.1\%$
2300	39.5	1.07	4.79	4.79	4.79	0.80	1.08	$\pm 12.1\%$
2450	39.2	1.80	4.66	4.66	4.66	0.90	1.08	$\pm 12.1\%$
2600	39.0	1.98	4.40	4.40	4.40	0.80	1.21	$\pm 12.1\%$

^B Frequency validity above 300 MHz of ± 100 MHz only applies for DASY v4.4 and higher (Page 2), else it is restricted to ± 50 MHz. The uncertainty is the RSS of ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below 300 MHz is $\pm 10, 25, 40, 50$ and 70 MHz for ConvF assessments at 30, 64, 128, 150 and 220 MHz respectively. Above 5 GHz frequency validity can be extended to ± 110 MHz.

^C At frequency below 3 GHz, the validity of tissue parameters (x and y) can be relaxed to $\pm 10\%$ if liquid compensation formula is applied to measured SAR values. At frequencies above 3 GHz, the validity of tissue parameters (x and y) is restricted to $\pm 5\%$. The uncertainty is the RSS of the ConvF uncertainty for indicated target tissue parameters.

^D Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than $\pm 1\%$ for frequencies below 3 GHz and below $\pm 2\%$ for the frequencies between 3-6 GHz at any distance larger than half the probe tip diameter from the boundary.

Certificate No: Z18-60398

Page 5 of 11



DASY/EASY – Parameters of Probe: ES3DV3 - SN: 3127

Calibration Parameter Determined in Body Tissue Simulating Media

f [MHz] ^B	Relative Permittivity ^C	Conductivity (S/m) ^C	ConvF X	ConvF Y	ConvF Z	Alpha ^D	Depth ^D (mm)	Unc ^E (k=2)
750	55.5	0.96	6.33	6.33	6.33	0.40	1.40	$\pm 12.1\%$
835	55.2	0.97	6.13	6.13	6.13	0.37	1.62	$\pm 12.1\%$
1810	53.3	1.52	4.78	4.78	4.78	0.65	1.27	$\pm 12.1\%$
2000	53.3	1.52	4.80	4.80	4.80	0.67	1.27	$\pm 12.1\%$
2300	52.9	1.81	4.46	4.46	4.46	0.90	1.16	$\pm 12.1\%$
2450	52.7	1.06	4.31	4.31	4.31	0.70	1.28	$\pm 12.1\%$
2600	52.5	2.10	4.14	4.14	4.14	0.90	1.10	$\pm 12.1\%$

^B Frequency validity above 300 MHz of ± 100 MHz only applies for DASY v4.4 and higher (Page 2), else it is restricted to ± 50 MHz. The uncertainty is the RSS of ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below 300 MHz is $\pm 10, 25, 40, 50$ and 70 MHz for ConvF assessments at 30, 64, 128, 150 and 220 MHz respectively. Above 5 GHz frequency validity can be extended to ± 110 MHz.

^C At frequency below 3 GHz, the validity of tissue parameters (x and y) can be relaxed to $\pm 10\%$ if liquid compensation formula is applied to measured SAR values. At frequencies above 3 GHz, the validity of tissue parameters (x and y) is restricted to $\pm 5\%$. The uncertainty is the RSS of the ConvF uncertainty for indicated target tissue parameters.


^D Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than $\pm 1\%$ for frequencies below 3 GHz and below $\pm 2\%$ for the frequencies between 3-6 GHz at any distance larger than half the probe tip diameter from the boundary.

Certificate No: Z18-60398

Page 6 of 11



**Frequency Response of E-Field
(TEM-Cell: if110 EXX, Waveguide: R22)**




Uncertainty of Frequency Response of E-field: $\pm 7.4\%$ (k=2)

Certificate No: Z18-60398

Page 7 of 11


Receiving Pattern (Φ), $\theta=0^\circ$



Certificate No: Z18-60398

Page 8 of 11


Dynamic Range f(SAR_{head}) (TEM cell, f = 900 MHz)




Certificate No: Z18-60398

Page 9 of 11

Conversion Factor Assessment



Deviation from Isotropy in Liquid



Certificate No: Z18-60398

Page 10 of 11

Appendix: Modulation Calibration Parameters

UID	Communication System Name	PAR	A dB	B dB-μV	C	VR mV	Unc ¹ (k=2)
0	CW	0.00	X 0.0	0.0	1.0	282.3	±2.5%
			Y 0.0	0.0	1.0	280.9	
			Z 0.0	0.0	1.0	275.1	
10012	IEEE 802.11b WiFi 2.4 GHz (DSSS, 1 Mbps)	1.87	X 2.77	68.02	18.46	143.0	±1.8%
			Y 2.75	68.05	18.52	145.0	
			Z 2.71	67.79	18.25	142.3	
10100	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	5.67	X 6.13	66.4	18.97	141.9	±1.9%
			Y 6.15	66.49	19.06	144.2	
			Z 6.09	66.32	18.90	140.9	
10108	LTE-FDD (SC-FDMA, 100% RB, 10 MHz, QPSK)	5.80	X 6.09	66.24	19.07	139.5	±1.9%
			Y 6.10	66.33	19.15	141.5	
			Z 6.05	66.19	19.05	138.0	
10154	LTE-FDD (SC-FDMA, 50% RB, 10 MHz, QPSK)	5.75	X 5.81	65.85	18.93	136.1	±1.9%
			Y 5.82	65.92	19.01	137.8	
			Z 5.79	65.89	18.97	134.7	
10169	LTE-FDD (SC-FDMA, 1 RB, 20 MHz, QPSK)	5.73	X 4.84	65.92	19.20	130.8	±1.9%
			Y 4.82	65.98	19.27	131.3	
			Z 4.80	66.00	19.29	129.1	
10175	LTE-FDD (SC-FDMA, 1 RB, 10 MHz, QPSK)	5.72	X 4.88	66.14	19.40	131.6	±1.9%
			Y 4.83	66.08	19.33	130.9	
			Z 4.79	66.02	19.29	129.3	
10297	LTE-FDD (SC-FDMA, 50% RB, 20 MHz, QPSK)	5.81	X 6.19	66.61	19.42	141.9	±1.9%
			Y 6.13	66.43	19.26	140.7	
			Z 6.14	66.52	19.33	139.6	

Certificate No: Z17-97142

Page 12 of 12

DAE4 Sn:720

In Collaboration with
TTL **s p e e a**
CALIBRATION LABORATORY

Add: No.31 Xuyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62594633-2512 Fax: +86-10-62594633-2504
E-mail: csl@chinafl.com Http://www.chinafl.cn

中国认可
国际互认
校准
CNAS L6570

Client: SRTC Certificate No: Z18-60399

CALIBRATION CERTIFICATE

Object: DAE4 - SN: 720

Calibration Procedure(s): FF-Z11-002-01
Calibration Procedure for the Data Acquisition Electronics (DAEx)

Calibration date: October 15, 2018

This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements(SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature(23±3)°C and humidity<70%.

Calibration Equipment used (M&TE critical for calibration)

Primary Standards	ID #	Cal Date(Calibrated by, Certificate No.)	Scheduled Calibration
Process Calibrator 753	1971018	20-Jun-18 (CTTL, No.J18X05034)	June-19

Calibrated by: Yu Zongying SAR Test Engineer
Reviewed by: Lin Hao SAR Test Engineer
Approved by: Qi Diaryuan SAR Project Leader

Signature: [Signatures]
Issued: October 17, 2018

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: Z18-60399

Page 1 of 3

In Collaboration with
TTL **s p e e a**
CALIBRATION LABORATORY

Add: No.31 Xuyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62594633-2512 Fax: +86-10-62594633-2504
E-mail: csl@chinafl.com Http://www.chinafl.cn

Glossary:

DAE: data acquisition electronics
Connector angle: information used in DASY system to align probe sensor X to the robot coordinate system.

Methods Applied and Interpretation of Parameters:

- DC Voltage Measurement:** Calibration Factor assessed for use in DASY system by comparison with a calibrated instrument traceable to national standards. The figure given corresponds to the full scale range of the voltmeter in the respective range.
- Connector angle:** The angle of the connector is assessed measuring the angle mechanically by a tool inserted. Uncertainty is not required.
- The report provide only calibration results for DAE, it does not contain other performance test results.

Certificate No: Z18-60399

Page 2 of 3

In Collaboration with
TTL **s p e e a**
CALIBRATION LABORATORY

Add: No.31 Xuyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62594633-2512 Fax: +86-10-62594633-2504
E-mail: csl@chinafl.com Http://www.chinafl.cn

DC Voltage Measurement

A/D Converter Resolution nominal
High Range: 1LSB = 6.1µV, full range = -100...+300 mV
Low Range: 1LSB = 61nV, full range = -1...+3mV
DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

Calibration Factors	X	Y	Z
High Range	403.343 ± 0.15% (k=2)	404.773 ± 0.15% (k=2)	403.205 ± 0.15% (k=2)
Low Range	3.95574 ± 0.7% (k=2)	3.95559 ± 0.7% (k=2)	3.95585 ± 0.7% (k=2)

Connector Angle

Connector Angle to be used in DASY system	29° ± 1°
---	----------

Certificate No: Z18-60399

Page 3 of 3

EX3DV4 Sn:3708 (1/7)

Calibration Laboratory of
Schmid & Partner
Engineering AG
Zugstrasse 43, 8034 Zurich, Switzerland

Accredited by the Swiss Accreditation Service (SAS)
The Swiss Accreditation Service is one of the signatories to the EA
Multilateral Agreement for the recognition of calibration certificates

Client: **SRTC (Auden)**

Accreditation No.: **SCS 0108**

Certificate No.: **EX3-3708_Oct18**

CALIBRATION CERTIFICATE

Object: **EX3DV4 - SN:3708**

Calibration procedure(s): **QA CAL-01-v9, QA CAL-12-v9, QA CAL-14-v4, QA CAL-25-v5,
QA CAL-25-v6
Calibration procedure for dosimetric E-field probes**

Calibration date: **October 22, 2018**

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI).
The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature (22 ± 0.2°C and humidity < 70%).

Calibration Equipment used (MPE critical for calibration)

Primary Standards	ID	Cal Data (Certificate No.)	Scheduled Calibration
Power meter NRP	SN: 104779	06-Apr-18 (No. 217-2087202923)	Apr-19
Power sensor NRP-231	SN: 103244	06-Apr-18 (No. 217-2087202923)	Apr-19
Reference 50 Ω Attenuator	SN: 103245	06-Apr-18 (No. 217-2087202923)	Apr-19
Reference Probe E8303V	SN: 59877 (270)	06-Apr-18 (No. 217-2087202923)	Apr-19
Reference Probe E8303V	SN: 3013	30-Dec-17 (No. 203-9113_Cert17)	Dec-18
DASY	SN: 680	21-Jan-17 (No. 040-4905_Cert17)	Dec-18

Secondary Standards	ID	Check Data (in house)	Scheduled Check
Power meter S44188	SN: 10841020214	06-Apr-18 (in house check Jun-18)	In house check Jun-20
Power sensor S4413A	SN: MY11083027	06-Apr-18 (in house check Jun-18)	In house check Jun-20
Power sensor S4413A	SN: 202110212	06-Apr-18 (in house check Jun-18)	In house check Jun-20
RF generator HP 8544C	SN: 10320405102	06-Apr-18 (in house check Jun-18)	In house check Jun-20
Network Analyser F5550A	SN: 10541090477	31-Mar-14 (in house check Oct-16)	In house check Oct-18

Calibrated by: **Gisela LAMAR** (Function: Laboratory Technician)

Approved by: **Kjetil PHILIP** (Function: Technical Manager)

Issued: October 23, 2018

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No.: EX3-3708_Oct18 Page 1 of 39

EX3DV4 - SN:3708 October 22, 2018

Probe EX3DV4

SN:3708

Manufactured: July 21, 2009
Calibrated: October 22, 2018

Calibrated for DASY/EASY Systems
(Note: non-compatible with DASY2 system!)

Certificate No.: EX3-3708_Oct18 Page 3 of 39

Calibration Laboratory of
Schmid & Partner
Engineering AG
Zugstrasse 43, 8034 Zurich, Switzerland

Accredited by the Swiss Accreditation Service (SAS)
The Swiss Accreditation Service is one of the signatories to the EA
Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 0108**

Glossary:

TSL: Issue simulating liquid
sensitivity in base space
sensitivity in TSL: NORM_{xy,z}
dCP: diode compression point
crest factor (100μs cycle) of the RF signal
modulation dependent inactivation parameters
A, B, C, D: a rotation around probe axis
Polarization: a rotation around probe axis
Connector Angle: information used in DASY system to align probe sensor X to the robot coordinate system

Calibration is Performed According to the Following Standards:

- IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013.
- IEC 62209-1, "Measurement procedure for the assessment of Specific Absorption Rate (SAR) from head-held and body-mounted devices used next to the ear (frequency range of 300 MHz to 8 GHz)", July 2016.
- IEC 62209-2, "Procedure to determine the Specific Absorption Rate (SAR) for wireless communication devices used in close proximity to the human body (frequency range of 30 MHz to 8 GHz)", March 2019.
- ISO 86564, "SAR Measurement Requirements for 100 MHz to 8 GHz"

Methods Applied and Interpretation of Parameters:

- NORM_{xy,z}: Assessed for E-field polarization $\theta = 0$ (if ≤ 900 MHz in TEM-cell; $f > 1800$ MHz: R22 waveguide). NORM_{xy,z} are only intermediate values, i.e., the uncertainties of NORM_{xy,z} does not affect the E-field uncertainty inside TSL (see below ConfF).
- NORM_{xy,z} = NORM_{xy,z} * frequency response (see Frequency Response Chart). This linearization is implemented in DASY4 software versions later than 4.2. The uncertainty of the frequency response is included in the stated uncertainty of ConfF.
- DCP_{xy,z}: DCP are numerical inactivation parameters assessed based on the data of power sweep with CW signal (no uncertainty required). DCP does not depend on frequency nor media.
- PAR: PAR is the Peak to Average Ratio that is not calibrated but determined based on the signal characteristics.
- A_{xy,z}, B_{xy,z}, C_{xy,z}, D_{xy,z}, VR_{xy,z}: A, B, C, D are numerical inactivation parameters assessed based on the data of power sweep for specific modulation signal. The parameters do not depend on frequency nor media. VR is the maximum calibration range expressed in dBm voltage across the diode.
- ConfF and Boundary Effect Parameters: Assessed in flat phantom using E-field (or Temperature Transfer Standard for $f \leq 300$ MHz) and inside waveguide using analytical field distributions based on power measurements for $f > 300$ MHz. The same setups are used for assessment of the parameters applied for boundary compensation (alpha, depth) of which typical uncertainty values are given. These parameters are used in DASY4 software to improve probe accuracy close to the boundary. The sensitivity in TSL corresponds to NORM_{xy,z} * ConfF whereby the uncertainty corresponds to that given for ConfF. A frequency dependent ConfF is used in DASY version 4.4 and higher which allows extending the validity from ± 50 MHz to ± 100 MHz.
- Spherical isotropy (3D deviation from isotropy): in a field of low gradients realized using a flat phantom exposed by a patch antenna.
- Sensor Offset: The sensor offset corresponds to the offset of virtual measurement center from the probe tip (on probe axis). No tolerance required.
- Connector Angle: The angle is assessed using the information gained by determining the NORM_{xy,z} (no uncertainty required).

Certificate No.: EX3-3708_Oct18 Page 2 of 39

EX3DV4 - SN:3708 October 22, 2018

DASY/EASY - Parameters of Probe: EX3DV4 - SN:3708

Basic Calibration Parameters

	Sensor X	Sensor Y	Sensor Z	Unc (k=2)
Norm (μV/m) ^{1,2}	0.20	0.35	0.42	± 10.1 %
DCP (mV)	95.4	103.6	100.4	

Modulation Calibration Parameters

URD	Communication System Name	A	B	C	D	VR	Unc ³ (k=2)
0	CW	X	0.0	0.0	1.0	0.00	134.7 ± 5.5 %
		Y	0.0	0.0	1.0	130.0	
		Z	0.0	0.0	1.0	146.0	

Note: For details on URD parameters see Appendix.

Sensor Model Parameters

	C1	C2	θ	T1	T2	T3	T4	T5	T6
ff	33.84	270.1	40.07	9.383	1.302	0.022	0.006	0.762	1.008
Y	40.04	291.1	34.06	15.47	0.801	0.012	1.756	0.156	1.005
Z	38.94	282.2	35.97	11.65	0.890	0.001	0.000	0.009	1.008

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

¹ The uncertainties of Norm X,Y,Z do not affect the E-field uncertainty inside TSL (see Pages 2 and 6).
² Numerical inactivation parameter: uncertainty not required.
³ Uncertainty is determined using the rule: deviation from linear response applying rectangular distribution and is expressed for the square of the field value.

Certificate No.: EX3-3708_Oct18 Page 4 of 39

EX3DV4 Sn:3708 (2/7)

EX3DV4- SN:3708

October 22, 2018

DASY/EASY - Parameters of Probe: EX3DV4 - SN:3708

Calibration Parameter Determined in Head Tissue Simulating Media

f (MHz)	Relative Permittivity ^a	Conductivity (S/m) ^b	Conc ^c X	Conc ^c Y	Conc ^c Z	Alpha ^d	Depth ^e (mm)	Unc (k=2)
450	43.5	0.87	9.79	9.79	9.79	0.14	1.20	± 13.3 %
750	41.9	0.89	9.59	9.59	9.59	0.54	0.90	± 12.0 %
835	41.5	0.90	9.16	9.16	9.16	0.51	0.82	± 12.0 %
1450	40.5	1.20	8.50	8.50	8.50	0.33	0.80	± 12.0 %
1750	40.1	1.37	8.20	8.20	8.20	0.28	0.94	± 12.0 %
1900	40.0	1.40	7.89	7.89	7.89	0.35	0.85	± 12.0 %
2000	40.0	1.40	7.86	7.86	7.86	0.34	0.80	± 12.0 %
2300	39.5	1.87	7.51	7.51	7.51	0.29	0.86	± 12.0 %
2450	39.2	1.80	7.13	7.13	7.13	0.38	0.86	± 12.0 %
2600	39.0	1.96	7.01	7.01	7.01	0.38	0.87	± 12.0 %
5200	35.0	4.68	5.48	5.48	5.48	0.40	1.80	± 13.1 %
5300	35.9	4.78	5.25	5.25	5.25	0.40	1.80	± 13.1 %
5600	35.5	5.07	4.84	4.84	4.84	0.40	1.80	± 13.1 %
5800	35.3	5.27	5.04	5.04	5.04	0.40	1.80	± 13.1 %

^a Frequency validity above 300 MHz of a 100 MHz only applies for DASY v4.4 and higher (see Page 2), when it is restricted to ± 10 MHz. The uncertainty is the RSS of the Conc^c uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below 300 MHz is a 10, 25, 40, 50 and 70 MHz for Conc^c assessments at 30, 64, 128, 150 and 220 MHz respectively. Above 5 GHz frequency validity can be extended to a 100 MHz.

^b At frequencies below 3 GHz, the validity of tissue parameters (u and v) can be relaxed to ± 10% if liquid compensation formula is applied to measured S₂₁ values. At frequencies above 3 GHz, the validity of tissue parameters (u and v) is restricted to ± 5%. The uncertainty is the RSS of the Conc^c uncertainty for indicated target tissue parameters.

^c Accuracy is determined during calibration. SPTAC warrants that the remaining deviation due to the boundary effect after compensation is always less than ± 1% for frequencies below 3 GHz and below ± 2% for frequencies between 3-6 GHz at any distance larger than half the probe tip diameter from the boundary.

Certificate No: EX3-3708_Oct18

Page 5 of 39

EX3DV4- SN:3708

October 22, 2018

DASY/EASY - Parameters of Probe: EX3DV4 - SN:3708

Calibration Parameter Determined in Body Tissue Simulating Media

f (MHz)	Relative Permittivity ^a	Conductivity (S/m) ^b	Conc ^c X	Conc ^c Y	Conc ^c Z	Alpha ^d	Depth ^e (mm)	Unc (k=2)
450	56.7	0.94	10.35	10.35	10.35	0.08	1.20	± 13.3 %
750	55.9	0.96	9.51	9.51	9.51	0.50	0.85	± 12.0 %
835	55.2	0.97	9.33	9.33	9.33	0.47	0.84	± 12.0 %
1450	54.0	1.30	7.84	7.84	7.84	0.38	0.80	± 12.0 %
1750	53.4	1.49	7.69	7.69	7.69	0.37	0.85	± 12.0 %
1900	53.3	1.52	7.56	7.56	7.56	0.42	0.84	± 12.0 %
2000	53.3	1.52	7.53	7.53	7.53	0.41	0.85	± 12.0 %
2300	52.9	1.81	7.34	7.34	7.34	0.39	0.88	± 12.0 %
2450	52.7	1.95	7.19	7.19	7.19	0.32	0.95	± 12.0 %
2600	52.5	2.16	7.14	7.14	7.14	0.32	0.95	± 12.0 %
5200	49.0	5.30	4.53	4.53	4.53	0.50	1.80	± 13.1 %
5300	49.9	5.42	4.41	4.41	4.41	0.50	1.80	± 13.1 %
5600	48.5	5.77	3.99	3.99	3.99	0.50	1.80	± 13.1 %
5800	48.2	6.00	4.21	4.21	4.21	0.50	1.80	± 13.1 %

^a Frequency validity above 300 MHz of a 100 MHz only applies for DASY v4.4 and higher (see Page 2), when it is restricted to ± 10 MHz. The uncertainty is the RSS of the Conc^c uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below 300 MHz is a 10, 25, 40, 50 and 70 MHz for Conc^c assessments at 30, 64, 128, 150 and 220 MHz respectively. Above 5 GHz frequency validity can be extended to a 100 MHz.

^b At frequencies below 3 GHz, the validity of tissue parameters (u and v) can be relaxed to ± 10% if liquid compensation formula is applied to measured S₂₁ values. At frequencies above 3 GHz, the validity of tissue parameters (u and v) is restricted to ± 5%. The uncertainty is the RSS of the Conc^c uncertainty for indicated target tissue parameters.

^c Accuracy is determined during calibration. SPTAC warrants that the remaining deviation due to the boundary effect after compensation is always less than ± 1% for frequencies below 3 GHz and below ± 2% for frequencies between 3-6 GHz at any distance larger than half the probe tip diameter from the boundary.


Certificate No: EX3-3708_Oct18

Page 6 of 39

EX3DV4- SN:3708

October 22, 2018

Frequency Response of E-Field (TEM-Cell:R110 EXX, Waveguide: R22)



Uncertainty of Frequency Response of E-field: ± 6.3% (k=2)



Certificate No: EX3-3708_Oct18

Page 7 of 39

EX3DV4- SN:3708

October 22, 2018

Receiving Pattern (φ), θ = 0°



Uncertainty of Axial Isotropy Assessment: ± 0.5% (k=2)

Certificate No: EX3-3708_Oct18



Page 8 of 39

EX3DV4 Sn:3708 (3/7)

EX3DV4-SN:3708

October 22, 2018

Dynamic Range f(SAR_{head}) (TEM cell, f_{max}=1900 MHz)



Uncertainty of Linearity Assessment: ± 0.6% (k=2)

Certificate No: EX3-3708_Oct18


Page 9 of 39

EX3DV4-SN:3708


October 22, 2018

Conversion Factor Assessment


f = 835 MHz, WGLS R9 (H₁ conf)



f = 1900 MHz, WGLS R22 (H₁ conf)



Deviation from Isotropy in Liquid Error (h, θ), f = 900 MHz



Uncertainty of Spherical Isotropy Assessment: ± 2.5% (k=2)

Certificate No: EX3-3708_Oct18

Page 10 of 39

EX3DV4-SN:3708

October 22, 2018

DASY/EASY - Parameters of Probe: EX3DV4 - SN:3708

Other Probe Parameters

Sensor Arrangement	Triangular
Connector Angle (°)	0.5
Mechanical Surface Detection Mode	enabled
Optical Surface Detection Mode	disabled
Probe Overall Length	337 mm
Probe Body Diameter	10 mm
Tip Length	9 mm
Tip Diameter	2.5 mm
Probe Tip to Sensor X Calibration Point	1 mm
Probe Tip to Sensor Y Calibration Point	1 mm
Probe Tip to Sensor Z Calibration Point	1 mm
Recommended Measurement Distance from Surface	1.4 mm

Certificate No: EX3-3708_Oct18

Page 11 of 39

EX3DV4 Sn:3708 (4/7)

EX3DV4- SN:3708

October 22, 2018

Appendix: Modulation Calibration Parameters

Mod	Communication System Name	A	B	C	D	E	F	G	H	I	J	K	L	M	N	O	P	Q	R	S	T	U	V	W	X	Y	Z	AA	AB	AC	AD	AE	AF	AG	AH	AI	AJ	AK	AL	AM	AN	AO	AP	AQ	AR	AS	AT	AU	AV	AW	AX	AY	AZ	BA	BB	BC	BD	BE	BF	BG	BH	BI	BJ	BK	BL	BM	BN	BO	BP	BQ	BR	BS	BT	BU	BV	BW	BX	BY	BZ	CA	CB	CC	CD	CE	CF	CG	CH	CI	CJ	CK	CL	CM	CN	CO	CP	CQ	CR	CS	CT	CU	CV	CW	CX	CY	CZ	DA	DB	DC	DD	DE	DF	DG	DH	DI	DJ	DK	DL	DM	DN	DO	DP	DQ	DR	DS	DT	DU	DV	DW	DX	DY	DZ	EA	EB	EC	ED	EE	EF	EG	EH	EI	EJ	EK	EL	EM	EN	EO	EP	EQ	ER	ES	ET	EU	EV	EW	EX	EY	EZ	FA	FB	FC	FD	FE	FF	FG	FH	FI	FJ	FK	FL	FM	FN	FO	FP	FQ	FR	FS	FT	FU	FV	FW	FX	FY	FZ	GA	GB	GC	GD	GE	GF	GG	GH	GI	GJ	GK	GL	GM	GN	GO	GP	GQ	GR	GS	GT	GU	GV	GW	GX	GY	GZ	HA	HB	HC	HD	HE	HF	HG	HH	HI	IJ	HK	HL	HM	HN	HO	HP	HQ	HR	HS	HT	HU	HV	HW	HX	HY	HZ	IA	IB	IC	ID	IE	IF	IG	IH	II	IJ	IK	IL	IM	IN	IO	IP	IQ	IR	IS	IT	IU	IV	IW	IX	IY	IZ	JA	JB	JC	JD	JE	JF	JG	JH	JI	IJ	JK	KL	LM	LN	LO	LP	LQ	LR	LS	LT	LU	LV	LW	LX	LY	LZ	MA	MB	MC	MD	ME	MF	MG	MH	MI	IJ	MK	ML	MM	MN	MO	MP	MQ	MR	MS	MT	MU	MV	MW	MX	MY	MZ	NA	NB	NC	ND	NE	NF	NG	NH	NI	IJ	NK	NL	NM	NN	NO	NP	NQ	NR	NS	NT	NU	NV	NW	NX	NY	NZ	OA	OB	OC	OD	OE	OF	OG	OH	OI	OJ	OK	OL	OM	ON	OO	OP	OQ	OR	OS	OT	OU	OV	OW	OX	OY	OZ	PA	PB	PC	PD	PE	PF	PG	PH	PI	IJ	PK	PL	PM	PN	PO	PP	PQ	PR	PS	PT	PU	PV	PW	PX	PY	PZ	QA	QB	QC	QD	QE	QF	QG	QH	QI	QJ	QK	QL	QM	QN	QO	QP	QQ	QR	QS	QT	QU	QV	QW	QX	QY	QZ	RA	RB	RC	RD	RE	RF	RG	RH	RI	IJ	RK	RL	RM	RN	RO	RP	RQ	RR	RS	RT	RU	RV	RW	RX	RY	RZ	SA	SB	SC	SD	SE	SF	SG	SH	SI	IJ	SK	SL	SM	SN	SO	SP	SQ	SR	SS	ST	SU	SV	SW	SX	SY	SZ	TA	TB	TC	TD	TE	TF	TG	TH	TI	IJ	TK	TL	TM	TN	TO	TP	TQ	TR	TS	TT	TU	TV	TW	TX	TY	TZ	UA	UB	UC	UD	UE	UF	UG	UH	UI	IJ	UK	UL	UM	UN	UO	UP	UQ	UR	US	UT	UU	UV	UW	UX	UY	UZ	VA	VB	VC	VD	VE	VF	VG	VH	VI	IJ	VK	VL	VM	VN	VO	VP	VQ	VR	VS	VT	VU	VV	VW	VX	VY	VZ	WA	WB	WC	WD	WE	WF	WG	WH	WI	IJ	WK	WL	WM	WN	WO	WP	WQ	WR	WS	WT	WU	WV	WW	WX	WY	WZ	XA	XB	XC	XD	XE	XF	YG	XH	XI	IJ	XK	XL	XM	XN	XO	XP	XQ	XR	XS	XT	XU	XV	XW	XX	XY	XZ	YA	YB	YC	YD	YE	YF	YG	YH	YI	IJ	YK	YL	YM	YN	YO	YP	YQ	YR	YS	YT	YU	YV	YW	YX	YY	YZ	ZA	ZB	ZC	ZD	ZE	ZF	ZG	ZH	ZI	IJ	ZK	ZL	ZM	ZN	ZO	ZP	ZQ	ZR	ZS	ZT	ZU	ZV	ZW	ZX	ZY	ZZ
0	CW	X	0.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00	1.00	0.00																																																																																																																																																																																																									

Certificate No: EX3-3708_Oct18 Page 12 of 39

EX3DV4- SN:3708		October 22, 2018																			
10001- CAB	IEEE 802.11b WFI 2.4 GHz (DSSS, 11 Mbps)	X	2.80	60.33	21.19	2.04	110.0	± 9.6 %													
		Y	3.78	84.50	20.28		110.0														
		Z	2.28	78.84	18.88		110.0														
		X	4.30	66.90	19.01	0.48	100.0	± 9.6 %													
		Y	4.55	68.86	18.90		100.0														
		Z	4.46	64.43	18.79		100.0														
10003- CAB	IEEE 802.11b WFI 5 GHz (OFDM, 48 Mbps)	X	4.38	68.68	16.55	0.72	100.0	± 9.6 %													
		Y	4.59	68.90	16.56		100.0														
		Z	4.52	68.92	16.79		100.0														
10004- CAB	IEEE 802.11a WFI 5 GHz (OFDM, 12 Mbps)	X	4.81	68.68	16.75	0.06	100.0	± 9.6 %													
		Y	4.82	67.17	16.77		100.0														
		Z	4.73	68.70	16.50		100.0														
10005- CAB	IEEE 802.11a WFI 5 GHz (OFDM, 18 Mbps)	X	4.63	67.94	16.81	1.21	100.0	± 9.6 %													
		Y	4.79	67.94	16.81		100.0														
		Z	4.61	66.63	16.58		100.0														
10006- CAB	IEEE 802.11a WFI 5 GHz (OFDM, 24 Mbps)	X	4.51	66.75	16.87	1.45	100.0	± 9.6 %													
		Y	4.70	67.04	16.69		100.0														
		Z	4.63	66.88	16.74		100.0														
10007- CAB	IEEE 802.11a WFI 5 GHz (OFDM, 36 Mbps)	X	4.82	67.08	16.66	2.04	100.0	± 9.6 %													
		Y	4.93	67.34	17.42		100.0														
		Z	4.85	66.92	17.42		100.0														
10008- CAB	IEEE 802.11a WFI 5 GHz (OFDM, 48 Mbps)	X	4.80	67.08	16.83	2.35	100.0	± 9.6 %													
		Y	4.93	67.30	17.58		100.0														
		Z	4.97	66.90	17.40		100.0														
10009- CAB	IEEE 802.11a WFI 5 GHz (OFDM, 54 Mbps)	X	4.92	67.07	17.62	2.67	100.0	± 9.6 %													
		Y	5.10	67.23	17.77		100.0														
		Z	5.05	66.88	17.82		100.0														
10011- CAB	IEEE 802.11g WFI 2.4 GHz (DSSS/OFDM, 4 Mbps)	X	4.69	66.77	17.34	1.99	100.0	± 9.6 %													
		Y	4.83	66.52	17.28		100.0														
		Z	4.78	66.60	17.33		100.0														
10012- CAB	IEEE 802.11g WFI 2.4 GHz (DSSS/OFDM, 12 Mbps)	X	4.84	67.04	17.63	2.30	100.0	± 9.6 %													
		Y	4.81	67.23	17.47		100.0														
		Z	4.75	66.87	17.38		100.0														
10013- CAB	IEEE 802.11g WFI 2.4 GHz (DSSS/OFDM, 18 Mbps)	X	4.74	67.28	17.87	2.83	100.0	± 9.6 %													
		Y	4.88	67.39	17.78		100.0														
		Z	4.83	67.08	17.80		100.0														
10014- CAB	IEEE 802.11g WFI 2.4 GHz (DSSS/OFDM, 24 Mbps)	X	4.88	67.32	17.93	3.30	100.0	± 9.6 %													
		Y	4.87	67.35	17.93		100.0														
10015- CAB	IEEE 802.11g WFI 2.4 GHz (DSSS/OFDM, 36 Mbps)	X	4.81	67.33	18.08	3.82	90.0	± 9.6 %													
		Y	4.91	67.40	18.20		90.0														
		Z	4.87	67.11	18.03		90.0														
10016- CAB	IEEE 802.11g WFI 2.4 GHz (DSSS/OFDM, 48 Mbps)	X	4.85	67.35	18.48	6.15	90.0	± 9.6 %													
		Y	4.95	67.35	18.34		90.0														
		Z	4.81	67.20	18.30		90.0														
10017- CAB	IEEE 802.11g WFI 2.4 GHz (DSSS/OFDM, 54 Mbps)	X	4.90	67.36	18.58	4.30	90.0	± 9.6 %													
		Y	4.96	67.36	18.45		90.0														
		Z	4.86	67.20	18.48		90.0														

Certificate No: EX3-3708_Oct18 Page 14 of 39

EX3DV4- SN:3708		October 22, 2018																			
10113- CAG	LTE-FDD (SC-FDMA, 100% RB, 15 MHz, 84-QAM)	X	2.71	67.18	19.85	0.00	158.0	x 9.6%													
		Z	2.87	68.02	19.17		150.0														
		X	2.77	66.54	19.23		150.0														
		X	2.46	68.82	19.75	0.00	150.0	x 9.6%													
10113- CAG	LTE-FDD (SC-FDMA, 100% RB, 5 MHz, 56-QAM)	X	2.80	69.72	16.79		150.0														
		Z	2.87	69.48	17.08		150.0														
10114- CAG	IEEE 802.11n HT Greenfield, 13.5 MHz, 64-QAM	X	4.45	67.01	16.53	0.00	150.0	x 9.6%													
		Z	5.02	67.35	16.48		150.0														
		Z	4.61	66.86	16.18		150.0														
10115- CAG	IEEE 802.11n HT Greenfield, 81 Mbps, 16-QAM	X	5.11	67.01	16.53	0.00	150.0	x 9.6%													
		Z	5.25	67.38	16.50		150.0														
		Z	5.18	66.89	16.21		150.0														
10116- CAG	IEEE 802.11n HT Greenfield, 120 Mbps, 64-QAM	X	4.92	67.15	16.62	0.00	150.0	x 9.6%													
		Z	5.09	67.63	16.48		150.0														
		Z	4.99	67.03	16.18		150.0														
10117- CAG	IEEE 802.11n HT MIMO, 13.5 MHz, 64-QAM	X	4.82	66.87	16.48	0.00	150.0	x 9.6%													
		Z	4.98	67.28	16.45		150.0														
		Z	4.88	66.87	16.18		150.0														
10118- CAG	IEEE 802.11n HT MIMO, 81 Mbps, 16-QAM	X	5.21	67.43	16.54	0.00	150.0	x 9.6%													
		Z	5.31	67.56	16.58		150.0														
		Z	5.23	67.13	16.31		150.0														
10119- CAG	IEEE 802.11n HT MIMO, 120 Mbps, 64-QAM	X	5.18	67.63	16.57	0.00	150.0	x 9.6%													
		Z	5.07	67.50	16.48		150.0														
		Z	4.99	67.03	16.18		150.0														
10140- CAG	LTE-FDD (SC-FDMA, 100% RB, 15 MHz, 16-QAM)	X	3.06	67.10	15.77	0.00	150.0	x 9.6%													
		Z	3.31	67.86	16.18		150.0														
		Z	3.31	68.07	16.45		150.0														
10141- CAG	LTE-FDD (SC-FDMA, 100% RB, 15 MHz, 64-QAM)	X	3.20	67.31	16.00	0.00	150.0	x 9.6%													
		Z	3.44	68.10	16.35		150.0														
10142- CAG	LTE-FDD (SC-FDMA, 100% RB, 3 MHz, QPSK)	X	2.32	66.56	10.93		150.0														
		Z	1.58	67.63	10.87		150.0	x 9.6%													
10143- CAG	LTE-FDD (SC-FDMA, 100% RB, 3 MHz, 16-QAM)	X	2.02	70.37	16.43		150.0														
		Z	1.63	68.79	14.18		150.0														
		X	1.85	67.47	13.73		150.0	x 9.6%													
10144- CAG	LTE-FDD (SC-FDMA, 100% RB, 3 MHz, 64-QAM)	X	2.06	70.80	16.42		150.0														
		Z	2.08	70.78	16.36		150.0														
10145- CAG	LTE-FDD (SC-FDMA, 100% RB, 1.4 MHz, QPSK)	X	1.65	64.41	11.60		150.0														
		X	2.15	67.28	14.13		150.0														
		Z	1.87	68.02	14.07		150.0														
10146- CAG	LTE-FDD (SC-FDMA, 100% RB, 1.4 MHz, 16-QAM)	X	0.97	68.00	9.00	0.00	150.0	x 9.6%													
		X	1.62	64.21	10.44		150.0														
		Z	0.77	67.08	7.95		150.0														
10147- CAG	LTE-FDD (SC-FDMA, 100% RB, 1.4 MHz, 16-QAM)	X	0.83	70.60	9.00	0.00	150.0	x 9.6%													
		X	1.48	63.60	9.05		150.0														
		Z	1.17	67.69	7.95		150.0														
		X	0.94	66.95	8.95		150.0	x 9.6%													
10148- CAG	LTE-FDD (SC-FDMA, 100% RB, 1.4 MHz, 16-QAM)	X	1.04	64.86	9.75		150.0														
		Z	1.24	62.17	8.36		150.0														

Certificate No: EX3-3708_Oct18 Page 15 of 39

EX3DV4 Sn:3708 (5/7)

EX3DV4-SN3708									
October 22, 2018									
10223-CAE	IEEE 802.11n (HT Mixed, 90 Mbps, 16-QAM)	X	5.00	67.07	18.58	0.00	100.0	±0.6 %	
		Y	5.23	67.44	18.54		100.0		
		Z	5.15	67.00	18.58		100.0		
10224-CAE	IEEE 802.11n (HT Mixed, 150 Mbps, 64-QAM)	X	4.84	67.01	18.49	0.00	100.0	±0.6 %	
		Y	4.59	67.37	18.42		100.0		
		Z	4.50	66.88	18.11		100.0		
10225-CAB	UMTS-FDD (HSPA+)	X	2.44	65.81	14.45	0.00	100.0	±0.6 %	
		Y	2.73	66.77	15.43		100.0		
		Z	2.58	65.53	14.45		100.0		
10226-CAA	LTE-TDD (SC-FDMA, 1 RB, 1.4 MHz, 16-QAM)	X	8.96	67.92	24.97	0.02	65.0	±0.6 %	
		Y	31.18	109.58	31.14		65.0		
		Z	8.89	68.89	25.48		65.0		
10227-CAA	LTE-TDD (SC-FDMA, 1 RB, 1.4 MHz, 64-QAM)	X	8.31	69.80	23.59	0.02	65.0	±0.6 %	
		Y	25.63	104.26	28.94		65.0		
		Z	8.57	67.84	24.30		65.0		
10228-CAA	LTE-TDD (SC-FDMA, 1 RB, 1.4 MHz, QPSK)	X	6.80	67.47	26.62	0.02	65.0	±0.6 %	
		Y	11.48	27.68	32.08		65.0		
		Z	6.33	66.07	26.42		65.0		
10229-CAE	LTE-TDD (SC-FDMA, 1 RB, 3 MHz, 16-QAM)	X	0.00	60.01	34.44	0.04	60.0	±0.6 %	
		Y	27.05	106.96	35.33		65.0		
		Z	0.36	67.57	25.02		65.0		
10230-CAE	LTE-TDD (SC-FDMA, 1 RB, 3 MHz, 64-QAM)	X	7.76	64.41	23.10	0.02	65.0	±0.6 %	
		Y	22.43	101.87	28.22		65.0		
		Z	8.06	65.94	23.27		65.0		
10231-CAE	LTE-TDD (SC-FDMA, 1 RB, 3 MHz, QPSK)	X	5.41	66.43	26.29	0.02	65.0	±0.6 %	
		Y	10.71	66.21	29.53		65.0		
		Z	6.56	65.18	26.01		65.0		
10232-CAF	LTE-TDD (SC-FDMA, 1 RB, 5 MHz, 16-QAM)	X	8.37	66.60	24.43	0.02	65.0	±0.6 %	
		Y	27.05	106.94	35.33		65.0		
		Z	8.38	67.35	25.01		65.0		
10233-CAE	LTE-TDD (SC-FDMA, 1 RB, 5 MHz, 64-QAM)	X	7.75	64.38	23.10	0.02	65.0	±0.6 %	
		Y	22.36	101.83	28.21		65.0		
		Z	8.04	65.91	23.86		65.0		
10234-CAE	LTE-TDD (SC-FDMA, 1 RB, 5 MHz, QPSK)	X	6.12	65.18	25.79	0.02	65.0	±0.6 %	
		Y	10.08	64.86	28.87		65.0		
		Z	8.85	64.36	25.03		65.0		
10235-CAF	LTE-TDD (SC-FDMA, 1 RB, 10 MHz, 16-QAM)	X	8.39	66.63	24.43	0.02	65.0	±0.6 %	
		Y	27.08	107.00	35.34		65.0		
		Z	8.35	67.57	25.02		65.0		
10236-CAF	LTE-TDD (SC-FDMA, 1 RB, 10 MHz, 64-QAM)	X	7.81	64.48	23.12	0.02	65.0	±0.6 %	
		Y	22.71	102.14	28.26		65.0		
		Z	8.11	66.03	23.90		65.0		
10237-CAF	LTE-TDD (SC-FDMA, 1 RB, 10 MHz, QPSK)	X	6.42	66.27	26.30	0.02	65.0	±0.6 %	
		Y	15.73	66.29	29.36		65.0		
		Z	6.06	65.19	25.02		65.0		
10238-CAF	LTE-TDD (SC-FDMA, 1 RB, 15 MHz, 16-QAM)	X	8.36	66.67	24.43	0.02	65.0	±0.6 %	
		Y	26.94	106.92	35.31		65.0		
		Z	8.33	67.52	25.00		65.0		

Certificate No: EX3-3708_Oct18

Page 20 of 39

EX3DV4-SN3708									
October 22, 2018									
10235-CAF	LTE-TDD (SC-FDMA, 10% RB, 15 MHz, QPSK)	X	6.62	76.10	20.55	3.98	65.0	±0.6 %	
		Y	6.54	78.26	21.42		65.0		
		Z	5.16	78.07	20.53		65.0		
10236-CAA	LTE-TDD (SC-FDMA, 100% RB, 1.4 MHz, 16-QAM)	X	2.45	63.21	18.47	3.98	65.0	±0.6 %	
		Y	3.51	67.37	12.58		65.0		
10237-CAA	LTE-TDD (SC-FDMA, 100% RB, 1.4 MHz, 64-QAM)	X	2.44	62.90	18.21	3.98	65.0	±0.6 %	
		Y	3.41	66.86	12.18		65.0		
		Z	3.56	65.78	17.83		65.0		
10238-CAA	LTE-TDD (SC-FDMA, 100% RB, 1.4 MHz, QPSK)	X	1.85	63.35	18.82	3.98	65.0	±0.6 %	
		Y	3.32	69.86	14.27		65.0		
		Z	2.73	67.24	12.82		65.0		
10239-CAE	LTE-TDD (SC-FDMA, 100% RB, 3 MHz, 16-QAM)	X	4.04	70.43	19.24	3.98	65.0	±0.6 %	
		Y	9.28	75.31	18.83		65.0		
		Z	4.62	72.23	17.83		65.0		
10240-CAE	LTE-TDD (SC-FDMA, 100% RB, 3 MHz, 64-QAM)	X	4.07	70.21	18.13	3.98	65.0	±0.6 %	
		Y	9.28	75.31	18.83		65.0		
		Z	4.62	72.23	17.83		65.0		
10241-CAE	LTE-TDD (SC-FDMA, 100% RB, 3 MHz, QPSK)	X	4.07	70.21	18.13	3.98	65.0	±0.6 %	
		Y	9.28	75.31	18.83		65.0		
		Z	4.62	72.23	17.83		65.0		
10242-CAF	LTE-TDD (SC-FDMA, 100% RB, 5 MHz, 16-QAM)	X	5.01	74.07	19.48	3.98	65.0	±0.6 %	
		Y	6.67	80.88	21.27		65.0		
		Z	5.33	72.24	19.86		65.0		
10243-CAF	LTE-TDD (SC-FDMA, 100% RB, 5 MHz, 64-QAM)	X	5.01	74.07	19.48	3.98	65.0	±0.6 %	
		Y	6.67	80.88	21.27		65.0		
		Z	5.33	72.24	19.86		65.0		
10243-CAF	LTE-TDD (SC-FDMA, 100% RB, 5 MHz, QPSK)	X	4.97	71.41	17.86	3.98	65.0	±0.6 %	
		Y	6.53	73.59	19.33		65.0		
		Z	5.02	72.30	18.64		65.0		
10244-CAF	LTE-TDD (SC-FDMA, 100% RB, 5 MHz, QPSK)	X	5.70	75.84	21.00	3.98	65.0	±0.6 %	
		Y	7.30	81.83	25.43		65.0		
		Z	5.81	78.88	21.13		65.0		
10245-CAF	LTE-TDD (SC-FDMA, 100% RB, 10 MHz, 16-QAM)	X	4.98	71.77	18.68	3.98	65.0	±0.6 %	
		Y	6.76	73.73	19.78		65.0		
		Z	5.37	72.49	19.12		65.0		
10246-CAF	LTE-TDD (SC-FDMA, 100% RB, 10 MHz, 64-QAM)	X	5.42	73.17	19.89	3.98	65.0	±0.6 %	
		Y	6.19	74.91	20.65		65.0		
		Z	5.72	73.47	20.01		65.0		
10247-CAF	LTE-TDD (SC-FDMA, 100% RB, 10 MHz, QPSK)	X	5.75	75.55	20.61	3.98	65.0	±0.6 %	
		Y	6.88	78.86	21.48		65.0		
		Z	5.05	76.58	20.85		65.0		
10248-CAF	LTE-TDD (SC-FDMA, 100% RB, 15 MHz, 16-QAM)	X	3.67	72.08	19.68	3.98	65.0	±0.6 %	
		Y	5.39	73.71	20.23		65.0		
		Z	3.88	72.55	19.72		65.0		
10249-CAF	LTE-TDD (SC-FDMA, 100% RB, 15 MHz, 64-QAM)	X	3.67	71.70	19.50	3.98	65.0	±0.6 %	
		Y	5.37	73.31	20.11		65.0		
		Z	3.89	72.18	19.61		65.0		
10250-CAF	LTE-TDD (SC-FDMA, 100% RB, 15 MHz, QPSK)	X	3.72	74.15	19.82	3.98	65.0	±0.6 %	
		Y	5.58	75.97	20.51		65.0		
		Z	4.01	74.38	19.88		65.0		

Certificate No: EX3-3708_Oct18

Page 22 of 39

EX3DV4-SN3708									
October 22, 2018									
10253-AAA	IEEE 802.11n WMMX (3T-16, 5mhz, 16QAM, PLUS)	X	4.86	67.09	18.50	4.98	50.0	±0.6 %	
		Y	4.83	66.81	17.92		50.0		
		Z	4.77	66.64	17.46		50.0		
10254-AAA	IEEE 802.11n WMMX (3T-16, 5mhz, 16QAM, PLUS)	X	4.60	66.87	17.82	4.17	50.0	±0.6 %	
		Y	4.64	65.92	17.45		50.0		
		Z	4.57	65.48	17.06		50.0		
10255-AAA	IEEE 802.11n WMMX (3T-16, 10mhz, 16QAM, PLUS, 18 symbols)	X	4.96	71.08	19.94	6.02	50.0	±0.6 %	
		Y	4.51	68.62	18.78		35.0		
		Z	4.50	68.96	19.48		35.0		
10256-AAA	IEEE 802.11n WMMX (3T-16, 10mhz, 10MHz, 64QAM, PLUS, 18 symbols)	X	4.69	68.93	19.42	6.02	35.0	±0.6 %	
		Y	4.49	67.30	19.20		35.0		
		Z	4.71	67.32	18.96		35.0		
10257-AAA	IEEE 802.11n WMMX (3T-16, 10mhz, 10MHz, QPSK, PLUS, 18 symbols)	X	4.82	69.24	19.38	6.02	35.0	±0.6 %	
		Y	4.60	67.21	18.18		35.0		
		Z	4.69	67.80	18.94		35.0		
10258-AAA	IEEE 802.11n WMMX (3T-16, 10mhz, 10MHz, 16QAM, PLUS)	X	4.94	69.40	19.09	6.02	35.0	±0.6 %	
		Y	4.60	67.78	19.36		35.0		
		Z	4.62	67.77	19.11		35.0		
10259-AAA	IEEE 802.11n WMMX (3T-16, 10mhz, 10MHz, 16QAM, AMC 2x3, 18 symbols)	X	4.90	68.88	19.51	6.02	35.0	±0.6 %	
		Y	4.73	67.44	19.31		35.0		
		Z	4.74	67.44	19.09		35.0		
10260-AAA	IEEE 802.11n WMMX (3T-16, 10mhz, 10MHz, QPSK, AMC 2x3, 18 symbols)	X	4.87	68.88	19.45	6.02	35.0	±0.6 %	
		Y	4.65	67.42	19.21		35.0		
		Z	4.68	67.43	18.99		35.0		
10261-AAA	LTE-FDD (SC-FDMA, 100% RB, 15 MHz, QPSK)	X	2.77	66.48	16.05	0.00	150.0	±0.6 %	
		Y	3.14	70.00	16.79		150.0		
		Z	2.75	67.58	15.39		150.0		
10262-AAA	IDEN 1.3	X	2.37	67.29	12.68	0.99	70.0	±0.6 %	
		Y	3.64	72.55	13.49		70.0		
		Z	2.67	66.89	14.17		70.0		
10263-AAA	IDEN 1.8	X	3.50	74.44	18.42	10.00	70.0	±0.6 %	
		Y	6.12	82.13	21.83		70.0		
		Z	4.18	78.04	19.55		70.0		
10264-AAA	IEEE 802.11n WFT 2.4 GHz (5555, 1 Mbps, 8000 duty cycle)	X	1.50	63.43	14.93	0.17	150.0	±0.6 %	
		Y	1.08	64.20	15.74		150.0		
		Z	1.27	65.77	14.10		150.0		
10265-AAA	IEEE 802.11n WFT 2.4 GHz (5555, 1 Mbps, 8000 duty cycle)	X	4.35	68.60	15.39	0.17	150.0	±0.6 %	
		Y	4.25	68.39	15.14		150.0		
		Z	4.40	68.39	15.14		150.0		
10266-AAA	IEEE 802.11n WFT 2.4 GHz (5555, 1 Mbps, 8000 duty cycle)	X	4.35	68.60	15.39	0.17	150.0	±0.6 %	
		Y	4.25	68.39	15.14		150.0		
		Z	4.40	68.39	15.14		150.0		
10267-AAA	IEEE 802.11n WFT 2.4 GHz (5555, 1 Mbps, 8000 duty cycle)	X	4.35	68.60	15.39	0.17	150.0	±0.6 %	
		Y	4.25	68.39	15.14		150.0		
		Z	4.40	68.39	15.14		150.0		
10268-AAA	IEEE 802.11n WFT 2.4 GHz (5555, 1 Mbps, 8000 duty cycle)	X	4.35	68.60	15.39	0.17	150.0	±0.6 %	
		Y	4.25	68.39	15.14		150.0		
		Z	4.40	68.39	15.14		150.0		
10269-AAA	IEEE 802.11n WFT 2.4 GHz (5555, 1 Mbps, 8000 duty cycle)	X	4.35	68.60	15.39	0.17	150.0	±0.6 %	
		Y	4.25	68.39	15.14		150.0		
		Z	4.40	68.39	15.14		150.0		
10270-AAA	IEEE 802.11n WFT 2.4 GHz (5555, 1 Mbps, 8000 duty cycle)	X	4.35	68.60	15.39	0.17	150.0	±0.6 %	
		Y	4.25	68.39	15.14		150.0		
		Z	4.40	68.39	15.14		150.0		
10271-AAA	IEEE 802.11n WFT 2.4 GHz (5555, 1 Mbps, 8000 duty cycle)	X	4.35	68.60	15.39	0.17	150.0	±0.6 %	
		Y	4.25	68.39	15.14		150.0		
		Z	4.40	68.39	15.14		150.0		
10272-AAA	IEEE 802.11n WFT 2.4 GHz (5555, 1 Mbps, 8000 duty cycle)	X	4.35	68.60	15.39	0.17	150.0	±0.6 %	
		Y	4.25	68.39	15.14		150.0		
		Z	4.40	68.39	15.14		150.0		
10273-AAA	IEEE 802.11n WFT 2.4 GHz (5555, 1 Mbps, 8000 duty cycle)	X	4.35	68.60	15.39	0.17	150.0	±0.6 %	
		Y	4.25	68.39	15.14		150.0		
		Z	4.40	68.39	15.14		150.0		
10274-AAA	IEEE 802.11n WFT 2.4 GHz (5555, 1 Mbps, 8000 duty cycle)	X	4.35	68.60	15.39	0.17	150.0	±0.6 %	
		Y	4.25	68.39	15.14		150.0		
		Z	4.40	68.39	15.14		150.0		
10275-AAA	IEEE 802.11n WFT 2.4 GHz (5555, 1 Mbps, 8000 duty cycle)	X	4.35	68.60	15.39	0.17	150.0	±0.6 %	
		Y	4.25	68.39	15.14		150.0		
		Z	4.40	68.39	15.14		150.0		
10276-AAA	IEEE 802.11n WFT 2.4 GHz (5555, 1 Mbps, 8000 duty cycle)	X	4.35	68.60	15.39	0.17	150.0	±0.6 %	
		Y	4.25	68.39	15.14		150.0		
		Z	4.40	68.39	15.14		150.0		
10277-AAA	IEEE 802.11n WFT 2.4 GHz (5555, 1 Mbps, 8000 duty cycle)	X	4.35	68.60	15.39	0.17	150.0	±0.6 %	
		Y	4.25	68.39	15.14		150.0		
		Z	4.40	68.39	15.14		150.0		
10278-AAA	IEEE 802.11n WFT 2.4 GHz (5555, 1 Mbps, 8000 duty cycle)	X	4.35	68.60	15.39	0.17	150.0	±0.6 %	
		Y	4.25	68.39	15.14		150.0		
		Z	4.40	68.39	15.14		150.0		
10279-AAA	IEEE 802.11n WFT 2.4 GHz (5555, 1 Mbps, 8000 duty cycle)	X	4.35	68.60	15.39	0.17	150.0	±0.6 %	
		Y	4.25	68.39	15.14		150.0		
		Z	4.40	68.39	15.14		150.0		
10280-AAA	IEEE 802.11n WFT 2.4 GHz (5555, 1 Mbps, 8000 duty cycle)	X	4.35	68.60	15.39	0.17	150.0	±0.6 %	
		Y	4.25	68.39	15.14		150.0		
		Z	4.40	68.39	15.14		150.0		
10281-AAA	IEEE 802.11n WFT 2.4 GHz (5555, 1 Mbps, 8000 duty cycle)	X	4.35	68.60	15.39	0.17	150.0	±0.6 %	
		Y	4.25	68.39	15.14		150.0		
		Z	4.40	68.39	15.14		150.0		
10282-AAA	IEEE 802.11n WFT 2.4 GHz (5555, 1 Mbps, 8000 duty cycle)	X	4.35	68.60	15.39	0.17	150.0	±0.6 %	
		Y	4.25	68.39	15.14		150.0		
		Z	4.40	68.39	15.14		150.0		
10283-AAA	IEEE 802.11n WFT 2.4 GHz (5555, 1 Mbps, 8000 duty cycle)	X	4.35	68.60	15.39	0.17	150.0	±0.6 %	
		Y	4.25	68.39	15.14		150.0		
		Z	4.40	68.39	15.14		150.0		
10284-AAA	IEEE 802.11n WFT 2.4 GHz (5555, 1 Mbps, 8000 duty cycle)	X	4.35	68.60	15.39	0.17	150.0	±0.6 %	
		Y	4.25	68.39	15.14		150.0		
		Z	4.40	68.39	15.14		150.0		
10285-AAA	IEEE 802.11n WFT 2.4 GHz (5555, 1 Mbps, 8000 duty cycle)	X	4.35	68.60	15.39	0.17	150.0	±0.6 %	
		Y	4.25	68.39	15.14		150.0		
		Z	4.40	68.39	15.14		150.0		
10286-AAA	IEEE 802.11n WFT 2.4 GHz (5555, 1 Mbps, 8000 duty cycle)	X	4.35	68.60	15.39	0.17	150.0	±0.6 %	
		Y	4.25	68.39	15.14		150.0		
		Z	4.40	68.39	15.14		150.0		
10287-AAA	IEEE 802.11n WFT 2.4 GHz (5555, 1 Mbps, 8000 duty cycle)	X	4.35	68.60	15.39	0.17	150.0	±0.6 %	
		Y	4.25	68.39	15.14		150.0		
		Z	4.40	68.39	15.14		150.0		
10288-AAA	IEEE 802.11n WFT 2.4 GHz (5555, 1 Mbps, 8000 duty cycle)	X	4.35	68.60	15.39	0.17	150.0	±0.6 %	
		Y	4.25	68.39	15.14		150.0		
		Z	4.40	68.39	15.14		150.0		
10289-AAA	IEEE 802.11n WFT 2.4 GHz (5555, 1 Mbps, 8000 duty cycle)	X	4.35	68.60	15.39	0.17	150.0	±0.6 %	
		Y	4.25	68.39	15.14		150.0		
		Z	4.40	68.39	15.14		150.0		
10290-AAA	IEEE 802.11n WFT 2.4 GHz (5555, 1 Mbps, 8000 duty cycle)	X	4.35	68.60	15.39	0.17	150.0	±0.6 %	
		Y	4.25	68.39	15.14		150.0		
		Z	4.40	68.39	15.14		150.0		
10291-AAA	IEEE 802.11n WFT 2.4 GHz (5555, 1 Mbps, 8000 duty cycle)	X	4.35	68.60	15.39	0.17	150.0	±0.6 %	
		Y	4.25	68.39	15.14		150.0		
		Z	4.40	68.39	15.14		150.0		
10292-AAA	IEEE 802.11n WFT 2.4 GHz (5555, 1 Mbps, 8000 duty cycle)	X	4.35	68.60	15.39	0.17	150.0	±0.6 %	
		Y	4.25	68.39	15.14		150.0		
		Z	4.40	68.39	15.14		150.0		
10293-AAA	IEEE 802.11n WFT 2.4 GHz (5555, 1 Mbps, 8000 duty cycle)	X	4.35	68.60	15.39	0.17	150.0	±0.6 %	
		Y	4.25	68.39	15.14		150.0		
		Z	4.40	68.39	15.14		150.0		
10294-AAA	IEEE 802.11n WFT 2.4 GHz (5555, 1 Mbps, 8000 duty cycle)	X	4.35	68.60	15.39	0.17	150.0	±0.6 %	
		Y	4.25	68.39	15.14		150.0		
		Z	4.40	68.39	15.14		150.0		
10295-AAA	IEEE 802.11n WFT 2.4 GHz (5555, 1 Mbps, 8000 duty cycle)	X	4.35	68.60	15.39	0.17	150.0	±0.6 %	
		Y	4.25	68.39	15.14		150.0		
		Z	4.40	68.39	15.14		150.0		
10296-AAA	IEEE 802.11n WFT 2.4 GHz (5555, 1 Mbps, 8000 duty cycle)	X	4.35	68.60	15.39	0.17	150.0	±0.6 %	
		Y	4.25	68.39	15.14		150.0		
		Z	4.40	68.39	15.14		150.0		
10297-AAA	IEEE 802.11n WFT 2.4 GHz (5555, 1 Mbps, 8000 duty cycle)	X	4.35	68.60	15.39	0.17	150.0	±0.6 %	
		Y	4.25	68.39	15.14		150.0		
		Z	4.40	68.39	15.14		150.0		
10298-AAA	IEEE 802.11n WFT 2.4 GHz (5555, 1 Mbps, 8000 duty cycle)	X	4.35	68.60	15.39	0.17	150.0	±0.6 %	
		Y	4.25	68.39	15.14		150.0		
		Z	4.40	68.39	15.14		150.0		
10299-AAA	IEEE 802.11n WFT 2.4 GHz (5555, 1 Mbps, 8000 duty cycle)	X	4.35	68.60	15.39	0.17	150.0	±0.6 %	
		Y	4.25	68.39	15.14		150.0		
		Z	4.40	68.39	15.14		150.0		
10300-AAA	IEEE 802.11n WFT 2.4 GHz (5555, 1 Mbps, 8000 duty cycle)	X	4.35	68.60	15.39	0.17	150.0	±0.6 %	
		Y	4.25	68.39	15.14		150.0		
		Z	4.40	68.39	15.14		150.0		

EX3DV4 Sn:3708 (6/7)

EX3DV4-SN:3708									
October 22, 2018									
10427-AAB	IEEE 802.11n (HT Greenfield, 150 Mbps, 64-QAM, UL, Subframe=3.4,7,8,9)	X	5.02	67.00	18.51	0.00	150.0	±0.6 %	
		Y	5.19	67.39	18.49		150.0		
		Z	5.11	66.94	18.20		150.0		
10430-AAD	LTE-FDD (FDDMA, 5 MHz, E-TM 3.1)	X	4.38	73.73	18.60	0.00	150.0	±0.6 %	
		Y	4.49	73.15	18.04		150.0		
		Z	4.10	71.38	17.94		150.0		
10431-AAD	LTE-FDD (FDDMA, 10 MHz, E-TM 3.1)	X	3.70	67.15	18.98	0.00	150.0	±0.6 %	
		Y	4.07	67.63	18.28		150.0		
		Z	3.92	66.88	18.14		150.0		
10432-AAC	LTE-FDD (FDDMA, 15 MHz, E-TM 3.1)	X	4.13	66.99	18.21	0.00	150.0	±0.6 %	
		Y	4.37	67.41	18.38		150.0		
		Z	4.28	66.79	18.20		150.0		
10433-AAC	LTE-FDD (FDDMA, 20 MHz, E-TM 3.1)	X	4.39	66.91	18.54	0.00	150.0	±0.6 %	
		Y	4.62	67.33	18.42		150.0		
		Z	4.51	66.77	18.04		150.0		
10434-AAA	W-CDMA (BS Test Model 1, 64 DPCH)	X	4.42	74.13	18.28	0.00	150.0	±0.6 %	
		Y	4.73	74.54	18.12		150.0		
		Z	4.17	72.09	17.89		150.0		
10438-AAF	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, QPSK, UL, Subframe=3.4,7,8,9)	X	5.40	80.51	20.00	0.00	150.0	±0.6 %	
		Y	100.00	117.25	27.58		80.0		
		Z	11.09	92.89	22.28		80.0		
10447-AAD	LTE-FDD (FDDMA, 5 MHz, E-TM 3.1, Clipping 44%)	X	2.96	66.95	18.33	0.00	150.0	±0.6 %	
		Y	3.37	67.75	18.49		150.0		
		Z	3.10	66.53	18.49		150.0		
10448-AAF	LTE-FDD (FDDMA, 10 MHz, E-TM 3.1, Clipping 44%)	X	3.06	66.96	18.53	0.00	150.0	±0.6 %	
		Y	3.93	67.43	18.16		150.0		
		Z	3.78	66.86	18.00		150.0		
10449-AAC	LTE-FDD (FDDMA, 15 MHz, E-TM 3.1, Clipping 44%)	X	3.97	66.81	18.10	0.00	150.0	±0.6 %	
		Y	4.50	67.56	18.28		150.0		
		Z	4.08	66.61	18.21		150.0		
10450-AAC	LTE-FDD (FDDMA, 20 MHz, E-TM 3.1, Clipping 44%)	X	4.18	66.60	18.18	0.00	150.0	±0.6 %	
		Y	4.46	67.13	18.25		150.0		
		Z	4.28	66.54	18.09		150.0		
10451-AAA	W-CDMA (BS Test Model 1, 64 DPCH, Clipping 44%)	X	2.70	65.89	19.23	0.00	150.0	±0.6 %	
		Y	3.22	67.79	18.95		150.0		
		Z	2.96	66.32	18.88		150.0		
10456-AAB	IEEE 802.11ac WFI (160MHz, 64-QAM, 99pc duty cycle)	X	4.09	67.79	18.92	0.00	150.0	±0.6 %	
		Y	6.10	68.03	18.60		150.0		
		Z	6.09	67.70	18.49		150.0		
10457-AAA	UMTS-FDD (DC-HSPA)	X	3.57	65.33	18.93	0.00	150.0	±0.6 %	
		Y	3.72	65.64	18.01		150.0		
		Z	3.64	65.12	18.61		150.0		
10458-AAA	CDMA2000 (1xEV-DO, Rev. B, 2.4Gbps)	X	3.31	69.88	18.53	0.00	150.0	±0.6 %	
		Y	4.24	73.24	18.12		150.0		
		Z	3.83	70.99	18.45		150.0		
10459-AAA	CDMA2000 (1xEV-DO, Rev. B, 3.1Gbps)	X	3.26	71.27	18.70	0.00	150.0	±0.6 %	
		Y	3.12	69.87	18.95		150.0		
		Z	4.91	69.06	17.98		150.0		

Certificate No: EX3-3708_Oct18

Page 26 of 39

EX3DV4-SN:3708									
October 22, 2018									
10477-AAF	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, 16-QAM, UL, Subframe=3.4,7,8,9)	X	0.86	60.00	7.77	3.23	80.0	±0.6 %	
		Y	0.86	60.00	7.37		80.0		
		Z	0.85	60.75	8.50		80.0		
10478-AAF	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, 64-QAM, UL, Subframe=3.4,7,8,9)	X	0.86	60.00	7.29	3.23	80.0	±0.6 %	
		Y	0.86	60.00	6.95		80.0		
		Z	0.89	60.00	7.80		80.0		
10479-AAA	LTE-TDD (SC-FDMA, 50% RB, 1.4 MHz, QPSK, UL, Subframe=3.4,7,8,9)	X	45.10	109.19	27.60	3.23	80.0	±0.6 %	
		Y	10.72	98.20	28.08		80.0		
		Z	6.80	84.41	21.78		80.0		
10485-AAA	LTE-TDD (SC-FDMA, 50% RB, 1.4 MHz, 16-QAM, UL, Subframe=3.4,7,8,9)	X	9.04	88.21	13.43	3.23	80.0	±0.6 %	
		Y	7.14	79.40	17.60		80.0		
		Z	3.88	72.88	18.52		80.0		
10481-AAA	LTE-TDD (SC-FDMA, 50% RB, 1.4 MHz, 64-QAM, UL, Subframe=3.4,7,8,9)	X	2.10	68.01	11.77	3.23	80.0	±0.6 %	
		Y	4.28	72.85	14.93		80.0		
		Z	2.95	68.94	13.69		80.0		
10482-AAF	LTE-TDD (SC-FDMA, 50% RB, 3 MHz, QPSK, UL, Subframe=3.4,7,8,9)	X	1.17	61.31	9.58	2.23	80.0	±0.6 %	
		Y	2.63	70.02	15.29		80.0		
		Z	1.70	64.80	12.42		80.0		
10483-AAF	LTE-TDD (SC-FDMA, 50% RB, 3 MHz, 16-QAM, UL, Subframe=3.4,7,8,9)	X	1.90	60.97	8.89	2.23	80.0	±0.6 %	
		Y	3.07	68.75	13.78		80.0		
		Z	2.34	65.62	12.29		80.0		
10484-AAF	LTE-TDD (SC-FDMA, 50% RB, 3 MHz, 64-QAM, UL, Subframe=3.4,7,8,9)	X	1.49	60.41	8.72	2.23	80.0	±0.6 %	
		Y	2.87	67.74	13.94		80.0		
		Z	2.25	64.87	11.98		80.0		
10485-AAF	LTE-TDD (SC-FDMA, 50% RB, 6 MHz, QPSK, UL, Subframe=3.4,7,8,9)	X	2.10	68.29	14.58	2.23	80.0	±0.6 %	
		Y	3.39	74.14	18.00		80.0		
		Z	2.30	68.88	13.38		80.0		
10486-AAF	LTE-TDD (SC-FDMA, 50% RB, 3 MHz, 16-QAM, UL, Subframe=3.4,7,8,9)	X	1.89	62.30	11.43	2.23	80.0	±0.6 %	
		Y	2.87	68.86	15.17		80.0		
		Z	2.31	65.30	13.27		80.0		
10487-AAF	LTE-TDD (SC-FDMA, 50% RB, 3 MHz, 64-QAM, UL, Subframe=3.4,7,8,9)	X	1.88	62.92	11.27	2.23	80.0	±0.6 %	
		Y	2.84	68.37	14.93		80.0		
		Z	2.33	68.19	13.13		80.0		
10488-AAF	LTE-TDD (SC-FDMA, 50% RB, 10 MHz, QPSK, UL, Subframe=3.4,7,8,9)	X	2.83	71.00	17.45	2.23	80.0	±0.6 %	
		Y	3.88	73.22	18.66		80.0		
		Z	2.85	69.54	16.89		80.0		
10489-AAF	LTE-TDD (SC-FDMA, 50% RB, 10 MHz, 16-QAM, UL, Subframe=3.4,7,8,9)	X	2.80	67.87	16.87	2.23	80.0	±0.6 %	
		Y	3.42	69.49	17.05		80.0		
		Z	2.88	67.34	15.84		80.0		
10490-AAF	LTE-TDD (SC-FDMA, 50% RB, 10 MHz, 64-QAM, UL, Subframe=3.4,7,8,9)	X	2.87	67.51	15.80	2.23	80.0	±0.6 %	
		Y	3.60	69.29	18.97		80.0		
		Z	3.07	67.17	15.82		80.0		
10491-AAF	LTE-TDD (SC-FDMA, 50% RB, 10 MHz, QPSK, UL, Subframe=3.4,7,8,9)	X	3.19	69.71	17.35	2.23	80.0	±0.6 %	
		Y	3.74	71.43	18.14		80.0		
		Z	3.30	67.45	16.42		80.0		
10492-AAF	LTE-TDD (SC-FDMA, 50% RB, 10 MHz, 16-QAM, UL, Subframe=3.4,7,8,9)	X	3.70	68.80	17.98		80.0		
		Z	3.40	68.98	16.20		80.0		

Certificate No: EX3-3708_Oct18

Page 28 of 39

EX3DV4-SN:3708									
October 22, 2018									
10509-AAE	LTE-TDD (SC-FDMA, 100% RB, 10 MHz, 64-QAM, UL, Subframe=3.4,7,8,9)	X	3.41	67.37	18.58	2.23	80.0	±0.6 %	
		Y	3.62	68.59	17.18		80.0		
		Z	3.82	67.03	18.39		80.0		
10509-AAF	LTE-TDD (SC-FDMA, 100% RB, 15 MHz, QPSK, UL, Subframe=3.4,7,8,9)	X	3.78	68.64	17.37	2.23	80.0	±0.6 %	
		Y	4.78	71.43	18.03		80.0		
		Z	3.81	69.14	18.94		80.0		
10510-AAF	LTE-TDD (SC-FDMA, 100% RB, 15 MHz, 16-QAM, UL, Subframe=3.4,7,8,9)	X	3.80	67.28	18.80	2.23	80.0	±0.6 %	
		Y	4.71	69.40	17.24		80.0		
		Z	3.91	67.15	18.93		80.0		
10511-AAF	LTE-TDD (SC-FDMA, 100% RB, 15 MHz, 64-QAM, UL, Subframe=3.4,7,8,9)	X	3.87	67.15	18.77	2.23	80.0	±0.6 %	
		Y	4.36	68.26	17.18		80.0		
		Z	3.88	66.69	18.50		80.0		
10512-AAF	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, QPSK, UL, Subframe=3.4,7,8,9)	X	3.83	70.67	17.64	2.23	80.0	±0.6 %	
		Y	4.57	72.98	18.50		80.0		
		Z	3.88	70.11	17.19		80.0		
10513-AAF	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, 16-QAM, UL, Subframe=3.4,7,8,9)	X	3.89	67.29	18.95	2.23	80.0	±0.6 %	
		Y	4.10	68.72	17.33		80.0		
		Z	3.70	67.28	18.97		80.0		
10514-AAF	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, 64-QAM, UL, Subframe=3.4,7,8,9)	X	3.74	67.09	18.78	2.23	80.0	±0.6 %	
		Y	4.12	69.32	17.20		80.0		
		Z	3.84	66.97	18.51		80.0		
10515-AAA	IEEE 802.11b WFI 2.4 GHz (DSSS, 2 Mbps, 99pc duty cycle)	X	0.81	83.02	14.44	0.00	150.0	±0.6 %	
		Y	0.98	84.09	15.32		150.0		
		Z	0.87	82.57	13.59		150.0		
10516-AAA	IEEE 802.11b WFI 2.4 GHz (DSSS, 5.5 Mbps, 99pc duty cycle)	X	0.89	78.74	18.87	0.00	150.0	±0.6 %	
		Y	0.80	77.81	20.56		150.0		
		Z	0.82	65.83	13.72		150.0		
10517-AAA	IEEE 802.11b WFI 2.4 GHz (DSSS, 11 Mbps, 99pc duty cycle)	X	0.86	85.37	15.03	0.00	150.0	±0.6 %	
		Y	0.83	86.70	16.38		150.0		
		Z	0.88	83.32	13.89		150.0		
10518-AAF	IEEE 802.11ah WFI 5 GHz (OFDM, 9 Mbps, 99pc duty cycle)	X	4.19	68.67	18.18	0.00	150.0	±0.6 %	
		Y	4.40	67.06	18.28		150.0		
		Z	4.30	66.49	18.88		150.0		
10519-AAF	IEEE 802.11ah WFI 5 GHz (OFDM, 12 Mbps, 99pc duty cycle)	X	4.30	66.84	18.27	0.00	150.0	±0.6 %	
		Y	4.46	67.23	18.37		150.0		
		Z	4.28	66.86	18.98		150.0		
10520-AAF	IEEE 802.11ah WFI 5 GHz (OFDM, 18 Mbps, 99pc duty cycle)	X	4.19	68.18	18.18	0.00	150.0	±0.6 %	
		Y	4.42	67.19	18.36		150.0		
		Z	4.30	66.95	18.99		150.0		
10521-AAF	IEEE 802.11ah WFI 5 GHz (OFDM, 24 Mbps, 99pc duty cycle)	X	4.19	68.18	18.18	0.00	150.0	±0.6 %	
		Y	4.42	67.19	18.36		150.0		
		Z	4.24	66.57	18.98		150.0		
10522-AAF	IEEE 802.11ah WFI 5 GHz (OFDM, 36 Mbps, 99pc duty cycle)	X	4.17	68.64	18.05	0.00	150.0	±0.6 %	
		Y	4.41	67.37	18.38		150.0		
		Z	4.28	66.57	18.97		150.0		

EX3DV4 Sn:3708 (7/7)

EX3DV4-SN3708									
October 22, 2018									
10007-AAB	IEEE 802.11ac WFI (20MHz, MCS0, 90pc duty cycle)	X	4.30	68.68	16.10	0.46	130.0	±9.6 %	
		Y	4.50	69.24	16.15		130.0		
		Z	4.40	69.72	16.80		130.0		
10008-AAB	IEEE 802.11ac WFI (20MHz, MCS1, 90pc duty cycle)	X	4.43	69.20	16.24	0.46	130.0	±9.6 %	
		Y	4.65	69.50	16.30		130.0		
		Z	4.54	69.56	15.90		130.0		
10009-AAB	IEEE 802.11ac WFI (20MHz, MCS2, 90pc duty cycle)	X	4.33	69.00	16.03	0.46	130.0	±9.6 %	
		Y	4.54	69.42	16.12		130.0		
		Z	4.44	69.87	15.76		130.0		
10010-AAB	IEEE 802.11ac WFI (20MHz, MCS3, 90pc duty cycle)	X	4.38	68.21	16.23	0.46	130.0	±9.6 %	
		Y	4.60	68.60	16.30		130.0		
		Z	4.48	69.00	15.59		130.0		
10011-AAB	IEEE 802.11ac WFI (20MHz, MCS4, 90pc duty cycle)	X	4.39	68.97	16.05	0.46	130.0	±9.6 %	
		Y	4.51	69.39	16.13		130.0		
		Z	4.40	69.84	15.77		130.0		
10012-AAB	IEEE 802.11ac WFI (20MHz, MCS5, 90pc duty cycle)	X	4.35	68.08	16.08	0.46	130.0	±9.6 %	
		Y	4.51	69.50	16.17		130.0		
		Z	4.39	69.98	15.80		130.0		
10013-AAB	IEEE 802.11ac WFI (20MHz, MCS6, 90pc duty cycle)	X	4.27	69.00	15.89	0.46	130.0	±9.6 %	
		Y	4.50	68.54	16.02		130.0		
		Z	4.38	69.79	15.65		130.0		
10014-AAB	IEEE 802.11ac WFI (20MHz, MCS7, 90pc duty cycle)	X	4.35	68.17	16.21	0.46	130.0	±9.6 %	
		Y	4.47	68.62	16.30		130.0		
		Z	4.36	69.03	15.93		130.0		
10015-AAB	IEEE 802.11ac WFI (20MHz, MCS8, 90pc duty cycle)	X	4.23	68.76	16.76	0.46	130.0	±9.6 %	
		Y	4.60	68.18	16.86		130.0		
		Z	4.38	69.85	15.52		130.0		
10016-AAB	IEEE 802.11ac WFI (40MHz, MCS0, 90pc duty cycle)	X	4.90	66.50	16.34	0.46	130.0	±9.6 %	
		Y	5.12	66.50	16.29		130.0		
		Z	5.05	66.10	16.02		130.0		
10017-AAB	IEEE 802.11ac WFI (40MHz, MCS1, 90pc duty cycle)	X	5.02	66.34	16.39	0.46	130.0	±9.6 %	
		Y	5.17	66.39	16.34		130.0		
		Z	5.10	66.39	16.07		130.0		
10018-AAB	IEEE 802.11ac WFI (40MHz, MCS2, 90pc duty cycle)	X	4.91	66.35	16.40	0.46	130.0	±9.6 %	
		Y	5.08	66.77	16.40		130.0		
		Z	5.00	66.31	16.11		130.0		
10019-AAB	IEEE 802.11ac WFI (40MHz, MCS3, 90pc duty cycle)	X	4.96	66.33	16.52	0.46	130.0	±9.6 %	
		Y	5.08	66.52	16.30		130.0		
		Z	5.01	66.09	15.80		130.0		
10020-AAB	IEEE 802.11ac WFI (40MHz, MCS4, 90pc duty cycle)	X	5.03	66.21	16.31	0.46	130.0	±9.6 %	
		Y	5.18	66.53	16.25		130.0		
		Z	5.09	66.11	15.89		130.0		
10021-AAB	IEEE 802.11ac WFI (40MHz, MCS5, 90pc duty cycle)	X	5.01	66.30	16.50	0.46	130.0	±9.6 %	
		Y	5.18	66.72	16.48		130.0		
		Z	5.10	66.27	16.20		130.0		
10022-AAB	IEEE 802.11ac WFI (40MHz, MCS6, 90pc duty cycle)	X	5.01	66.42	16.49	0.46	130.0	±9.6 %	
		Y	5.18	66.83	16.52		130.0		
		Z	5.09	66.37	16.24		130.0		

Certificate No: EX3-3708_Out18

Page 36 of 38

EX3DV4-SN3708									
October 22, 2018									
10030-AAC	IEEE 802.11ac WFI (160MHz, MCS0, 90pc duty cycle)	X	5.68	66.78	16.50	0.46	130.0	±9.6 %	
		Y	5.96	67.16	16.45		130.0		
		Z	5.80	66.76	16.22		130.0		
10040-AAC	IEEE 802.11ac WFI (160MHz, MCS4, 90pc duty cycle)	X	5.19	66.53	16.33	0.46	130.0	±9.6 %	
		Y	5.88	67.13	16.30		130.0		
		Z	5.88	66.72	16.13		130.0		
10041-AAC	IEEE 802.11ac WFI (160MHz, MCS5, 90pc duty cycle)	X	5.97	66.66	16.51	0.46	130.0	±9.6 %	
		Y	6.01	67.09	16.38		130.0		
		Z	5.98	66.74	16.16		130.0		
10042-AAC	IEEE 802.11ac WFI (160MHz, MCS6, 90pc duty cycle)	X	5.94	66.62	16.72	0.46	130.0	±9.6 %	
		Y	6.08	67.38	16.68		130.0		
		Z	5.92	66.88	16.48		130.0		
10043-AAC	IEEE 802.11ac WFI (160MHz, MCS7, 90pc duty cycle)	X	5.76	66.59	16.43	0.46	130.0	±9.6 %	
		Y	6.29	67.02	16.30		130.0		
		Z	6.03	66.65	16.18		130.0		
10044-AAC	IEEE 802.11ac WFI (160MHz, MCS8, 90pc duty cycle)	X	5.93	66.71	16.32	0.46	130.0	±9.6 %	
		Y	6.36	67.32	16.50		130.0		
		Z	6.07	66.89	16.31		130.0		
10046-AAC	IEEE 802.11ac WFI (160MHz, MCS9, 90pc duty cycle)	X	6.16	67.41	16.53	0.46	130.0	±9.6 %	
		Y	6.27	67.25	16.48		130.0		
		Z	6.04	66.85	16.31		130.0		
10046-AAF	LTE-TDD (SC-FDMA, 1 RB, 5 MHz, QPSK, UL, Subframe=2.7)	X	9.04	62.83	31.00	0.30	60.0	±9.6 %	
		Y	16.22	106.00	36.65		60.0		
		Z	10.43	96.01	32.53		60.0		
10047-AAF	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, QPSK, UL, Subframe=2.7)	X	9.34	91.27	30.60	0.30	60.0	±9.6 %	
		Y	13.94	103.45	34.79		60.0		
		Z	9.34	94.22	31.85		60.0		
10048-AAA	CDMA2000 (1X Advanced)	X	0.32	90.00	5.03	0.00	150.0	±9.6 %	
		Y	0.59	93.51	9.94		150.0		
		Z	0.44	90.36	7.17		150.0		
10052-AAD	LTE-TDD (OFDMA, 5 MHz, E-TM 3.1, CQI=9, 44%)	X	3.20	66.29	16.67	2.23	60.0	±9.6 %	
		Y	3.56	67.36	16.44		60.0		
		Z	3.26	66.84	15.59		60.0		
10053-AAD	LTE-TDD (OFDMA, 10 MHz, E-TM 3.1, CQI=9, 44%)	X	3.75	66.70	16.24	2.23	60.0	±9.6 %	
		Y	4.08	68.63	16.60		60.0		
		Z	3.87	65.57	16.04		60.0		
10054-AAD	LTE-TDD (OFDMA, 15 MHz, E-TM 3.1, CQI=9, 44%)	X	3.81	66.39	16.33	2.23	60.0	±9.6 %	
		Y	4.08	68.15	16.62		60.0		
		Z	3.89	65.26	16.10		60.0		
10055-AAD	LTE-TDD (OFDMA, 20 MHz, E-TM 3.1, CQI=9, 44%)	X	3.89	65.27	16.38	2.23	60.0	±9.6 %	
		Y	4.12	68.08	16.65		60.0		
		Z	3.95	65.32	16.15		60.0		
10065-AAA	Pulse Waveform (200Hz, 10%)	X	3.95	69.00	12.97	10.00	50.0	±9.6 %	
		Y	7.30	77.55	16.50		50.0		
		Z	6.32	75.95	16.50		50.0		
10069-AAA	Pulse Waveform (200Hz, 20%)	X	4.20	68.91	13.74	6.69	50.0	±9.6 %	
		Y	10.68	93.57	16.50		50.0		
		Z	8.16	81.85	16.51		50.0		

Certificate No: EX3-3708_Out18

Page 36 of 39

750V3 Sn:1101 (1/2)

In Collaboration with
TTL **SPEAG** **ILAC-MRA** **CNAS** 中国认可
CALIBRATION LABORATORY 校准 校准
Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2304
E-mail: cti@china-test.com http://www.china-test.cn

Client: **SRTC** Certificate No: **Z17-97134**

CALIBRATION CERTIFICATE

Object: **D750V3 - SN: 1101**

Calibration Procedure(s): **FF-Z11-003-01**
Calibration Procedures for dipole validation kits

Calibration date: **September 13, 2017**

This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements(SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature(22±3)°C and humidity<70%.

Calibration Equipment used (MATE critical for calibration)

Primary Standards	ID #	Cal Date/Calibrated by: Certificate No.	Scheduled Calibration
Power Meter NRVD	102196	02-Mar-17 (CTTL, No.J17X01254)	Mar-18
Power sensor NRV-Z5	100596	02-Mar-17 (CTTL, No.J17X01254)	Mar-18
Reference Probe EX30V4	SN 7433	26-Sep-16(SPEAG, No EX3-7433_Sep16)	Sep-17
DAE4	SN 1331	19-Jan-17(CTTL-SPEAG, No Z17-97015)	Jan-18

Secondary Standards	ID #	Cal Date/Calibrated by: Certificate No.	Scheduled Calibration
Signal Generator E4438C	MY49071430	13-Jan-17 (CTTL, No.J17X00286)	Jan-18
Network Analyzer E5071C	MY46111013	13-Jan-17 (CTTL, No.J17X00285)	Jan-18

Calibrated by: **Zhao Jing** SAR Test Engineer Signature:

Reviewed by: **Yu Zongying** SAR Test Engineer Signature:

Approved by: **Qi Dianyuan** SAR Project Leader Signature:

Issued: September 16, 2017

This calibration certificate shall not be reproduced except in full without written approval of the laboratory

Certificate No: Z17-97134

Page 1 of 8

In Collaboration with
TTL **SPEAG** **ILAC-MRA** **CNAS** 中国认可
CALIBRATION LABORATORY 校准 校准
Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2304
E-mail: cti@china-test.com http://www.china-test.cn

Glossary:
TSL: tissue simulating liquid
ConvF: sensitivity in TSL / NORMx,y,z
N/A: not applicable or not measured

Calibration is Performed According to the Following Standards:
a) IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013
b) IEC 62209-1, "Measurement procedure for assessment of specific absorption rate of human exposure to radio frequency fields from hand-held and body-mounted wireless communication devices- Part 1: Device used next to the ear (Frequency range of 300MHz to 6GHz)", July 2016
c) IEC 62209-2, "Procedure to measure the Specific Absorption Rate (SAR) For wireless communication devices used in close proximity to the human body (frequency range of 30MHz to 6GHz)", March 2010
d) KDB85664, SAR Measurement Requirements for 100 MHz to 6 GHz

Additional Documentation:
e) DASY4/5 System Handbook

Methods Applied and Interpretation of Parameters:
• **Measurement Conditions:** Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
• **Antenna Parameters with TSL:** The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
• **Feed Point Impedance and Return Loss:** These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
• **Electrical Delay:** One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
• **SAR measured:** SAR measured at the stated antenna input power.
• **SAR normalized:** SAR as measured, normalized to an input power of 1 W at the antenna connector.
• **SAR for nominal TSL parameters:** The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution Corresponds to a coverage probability of approximately 95%.

Certificate No: Z17-97134

Page 2 of 8

In Collaboration with
TTL **SPEAG** **ILAC-MRA** **CNAS** 中国认可
CALIBRATION LABORATORY 校准 校准
Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2304
E-mail: cti@china-test.com http://www.china-test.cn

Measurement Conditions
DASY system configuration, as far as not given on page 1

DASY Version	DASY52	52.10.0.1448
Extrapolation	Advanced Extrapolation	
Phantom	Triple Flat Phantom 5.1C	
Distance Dipole Center - TSL	15 mm	with Spacer
Zoom Scan Resolution	dx, dy, dz = 5 mm	
Frequency	750 MHz ± 1 MHz	

Head TSL parameters
The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	41.9	0.89 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	41.5 ± 6 %	0.88 mho/m ± 6 %
Head TSL temperature change during test	<1.0 °C	—	—

SAR result with Head TSL

	Condition	
SAR averaged over 1 cm ³ (1 g) of Head TSL	250 mW input power	2.05 mW / g
SAR measured	normalized to 1W	8.26 mW / g ± 18.6 % (k=2)
SAR for nominal Head TSL parameters	Condition	
SAR averaged over 10 cm ³ (10 g) of Head TSL	250 mW input power	1.34 mW / g
SAR measured	normalized to 1W	6.39 mW / g ± 18.7 % (k=2)

Body TSL parameters
The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Body TSL parameters	22.0 °C	55.5	0.96 mho/m
Measured Body TSL parameters	(22.0 ± 0.2) °C	55.4 ± 6 %	0.95 mho/m ± 6 %
Body TSL temperature change during test	<1.0 °C	—	—

SAR result with Body TSL

	Condition	
SAR averaged over 1 cm ³ (1 g) of Body TSL	250 mW input power	2.15 mW / g
SAR measured	normalized to 1W	8.69 mW / g ± 18.8 % (k=2)
SAR for nominal Body TSL parameters	Condition	
SAR averaged over 10 cm ³ (10 g) of Body TSL	250 mW input power	1.42 mW / g
SAR measured	normalized to 1W	6.73 mW / g ± 18.7 % (k=2)

Certificate No: Z17-97134

Page 3 of 8

In Collaboration with
TTL **SPEAG** **ILAC-MRA** **CNAS** 中国认可
CALIBRATION LABORATORY 校准 校准
Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2304
E-mail: cti@china-test.com http://www.china-test.cn

Appendix (Additional assessments outside the scope of CNAS L0570)

Antenna Parameters with Head TSL

Impedance, transformed to feed point	53.90 ± 0.24jΩ
Return Loss	- 28.4dB

Antenna Parameters with Body TSL

Impedance, transformed to feed point	52.0Q; 2.22jΩ
Return Loss	- 30.9dB

General Antenna Parameters and Design

Electrical Delay (one direction)	1.136 ns

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard.

No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

Additional EUT Data

Manufactured by	SPEAG

Certificate No: Z17-97134

Page 4 of 8