



Dräger Polytron[®] 6100 EC WL, Dräger Polytron[®] Repeater ISA

Notes on Approval

Device marking

1 Device marking



Model	Product	Type of communication
RC001	Polytron 6100 EC WL Polytron Repeater ISA	ISA100

Model	Product	Type of communication
RC002	Polytron 6100 EC WL Polytron Repeater ISA	ISA100
RC003	Polytron 6100 EC WL Polytron Repeater Wire-lessHART	HART

Serial Number key

The third letter of the serial number specifies the manufacturing year: M = 2019, N = 2020, P = 2021, R = 2022, S = 2023, T = 2024, U = 2025, W = 2026, X = 2027, Y = 2028, Z = 2029, etc. (Letters G, I, O, Q are omitted)

Example: Serial Number ARMB-0001: the third letter is M, which means that the unit was manufactured in 2019.

Radio

Max. radiated power:

Bluetooth® LE: <20 dBm EIRP

ISA100: <20 dBm EIRP

Operating frequency:

Bluetooth® LE: 2402-2480 MHz

ISA100: 2405-2480 MHz

Notes on Approval | Dräger Polytron® 6100 EC WL, Dräger Polytron® Repeater ISA



Downloaded by [University of California, San Diego] at 19:28 11 September 2015

Dräger Safety AG & Co. KGaA, Feuerbach 1, 65000 Ludwig, Germany

Discontinued Type ETR 60[™] (Polygon ETR 60 W)
 Gas Detector Instrument type ETR 60[™] (Polygon ETR 60 W)

and the EU-Parliamentary Election: Evidence
in Accordance with the EU-Typing Convention?
Eckstein

© 2003 Blackwell Publishing Ltd
Journal of Internal Medicine 253: 103–110

and not the biological functions under knowledge the adaptation. However, the organism has to be complex and the literary function is reproduction of the total condition.

[illegible]

Dr. Peter Torsberg and
Dr. John Torsberg are
the only pediatric eye
surgeons in the
Midwest. Dr. Torsberg
is a Fellow of the
American Academy of
Ophthalmology.

11th Ave. South, Room 1100

For and Robert (2009) at
 Place and the (1999) with
 (1999) with

Mr. [Signature]
 Director, [Signature]
 Department of [Signature]
 State of [Signature]
 [Signature]
 [Signature]
 [Signature]

Alex H.

3 USA

Changes or modifications not expressly approved by the party responsible for compliance could void the user's authority to operate the equipment.

This device complies with part 15 of the FCC Rules. Operation is subject to the following two conditions:

- (1) This device may not cause harmful interference, and
- (2) this device must accept any interference received, including interference that may cause undesired operation.

[i] This equipment has been tested and found to comply with the limits for a Class B digital device, pursuant to part 15 of the FCC Rules. These limits are designed to provide reasonable protection against harmful interference in a residential installation. This equipment generates, uses and can radiate radio frequency energy and, if not installed and used in accordance with the instructions, may cause harmful interference to radio communications.

However, there is no guarantee that interference will not occur in a particular installation. If this equipment does cause harmful interference to radio or television reception, which can be determined by turning the equipment off and on, the user is encouraged to try to correct the interference by one or more of the following measures:

- Reorient or relocate the receiving antenna.
- Increase the separation between the equipment and receiver.
- Connect the equipment into an outlet on a circuit different from that to which the receiver is connected.
- Consult the dealer or an experienced radio/TV technician for help.

FCC responsible party:

Draeger Inc.

7256 S. Sam Houston W. Parkway

Suite 100

Houston, TX 77085 USA

phone: +1 346-802-6111

e-mail: DIHouston.Approvals@draeger.com

4 Canada

This device contains licence-exempt transmitter(s)/receiver(s) that comply with Innovation, Science and Economic Development Canada's licence-exempt RSS(s). Operation is subject to the following two conditions:

1. This device may not cause interference.
2. This device must accept any interference, including interference that may cause undesired operation of the device.

L'émetteur/récepteur exempt de licence contenu dans le présent appareil est conforme aux CNR d'Innovation, Sciences et Développement économique Canada applicables aux appareils radio exempts de licence. L'exploitation est autorisée aux deux conditions suivantes :

1. L'appareil ne doit pas produire de brouillage;
2. L'appareil doit accepter tout brouillage radioélectrique subi, même si le brouillage est susceptible d'en compromettre le fonctionnement.

■ Manufacturer
Dräger Safety AG & Co. KGaA
Revalstraße 1
D-23560 Lübeck
Germany
+49 451 8 82-0

9300580 – 4683.630 enUS
© **Dräger Safety AG & Co. KGaA**
Edition: 01 – 2021-10
Subject to alterations
www.draeger.com

