

## Appendix C for KSCR220800133001

## Calibration Certificate

Object	Apply	No	Model	SN	Calibration Date
Dipole	<input type="checkbox"/>	1	CLA150	4025	2021/04/26
	<input type="checkbox"/>	2	D450V3	1103	2021/04/21
	<input type="checkbox"/>	3	D750V3	1188	2022/03/29
	<input type="checkbox"/>	4	D835V2	4d114	2022/03/31
	<input type="checkbox"/>	5	D900V2	1d079	2022/06/07
	<input type="checkbox"/>	6	D1800V2	2d170	2022/03/31
	<input type="checkbox"/>	7	D1900V2	5d1136	2022/06/07
	<input type="checkbox"/>	8	D2000V2	1041	2022/06/06
	<input type="checkbox"/>	9	D2300V2	1096	2022/03/31
	<input checked="" type="checkbox"/>	10	D2450V2	817	2022/04/01
	<input type="checkbox"/>	11	D2600V2	1158	2022/03/31
	<input type="checkbox"/>	12	D5GHzV2	1095	2022/06/01
DAE	<input checked="" type="checkbox"/>	13	DAE4	1245	2022/05/30
Probe	<input checked="" type="checkbox"/>	14	EX3DV4	7346	2022/03/30



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: [CN.Doccheck@sgs.com](mailto:CN.Doccheck@sgs.com)

No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300  
中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgsgroup.com.cn  
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

## 1 Dipole

## 1.1 CLA150 - SN 4025

Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland		S Schweizerischer Kalibrierdienst Service suisse d'étalonnage C Servizio svizzero di taratura S Swiss Calibration Service	
Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates		Accreditation No.: SCS 0108	
Client: SGS-CN (Auden)		Certificate No.: CLA150-4025_Apr21	
<b>CALIBRATION CERTIFICATE</b>			
Object: CLA150 - SN: 4025			
Calibration procedure(s): QA CAL-15.v9 Calibration Procedure for SAR Validation Sources below 700 MHz			
Calibration date: April 26, 2021			
This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.			
All calibrations have been conducted in the closed laboratory facility: environment temperature (22 ± 3)°C and humidity < 70%.			
Calibration Equipment used (M&T: critical for calibration)			
Primary Standards	ID #	Cal Date (Certificate No.)	Scheduled Calibration
Power meter NRP	SN: 10476	09-Apr-21 (No. 217-03201/03202)	Apr-22
Power sensor NRP Z91	SN: 10344	09-Apr-21 (No. 217-03201)	Apr-22
Power sensor NRP Z91	SN: 10345	09-Apr-21 (No. 217-03202)	Apr-22
Reference 20 dB Attenuator	SN: C2382 (200)	09-Apr-21 (No. 217-03343)	Apr-22
Type-N mismatch combination	SN: 31092 / 0037	09-Apr-21 (No. 217-03344)	Apr-22
Reference Probe EXENNA	SN: 3877	30-Dec-20 (No. C3X-3877_Dec20)	Dec-21
EXENNA	SN: 664	26-Jan-20 (No. D4M5-654_Jan20)	Jun-21
Secondary Standards	ID #	Check Date (in house)	Scheduled Check
Power meter S44185	SN: G841203074	06-Apr-16 (in house check Jun-20)	In house check: Jun-22
Power sensor E4113A	SN: MY4148067	06-Apr-16 (in house check Jun-20)	In house check: Jun-22
Power sensor E4113A	SN: 000110210	06-Apr-16 (in house check Jun-20)	In house check: Jun-22
RF generator HP 8440D	SN: US440107170	04-Aug-09 (in house check Jun-20)	In house check: Jun-21
Network Analyzer Agilent E8363A	SN: US41080477	31-Mar-14 (in house check Oct-20)	In house check: Oct-21
Calibrated by:	Name: Jeffrey Katman	Function: Laboratory Technician	Signature:
Approved by:	Name: Kaja Polovic	Function: Technical Manager	Signature:
This calibration certificate shall not be reproduced except in full without written approval of the laboratory.			
Issued: April 26, 2021			
Certificate No.: CLA150-4025_Apr21 Page 1 of 6			

Measurement Conditions	
DASY system configuration, as far as not given on page 1.	
DASY Version	V52.10.4
Extrapolation	Advanced Extrapolation
Phantom	ELI Flat Phantom
EUT Positioning	Touch Position
Zoom Scan Resolution	dx, dy = 4.0 mm, dz = 1.4 mm
Frequency	150 MHz ± 1 MHz
Graded Ratio = 1.4 (z direction)	


Head TSL parameters	
The following parameters and calculations were applied.	
Temperature	22.0 °C
Permittivity	52.3
Conductivity	0.75 mho/m
Nominal Head TSL parameters	(22.0 ± 0.2) °C
Measured Head TSL parameters	51.1 ± 6 %
Head TSL temperature change during test	< 0.5 °C

SAR result with Head TSL	
SAR averaged over 1 cm³ (1 g) of Head TSL	Condition
SAR measured	1 W input power
SAR for nominal Head TSL parameters	normalized to 1W
	3.88 W/kg ± 19.4 % (k=2)
SAR averaged over 10 cm³ (10 g) of Head TSL	condition
SAR measured	1 W input power
SAR for nominal Head TSL parameters	normalized to 1W
	2.59 W/kg ± 18.0 % (k=2)

Appendix (Additional assessments outside the scope of SCS 0108)	
Antenna Parameters with Head TSL	
Impedance, transformed to feed point	47.9 Ω ± 1.5 Ω
Return Loss	-31.4 dB
Additional EUT Data	
Manufactured by	SPLEAG

Certificate No.: CLA150-4025_Apr21	
Page 3 of 6	

Certificate No.: CLA150-4025_Apr21	
Page 4 of 6	



## 1.2 D450V3 - SN 1103

**Calibration Laboratory of Schmid & Partner Engineering AG**  
Zugstrasse 43, 8004 Zurich, Switzerland

Accredited by the Swiss Accreditation Service (SAS)  
The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Client: **SGS-CN (Aude)** Certificate No: **D450V3-1103\_Apr21**

**CALIBRATION CERTIFICATE**

Object: **D450V3 - SN: 1103**

Calibration procedure(s): **QA CAL-15-V9**  
Calibration Procedure for SAR Validation Sources below 700 MHz

Calibration date: **April 21, 2021**

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature (20 ± 2) °C and humidity < 70%.

Calibration Equipment used (MPE critical for calibration)

Primary Standards	ID #	Cal Date (Certificate No.)	Scheduled Calibration
Power meter NRP	SN: 104778	09-Apr-21 (No. 217-03201/03202)	Apr-22
Power sensor NRP-291	SN: 102344	09-Apr-21 (No. 217-03201)	Apr-22
Power sensor NRP-291	SN: 102345	09-Apr-21 (No. 217-03202)	Apr-22
Reference 20 dB Attenuator	SN: CG2852 (200)	09-Apr-21 (No. 217-03345)	Apr-22
Type-N mismatch combination	SN: 310827 / 06327	09-Apr-21 (No. 217-03344)	Apr-22
Reference Probe EX3DV4	SN: 3877	30-Dec-20 (No. EX3-2877 Dec20)	Dec-21
DA54	SN: 654	26-Jun-20 (No. DA54-654 Jun20)	Jun-21

Secondary Standards

ID #	Cal Date (In House)	Scheduled Check
Power meter E4418B	SN: GB41200274 06-Apr-16 (In house check Jun-20)	In house check Jun-22
Power sensor E4412A	SN: MY41496027 06-Apr-16 (In house check Jun-20)	In house check Jun-22
Power sensor E4412A	SN: 00018010 06-Apr-16 (In house check Jun-20)	In house check Jun-22
RF generator HP 8446C	SN: U03460.01700 06-Aug-99 (In house check Jun-20)	In house check Jun-22
Network Analyzer Agilent E8358A	SN: U841980477 31-Mar-14 (In house check Oct-20)	In house check Oct-21

Calibrated by: **Christoph Leubner** Function: **Laboratory Technician**

Approved by: **Kelly Polovic** Technical Manager

This calibration certificate shall not be reproduced except in full without written approval of the laboratory. Issued: April 23, 2021

Certificate No: D450V3-1103\_Apr21 Page 1 of 6

**Calibration Laboratory of Schmid & Partner Engineering AG**  
Zugstrasse 43, 8004 Zurich, Switzerland

Accredited by the Swiss Accreditation Service (SAS)  
The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Accreditation No: **SCS 0106**

**Glossary:**

TSL: Issue simulating liquid sensitivity in TSL / NORM x,y,z

ConvF: not applicable or not measured

N/A: not applicable or not measured

**Calibration is Performed According to the Following Standards:**

- IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013
- IEC 62209-1, "Measurement procedure for the assessment of Specific Absorption Rate (SAR) from hand-held and body-mounted devices used next to the ear (frequency range of 300 MHz to 6 GHz)", July 2016
- IEC 62209-2, "Procedure to determine the Specific Absorption Rate (SAR) for wireless communication devices used in close proximity to the human body (frequency range of 30 MHz to 6 GHz)", March 2010
- KDB 665664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

**Additional Documentation:**

- DASY4/5 System Handbook

**Methods Applied and Interpretation of Parameters:**

- Measurement Conditions:** Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL:** The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss:** These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay:** One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured:** SAR measured at the stated antenna input power.
- SAR normalized:** SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters:** The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

Certificate No: D450V3-1103\_Apr21 Page 2 of 6



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: [CN.Doccheck@sgs.com](mailto:CN.Doccheck@sgs.com)

No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300  
中国·江苏·昆山市留学院创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgsgroup.com.cn  
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com



Measurement Conditions		
DASY system configuration, as far as not given on page 1.		
DASY Version	DASY5	V52.10.4
Extrapolation	Advanced Extrapolation	
Phantom	ELJ4 Flat Phantom	Shell thickness: $2 \pm 0.2$ mm
Distance Dipole Center - TSL	15 mm	with Spacer
Zoom Scan Resolution	$d_x, d_y, d_z \approx 5$ mm	
Frequency	450 MHz $\pm 1$ MHz	

Head TSL parameters			
The following parameters and calculations were applied:			
	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	43.5	0.57 mho/m
Measured Head TSL parameters	(22.0 $\pm$ 0.2) °C	43.1 $\pm$ 6 %	0.57 mho/m $\pm$ 6 %
Head TSL temperature change during test	< 0.5 °C	---	---

SAR result with Head TSL		
SAR averaged over 1 cm <sup>3</sup> (1 g) of Head TSL	Condition	
SAR measured	250 mW input power	1.14 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	4.55 W/kg $\pm$ 18.1 % (k=2)

SAR result with Head TSL		
SAR averaged over 10 cm <sup>3</sup> (10 g) of Head TSL	Condition	
SAR measured	250 mW input power	0.757 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	3.06 W/kg $\pm$ 17.6 % (k=2)

Appendix (Additional assessments outside the scope of SCS 0108)		
Antenna Parameters with Head TSL		
Impedance, transformed to feed point	57.1 $\Omega$ - j2.8 $\Omega$	
Return Loss	> 23.0 dB	

General Antenna Parameters and Design	
Electrical Delay (one direction)	1.346 ns

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard overnight coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when twisted according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard.

No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

Additional EUT Data	
Manufactured by	SPEAG

Certificate No: D450V3-1103\_Apr21

Page 5 of 6

Certificate No: D450V3-1103\_Apr21

Page 4 of 6

## DASY5 Validation Report for Head TSL

Date: 21.04.2021

Test Laboratory: SPEAG, Zurich, Switzerland

DUT: Dipole 450 MHz; Type: D450V3; Serial: D450V3 - SN:1103

Communication System: UID 0 - CW; Frequency: 450 MHz

Medium parameters used:  $f = 450$  MHz;  $\alpha = 0.87$  S/m;  $\epsilon_r = 43.1$ ;  $\rho = 1000$  kg/m<sup>3</sup>

Phantom section: Flat Section

Measurement Standard: DASY5 (IEC/IEC/ANSI C63.19-2011)

## DASY52 Configuration:

- Probe: EX3DV4 - SN3877; ConvF(10.64, 10.64, 10.64) @ 450 MHz; Calibrated: 30.12.2020
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn654; Calibrated: 26.06.2020
- Phantom: ELJ v4.0; Type: QDOVA001BB; Serial: TP:1003
- DASY52 52.10.4(1527); SEMCAD X 14.6.14(7483)

## Dipole Calibration for Head Tissue/d=15mm, Pin=250mW/Zoom Scan (7x7x7)/Cube 0:

Measurement grid:  $d_x=5$ mm,  $d_y=5$ mm,  $d_z=5$ mm

Reference Value = 39.18 V/m; Power Drift = -0.08 dB


Peak SAR (extrapolated) = 1.76 W/kg

SAR(1 g) = 1.14 W/kg; SAR(10 g) = 0.767 W/kg

Smallest distance from peaks to all points 3 dB below: Larger than measurement grid

Ratio of SAR at M2 to SAR at M1 = 64.9%

Maximum value of SAR (measured) = 1.53 W/kg



Certificate No: D450V3-1103\_Apr21

Page 5 of 6

Certificate No: D450V3-1103\_Apr21

Page 6 of 6

Compliance Certification Services (Kunshan)  
EMC Laboratory

Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: [CN.Doccheck@sgs.com](mailto:CN.Doccheck@sgs.com)

No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300  
中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 [www.sgsgroup.com.cn](http://www.sgsgroup.com.cn)  
t(86-512)57355888 f(86-512)57370818 [sgs.china@sgs.com](mailto:sgs.china@sgs.com)

## 1.3 D750V3 - SN 1188


TTL Speag CALIBRATION LABORATORY		CAICT																					
Add: No.52 HuaYuanbei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2112 Fax: +86-10-42304633-2504 E-mail: cti@chinaetl.com http://www.chinaetl.cn		Add: No.52 HuaYuanbei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@chinaetl.com http://www.chinaetl.cn																					
Client: <b>SGS-CN</b>		Certificate No: <b>Z22-60103</b>																					
<b>CALIBRATION CERTIFICATE</b>																							
Object: <b>D750V3 - SN: 1188</b>																							
Calibration Procedure(s): <b>FF-Z11-003-01</b> Calibration Procedures for dipole validation kits																							
Calibration date: <b>March 28, 2022</b>																							
This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.																							
All calibrations have been conducted in the closed laboratory facility: environment temperature (22±3)°C and humidity<70%.																							
Calibration Equipment used (M&TE critical for calibration)																							
<table border="1"><thead><tr><th>Primary Standards</th><th>ID #</th><th>Cal Date (Calibrated by: Certificate No.)</th><th>Scheduled Calibration</th></tr></thead><tbody><tr><td>Power Meter NRP2</td><td>104277</td><td>24-Sep-21 (CTTL No.J21X08328)</td><td>Sep-22</td></tr><tr><td>Power sensor NRP88</td><td>104291</td><td>24-Sep-21 (CTTL No.J21X08328)</td><td>Sep-22</td></tr><tr><td>Reference Probe EX30V4</td><td>SN 7307</td><td>26-May-21(SPEAG No EX3-7307_May21)</td><td>May-22</td></tr><tr><td>DAE4</td><td>SN 1556</td><td>12-Jan-22(CTTL-SPEAG No.Z22-60007)</td><td>Jan-23</td></tr></tbody></table>				Primary Standards	ID #	Cal Date (Calibrated by: Certificate No.)	Scheduled Calibration	Power Meter NRP2	104277	24-Sep-21 (CTTL No.J21X08328)	Sep-22	Power sensor NRP88	104291	24-Sep-21 (CTTL No.J21X08328)	Sep-22	Reference Probe EX30V4	SN 7307	26-May-21(SPEAG No EX3-7307_May21)	May-22	DAE4	SN 1556	12-Jan-22(CTTL-SPEAG No.Z22-60007)	Jan-23
Primary Standards	ID #	Cal Date (Calibrated by: Certificate No.)	Scheduled Calibration																				
Power Meter NRP2	104277	24-Sep-21 (CTTL No.J21X08328)	Sep-22																				
Power sensor NRP88	104291	24-Sep-21 (CTTL No.J21X08328)	Sep-22																				
Reference Probe EX30V4	SN 7307	26-May-21(SPEAG No EX3-7307_May21)	May-22																				
DAE4	SN 1556	12-Jan-22(CTTL-SPEAG No.Z22-60007)	Jan-23																				
<table border="1"><thead><tr><th>Secondary Standards</th><th>ID #</th><th>Cal Date (Calibrated by: Certificate No.)</th><th>Scheduled Calibration</th></tr></thead><tbody><tr><td>Signal Generator S4438C</td><td>MY48071430</td><td>13-Jan-22 (CTTL No.J22X00409)</td><td>Jan-23</td></tr><tr><td>Network Analyzer E5071C</td><td>MY48110673</td><td>14-Jan-22 (CTTL No.J22X00409)</td><td>Jan-23</td></tr></tbody></table>				Secondary Standards	ID #	Cal Date (Calibrated by: Certificate No.)	Scheduled Calibration	Signal Generator S4438C	MY48071430	13-Jan-22 (CTTL No.J22X00409)	Jan-23	Network Analyzer E5071C	MY48110673	14-Jan-22 (CTTL No.J22X00409)	Jan-23								
Secondary Standards	ID #	Cal Date (Calibrated by: Certificate No.)	Scheduled Calibration																				
Signal Generator S4438C	MY48071430	13-Jan-22 (CTTL No.J22X00409)	Jan-23																				
Network Analyzer E5071C	MY48110673	14-Jan-22 (CTTL No.J22X00409)	Jan-23																				
Calibrated by: <b>Zhao Jing</b> SAR Test Engineer		Signature																					
Reviewed by: <b>Lin Hao</b> SAR Test Engineer		Signature																					
Approved by: <b>Qi Dianyan</b> SAR Project Leader		Signature																					
Issued: April 3, 2022																							
This calibration certificate shall not be reproduced except in full without written approval of the laboratory.																							
Certificate No: Z22-60103		Page 1 of 5																					

TTL Speag CALIBRATION LABORATORY		CAICT	
Add: No.52 HuaYuanbei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@chinaetl.com http://www.chinaetl.cn		Add: No.52 HuaYuanbei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@chinaetl.com http://www.chinaetl.cn	
<b>Measurement Conditions</b> DASY system configuration, as far as not given on page 1:			
DASY Version		V52.10.4	
Extrapolation		Advanced Extrapolation	
Phantom		Triple Flat Phantom 5.1C	
Distance Dipole Center - TSL		15 mm with Spacer	
Zoom Scan Resolution		dx, dy, dz = 5 mm	
Frequency		750 MHz ± 1 MHz	
<b>Head TSL parameters</b> The following parameters and calculations were applied:			
Temperature		22.0 °C	
Permittivity		42.0	
Conductivity		0.90 mho/m	
Nominal Head TSL parameters		(22.0 ± 0.2) °C	
Measured Head TSL parameters		(22.0 ± 0.2) °C	
Head TSL temperature change during test		<1.0 °C	
<b>SAR result with Head TSL</b>			
SAR averaged over 1 cm <sup>2</sup> (1 g) of Head TSL		Condition	
SAR measured		250 mW input power	
SAR for nominal Head TSL parameters		normalized to 1W	
SAR averaged over 10 cm <sup>2</sup> (10 g) of Head TSL		Condition	
SAR measured		250 mW input power	
SAR for nominal Head TSL parameters		normalized to 1W	

TTL Speag CALIBRATION LABORATORY		CAICT	
Add: No.52 HuaYuanbei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@chinaetl.com http://www.chinaetl.cn		Add: No.52 HuaYuanbei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@chinaetl.com http://www.chinaetl.cn	
<b>Glossary:</b>			
TSL	tissue simulating liquid		
ConvF	sensitivity in TSL / NORMx.y.z		
N/A	not applicable or not measured		
<b>Calibration is Performed According to the Following Standards:</b>			
a) IEC/IEEE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices- Part 1528: Human Models, instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020			
b) KDB 855664, "SAR Measurement Requirements for 100 MHz to 6 GHz"			
<b>Additional Documentation:</b>			
c) DASY4/5 System Handbook			
<b>Methods Applied and Interpretation of Parameters:</b>			
• <b>Measurement Conditions:</b> Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.			
• <b>Antenna Parameters with TSL:</b> The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.			
• <b>Feed Point Impedance and Return Loss:</b> These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.			
• <b>Electrical Delay:</b> One-way delay between the SMA connector and the antenna feed point. No uncertainty required.			
• <b>SAR measured:</b> SAR measured at the stated antenna input power.			
• <b>SAR normalized:</b> SAR as measured, normalized to an input power of 1 W at the antenna connector.			
• <b>SAR for nominal TSL parameters:</b> The measured TSL parameters are used to calculate the nominal SAR result.			
The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.			
Certificate No: Z22-60103		Page 2 of 5	

TTL Speag CALIBRATION LABORATORY		CAICT	
Add: No.52 HuaYuanbei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@chinaetl.com http://www.chinaetl.cn		Add: No.52 HuaYuanbei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@chinaetl.com http://www.chinaetl.cn	
<b>Appendix (Additional assessments outside the scope of CNAS L0570)</b>			
<b>Antenna Parameters with Head TSL</b>			
Impedance, transformed to feed point		53.60 - 1.13jΩ	
Return Loss		-28.7dB	
<b>General Antenna Parameters and Design</b>			
Electrical Delay (one direction)		0.947 ns	
After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.			
The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small and caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard.			
No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.			
<b>Additional EUT Data</b>			
Manufactured by		SPEAG	
Certificate No: Z22-60103		Page 4 of 5	





## 1.4 D835V2 - SN 4d114

**TTL** In Collaboration with **CAICT**  
CALIBRATION LABORATORY

Add: No.52 HuaYuanbei Road, Haidian District, Beijing, 100191, China  
Tel: +86-10-62506633-2079 Fax: +86-10-62506633-2504  
E-mail: cti@china.ttl.com http://www.chinatitl.cn

Client: **SGS-CN** Certificate No: **Z22-60104**

**CALIBRATION CERTIFICATE**

Object: **D835V2 - SN: 4d114**

Calibration Procedure(s): **FF-211-003-01**  
Calibration Procedures for dipole validation kits

Calibration date: **March 31, 2022**

This Calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature (22±3)°C and humidity <70%.

Calibration Equipment used (M&TE critical for calibration)

Primary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration
Power Meter NRP2	106277	24-Sep-21 (CTTL No.J21X08326)	Sep-22
Power sensor NRPBS	104261	24-Sep-21 (CTTL No.J21X08326)	Sep-22
Reference Probe EX3DV4	SN 7307	26-May-21 (SPEAG No.EX3-7307_May21)	May-22
DAE4	SN 1556	12-Jan-22 (CTTL-SPEAG No.Z22-60007)	Jan-23

Secondary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration
Signal Generator E4438C	MY46071430	13-Jan-22 (CTTL No.J22X00409)	Jan-23
Network Analyzer E5071C	MY46110673	14-Jan-22 (CTTL No.J22X00409)	Jan-23

Calibrated by: **Zhao Jing** SAR Test Engineer

Reviewed by: **Lin Hao** SAR Test Engineer

Approved by: **Qi Dianyan** SAR Project Leader

Issued: April 6, 2022

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: Z22-60104 Page 1 of 6

**TTL** In Collaboration with **CAICT**  
CALIBRATION LABORATORY

Add: No.52 HuaYuanbei Road, Haidian District, Beijing, 100191, China  
Tel: +86-10-62506633-2079 Fax: +86-10-62506633-2504  
E-mail: cti@china.ttl.com http://www.chinatitl.cn

**Glossary:**

TSL: tissue simulating liquid  
ConvF: sensitivity in TSL / NORMx.y.z  
N/A: not applicable or not measured

**Calibration is Performed According to the Following Standards:**

a) IEC/IEEE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices- Part 1528: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020  
b) KDB 865864, "SAR Measurement Requirements for 100 MHz to 6 GHz"

**Additional Documentation:**

c) DASY4/5 System Handbook

**Methods Applied and Interpretation of Parameters:**

- Measurement Conditions:** Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL:** The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss:** These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay:** One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured:** SAR measured at the stated antenna input power.
- SAR normalized:** SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters:** The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution Corresponds to a coverage probability of approximately 95%.

Certificate No: Z22-60104 Page 2 of 6





Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic formal documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing/inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: [CN.Doccheck@sgs.com](mailto:CN.Doccheck@sgs.com)

No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300  
中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 [www.sgsgroup.com.cn](http://www.sgsgroup.com.cn)  
t(86-512)57355888 f(86-512)57370818 [sgs.china@sgs.com](mailto:sgs.china@sgs.com)

<p><b>TTL</b> In Collaboration with <b>CAICT</b> CALIBRATION LABORATORY</p> <p>Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com http://www.china.ttl.com</p> <p><b>Measurement Conditions</b> DASY system configuration, as far as not given on page 1.</p> <table border="1"> <tr> <td>DASY Version</td> <td>DASY52</td> <td>VS2 10.4</td> </tr> <tr> <td>Extrapolation</td> <td>Advanced Extrapolation</td> <td></td> </tr> <tr> <td>Phantom</td> <td>Triple Flat Phantom 5.1C</td> <td></td> </tr> <tr> <td>Distance Dipole Center - TSL</td> <td>15 mm</td> <td>with Spacer</td> </tr> <tr> <td>Zoom Scan Resolution</td> <td>dx, dy, dz = 5 mm</td> <td></td> </tr> <tr> <td>Frequency</td> <td>835 MHz ± 1 MHz</td> <td></td> </tr> </table> <p><b>Head TSL parameters</b> The following parameters and calculations were applied.</p> <table border="1"> <tr> <th></th> <th>Temperature</th> <th>Permittivity</th> <th>Conductivity</th> </tr> <tr> <td>Nominal Head TSL parameters</td> <td>22.0 °C</td> <td>41.5</td> <td>0.90 mho/m</td> </tr> <tr> <td>Measured Head TSL parameters</td> <td>(22.0 ± 0.2) °C</td> <td>41.0 ± 6 %</td> <td>0.91 mho/m ± 6 %</td> </tr> <tr> <td>Head TSL temperature change during test</td> <td>&lt;1.0 °C</td> <td>---</td> <td>---</td> </tr> </table> <p><b>SAR result with Head TSL</b></p> <table border="1"> <tr> <td>SAR averaged over 1 cm<sup>3</sup> (1 g) of Head TSL</td> <td>Condition</td> <td></td> </tr> <tr> <td>SAR measured</td> <td>250 mW input power</td> <td>2.37 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>9.40 W/kg ± 18.6 % (k=2)</td> </tr> <tr> <td>SAR averaged over 10 cm<sup>3</sup> (10 g) of Head TSL</td> <td>Condition</td> <td></td> </tr> <tr> <td>SAR measured</td> <td>250 mW input power</td> <td>1.54 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>6.12 W/kg ± 18.7 % (k=2)</td> </tr> </table> <p>Certificate No: Z22-60104 Page 3 of 6</p>	DASY Version	DASY52	VS2 10.4	Extrapolation	Advanced Extrapolation		Phantom	Triple Flat Phantom 5.1C		Distance Dipole Center - TSL	15 mm	with Spacer	Zoom Scan Resolution	dx, dy, dz = 5 mm		Frequency	835 MHz ± 1 MHz			Temperature	Permittivity	Conductivity	Nominal Head TSL parameters	22.0 °C	41.5	0.90 mho/m	Measured Head TSL parameters	(22.0 ± 0.2) °C	41.0 ± 6 %	0.91 mho/m ± 6 %	Head TSL temperature change during test	<1.0 °C	---	---	SAR averaged over 1 cm <sup>3</sup> (1 g) of Head TSL	Condition		SAR measured	250 mW input power	2.37 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	9.40 W/kg ± 18.6 % (k=2)	SAR averaged over 10 cm <sup>3</sup> (10 g) of Head TSL	Condition		SAR measured	250 mW input power	1.54 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	6.12 W/kg ± 18.7 % (k=2)	<p><b>TTL</b> In Collaboration with <b>CAICT</b> CALIBRATION LABORATORY</p> <p>Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com http://www.china.ttl.com</p> <p><b>Appendix (Additional assessments outside the scope of CNAS L0570)</b></p> <p><b>Antenna Parameters with Head TSL</b></p> <table border="1"> <tr> <td>Impedance, transformed to feed point</td> <td>48.70 - j22.0Q</td> </tr> <tr> <td>Return Loss</td> <td>-25.3dB</td> </tr> </table> <p><b>General Antenna Parameters and Design</b></p> <table border="1"> <tr> <td>Electrical Delay (one direction)</td> <td>1.307 ns</td> </tr> </table> <p>After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.</p> <p>The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard.</p> <p>No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.</p> <p><b>Additional EUT Data</b></p> <table border="1"> <tr> <td>Manufactured by</td> <td>SPEAG</td> </tr> </table> <p>Certificate No: Z22-60104 Page 4 of 6</p>	Impedance, transformed to feed point	48.70 - j22.0Q	Return Loss	-25.3dB	Electrical Delay (one direction)	1.307 ns	Manufactured by	SPEAG
DASY Version	DASY52	VS2 10.4																																																											
Extrapolation	Advanced Extrapolation																																																												
Phantom	Triple Flat Phantom 5.1C																																																												
Distance Dipole Center - TSL	15 mm	with Spacer																																																											
Zoom Scan Resolution	dx, dy, dz = 5 mm																																																												
Frequency	835 MHz ± 1 MHz																																																												
	Temperature	Permittivity	Conductivity																																																										
Nominal Head TSL parameters	22.0 °C	41.5	0.90 mho/m																																																										
Measured Head TSL parameters	(22.0 ± 0.2) °C	41.0 ± 6 %	0.91 mho/m ± 6 %																																																										
Head TSL temperature change during test	<1.0 °C	---	---																																																										
SAR averaged over 1 cm <sup>3</sup> (1 g) of Head TSL	Condition																																																												
SAR measured	250 mW input power	2.37 W/kg																																																											
SAR for nominal Head TSL parameters	normalized to 1W	9.40 W/kg ± 18.6 % (k=2)																																																											
SAR averaged over 10 cm <sup>3</sup> (10 g) of Head TSL	Condition																																																												
SAR measured	250 mW input power	1.54 W/kg																																																											
SAR for nominal Head TSL parameters	normalized to 1W	6.12 W/kg ± 18.7 % (k=2)																																																											
Impedance, transformed to feed point	48.70 - j22.0Q																																																												
Return Loss	-25.3dB																																																												
Electrical Delay (one direction)	1.307 ns																																																												
Manufactured by	SPEAG																																																												
<p><b>TTL</b> In Collaboration with <b>CAICT</b> CALIBRATION LABORATORY</p> <p>Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com http://www.china.ttl.com</p> <p><b>DASY5 Validation Report for Head TSL</b> Test Laboratory: CTTL, Beijing, China Date: 2022-03-31</p> <p><b>DUT: Dipole 835 MHz; Type: D835V2; Serial: D835V2 - SN: 4d114</b> Communication System: UID 0, CW; Frequency: 835 MHz; Duty Cycle: 1:1 Medium parameters used: f = 835 MHz; σ = 0.907 S/m; ε<sub>r</sub> = 40.98; ρ = 1000 kg/m<sup>3</sup> Phantom section: Right Section Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007) DASY5 Configuration:</p> <ul style="list-style-type: none"> <li>Probe: EX3DV4 - SN7307; ConvF(10.13, 10.13, 10.13) @ 835 MHz; Calibrated: 2021-05-26</li> <li>Sensor-Surface: 1.4mm (Mechanical Surface Detection)</li> <li>Electronics: DA-E4 Sn1556; Calibrated: 2022-01-12</li> <li>Phantom: MFP_V5.1C (20deg probe tilt); Type: QD 000 P51 Cx; Serial: 1062</li> <li>DASY52 52.10.4(1535); SEMCAD X 14.6(147501)</li> </ul> <p><b>Dipole Calibration/Zoom Scan (7x7x7) (7x7x7) Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm</b> Reference Value = 57.88 V/m; Power Drift = 0.04 dB Peak SAR (extrapolated) = 3.56 W/kg SAR(1 g) = 2.37 W/kg; SAR(10 g) = 1.54 W/kg Smallest distance from peaks to all points 3 dB below = 15.8 mm Ratio of SAR at M2 to SAR at M1 = 66.2% Maximum value of SAR (measured) = 3.17 W/kg</p>  <p>0 dB = 3.17 W/kg = 5.01 dBW/kg</p> <p>Certificate No: Z22-60104 Page 5 of 6</p>	<p><b>TTL</b> In Collaboration with <b>CAICT</b> CALIBRATION LABORATORY</p> <p>Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com http://www.china.ttl.com</p> <p><b>Impedance Measurement Plot for Head TSL</b></p>  <p>Certificate No: Z22-60104 Page 6 of 6</p>																																																												



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: [CN.Doccheck@sgs.com](mailto:CN.Doccheck@sgs.com)  
No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300  
中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgs.com.cn  
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com



## 1.5 D900V2 - SN 1d079

In Collaboration with <b>TTL</b> <b>s p e a g</b> CALIBRATION LABORATORY		In Collaboration with <b>CAICT</b> CALIBRATION LAB																					
Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2117 E-mail: cti@sgs.com.cn		Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2117 E-mail: cti@sgs.com.cn																					
Client: <b>SGS-CN</b>		Certificate No: <b>Z22-60184</b>																					
<b>CALIBRATION CERTIFICATE</b>																							
Object: <b>D900V2 - SN: 1d079</b>																							
Calibration Procedure(s): <b>FF-Z11-003-01</b>																							
Calibration date: <b>June 7, 2022</b>																							
This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.																							
All calibrations have been conducted in the closed laboratory facility: environment temperature (23±3)°C and humidity <70%.																							
Calibration Equipment used (MATE critical for calibration)																							
<table border="1"><thead><tr><th>Primary Standards</th><th>ID #</th><th>Cal Date (Calibrated by Certificate No.)</th><th>Scheduled Calibration</th></tr></thead><tbody><tr><td>Power Meter NRP2</td><td>106277</td><td>24-Sep-21 (CTTL No.J21X08326)</td><td>Sep-22</td></tr><tr><td>Power sensor NRP85</td><td>104291</td><td>24-Sep-21 (CTTL No.J21X08326)</td><td>Sep-22</td></tr><tr><td>Reference Probe EX3DV4</td><td>SN 7464</td><td>28-Jan-22 (SPEAG No.EK3-7464_Jan22)</td><td>Jan-23</td></tr><tr><td>DAE4</td><td>SN 1556</td><td>12-Jan-22 (CTTL-SPEAG No.Z22-60007)</td><td>Jan-23</td></tr></tbody></table>				Primary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration	Power Meter NRP2	106277	24-Sep-21 (CTTL No.J21X08326)	Sep-22	Power sensor NRP85	104291	24-Sep-21 (CTTL No.J21X08326)	Sep-22	Reference Probe EX3DV4	SN 7464	28-Jan-22 (SPEAG No.EK3-7464_Jan22)	Jan-23	DAE4	SN 1556	12-Jan-22 (CTTL-SPEAG No.Z22-60007)	Jan-23
Primary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration																				
Power Meter NRP2	106277	24-Sep-21 (CTTL No.J21X08326)	Sep-22																				
Power sensor NRP85	104291	24-Sep-21 (CTTL No.J21X08326)	Sep-22																				
Reference Probe EX3DV4	SN 7464	28-Jan-22 (SPEAG No.EK3-7464_Jan22)	Jan-23																				
DAE4	SN 1556	12-Jan-22 (CTTL-SPEAG No.Z22-60007)	Jan-23																				
<table border="1"><thead><tr><th>Secondary Standards</th><th>ID #</th><th>Cal Date (Calibrated by Certificate No.)</th><th>Scheduled Calibration</th></tr></thead><tbody><tr><td>Signal Generator E4438C</td><td>MY48071430</td><td>13-Jan-22 (CTTL No.J22X00409)</td><td>Jan-23</td></tr><tr><td>Network Analyzer E5071C</td><td>MY48110673</td><td>14-Jan-22 (CTTL No.J22X00409)</td><td>Jan-23</td></tr></tbody></table>				Secondary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration	Signal Generator E4438C	MY48071430	13-Jan-22 (CTTL No.J22X00409)	Jan-23	Network Analyzer E5071C	MY48110673	14-Jan-22 (CTTL No.J22X00409)	Jan-23								
Secondary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration																				
Signal Generator E4438C	MY48071430	13-Jan-22 (CTTL No.J22X00409)	Jan-23																				
Network Analyzer E5071C	MY48110673	14-Jan-22 (CTTL No.J22X00409)	Jan-23																				
Calibrated by: <b>Zhao Jing</b> SAR Test Engineer																							
Reviewed by: <b>Lin Hao</b> SAR Test Engineer																							
Approved by: <b>Qi Dianyan</b> SAR Project Leader																							
Issued: June 13, 2022																							
This calibration certificate shall not be reproduced except in full without written approval of the laboratory.																							
Certificate No: Z22-60184		Page 1 of 6																					

In Collaboration with <b>TTL</b> <b>s p e a g</b> CALIBRATION LABORATORY		In Collaboration with <b>CAICT</b> CALIBRATION LAB	
Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2117 E-mail: cti@sgs.com.cn		Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2117 E-mail: cti@sgs.com.cn	
Certificate No: Z22-60184		Page 2 of 6	

In Collaboration with <b>TTL</b> <b>s p e a g</b> CALIBRATION LABORATORY		In Collaboration with <b>CAICT</b> CALIBRATION LAB	
Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2117 E-mail: cti@sgs.com.cn		Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2117 E-mail: cti@sgs.com.cn	
Certificate No: Z22-60184		Page 3 of 6	

In Collaboration with <b>TTL</b> <b>s p e a g</b> CALIBRATION LABORATORY		In Collaboration with <b>CAICT</b> CALIBRATION LAB	
Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2117 E-mail: cti@sgs.com.cn		Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2117 E-mail: cti@sgs.com.cn	
Certificate No: Z22-60184		Page 4 of 6	

**Measurement Conditions**  
DASY system configuration, as far as not given on page 1.

DASY Version	DASY2	52.10.4
Extrapolation	Advanced Extrapolation	
Phantom	Triple Flat Phantom 5.1C	
Distance Dipole Center - TSL	15 mm	with Spacer
Zoom Scan Resolution	0k, 0y, 0z = 5 mm	
Frequency	900 MHz ± 1 MHz	

**Head TSL parameters**  
The following parameters and calculations were applied.

Temperature	Permittivity	Conductivity	
Nominal Head TSL parameters	22.0 °C	41.5	0.97 nS/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	42.1 ± 6 %	0.98 nS/m ± 6 %
Head TSL temperature change during test	+1.0 °C	---	---

**SAR result with Head TSL**

SAR averaged over 1 cm <sup>3</sup> (1 g) of Head TSL	Condition	270 W/kg
SAR measured	250 mW input power	2.70 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	11.0 W/kg ± 18.5 % (k=2)
SAR averaged over 19 cm <sup>3</sup> (10 g) of Head TSL	Condition	1.73 W/kg
SAR measured	250 mW input power	1.73 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	7.09 W/kg ± 18.7 % (k=2)

**Glossary:**  
TSL: tissue simulating liquid  
ConvF: sensitivity in TSL / NORM<sub>x,y,z</sub>  
N/A: not applicable or not measured

**Calibration is Performed According to the Following Standards:**  
a) IEC/IEEE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices- Part 1528: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020  
b) KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"  
c) DASY4/5 System Handbook

**Additional Documentation:**  
c) DASY4/5 System Handbook

**Methods Applied and Interpretation of Parameters:**  
• **Measurement Conditions:** Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.  
• **Antenna Parameters with TSL:** The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.  
• **Feed Point Impedance and Return Loss:** These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.  
• **Electrical Delay:** One-way delay between the SMA connector and the antenna feed point. No uncertainty required.  
• **SAR measured:** SAR measured at the stated antenna input power.  
• **SAR normalized:** SAR as measured, normalized to an input power of 1 W at the antenna connector.  
• **SAR for nominal TSL parameters:** The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

**Appendix (Additional assessments outside the scope of CNAS L0570)**

**Antenna Parameters with Head TSL**

Impedance, transformed to feed point	48.10 - 8.49jΩ
Return Loss	-23.3 dB

**General Antenna Parameters and Design**

Electrical Delay (one direction)	1.312 ns
----------------------------------	----------

After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.

**Additional EUT Data**

Manufactured by	SPEAG
-----------------	-------




Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic formal documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com  
No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300  
中国·江苏·昆山市留学院创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgs.com.cn  
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com





## 1.6 D1800V2 - SN 2d170

**TTL** In Collaboration with **CAICT**  
CALIBRATION LABORATORY

Address: No.52 HuaYuanbei Road, Haidian District, Beijing, 100191, China  
Tel: +86-10-42306033-2117 Fax: +86-10-42306033-2117  
E-mail: cti@china.ttl.com http://www.china.ttl.com

Client: **SGS-CN** Certificate No: **Z22-60105**

**CALIBRATION CERTIFICATE**

Object: **D1800V2 - SN: 2d170**

Calibration Procedure(s): **FF-Z11-003-01**  
Calibration Procedures for dipole validation kits

Calibration date: **March 31, 2022**

This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility, environment temperature (22±3)°C and humidity <70%.

Calibration Equipment used (M&TE critical for calibration)

Primary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration
Power Meter NRP2	106277	24-Sep-21 (CTTL No.J21X08326)	Sep-22
Power sensor NRP8S	104291	24-Sep-21 (CTTL No.J21X08326)	Sep-22
Reference Probe EX3DV4	SN 7307	26-May-21 (SPEAG No.EX3-7307_May21)	May-22
DAE4	SN 1556	12-Jan-22 (CTTL-SPEAG No.Z22-60007)	Jan-23

Secondary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration
Signal Generator E4438C	MY49071430	13-Jan-22 (CTTL No.J22X00409)	Jan-23
Network Analyzer E5071C	MY46110573	14-Jan-22 (CTTL No.J22X00409)	Jan-23

Calibrated by: **Zhao Jing** SAR Test Engineer

Reviewed by: **Lin Hao** SAR Test Engineer

Approved by: **Qi Dianyan** SAR Project Leader

Issued April 6, 2022

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: Z22-60105 Page 1 of 6

**TTL** In Collaboration with **CAICT**  
CALIBRATION LABORATORY

Address: No.52 HuaYuanbei Road, Haidian District, Beijing, 100191, China  
Tel: +86-10-42306033-2079 Fax: +86-10-42306033-2504  
E-mail: cti@china.ttl.com http://www.china.ttl.com

**Glossary:**

TSL: tissue simulating liquid  
ConvF: sensitivity in TSL / NORMx,y,z  
N/A: not applicable or not measured

**Calibration is Performed According to the Following Standards:**

a) IEC/IEEE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices, Part 1528: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020  
b) KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

**Additional Documentation:**

c) DASY4/5 System Handbook

**Methods Applied and Interpretation of Parameters:**

- Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss: These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay: One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured: SAR measured at the stated antenna input power.
- SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

Certificate No: Z22-60105 Page 2 of 6



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.


Attention: To check the authenticity of testing/inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com


No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300  
中国·江苏·昆山市留学院创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgsgroup.com.cn  
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

TTL Speaq CALIBRATION LABORATORY		CAICT	
Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com http://www.china.ttl.com		Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com http://www.china.ttl.com	
<b>Measurement Conditions</b> DASY system configuration, as far as not given on page 1.			
DASY Version	DASY52	52.10.4	
Extrapolation	Advanced Extrapolation		
Phantom	Triple Flat Phantom 5.1C		
Distance Dipole Center - TSL	10 mm	with Spacer	
Zoom Scan Resolution	dx, dy, dz = 5 mm		
Frequency	1800 MHz ± 1 MHz		
<b>Head TSL parameters</b> The following parameters and calculations were applied.			
	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	40.0	1.40 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	40.8 ± 6 %	1.41 mho/m ± 6 %
Head TSL temperature change during test	<1.0 °C	---	---
<b>SAR result with Head TSL</b>			
SAR averaged over 1 cm <sup>3</sup> (1 g) of Head TSL	Condition		
SAR measured	250 mW input power	9.73 W/kg	
SAR for nominal Head TSL parameters	normalized to 1W	38.9 W/kg ± 18.8 % (k=2)	
SAR averaged over 10 cm <sup>3</sup> (10 g) of Head TSL	Condition		
SAR measured	250 mW input power	5.11 W/kg	
SAR for nominal Head TSL parameters	normalized to 1W	20.4 W/kg ± 18.7 % (k=2)	
Certificate No: Z22-60105 Page 3 of 6			

TTL Speaq CALIBRATION LABORATORY		CAICT	
Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com http://www.china.ttl.com		Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com http://www.china.ttl.com	
<b>Appendix (Additional assessments outside the scope of CNAS L0570)</b>			
<b>Antenna Parameters with Head TSL</b>			
Impedance, transformed to feed point	47.90-2.54jΩ		
Return Loss	-29.4dB		
<b>General Antenna Parameters and Design</b>			
Electrical Delay (one direction)	1.116 ns		
After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.			
The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.			
<b>Additional EUT Data</b>			
Manufactured by	SPEAG		
Certificate No: Z22-60105 Page 4 of 6			

TTL Speaq CALIBRATION LABORATORY		CAICT	
Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com http://www.china.ttl.com		Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com http://www.china.ttl.com	
<b>DASY5 Validation Report for Head TSL</b> Test Laboratory: CTIL, Beijing, China DUT: Dipole 1800 MHz; Type: D1800V2; Serial: D1800V2 - SN: 2d170 Communication System: UID 0, CW; Frequency: 1800 MHz; Duty Cycle: 1:1 Medium parameters used: f = 1800 MHz; σ = 1.411 S/m; ε = 40.62; ρ = 1000 kg/m <sup>3</sup> Phantom section: Right Section Measurement Standard: DASY5 (IEEE/ANSI C63.19-2007) DASY5 Configuration: <ul style="list-style-type: none"><li>Probe: EX3DV4 - SN7307; ConvF(8.34, 8.34, 8.34) @ 1800 MHz; Calibrated: 2021-05-26</li><li>Sensor-Surface: 1.4mm (Mechanical Surface Detection)</li><li>Electronics: DAE4 Sn1556; Calibrated: 2022-01-12</li><li>Phantom: MFP, V5.1C (20deg probe tilt); Type: QD 000 P51 Cx; Serial: 1062</li><li>DASY52 52.10.4(1535); SEMCAD X 14.6.14(7501)</li></ul> <b>Dipole Calibration/Zoom Scan (7x7x7) (7x7x7) Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm</b> Reference Value = 98.14 V/m; Power Drift = 0.03 dB Peak SAR (extrapolated) = 18.2 W/kg SAR(1 g) = 9.73 W/kg; SAR(10 g) = 5.11 W/kg Smallest distance from peaks to all points 3 dB below = 10 mm Ratio of SAR at M2 to SAR at M1 = 54% Maximum value of SAR (measured) = 15.2 W/kg			
			
Certificate No: Z22-60105 Page 5 of 6			

TTL Speaq CALIBRATION LABORATORY		CAICT	
Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com http://www.china.ttl.com		Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com http://www.china.ttl.com	
<b>Impedance Measurement Plot for Head TSL</b>			
			
Certificate No: Z22-60105 Page 6 of 6			



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing/inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: [CN.Doccheck@sgs.com](mailto:CN.Doccheck@sgs.com)  
No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300  
中国·江苏·昆山市留学院创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 [www.sgs.com](http://www.sgs.com)  
t(86-512)57355888 f(86-512)57370818 [sgs.china@sgs.com](mailto:sgs.china@sgs.com)



## 1.7 D1900V2 - SN 5d136

TTL S p e a g CALIBRATION LABORATORY		CAICT																					
Add: No.52 Huayuan Road, Haidian District, Beijing, 100191, China Tel: +86-10-4220633-2117 E-mail: cti@china.com.cn		Add: No.52 Huayuan Road, Haidian District, Beijing, 100191, China Tel: +86-10-4220633-2117 E-mail: cti@china.com.cn																					
Client: SGS-CN		Certificate No: Z22-60185																					
<b>CALIBRATION CERTIFICATE</b>																							
Object: D1900V2 - SN: 5d136																							
Calibration Procedure(s): FF-Z11-003-01 Calibration Procedures for dipole validation kits																							
Calibration date: June 7, 2022																							
This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.																							
All calibrations have been conducted in the closed laboratory facility: environment temperature (23±1)°C and humidity <70%.																							
Calibration Equipment used (M&TE critical for calibration)																							
<table border="1"><thead><tr><th>Primary Standards</th><th>ID #</th><th>Cal Date (Calibrated by Certificate No.)</th><th>Scheduled Calibration</th></tr></thead><tbody><tr><td>Power Meter NRP2</td><td>106277</td><td>24-Sep-21 (CTTL No.J21X08326)</td><td>Sep-22</td></tr><tr><td>Power sensor NRP6S</td><td>104291</td><td>24-Sep-21 (CTTL No.J21X08326)</td><td>Sep-22</td></tr><tr><td>Reference Probe EXSDV4</td><td>SN 7464</td><td>28-Jan-22 (SPEAG No EX3-7464_Jan22)</td><td>Jan-23</td></tr><tr><td>DAE4</td><td>SN 1656</td><td>12-Jan-22 (CTTL-SPEAG No Z22-60007)</td><td>Jan-23</td></tr></tbody></table>				Primary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration	Power Meter NRP2	106277	24-Sep-21 (CTTL No.J21X08326)	Sep-22	Power sensor NRP6S	104291	24-Sep-21 (CTTL No.J21X08326)	Sep-22	Reference Probe EXSDV4	SN 7464	28-Jan-22 (SPEAG No EX3-7464_Jan22)	Jan-23	DAE4	SN 1656	12-Jan-22 (CTTL-SPEAG No Z22-60007)	Jan-23
Primary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration																				
Power Meter NRP2	106277	24-Sep-21 (CTTL No.J21X08326)	Sep-22																				
Power sensor NRP6S	104291	24-Sep-21 (CTTL No.J21X08326)	Sep-22																				
Reference Probe EXSDV4	SN 7464	28-Jan-22 (SPEAG No EX3-7464_Jan22)	Jan-23																				
DAE4	SN 1656	12-Jan-22 (CTTL-SPEAG No Z22-60007)	Jan-23																				
<table border="1"><thead><tr><th>Secondary Standards</th><th>ID #</th><th>Cal Date (Calibrated by Certificate No.)</th><th>Scheduled Calibration</th></tr></thead><tbody><tr><td>Signal Generator E4438C</td><td>MY48071430</td><td>13-Jan-22 (CTTL No.J22X00409)</td><td>Jan-23</td></tr><tr><td>Network Analyzer E5071C</td><td>MY48110073</td><td>14-Jan-22 (CTTL No.J22X00409)</td><td>Jan-23</td></tr></tbody></table>				Secondary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration	Signal Generator E4438C	MY48071430	13-Jan-22 (CTTL No.J22X00409)	Jan-23	Network Analyzer E5071C	MY48110073	14-Jan-22 (CTTL No.J22X00409)	Jan-23								
Secondary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration																				
Signal Generator E4438C	MY48071430	13-Jan-22 (CTTL No.J22X00409)	Jan-23																				
Network Analyzer E5071C	MY48110073	14-Jan-22 (CTTL No.J22X00409)	Jan-23																				
Calibrated by: Name: Zhao Jing, Function: SAR Test Engineer, Signature: [Signature]																							
Reviewed by: Lin Hao, SAR Test Engineer, Signature: [Signature]																							
Approved by: Qi Diqian, SAR Project Leader, Signature: [Signature]																							
Issued: June 13, 2022																							
This calibration certificate shall not be reproduced except in full without written approval of the laboratory.																							
Certificate No: Z22-60185		Page 1 of 6																					

TTL S p e a g CALIBRATION LABORATORY		CAICT	
Add: No.52 Huayuan Road, Haidian District, Beijing, 100191, China Tel: +86-10-4220633-2117 E-mail: cti@china.com.cn		Add: No.52 Huayuan Road, Haidian District, Beijing, 100191, China Tel: +86-10-4220633-2117 E-mail: cti@china.com.cn	
Measurement Conditions		Appendix (Additional assessments outside the scope of CNAS L0570)	
DASY Version: DASY52, 52.10.4		Antenna Parameters with Head TSL	
Extrapolation: Advanced Extrapolation		Impedance, transformed to feed point: 51.02 ± 7.58Ω	
Phantom: Triple Flat Phantom 5.1C		Return Loss: -22.4dB	
Distance Dipole Center - TSL: 10 mm, with Spacer		General Antenna Parameters and Design	
Zoom Scan Resolution: 6x, 6y, 6z = 5 mm		Electrical Delay (one direction): 1.109 ns	
Frequency: 1900 MHz ± 1 MHz		After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.	
Head TSL parameters		The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC signals. On some of the dipoles, small and caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.	
The following parameters and calculations were applied:		Additional EUT Data	
Nominal Head TSL parameters: Temperature: 22.0 °C, Permittivity: 40.0, Conductivity: 1.40 mhos/m		Manufactured by: SPEAG	
Measured Head TSL parameters: (22.0 ± 0.5) °C, 39.9 ± 6 %, 1.39 mhos/m ± 6 %			
Head TSL temperature change during test: <1.0 °C			
SAR result with Head TSL			
SAR averaged over 1 cm² (1 g) of Head TSL			
SAR measured: 250 mW input power, 9.66 W/kg			
SAR for nominal Head TSL parameters: normalized to 1W, 40.0 W/kg ± 16.8 % (k=2)			
SAR averaged over 10 cm² (10 g) of Head TSL			
SAR measured: 250 mW input power, 5.18 W/kg			
SAR for nominal Head TSL parameters: normalized to 1W, 20.8 W/kg ± 16.7 % (k=2)			
Certificate No: Z22-60185		Page 3 of 6	

TTL S p e a g CALIBRATION LABORATORY		CAICT	
Add: No.52 Huayuan Road, Haidian District, Beijing, 100191, China Tel: +86-10-4220633-2117 E-mail: cti@china.com.cn		Add: No.52 Huayuan Road, Haidian District, Beijing, 100191, China Tel: +86-10-4220633-2117 E-mail: cti@china.com.cn	
Glossary:		TSL: tissue simulating liquid	
ComF: sensitivity in TSL / NORMx.y.z			
NA: not applicable or not measured			
Calibration is Performed According to the Following Standards:			
a) IEC/IEEE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices- Part 1528: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020			
b) KDB 865984, "SAR Measurement Requirements for 100 MHz to 6 GHz"			
Additional Documentation:			
c) DASY4/5 System Handbook			
Methods Applied and Interpretation of Parameters:			
• Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.			
• Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.			
• Feed Point Impedance and Return Loss: These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.			
• Electrical Delay: One-way delay between the SMA connector and the antenna feed point. No uncertainty required.			
• SAR measured: SAR measured at the stated antenna input power.			
• SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector.			
• SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.			
The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.			
Certificate No: Z22-60185		Page 2 of 6	




Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300  
中国·江苏·昆山市留学院创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgs.com.cn  
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com





## 1.8 D2000V2 - SN 1041

**TTLSPEAG** Calibration Laboratory  
Add: No.52 Huayuan Road, Haidian District, Beijing, 100191, China  
Tel: +86-10-62506033-3117  
E-mail: cti@ttspeag.com  
http://www.ttspeag.com

**CAICT**

**Client:** SGS-CN  
**Certificate No:** Z22-60186

**CALIBRATION CERTIFICATE**

**Object:** D2000V2 - SN: 1041

**Calibration Procedure(s):** FF-Z11-003-01  
Calibration Procedures for dipole validation kits

**Calibration date:** June 8, 2022

This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (B). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility; environment temperature (22±3)°C and humidity <70%.

**Calibration Equipment used (M&TE critical for calibration)**

Primary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration
Power Meter NRP2	106277	24-Sep-21 (CTTL No.J21X08328)	Sep-22
Power sensor NRP8S	104291	24-Sep-21 (CTTL No.J21X08328)	Sep-22
Reference Probe EX3DV4	SN 7464	26-Jan-22(SPEAG No EX3-7464-Jan22)	Jan-23
DAE4	SN 1556	12-Jan-22(CTTL-SPEAG No Z22-60007)	Jan-23

Secondary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration
Signal Generator E4438C	MY49071430	13-Jan-22 (CTTL No.Z22X00409)	Jan-23
Network Analyzer E5071C	MY48110673	14-Jan-22 (CTTL No.Z22X00409)	Jan-23

**Calibrated by:** Zhao Jing SAR Test Engineer  
**Reviewed by:** Lin Hao SAR Test Engineer  
**Approved by:** Qi Danyuan SAR Project Leader

Issued: June 13, 2022

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: Z22-60186

**Page 2 of 6:**

**TTLSPEAG** Calibration Laboratory  
Add: No.52 Huayuan Road, Haidian District, Beijing, 100191, China  
Tel: +86-10-62506033-3117  
E-mail: cti@ttspeag.com  
http://www.ttspeag.com

**CAICT**

**Glossary:**  
TSL: Issue simulating liquid  
Comp: sensitivity in TSL / NORMx.y.z  
N/A: not applicable or not measured

**Calibration is Performed According to the Following Standards:**  
a) IEC/IEEE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices- Part 1528: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020  
b) K95 865964, "SAR Measurement Requirements for 100 MHz to 6 GHz"  
c) DASYS4/S System Handbook

**Methods Applied and Interpretation of Parameters:**  
• **Measurement Conditions:** Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.  
• **Antenna Parameters with TSL:** The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.  
• **Feed Point Impedance and Return Loss:** These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.  
• **Electrical Delay:** One-way delay between the SMA connector and the antenna feed point. No uncertainty required.  
• **SAR measured:** SAR measured at the stated antenna input power.  
• **SAR normalized:** SAR as measured, normalized to an input power of 1 W at the antenna connector.  
• **SAR for nominal TSL parameters:** The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution Corresponds to a coverage probability of approximately 95%.

Certificate No: Z22-60186



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic formal documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.


Attention: To check the authenticity of testing/inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com


No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300  
中国·江苏·昆山市留学院创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgsgroup.com.cn  
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

In Collaboration with <b>TTL</b> CALIBRATION LABORATORY		<b>CAICT</b>	
Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62306833-2117 E-mail: ott@china.ttl.com http://www.caict.ac.cn			
<b>Measurement Conditions</b> The following parameters and calculations were applied:			
DASY Version	DASY52	52.10.4	
Extrapolation	Advanced Extrapolation		
Phantom	Triple Flat Phantom 5.1C		
Distance Dipole Center - TSL	10 mm	with Spacer	
Zoom Scan Resolution	dx, dy, dz = 5 mm		
Frequency	2000 MHz ± 1 MHz		
<b>Head TSL parameters</b> The following parameters and calculations were applied:			
Nominal Head TSL parameters	Temperature	Permittivity	Conductivity
	22.0 °C	40.0	1.40 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	40.2 ± 6 %	1.39 mho/m ± 6 %
Head TSL temperature change during test	<1.0 °C	---	---
<b>SAR result with Head TSL</b>			
SAR averaged over 1 cm <sup>3</sup> (1 g) of Head TSL	Condition		
SAR measured	250 mW input power	10.4 W/kg	
SAR for nominal Head TSL parameters	normalized to 1W	41.8 W/kg ± 16.8 % (n=2)	
SAR averaged over 10 cm <sup>3</sup> (10 g) of Head TSL	Condition		
SAR measured	250 mW input power	5.30 W/kg	
SAR for nominal Head TSL parameters	normalized to 1W	21.3 W/kg ± 18.7 % (n=2)	
Certificate No: Z22-60186 Page 3 of 6			

In Collaboration with <b>TTL</b> CALIBRATION LABORATORY		<b>CAICT</b>	
Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62306833-2117 E-mail: ott@china.ttl.com http://www.caict.ac.cn			
<b>Appendix (Additional assessments outside the scope of CNAS L0570)</b>			
<b>Antenna Parameters with Head TSL</b>			
Impedance, transformed to feed point	48.4Ω ± 0.74Ω		
Return Loss	-34.9dB		
<b>General Antenna Parameters and Design</b>			
Electrical Delay (one direction)	1.088 ns		
After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.			
The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small and caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard.			
No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.			
<b>Additional EUT Data</b>			
Manufactured by	SPEAG		
Certificate No: Z22-60186 Page 4 of 6			

In Collaboration with <b>TTL</b> CALIBRATION LABORATORY		<b>CAICT</b>	
Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62306833-2117 E-mail: ott@china.ttl.com http://www.caict.ac.cn			
<b>DASY5 Validation Report for Head TSL</b> Test Laboratory: CTTL, Beijing, China DUT: Dipole 2000 MHz; Type: D2000V2; Serial: D2000V2 - SN: 1041 Communication System: UTD 0, CW; Frequency: 2000 MHz; Duty Cycle: 1:1 Medium parameters used: f = 2000 MHz; σ = 1.392 S/m; ε <sub>r</sub> = 40.21; ρ = 1000 kg/m <sup>3</sup> Phantom section: Right Section Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007) DASY5 Configuration: <ul style="list-style-type: none"><li>Probe: EX3DV4 - SN7464; ConvF(8.2, 8.2) @ 2000 MHz; Calibrated: 2022-01-26</li><li>Sensor-Surface: 1.4mm (Mechanical Surface Detection)</li><li>Electronics: DA64 Sn1556; Calibrated: 2022-01-12</li><li>Phantom: MPP_V5.1C (20deg probe tilt); Type: QD 000 P51 Cx; Serial: 1062</li><li>DASY52 52.10.4(1555); SEMCAD X 14.6.14(7501)</li></ul>			
Dipole Calibration/Zoom Scan (7x7x7) (7x7x7) Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm Reference Value = 103.4 V/m; Power Drift = 0.03 dB Peak SAR (extrapolated) = 19.6 W/kg SAR(1 g) = 10.4 W/kg; SAR(10 g) = 5.3 W/kg Smallest distance from peaks to all points 3 dB below = 9.1 mm Ratio of SAR at M2 to SAR at M1 = 51.6% Maximum value of SAR (measured) = 16.3 W/kg			
			
Certificate No: Z22-60186 Page 5 of 6			

In Collaboration with <b>TTL</b> CALIBRATION LABORATORY		<b>CAICT</b>	
Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62306833-2117 E-mail: ott@china.ttl.com http://www.caict.ac.cn			
<b>Impedance Measurement Plot for Head TSL</b>			
			
Certificate No: Z22-60186 Page 6 of 6			



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com  
No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300  
中国·江苏·昆山市留学院创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgs.com.cn  
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

## 1.9 D2300V2 - SN 1096


TTL Speaq CALIBRATION LABORATORY		CAICT	
Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191 Tel: +86-10-42304633-2512 Fax: +86-10-42304633-2504 E-mail: cti@chinaatl.com http://www.chinaatl.cn		Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@chinaatl.com http://www.chinaatl.cn	
Client: SGS-CN		Certificate No: Z22-60106	
<b>CALIBRATION CERTIFICATE</b>			
Object	D2300V2 - SN: 1096		
Calibration Procedure(s)	FF-Z11-003-01 Calibration Procedures for dipole validation kits		
Calibration date:	March 31, 2022		
This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.			
All calibrations have been conducted in the closed laboratory facility: environment temperature (22±3)°C and humidity <70%.			
Calibration Equipment used (M&TE critical for calibration)			
Primary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration
Power Meter NRP2	108277	24-Sep-21 (CTTL No.J21X08328)	Sep-22
Power sensor NRPBS	104291	24-Sep-21 (CTTL No.J21X08328)	Sep-22
Reference Probe EX30V4	SN 7307	26-May-21 (SPEAG No.EK3-7307_May21)	May-22
DAE4	SN 1556	12-Jan-22 (CTTL-SPEAG No.Z22-60007)	Jan-23
Secondary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration
Signal Generator E4438C	MY49071430	13-Jan-22 (CTTL No.J22X00406)	Jan-23
Network Analyzer E5071C	MY48110673	14-Jan-22 (CTTL No.J22X00406)	Jan-23
Calibrated by:	Name	Function	Signature
	Zhao Jing	SAR Test Engineer	
Reviewed by:	Lin Hao	SAR Test Engineer	
Approved by:	Qi Diqian	SAR Project Leader	
Issued: April 6, 2022			
This calibration certificate shall not be reproduced except in full without written approval of the laboratory.			
Certificate No: Z22-60106		Page 1 of 6	

TTL Speaq CALIBRATION LABORATORY		CAICT	
Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@chinaatl.com http://www.chinaatl.cn		Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@chinaatl.com http://www.chinaatl.cn	
<b>Measurement Conditions</b> DASY system configuration, as far as not given on page 1:			
DASY Version	DASY2	52.10.4	
Extrapolation	Advanced Extrapolation		
Phantom	Triple Flat Phantom 5.1C		
Distance Dipole Center - TSL	10 mm	with Spacer	
Zoom Scan Resolution	dx, dy, dz = 5 mm		
Frequency	2300 MHz ± 1 MHz		
<b>Head TSL parameters</b> The following parameters and calculations were applied:			
	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	39.5	1.67 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	39.8 ± 6 %	1.70 mho/m ± 6 %
Head TSL temperature change during test	<1.0 °C	---	---
<b>SAR result with Head TSL</b>			
SAR averaged over 1 cm <sup>2</sup> (1 g) of Head TSL	Condition		
SAR measured	250 mW input power	12.4 W/kg	
SAR for nominal Head TSL parameters	normalized to 1W	49.2 W/kg ± 18.8 % (k=2)	
SAR averaged over 10 cm <sup>2</sup> (10 g) of Head TSL	Condition		
SAR measured	250 mW input power	5.88 W/kg	
SAR for nominal Head TSL parameters	normalized to 1W	23.4 W/kg ± 18.7 % (k=2)	
Certificate No: Z22-60106		Page 3 of 6	

TTL Speaq CALIBRATION LABORATORY		CAICT	
Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@chinaatl.com http://www.chinaatl.cn		Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@chinaatl.com http://www.chinaatl.cn	
<b>Glossary:</b>			
TSL	tissue simulating liquid		
ConvF	sensitivity in TSL / NCRMx,y,z		
N/A	not applicable or not measured		
<b>Calibration is Performed According to the Following Standards:</b>			
a) IEC/IEEE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices- Part 1528: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020			
b) KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"			
<b>Additional Documentation:</b>			
c) DASY4/5 System Handbook			
<b>Methods Applied and Interpretation of Parameters:</b>			
• <b>Measurement Conditions:</b> Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.			
• <b>Antenna Parameters with TSL:</b> The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.			
• <b>Feed Point Impedance and Return Loss:</b> These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.			
• <b>Electrical Delay:</b> One-way delay between the SMA connector and the antenna feed point. No uncertainty required.			
• <b>SAR measured:</b> SAR measured at the stated antenna input power.			
• <b>SAR normalized:</b> SAR as measured, normalized to an input power of 1 W at the antenna connector.			
• <b>SAR for nominal TSL parameters:</b> The measured TSL parameters are used to calculate the nominal SAR result.			
The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.			
Certificate No: Z22-60106		Page 2 of 6	

TTL Speaq CALIBRATION LABORATORY		CAICT	
Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@chinaatl.com http://www.chinaatl.cn		Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@chinaatl.com http://www.chinaatl.cn	
<b>Appendix (Additional assessments outside the scope of CNAS L0570)</b>			
<b>Antenna Parameters with Head TSL</b>			
Impedance, transformed to feed point	49.20 - 4.56jΩ		
Return Loss	- 26.6dB		
<b>General Antenna Parameters and Design</b>			
Electrical Delay (one direction)	1.083 ns		
After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.			
The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.			
<b>Additional EUT Data</b>			
Manufactured by	SPEAG		
Certificate No: Z22-60106		Page 4 of 6	





## 1.10 D2450V2 - SN 817

**TTL** In Collaboration with **CAICT**  
CALIBRATION LABORATORY

Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China  
Tel: +86-10-62504633-2079 Fax: +86-10-62504633-2504  
E-mail: cti@china.ttl.com http://www.china.ttl.com

Client: **SGS-CN** Certificate No: **Z22-60107**

**CALIBRATION CERTIFICATE**

Object: **D2450V2 - SN: 817**

Calibration Procedure(s): **FF-Z11-003-Q1**  
Calibration Procedures for dipole validation kits

Calibration date: **April 1, 2022**

This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature (22±3)°C and humidity <70%.

Calibration Equipment used (M&E critical for calibration)

Primary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration
Power Meter: NRP2	106277	24-Sep-21 (CTTL No.J21X08320)	Sep-22
Power sensor: NRP8S	104291	24-Sep-21 (CTTL No.J21X08320)	Sep-22
Reference Probe EX3DV4	SN 7307	26-May-21 (SPEAG No EX3-7307_May21)	May-22
DAE4	SN 1556	12-Jan-22 (CTTL-SPEAG No Z22-60007)	Jan-23

Secondary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration
Signal Generator E4438C	MY49071430	13-Jan-22 (CTTL No.J22X00406)	Jan-23
Network Analyzer E5071C	MY46110873	14-Jan-22 (CTTL No.J22X00406)	Jan-23

Calibrated by: **Zhao Jing** SAR Test Engineer  
Reviewed by: **Lin Hao** SAR Test Engineer  
Approved by: **Qi Dianyan** SAR Project Leader

Signature: [Signatures]  
Issued: April 6, 2022

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: Z22-60107 Page 1 of 6

**TTL** In Collaboration with **CAICT**  
CALIBRATION LABORATORY

Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China  
Tel: +86-10-62504633-2079 Fax: +86-10-62504633-2504  
E-mail: cti@china.ttl.com http://www.china.ttl.com

Glossary:  
TSL: tissue simulating liquid  
ConvF: sensitivity in TSL / NORMx.y.z  
N/A: not applicable or not measured

**Calibration is Performed According to the Following Standards:**

a) IEC/IEEE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices-Part 1528: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020  
b) KDB 865864, "SAR Measurement Requirements for 100 MHz to 6 GHz"

**Additional Documentation:**  
c) DASY4/5 System Handbook

**Methods Applied and Interpretation of Parameters:**

- Measurement Conditions:** Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL:** The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss:** These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay:** One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured:** SAR measured at the stated antenna input power.
- SAR normalized:** SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters:** The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution Corresponds to a coverage probability of approximately 95%.

Certificate No: Z22-60107 Page 2 of 6



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing/inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: [CN.Doccheck@sgs.com](mailto:CN.Doccheck@sgs.com)


No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300  
中国·江苏·昆山市留学院创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgsgroup.com.cn  
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

TTL Speaq CALIBRATION LABORATORY		CAICT	
Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@chinaatl.com http://www.chinaatl.cn		Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@chinaatl.com http://www.chinaatl.cn	
<b>Measurement Conditions</b> DASY system configuration, as far as not given on page 1.			
DASY Version	DASY52	52.10.4	
Extrapolation	Advanced Extrapolation		
Phantom	Triple Flat Phantom 5.1C		
Distance Dipole Center - TSL	10 mm	with Spacer	
Zoom Scan Resolution	dx, dy, dz = 5 mm		
Frequency	2450 MHz ± 1 MHz		
<b>Head TSL parameters</b> The following parameters and calculations were applied.			
	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	39.2	1.80 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	39.5 ± 6 %	1.79 mho/m ± 6 %
Head TSL temperature change during test	<1.0 °C	---	---
<b>SAR result with Head TSL</b>			
SAR averaged over 1 cm <sup>3</sup> (1 g) of Head TSL	Condition		
SAR measured	250 mW input power	13.2 W/kg	
SAR for nominal Head TSL parameters	normalized to 1W	53.0 W/kg ± 18.8 % (k=2)	
SAR averaged over 10 cm <sup>3</sup> (10 g) of Head TSL	Condition		
SAR measured	250 mW input power	6.15 W/kg	
SAR for nominal Head TSL parameters	normalized to 1W	24.7 W/kg ± 18.7 % (k=2)	
<b>Appendix (Additional assessments outside the scope of CNAS L0570)</b>			
<b>Antenna Parameters with Head TSL</b>			
Impedance, transformed to feed point	52.10 ± 3.20 Ω		
Return Loss	-28.5 dB		
<b>General Antenna Parameters and Design</b>			
Electrical Delay (one direction)	1.066 ns		
After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.			
The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.			
<b>Additional EUT Data</b>			
Manufactured by	SPEAG		


Certificate No: Z22-60107 Page 3 of 6

Certificate No: Z22-60107 Page 4 of 6

TTL Speaq CALIBRATION LABORATORY		CAICT	
Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@chinaatl.com http://www.chinaatl.cn		Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@chinaatl.com http://www.chinaatl.cn	
<b>DASY5 Validation Report for Head TSL</b> Test Laboratory: CTTL, Beijing, China Date: 2022-04-01 DUT: Dipole 2450 MHz; Type: D2450V2; Serial: D2450V2 - SN: 817 Communication System: UTD 0, CW; Frequency: 2450 MHz; Duty Cycle: 1:1 Medium parameters used: f = 2450 MHz; σ = 1.79 S/m; ε <sub>r</sub> = 39.52; ρ = 1000 kg/m <sup>3</sup> Phantom section: Right Section Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007) DASY5 Configuration:			
<ul style="list-style-type: none"><li>Probe: EX3DV4 - SN7307; ConvF(7.75, 7.75, 7.75) @ 2450 MHz; Calibrated: 2021-05-26</li><li>Sensor-Surface: 1.4mm (Mechanical Surface Detection)</li><li>Electronics: DA14 Sni556; Calibrated: 2022-01-12</li><li>Phantom: MFP_V5.1C (20deg probe tilt); Type: QD 000 P51 Cx; Serial: 1062</li><li>DASY52 52.10.4(1535); SEMCAD X 14.6.14(7501)</li></ul>			
<b>Dipole Calibration Zoom Scan (7x7x7) (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm</b> Reference Value = 104.6 V/m; Power Drift = -0.03 dB Peak SAR (extrapolated) = 27.0 W/kg SAR(1 g) = 13.2 W/kg; SAR(10 g) = 6.15 W/kg Smallest distance from peaks to all points 3 dB below = 8.9 mm Ratio of SAR at M2 to SAR at M1 = -49.2% Maximum value of SAR (measured) = 22.1 W/kg			
			

Certificate No: Z22-60107 Page 5 of 6

Certificate No: Z22-60107 Page 6 of 6

TTL Speaq CALIBRATION LABORATORY		CAICT	
Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@chinaatl.com http://www.chinaatl.cn		Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@chinaatl.com http://www.chinaatl.cn	
<b>Impedance Measurement Plot for Head TSL</b>			
			

Certificate No: Z22-60107 Page 6 of 6



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing/inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300  
中国·江苏·昆山市留学院创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgsgroup.com.cn  
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com



## 1.11 D2600V2 - SN 1158


TTL S p e a g CALIBRATION LABORATORY		CAICT	
Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191 Tel: +86-10-42304633-2512 Fax: +86-10-42304633-2504 E-mail: cti@chinaatl.com http://www.chinaatl.cn		Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@chinaatl.com http://www.chinaatl.cn	
Client: SGS-CN		Certificate No: Z22-60108	
<b>CALIBRATION CERTIFICATE</b>			
Object: D2600V2 - SN: 1158			
Calibration Procedure(s): FF-Z11-003-01 Calibration Procedures for dipole validation kits			
Calibration date: March 31, 2022			
This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.			
All calibrations have been conducted in the closed laboratory facility: environment temperature (22±3)°C and humidity<70%.			
Calibration Equipment used (M&TE critical for calibration)			
Primary Standards		Cal Date (Calibrated by Certificate No.) Scheduled Calibration	
Power Meter NRP2	102377	24-Sep-21 (CTTL No.J21X08326)	Sep-22
Power sensor NRP8S	104291	24-Sep-21 (CTTL No.J21X08326)	Sep-22
Reference Probe EX3DVA	SN 7307	26-May-21 (SPEAG No EX3-7307_May21)	May-22
D4E4	SN 1556	12-Jan-22 (CTTL-SPEAG No Z22-60007)	Jan-23
Secondary Standards		Cal Date (Calibrated by Certificate No.) Scheduled Calibration	
Signal Generator E4438C	MY49071430	13-Jan-22 (CTTL No.J22X00406)	Jan-23
Network Analyzer E5071C	MY48110673	14-Jan-22 (CTTL No.J22X00406)	Jan-23
Calibrated by: Zhao Jing SAR Test Engineer		Signature	
Reviewed by: Lin Hao SAR Test Engineer		Signature	
Approved by: Qi Diaryuan SAR Project Leader		Signature	
Issued: April 6, 2022			
This calibration certificate shall not be reproduced except in full without written approval of the laboratory.			
Certificate No: Z22-60108		Page 1 of 6	

TTL S p e a g CALIBRATION LABORATORY		CAICT	
Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@chinaatl.com http://www.chinaatl.cn		Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@chinaatl.com http://www.chinaatl.cn	
Measurement Conditions DASY system configuration, as far as not given on page 1:		Appendix (Additional assessments outside the scope of CNAS L0570)	
DASY Version: DASY2		Antenna Parameters with Head TSL	
Extrapolation: Advanced Extrapolation		Impedance, transformed to feed point: 49.90-6.49jΩ	
Phantom: Triple Flat Phantom 5.1C		Return Loss: -23.8dB	
Distance Dipole Center - TSL: 10 mm with Spacer		General Antenna Parameters and Design	
Zoom Scan Resolution: 0.1, 0.2, 0.5 mm		Electrical Delay (one direction): 1.053 ns	
Frequency: 2600 MHz ± 1 MHz		After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.	
Head TSL parameters The following parameters and calculations were applied:		The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.	
Nominal Head TSL parameters: 22.0 °C, 39.0, 1.96 mho/m		Additional EUT Data	
Measured Head TSL parameters: (22.0 ± 0.2) °C, 38.7 ± 6 %, 1.96 mho/m ± 6 %		Manufactured by: SPEAG	
Head TSL temperature change during test: <1.0 °C, ---, ---			
SAR result with Head TSL			
SAR averaged over 1 cm² (1 g) of Head TSL			
SAR measured: 250 mW input power, 13.7 W/kg			
SAR for nominal Head TSL parameters: normalized to 1W, 54.8 W/kg ± 18.8 % (k=2)			
SAR averaged over 10 cm² (10 g) of Head TSL			
SAR measured: 250 mW input power, 6.12 W/kg			
SAR for nominal Head TSL parameters: normalized to 1W, 24.8 W/kg ± 18.7 % (k=2)			
Certificate No: Z22-60108		Page 3 of 6	

TTL S p e a g CALIBRATION LABORATORY		CAICT	
Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@chinaatl.com http://www.chinaatl.cn		Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@chinaatl.com http://www.chinaatl.cn	
Glossary:		Glossary:	
TSL: tissue simulating liquid		TSL: tissue simulating liquid	
ConvF: sensitivity in TSL / NORMx.y.z		ConvF: sensitivity in TSL / NORMx.y.z	
N/A: not applicable or not measured		N/A: not applicable or not measured	
Calibration is Performed According to the Following Standards:		Calibration is Performed According to the Following Standards:	
a) IEC/IEEE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices-Part 1528: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020		a) IEC/IEEE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices-Part 1528: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020	
b) KDB 865864, "SAR Measurement Requirements for 100 MHz to 6 GHz"		b) KDB 865864, "SAR Measurement Requirements for 100 MHz to 6 GHz"	
Additional Documentation:		Additional Documentation:	
c) DASY4/S System Handbook		c) DASY4/S System Handbook	
Methods Applied and Interpretation of Parameters:		Methods Applied and Interpretation of Parameters:	
• Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.		• Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.	
• Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.		• Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.	
• Feed Point Impedance and Return Loss: These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.		• Feed Point Impedance and Return Loss: These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.	
• Electrical Delay: One-way delay between the SMA connector and the antenna feed point. No uncertainty required.		• Electrical Delay: One-way delay between the SMA connector and the antenna feed point. No uncertainty required.	
• SAR measured: SAR measured at the stated antenna input power.		• SAR measured: SAR measured at the stated antenna input power.	
• SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector.		• SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector.	
• SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.		• SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.	
The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution Corresponds to a coverage probability of approximately 95%.		The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution Corresponds to a coverage probability of approximately 95%.	
Certificate No: Z22-60108		Page 2 of 6	

TTL S p e a g CALIBRATION LABORATORY		CAICT	
Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@chinaatl.com http://www.chinaatl.cn		Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@chinaatl.com http://www.chinaatl.cn	
Measurement Conditions DASY system configuration, as far as not given on page 1:		Appendix (Additional assessments outside the scope of CNAS L0570)	
DASY Version: DASY2		Antenna Parameters with Head TSL	
Extrapolation: Advanced Extrapolation		Impedance, transformed to feed point: 49.90-6.49jΩ	
Phantom: Triple Flat Phantom 5.1C		Return Loss: -23.8dB	
Distance Dipole Center - TSL: 10 mm with Spacer		General Antenna Parameters and Design	
Zoom Scan Resolution: 0.1, 0.2, 0.5 mm		Electrical Delay (one direction): 1.053 ns	
Frequency: 2600 MHz ± 1 MHz		After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.	
Head TSL parameters The following parameters and calculations were applied:		The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.	
Nominal Head TSL parameters: 22.0 °C, 39.0, 1.96 mho/m		Additional EUT Data	
Measured Head TSL parameters: (22.0 ± 0.2) °C, 38.7 ± 6 %, 1.96 mho/m ± 6 %		Manufactured by: SPEAG	
Head TSL temperature change during test: <1.0 °C, ---, ---			
SAR result with Head TSL			
SAR averaged over 1 cm² (1 g) of Head TSL			
SAR measured: 250 mW input power, 13.7 W/kg			
SAR for nominal Head TSL parameters: normalized to 1W, 54.8 W/kg ± 18.8 % (k=2)			
SAR averaged over 10 cm² (10 g) of Head TSL			
SAR measured: 250 mW input power, 6.12 W/kg			
SAR for nominal Head TSL parameters: normalized to 1W, 24.8 W/kg ± 18.7 % (k=2)			
Certificate No: Z22-60108		Page 3 of 6	





## 1.12 D5GHzV2 - SN 1095

**TTL S p e a g**  
CALIBRATION LABORATORY

Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China  
Tel: +86-10-62104633-2079 Fax: +86-10-62104633-2504  
E-mail: cti@china.ttl.com http://www.china.ttl.com

**CAICT**

**Client: SGS-CN Certificate No: Z22-60187**

**CALIBRATION CERTIFICATE**

Object: D5GHzV2 - SN: 1095

Calibration Procedure(s): FF-Z11-003-01  
Calibration Procedures for dipole validation kits

Calibration date: June 1, 2022

This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature (23±3)°C and humidity <70%.

Calibration Equipment used (M&TE critical for calibration)

Primary Standards	ID #	Cal Date (Calibrated by: Certificate No.)	Scheduled Calibration
Power Meter NRP2	106277	24-Sep-21 (CTTL No.J21008328)	Sep-22
Power sensor NRP4S	104201	24-Sep-21 (CTTL No.J21008328)	Sep-22
Reference Probe EX3DV4	SN 7464	26-Jan-22(SPEAG No EX3-7464, Jan22)	Jan-23
DAE4	SN 1556	12-Jan-22(CTTL-SPEAG No.Z22-60007)	Jan-23

Secondary Standards	ID #	Cal Date (Calibrated by: Certificate No.)	Scheduled Calibration
Signal Generator E4438C	MY48071430	13-Jan-22 (CTTL No.J22000408)	Jan-23
Network Analyzer E5071C	MY48110673	14-Jan-22 (CTTL No.J22000406)	Jan-23

Calibrated by: Zhao Jing SAR Test Engineer

Reviewed by: Lin Hao SAR Test Engineer

Approved by: Qi Dianyan SAR Project Leader

Issued: June 6, 2022

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: Z22-60187 Page 1 of 10

**TTL S p e a g**  
CALIBRATION LABORATORY

Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China  
Tel: +86-10-62104633-2079 Fax: +86-10-62104633-2504  
E-mail: cti@china.ttl.com http://www.china.ttl.com

**CAICT**

**Glossary:**

TSL Issue simulating liquid

ComF sensitivity in TSL / NORMx,y,z

N/A not applicable or not measured

**Calibration is Performed According to the Following Standards:**

a) IEC/IEEE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices-Part 1528: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020

b) KDB 665664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

**Additional Documentation:**

c) DASY5 System Handbook

**Methods Applied and Interpretation of Parameters:**

- Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss: These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay: One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured: SAR measured at the stated antenna input power.
- SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution Corresponds to a coverage probability of approximately 95%.

Certificate No: Z22-60187 Page 2 of 10



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic formal documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing/inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10, Weyue Road, Innovation Park, Kunshan, Jiangsu, China 215300  
中国·江苏·昆山市留学院创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgs.com.cn  
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

In Collaboration with <b>TTL</b> <b>s p e a g</b> CALIBRATION LABORATORY		<b>CAICT</b>	
Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62362117 E-mail: csl@sgs.com.cn http://www.caict.ac.cn		Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62362117 E-mail: csl@sgs.com.cn http://www.caict.ac.cn	
<b>Measurement Conditions</b> DASY system configuration, as for as not given on page 1.			
DASY Version	DASY2	52.10.4	
Extrapolation	Advanced Extrapolation		
Phantom	Triple Flat Phantom 5.1C		
Distance Dipole Center - TSL	10 mm	with Spacer	
Zoom Scan Resolution	dx, dy = 4 mm, dz = 1.4 mm	Graded Ratio = 1.4 (Z direction)	
Frequency	5200 MHz ± 1 MHz 5300 MHz ± 1 MHz 5500 MHz ± 1 MHz 5600 MHz ± 1 MHz 5800 MHz ± 1 MHz		
<b>Head TSL parameters at 5200MHz</b> The following parameters and calculations were applied.			
Nominal Head TSL parameters	Temperature	Permittivity	Conductivity
	22.0 °C	35.9	4.76 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	35.2 ± 6 %	4.73 mho/m ± 6 %
Head TSL temperature change during test	<1.0 °C	---	---
<b>SAR result with Head TSL at 5200MHz</b>			
SAR averaged over 1 cm <sup>3</sup> (1 g) of Head TSL	Condition		
SAR measured	250 mW input power	7.79 W/kg	
SAR for nominal Head TSL parameters	normalized to 1W	77.8 W/kg ± 24.4 % (k=2)	
SAR averaged over 10 cm <sup>3</sup> (10 g) of Head TSL	Condition		
SAR measured	250 mW input power	2.22 W/kg	
SAR for nominal Head TSL parameters	normalized to 1W	22.1 W/kg ± 24.2 % (k=2)	
Certificate No: Z22-60187 Page 3 of 10			

In Collaboration with <b>TTL</b> <b>s p e a g</b> CALIBRATION LABORATORY		<b>CAICT</b>	
Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62362117 E-mail: csl@sgs.com.cn http://www.caict.ac.cn		Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62362117 E-mail: csl@sgs.com.cn http://www.caict.ac.cn	
<b>Head TSL parameters at 5300MHz</b> The following parameters and calculations were applied.			
Nominal Head TSL parameters	Temperature	Permittivity	Conductivity
	22.0 °C	35.9	4.76 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	35.2 ± 6 %	4.73 mho/m ± 6 %
Head TSL temperature change during test	<1.0 °C	---	---
<b>SAR result with Head TSL at 5300MHz</b>			
SAR averaged over 1 cm <sup>3</sup> (1 g) of Head TSL	Condition		
SAR measured	100 mW input power	7.94 W/kg	
SAR for nominal Head TSL parameters	normalized to 1W	79.1 W/kg ± 24.4 % (k=2)	
SAR averaged over 10 cm <sup>3</sup> (10 g) of Head TSL	Condition		
SAR measured	100 mW input power	2.27 W/kg	
SAR for nominal Head TSL parameters	normalized to 1W	22.6 W/kg ± 24.2 % (k=2)	
<b>Head TSL parameters at 5500MHz</b> The following parameters and calculations were applied.			
Nominal Head TSL parameters	Temperature	Permittivity	Conductivity
	22.0 °C	35.8	4.96 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	34.8 ± 6 %	4.94 mho/m ± 6 %
Head TSL temperature change during test	<1.0 °C	---	---
<b>SAR result with Head TSL at 5500MHz</b>			
SAR averaged over 1 cm <sup>3</sup> (1 g) of Head TSL	Condition		
SAR measured	100 mW input power	8.29 W/kg	
SAR for nominal Head TSL parameters	normalized to 1W	82.6 W/kg ± 24.4 % (k=2)	
SAR averaged over 10 cm <sup>3</sup> (10 g) of Head TSL	Condition		
SAR measured	100 mW input power	2.34 W/kg	
SAR for nominal Head TSL parameters	normalized to 1W	23.3 W/kg ± 24.2 % (k=2)	
Certificate No: Z22-60187 Page 4 of 10			

In Collaboration with <b>TTL</b> <b>s p e a g</b> CALIBRATION LABORATORY		<b>CAICT</b>	
Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62362117 E-mail: csl@sgs.com.cn http://www.caict.ac.cn		Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62362117 E-mail: csl@sgs.com.cn http://www.caict.ac.cn	
<b>Head TSL parameters at 5600MHz</b> The following parameters and calculations were applied.			
Nominal Head TSL parameters	Temperature	Permittivity	Conductivity
	22.0 °C	35.5	5.07 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	34.7 ± 6 %	5.05 mho/m ± 6 %
Head TSL temperature change during test	<1.0 °C	---	---
<b>SAR result with Head TSL at 5600MHz</b>			
SAR averaged over 1 cm <sup>3</sup> (1 g) of Head TSL	Condition		
SAR measured	100 mW input power	8.12 W/kg	
SAR for nominal Head TSL parameters	normalized to 1W	80.8 W/kg ± 24.4 % (k=2)	
SAR averaged over 10 cm <sup>3</sup> (10 g) of Head TSL	Condition		
SAR measured	100 mW input power	2.30 W/kg	
SAR for nominal Head TSL parameters	normalized to 1W	22.9 W/kg ± 24.2 % (k=2)	
<b>Head TSL parameters at 5800MHz</b> The following parameters and calculations were applied.			
Nominal Head TSL parameters	Temperature	Permittivity	Conductivity
	22.0 °C	35.3	5.27 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	34.4 ± 6 %	5.25 mho/m ± 6 %
Head TSL temperature change during test	<1.0 °C	---	---
<b>SAR result with Head TSL at 5800MHz</b>			
SAR averaged over 1 cm <sup>3</sup> (1 g) of Head TSL	Condition		
SAR measured	100 mW input power	7.71 W/kg	
SAR for nominal Head TSL parameters	normalized to 1W	76.7 W/kg ± 24.4 % (k=2)	
SAR averaged over 10 cm <sup>3</sup> (10 g) of Head TSL	Condition		
SAR measured	100 mW input power	2.16 W/kg	
SAR for nominal Head TSL parameters	normalized to 1W	21.5 W/kg ± 24.2 % (k=2)	
Certificate No: Z22-60187 Page 5 of 10			

In Collaboration with <b>TTL</b> <b>s p e a g</b> CALIBRATION LABORATORY		<b>CAICT</b>	
Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62362117 E-mail: csl@sgs.com.cn http://www.caict.ac.cn		Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62362117 E-mail: csl@sgs.com.cn http://www.caict.ac.cn	
<b>Appendix (Additional assessments outside the scope of CNAS L0570)</b>			
<b>Antenna Parameters with Head TSL at 5200MHz</b>			
Impedance, transformed to feed point	48.1D-5.03jΩ		
Return Loss	-23.6dB		
<b>Antenna Parameters with Head TSL at 5300MHz</b>			
Impedance, transformed to feed point	47.8D-2.42jΩ		
Return Loss	-28.5dB		
<b>Antenna Parameters with Head TSL at 5500MHz</b>			
Impedance, transformed to feed point	50.3D-4.26jΩ		
Return Loss	-27.4dB		
<b>Antenna Parameters with Head TSL at 5600MHz</b>			
Impedance, transformed to feed point	54.5D-4.80jΩ		
Return Loss	-24.0dB		
<b>Antenna Parameters with Head TSL at 5800MHz</b>			
Impedance, transformed to feed point	51.5D-5.61jΩ		
Return Loss	-24.9dB		
Certificate No: Z22-60187 Page 6 of 10			




Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com





No.10, Welye Road, Innovation Park, Kunshan, Jiangsu, China 215300  
中国·江苏·昆山市留学院创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgs.com.cn  
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

<p>In Collaboration with <b>TTL</b> <b>s p e a q</b> CALIBRATION LABORATORY</p> <p>CAICT</p> <p>Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62302117 E-mail: cti@chinaetl.com http://www.caict.ac.cn</p> <p><b>General Antenna Parameters and Design</b></p> <p>Electrical Delay (one direction) 1.101 ns</p> <p>After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.</p> <p>The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard.</p> <p>No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.</p> <p><b>Additional EUT Data</b></p> <p>Manufactured by SPEAG</p> <p>Certificate No: Z22-60187 Page 7 of 10</p>	<p>In Collaboration with <b>TTL</b> <b>s p e a q</b> CALIBRATION LABORATORY</p> <p>CAICT</p> <p>Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62302117 E-mail: cti@chinaetl.com http://www.caict.ac.cn</p> <p><b>DASY5 Validation Report for Head TSL</b></p> <p>Test Laboratory: CTTL, Beijing, China Date: 2022-06-01</p> <p><b>DUT: Dipole 5GHz; Type: D5GHzV2; Serial: D5GHzV2 - SN: 1095</b></p> <p>Communication System: CW; Frequency: 5200 MHz; Frequency: 5300 MHz; Frequency: 5500 MHz; Frequency: 5600 MHz; Frequency: 5800 MHz; Duty Cycle: 1:1 Medium parameters used: <math>f = 5200 \text{ MHz}</math>; <math>\sigma = 4.62 \text{ S/m}</math>; <math>\epsilon_r = 35.19</math>; <math>\rho = 1000 \text{ kg/m}^3</math> Medium parameters used: <math>f = 5300 \text{ MHz}</math>; <math>\sigma = 4.73 \text{ S/m}</math>; <math>\epsilon_r = 35.19</math>; <math>\rho = 1000 \text{ kg/m}^3</math> Medium parameters used: <math>f = 5500 \text{ MHz}</math>; <math>\sigma = 4.939 \text{ S/m}</math>; <math>\epsilon_r = 34.83</math>; <math>\rho = 1000 \text{ kg/m}^3</math> Medium parameters used: <math>f = 5600 \text{ MHz}</math>; <math>\sigma = 5.051 \text{ S/m}</math>; <math>\epsilon_r = 34.89</math>; <math>\rho = 1000 \text{ kg/m}^3</math> Medium parameters used: <math>f = 5800 \text{ MHz}</math>; <math>\sigma = 5.247 \text{ S/m}</math>; <math>\epsilon_r = 34.42</math>; <math>\rho = 1000 \text{ kg/m}^3</math> Phantom section: Right Section Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007) DASY5 Configuration:</p> <ul style="list-style-type: none"><li>Probe: EX3DV4 - SN7484; ConvF(5.6, 5.6, 5.6) @ 5200 MHz; ConvF(5.32, 5.32, 5.32) @ 5300 MHz; ConvF(5.11, 5.11, 5.11) @ 5500 MHz; ConvF(4.91, 4.91, 4.91) @ 5600 MHz; ConvF(5, 5, 5) @ 5800 MHz; Calibrated: 2022-01-26</li><li>Sensor-Surface: 1.4mm (Mechanical Surface Detection)</li><li>Electronics: DA64 Sn1556; Calibrated: 2022-01-12</li><li>Phantom: MPF_V5.1C (20deg probe tilt); Type: QD 000 P51 Cx; Serial: 1062</li><li>DASY52 52.10.4(1535); SEMCAD X 14.6.14(7501)</li></ul> <p><b>Dipole Calibration /Pin=100mW, d=10mm, f=5200 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm</b> Reference Value = 60.80 V/m; Power Drift = -0.06 dB Peak SAR (extrapolated) = 29.8 W/kg SAR(1 g) = 7.79 W/kg; SAR(10 g) = 2.22 W/kg Smallest distance from peaks to all points 3 dB below = 7.2 mm Ratio of SAR at M2 to SAR at M1 = 66.8% Maximum value of SAR (measured) = 18.3 W/kg</p> <p><b>Dipole Calibration /Pin=100mW, d=10mm, f=5300 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm</b> Reference Value = 61.08 V/m; Power Drift = -0.07 dB Peak SAR (extrapolated) = 31.5 W/kg SAR(1 g) = 7.94 W/kg; SAR(10 g) = 2.27 W/kg Smallest distance from peaks to all points 3 dB below = 7.2 mm Ratio of SAR at M2 to SAR at M1 = 65.5% Maximum value of SAR (measured) = 19.0 W/kg</p> <p>Certificate No: Z22-60187 Page 8 of 10</p>
<p>In Collaboration with <b>TTL</b> <b>s p e a q</b> CALIBRATION LABORATORY</p> <p>CAICT</p> <p>Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62302117 E-mail: cti@chinaetl.com http://www.caict.ac.cn</p> <p><b>Dipole Calibration /Pin=100mW, d=10mm, f=5500 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm</b> Reference Value = 61.92 V/m; Power Drift = -0.08 dB Peak SAR (extrapolated) = 34.7 W/kg SAR(1 g) = 8.28 W/kg; SAR(10 g) = 2.34 W/kg Smallest distance from peaks to all points 3 dB below = 7.2 mm Ratio of SAR at M2 to SAR at M1 = 63.9% Maximum value of SAR (measured) = 20.2 W/kg</p> <p><b>Dipole Calibration /Pin=100mW, d=10mm, f=5600 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm</b> Reference Value = 65.08 V/m; Power Drift = -0.07 dB Peak SAR (extrapolated) = 35.2 W/kg SAR(1 g) = 8.12 W/kg; SAR(10 g) = 2.3 W/kg Smallest distance from peaks to all points 3 dB below = 7.2 mm Ratio of SAR at M2 to SAR at M1 = 62.5% Maximum value of SAR (measured) = 19.1 W/kg</p> <p><b>Dipole Calibration /Pin=100mW, d=10mm, f=5800 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm</b> Reference Value = 62.13 V/m; Power Drift = -0.06 dB Peak SAR (extrapolated) = 34.8 W/kg SAR(1 g) = 7.71 W/kg; SAR(10 g) = 2.16 W/kg Smallest distance from peaks to all points 3 dB below = 7.2 mm Ratio of SAR at M2 to SAR at M1 = 61.6% Maximum value of SAR (measured) = 18.7 W/kg</p> <p>0 dB = 16.7 W/kg = 12.72 dBW/kg</p> <p>Certificate No: Z22-60187 Page 9 of 10</p>	<p>In Collaboration with <b>TTL</b> <b>s p e a q</b> CALIBRATION LABORATORY</p> <p>CAICT</p> <p>Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62302117 E-mail: cti@chinaetl.com http://www.caict.ac.cn</p> <p><b>Impedance Measurement Plot for Head TSL</b></p>  <p>Certificate No: Z22-60187 Page 10 of 10</p>



## 2 DAE4 - SN 1245

<p>Schmid &amp; Partner Engineering AG Zugstrasse 43, 8004 Zurich, Switzerland Phone +41 44 245 9700, Fax +41 44 245 9770 www.spgs.ch, info@spgs.ch</p> <p style="text-align: center;"><b>s p e a g</b></p> <p style="text-align: center;"><b>IMPORTANT NOTICE</b></p> <p><b>USAGE OF THE DAE4</b></p> <p>The DAE unit is a delicate, high precision instrument and requires careful treatment by the user. There are no serviceable parts inside the DAE. Special attention shall be given to the following points:</p> <p><b>Battery Exchange:</b> The battery cover of the DAE4 unit is fixed using a screw, over tightening the screw may cause the threads inside the DAE to wear out.</p> <p><b>Shipping of the DAE:</b> Before shipping the DAE to SPEAG for calibration, remove the batteries and pack the DAE in an anti-static bag. This anti-static bag shall then be packed into a larger box or container which protects the DAE from impacts during transportation. The package shall be marked to indicate that a fragile instrument is inside.</p> <p><b>E-Stop Failures:</b> Touch detection may be malfunctioning due to broken magnets in the E-stop. Rough handling of the E-stop may lead to damage of these magnets. Touch and collision errors are often caused by dust and dirt accumulated in the E-stop. To prevent E-stop failure, the customer shall always mount the probe to the DAE carefully and keep the DAE unit in a non-dusty environment if not used for measurements.</p> <p><b>Repair:</b> Minor repairs are performed at no extra cost during the annual calibration. However, SPEAG reserves the right to charge for any repair especially if rough unprofessional handling caused the defect.</p> <p><b>DASY Configuration Files:</b> Since the exact values of the DAE input resistances, as measured during the calibration procedure of a DAE unit, are not used by the DASY software, a nominal value of 200 MOhm is given in the corresponding configuration file.</p> <p><b>Important Note:</b> Warranty and calibration is void if the DAE unit is disassembled partly or fully by the Customer.</p> <p><b>Important Note:</b> Never attempt to grease or oil the E-stop assembly. Cleaning and readjusting of the E-stop assembly is allowed by certified SPEAG personnel only and is part of the annual calibration procedure.</p> <p><b>Important Note:</b> To prevent damage of the DAE probe connector pins, use great care when installing the probe to the DAE. Carefully connect the probe with the connector notch oriented in the mating position. Avoid any rotational movement of the probe body versus the DAE while turning the locking nut of the connector. The same care shall be used when disconnecting the probe from the DAE.</p> <p>TN_EH190306AE DAE4.docx 07.03.2019</p>	<p>Calibration Laboratory of Schmid &amp; Partner Engineering AG Zugstrasse 43, 8004 Zurich, Switzerland</p> <p> S Schweizerischer Kalibrierdienst C Service suisse d'étalonnage S Service suisse d'étalonnage S Swiss Calibration Service</p> <p>Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates</p> <p>Accreditation No.: SCS 0108</p> <p>Client: <b>SGS-CN (Auden)</b> Certificate No.: <b>DAE4-1245_May22</b></p> <p><b>CALIBRATION CERTIFICATE</b></p> <p>Object: <b>DAE4 - SD 000 D04 BM - SN: 1245</b></p> <p>Calibration procedure(s): <b>QA CAL-06 v30</b> Calibration procedure for the data acquisition electronics (DAE)</p> <p>Calibration date: <b>May 30, 2022</b></p> <p>This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.</p> <p>All calibrations have been conducted in the closed laboratory facility: environment temperature (22 ± 3) °C and humidity &lt; 70%.</p> <p>Calibration Equipment used (MATE critical for calibration)</p> <table border="1"><thead><tr><th>Primary Standards</th><th>ID #</th><th>Cal Date (Certificate No.)</th><th>Scheduled Calibration</th></tr></thead><tbody><tr><td>Kethley Multimeter Type 2001</td><td>SN: 0810276</td><td>31-Aug-21 (No.31368)</td><td>Aug-22</td></tr></tbody></table> <table border="1"><thead><tr><th>Secondary Standards</th><th>ID #</th><th>Check Date (in house)</th><th>Scheduled Check</th></tr></thead><tbody><tr><td>Auto DAE Calibration Unit</td><td>SE LWS 003 AA 1001</td><td>24-Jan-22 (in house check)</td><td>In house check: Jan-23</td></tr><tr><td>Calibrator Blue V2.1</td><td>SE LWS 006 AA 1002</td><td>24-Jan-22 (in house check)</td><td>In house check: Jan-23</td></tr></tbody></table> <p>Calibrated by: <b>Dominique Stettin</b> Function: <b>Laboratory Technician</b> Signature: </p> <p>Approved by: <b>Ben Kohn</b> Technical Manager Signature: </p> <p>This calibration certificate shall not be reproduced except in full without written approval of the laboratory. Issued: May 30, 2022</p> <p>Certificate No: DAE4-1245_May22 Page 1 of 5</p>	Primary Standards	ID #	Cal Date (Certificate No.)	Scheduled Calibration	Kethley Multimeter Type 2001	SN: 0810276	31-Aug-21 (No.31368)	Aug-22	Secondary Standards	ID #	Check Date (in house)	Scheduled Check	Auto DAE Calibration Unit	SE LWS 003 AA 1001	24-Jan-22 (in house check)	In house check: Jan-23	Calibrator Blue V2.1	SE LWS 006 AA 1002	24-Jan-22 (in house check)	In house check: Jan-23
Primary Standards	ID #	Cal Date (Certificate No.)	Scheduled Calibration																		
Kethley Multimeter Type 2001	SN: 0810276	31-Aug-21 (No.31368)	Aug-22																		
Secondary Standards	ID #	Check Date (in house)	Scheduled Check																		
Auto DAE Calibration Unit	SE LWS 003 AA 1001	24-Jan-22 (in house check)	In house check: Jan-23																		
Calibrator Blue V2.1	SE LWS 006 AA 1002	24-Jan-22 (in house check)	In house check: Jan-23																		
<p>Calibration Laboratory of Schmid &amp; Partner Engineering AG Zugstrasse 43, 8004 Zurich, Switzerland</p> <p> S Schweizerischer Kalibrierdienst C Service suisse d'étalonnage S Service suisse d'étalonnage S Swiss Calibration Service</p> <p>Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates</p> <p>Accreditation No.: SCS 0108</p> <p><b>Glossary</b></p> <p>DAE data acquisition electronics Connector angle information used in DASY system to align probe sensor X to the robot coordinate system.</p> <p><b>Methods Applied and Interpretation of Parameters</b></p> <ul style="list-style-type: none"><li>• <b>DC Voltage Measurement:</b> Calibration Factor assessed for use in DASY system by comparison with a calibrated instrument traceable to national standards. The figure given corresponds to the full scale range of the voltmeter in the respective range.</li><li>• <b>Connector angle:</b> The angle of the connector is assessed measuring the angle mechanically by a tool inserted. Uncertainty is not required.</li><li>• The following parameters as documented in the Appendix contain technical information as a result from the performance test and require no uncertainty.</li><li>• <b>DC Voltage Measurement Linearity:</b> Verification of the Linearity at +10% and -10% of the nominal calibration voltage. Influence of offset voltage is included in this measurement.</li><li>• <b>Common mode sensitivity:</b> Influence of a positive or negative common mode voltage on the differential measurement.</li><li>• <b>Channel separation:</b> Influence of a voltage on the neighbor channels not subject to an input voltage.</li><li>• <b>AD Converter Values with inputs shorted:</b> Values on the internal AD converter corresponding to zero input voltage.</li><li>• <b>Input Offset Measurement:</b> Output voltage and statistical results over a large number of zero voltage measurements.</li><li>• <b>Input Offset Current:</b> Typical value for information; Maximum channel input offset current, not considering the input resistance.</li><li>• <b>Input resistance:</b> Typical value for information; DAE input resistance at the connector, during internal auto-zeroing and during measurement.</li><li>• <b>Low Battery Alarm Voltage:</b> Typical value for information. Below this voltage, a battery alarm signal is generated.</li><li>• <b>Power consumption:</b> Typical value for information. Supply currents in various operating modes.</li></ul> <p>Certificate No: DAE4-1245_May22 Page 2 of 5</p>	<p><b>DC Voltage Measurement</b></p> <p>AD - Converter Resolution nominal High Range: 1LSB = 6.1µV, full range = -190...+320 mV Low Range: 1LSB = 61µV, full range = -1...+3mV DASY measurement parameters: Auto Zero-Time: 3 sec; Measuring time: 3 sec</p> <table border="1"><thead><tr><th>Calibration Factors</th><th>X</th><th>Y</th><th>Z</th></tr></thead><tbody><tr><td>High Range</td><td>405.265 ± 0.02% (k=2)</td><td>403.974 ± 0.02% (k=2)</td><td>406.092 ± 0.02% (k=2)</td></tr><tr><td>Low Range</td><td>3.99534 ± 1.50% (k=2)</td><td>3.99508 ± 1.50% (k=2)</td><td>4.01015 ± 1.50% (k=2)</td></tr></tbody></table> <p><b>Connector Angle</b></p> <table border="1"><thead><tr><th>Connector Angle to be used in DASY system</th><th>30.0° ± 1°</th></tr></thead></table> <p>Certificate No: DAE4-1245_May22 Page 3 of 5</p>	Calibration Factors	X	Y	Z	High Range	405.265 ± 0.02% (k=2)	403.974 ± 0.02% (k=2)	406.092 ± 0.02% (k=2)	Low Range	3.99534 ± 1.50% (k=2)	3.99508 ± 1.50% (k=2)	4.01015 ± 1.50% (k=2)	Connector Angle to be used in DASY system	30.0° ± 1°						
Calibration Factors	X	Y	Z																		
High Range	405.265 ± 0.02% (k=2)	403.974 ± 0.02% (k=2)	406.092 ± 0.02% (k=2)																		
Low Range	3.99534 ± 1.50% (k=2)	3.99508 ± 1.50% (k=2)	4.01015 ± 1.50% (k=2)																		
Connector Angle to be used in DASY system	30.0° ± 1°																				



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic formal documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: [CN.Doccheck@sgs.com](mailto:CN.Doccheck@sgs.com)  
No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300  
中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 [www.sgsgroup.com.cn](http://www.sgsgroup.com.cn)  
t(86-512)57355888 f(86-512)57370818 [sgs.china@sgs.com](mailto:sgs.china@sgs.com)

## Appendix (Additional assessments outside the scope of SCS0108)

## 1. DC Voltage Linearity

High Range	Reading (µV)	Difference (µV)	Error (%)
Channel X + Input	19994.45	1.52	0.00
Channel X + Input	20004.58	2.22	0.01
Channel X - Input	-20001.14	1.12	-0.01
Channel Y + Input	19994.72	1.58	0.00
Channel Y + Input	20001.22	-1.00	-0.00
Channel Y - Input	-20003.05	-1.57	0.01
Channel Z + Input	19992.44	0.19	0.00
Channel Z + Input	20003.09	0.58	0.00
Channel Z - Input	-20001.73	-0.27	0.00
Low Range	Reading (µV)	Difference (µV)	Error (%)
Channel X + Input	2001.91	0.41	0.02
Channel X + Input	202.54	0.65	0.32
Channel X - Input	-197.86	0.07	-0.04
Channel Y + Input	2002.05	0.58	0.03
Channel Y + Input	201.27	-0.57	-0.28
Channel Y - Input	-196.23	-0.06	0.03
Channel Z + Input	2001.98	0.08	0.00
Channel Z + Input	200.09	-1.53	-0.76
Channel Z - Input	-199.85	-1.57	0.79

## 2. Common mode sensitivity

Channel	Common mode Input Voltage (mV)	High Range Average Reading (µV)	Low Range Average Reading (µV)
Channel X	200	-5.87	-7.69
	-200	9.12	7.79
Channel Y	200	-8.68	-9.28
	-200	8.52	6.36
Channel Z	200	-5.36	-5.60
	-200	3.58	3.06

## 3. Channel separation

Channel	Input Voltage (mV)	Channel X (µV)	Channel Y (µV)	Channel Z (µV)
Channel X	200	-	4.07	-3.14
Channel Y	200	9.36	-	4.27
Channel Z	200	10.11	7.14	-

Certificate No: DAE4-1245\_May22

Page 4 of 5

## 4. AD-Converter Values with inputs shorted

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

	High Range (LSB)	Low Range (LSB)
Channel X	15984	17040
Channel Y	16562	16768
Channel Z	16035	15968

## 5. Input Offset Measurement

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

Input (mV)	Average (µV)	min. Offset (µV)	max. Offset (µV)	Std. Deviation (µV)
Channel X	1.00	-0.15	1.93	0.45
Channel Y	-0.18	-1.28	0.94	0.45
Channel Z	-0.58	-2.81	0.58	0.60

## 6. Input Offset Current

Nominal input circuitry offset current on all channels: &lt;25nA

## 7. Input Resistance (Typical values for information)

	Zeroing (kOhm)	Measuring (MOhm)
Channel X	200	200
Channel Y	200	200
Channel Z	200	200

## 8. Low Battery Alarm Voltage (Typical values for information)

Typical values	Alarm Level (VDC)
Supply (+ Vcc)	+7.9
Supply (- Vcc)	-7.6

## 9. Power Consumption (Typical values for information)

Typical values	Switched off (mA)	Stand by (mA)	Transmitting (mA)
Supply (+ Vcc)	+0.01	+8	+14
Supply (- Vcc)	-0.01	-8	-9

Certificate No: DAE4-1245\_May22

Page 5 of 5

## 3 EX3DV4 - SN 7346

<p>Calibration Laboratory of Schmid &amp; Partner Engineering AG Zugstrasse 65, 8044 Zurich, Switzerland</p> <p>Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates</p> <p>Client: Auden Certificate No: EX3-7346_Mar22</p> <p>Accreditation No: SCS 0108</p>	<p>Calibration Laboratory of Schmid &amp; Partner Engineering AG Zugstrasse 65, 8044 Zurich, Switzerland</p> <p>Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates</p> <p>Accreditation No: SCS 0108</p>																																															
<p><b>CALIBRATION CERTIFICATE</b></p> <p>Object: EX3DV4 - SN 7346</p> <p>Calibration procedure(s): QA CAL-01 v9; QA CAL-14 v6; QA CAL-23 v5; QA CAL-25 v7 Calibration procedure for domestic E-field probes</p> <p>Calibration date: March 30, 2022</p> <p>This calibration certificate documents the traceability to national standards, which include the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.</p> <p>All calibrations have been conducted in the closed laboratory facility: environment temperature (22 ± 3) °C and humidity &lt; 70%.</p> <p>Calibration Equipment used (MATE critical for calibration):</p> <table><tr><th>Primary Standards</th><th>ID</th><th>Cal Date (Certificate No.)</th><th>Scheduled Calibration</th></tr><tr><td>Power meter NRP</td><td>SN: 10478</td><td>09-Apr-21 (No. 217-03201/03202)</td><td>Apr-22</td></tr><tr><td>Power sensor NRP-291</td><td>SN: 10304</td><td>09-Apr-21 (No. 217-03201)</td><td>Apr-22</td></tr><tr><td>Power sensor NRP-291</td><td>SN: 10304</td><td>09-Apr-21 (No. 217-03202)</td><td>Apr-22</td></tr><tr><td>Reference 20 dB attenuator</td><td>SN: C22052 (20b)</td><td>09-Apr-21 (No. 217-03203)</td><td>Apr-22</td></tr><tr><td>DAE4</td><td>SN: 660</td><td>13-Dec-21 (No. DAE4-485; DAE21)</td><td>Dec-22</td></tr><tr><td>Reference Probe ES302</td><td>SN: 3013</td><td>27-Dec-21 (No. ES3-3013 Dec21)</td><td>Dec-22</td></tr></table> <p>Secondary Standards</p> <table><tr><th>ID</th><th>Check Date (in house)</th><th>Scheduled Calibration</th></tr><tr><td>Power meter E4413B</td><td>SN: G84123074 06-Apr-16 (in house check Jun-20)</td><td>In house check Jun-22</td></tr><tr><td>Power sensor E4413A</td><td>SN: MY4149887 06-Apr-16 (in house check Jun-20)</td><td>In house check Jun-22</td></tr><tr><td>Power sensor E4413A</td><td>SN: 06011010 06-Apr-16 (in house check Jun-20)</td><td>In house check Jun-22</td></tr><tr><td>RF generator HP 8640C</td><td>SN: US342011109 04-Apr-09 (in house check Jun-20)</td><td>In house check Jun-22</td></tr><tr><td>Network Analyzer E8309A</td><td>SN: US41050477 31-Mar-14 (in house check Oct-20)</td><td>In house check Oct-22</td></tr></table> <p>Calibrated by: Name: Jörn Kuster Function: Laboratory Technician Signature: [Signature]</p> <p>Approved by: Sven Kuster Deputy Manager Signature: [Signature]</p> <p>This calibration certificate shall not be reproduced except in full without written approval of the laboratory.</p> <p>Issued: March 31, 2022</p> <p>Certificate No: EX3-7346_Mar22 Page 1 of 24</p>		Primary Standards	ID	Cal Date (Certificate No.)	Scheduled Calibration	Power meter NRP	SN: 10478	09-Apr-21 (No. 217-03201/03202)	Apr-22	Power sensor NRP-291	SN: 10304	09-Apr-21 (No. 217-03201)	Apr-22	Power sensor NRP-291	SN: 10304	09-Apr-21 (No. 217-03202)	Apr-22	Reference 20 dB attenuator	SN: C22052 (20b)	09-Apr-21 (No. 217-03203)	Apr-22	DAE4	SN: 660	13-Dec-21 (No. DAE4-485; DAE21)	Dec-22	Reference Probe ES302	SN: 3013	27-Dec-21 (No. ES3-3013 Dec21)	Dec-22	ID	Check Date (in house)	Scheduled Calibration	Power meter E4413B	SN: G84123074 06-Apr-16 (in house check Jun-20)	In house check Jun-22	Power sensor E4413A	SN: MY4149887 06-Apr-16 (in house check Jun-20)	In house check Jun-22	Power sensor E4413A	SN: 06011010 06-Apr-16 (in house check Jun-20)	In house check Jun-22	RF generator HP 8640C	SN: US342011109 04-Apr-09 (in house check Jun-20)	In house check Jun-22	Network Analyzer E8309A	SN: US41050477 31-Mar-14 (in house check Oct-20)	In house check Oct-22	<p><b>Glossary:</b></p> <p>TSL: Issue simulating liquid sensitivity in free space</p> <p>NORM<sub>M,y,z</sub>: sensitivity in TSL / NORM<sub>M,y,z</sub></p> <p>Comp: diode compression point</p> <p>CF: crest factor (10 duty cycle) of the RF signal</p> <p>A, B, C, D: modulation dependent linearization parameters</p> <p>Polarization: ϕ rotation around probe axis</p> <p>Polarization: θ rotation around an axis that is in the plane normal to probe axis (at measurement center). i.e., θ = 0 is normal to probe axis</p> <p>Connector Angle: information used in DASY system to align probe sensor X to the robot coordinate system</p> <p><b>Calibration is Performed According to the Following Standards:</b></p> <p>a) IEC/IEEE 62208-1:2018 "Measurement Procedure For The Assessment Of Specific Absorption Rate Of Human Exposure To Radio Frequency Fields From Hand-Held And Body-Worn Wireless Communication Devices - Part 1:333: Human Models, Instrumentation And Procedures (Frequency Range of 4 MHz to 10 GHz)" October 2020</p> <p>b) KOB 805664 "SAR Measurement Requirements for 100 MHz to 6 GHz"</p> <p><b>Methods Applied and Interpretation of Parameters:</b></p> <ul style="list-style-type: none"><li>NORM<sub>M,y,z</sub>: Assessed for E-field polarization θ = 0 (f = 800 MHz in TEM-cell; f = 1800 MHz: R22 waveguide). NORM<sub>M,y,z</sub> are only intermediate values, i.e., the uncertainties of NORM<sub>M,y,z</sub> do not affect the E-field uncertainty inside TSL (see below Conf).</li><li>NORM<sub>M,y,z</sub> = NORM<sub>M,y,z</sub> * Frequency response (see Frequency Response Chart). This linearization is implemented in DASY4 software version later than 4.2. The uncertainty of the frequency response is included in the stated uncertainty of Conf.</li><li>DCP<sub>M,y,z</sub>: DCP are numerical linearization parameters assessed based on the data of power sweep with CW signal (no uncertainty required). DCP does not depend on frequency nor media.</li><li>PAC: PAC is the Peak to Average Ratio that is not calibrated but determined based on the signal characteristics.</li><li>Al<sub>M,y,z</sub>, B<sub>M,y,z</sub>, C<sub>M,y,z</sub>, D<sub>M,y,z</sub>, V<sub>M,y,z</sub>: A, B, C, D are numerical linearization parameters assessed based on the data of power sweep for specific modulation signal. The parameters do not depend on frequency nor media. V<sub>M,y,z</sub> is the maximum calibration range expressed in RMS voltage across the diode.</li><li>Conf<sub>M</sub> and Boundary Effect Parameters: Assessed in far phantom using E-field (or Temperature Transfer Standard for f &lt; 800 MHz) and inside waveguide using analytical field distributions based on power measurements for f &gt; 800 MHz. The same values are used for assessment of the parameters applied for boundary compensation (alpha, depth) of which typical uncertainty values are given. These parameters are to NORM<sub>M,y,z</sub> * Conf<sub>M</sub> whereby the uncertainty corresponds to that given for Conf<sub>M</sub>. A frequency dependent Conf<sub>M</sub> is used in DASY version 4.4 and higher which allows extending the validity from ± 50 MHz to ± 100 MHz.</li><li>Spherical isotropy / DCP deviation from isotropy: in a field of low gradients realized using a flat phantom exposed by a patch antenna.</li><li>Sensor Offset: The sensor offset corresponds to the offset of virtual measurement center from the probe tip (on probe axis). No tolerance required.</li><li>Connector Angle: The angle is assessed using the information gained by determining the NORM<sub>M</sub> (no uncertainty required).</li></ul> <p>Certificate No: EX3-7346_Mar22 Page 2 of 24</p>
Primary Standards	ID	Cal Date (Certificate No.)	Scheduled Calibration																																													
Power meter NRP	SN: 10478	09-Apr-21 (No. 217-03201/03202)	Apr-22																																													
Power sensor NRP-291	SN: 10304	09-Apr-21 (No. 217-03201)	Apr-22																																													
Power sensor NRP-291	SN: 10304	09-Apr-21 (No. 217-03202)	Apr-22																																													
Reference 20 dB attenuator	SN: C22052 (20b)	09-Apr-21 (No. 217-03203)	Apr-22																																													
DAE4	SN: 660	13-Dec-21 (No. DAE4-485; DAE21)	Dec-22																																													
Reference Probe ES302	SN: 3013	27-Dec-21 (No. ES3-3013 Dec21)	Dec-22																																													
ID	Check Date (in house)	Scheduled Calibration																																														
Power meter E4413B	SN: G84123074 06-Apr-16 (in house check Jun-20)	In house check Jun-22																																														
Power sensor E4413A	SN: MY4149887 06-Apr-16 (in house check Jun-20)	In house check Jun-22																																														
Power sensor E4413A	SN: 06011010 06-Apr-16 (in house check Jun-20)	In house check Jun-22																																														
RF generator HP 8640C	SN: US342011109 04-Apr-09 (in house check Jun-20)	In house check Jun-22																																														
Network Analyzer E8309A	SN: US41050477 31-Mar-14 (in house check Oct-20)	In house check Oct-22																																														



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic formal documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing/inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: [CN.Doccheck@sgs.com](mailto:CN.Doccheck@sgs.com)

No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300  
中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgsgroup.com.cn  
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

EX3DV4 - SN:7346									
March 30, 2022									
DASY/EASY - Parameters of Probe: EX3DV4 - SN:7346									
<b>Basic Calibration Parameters</b>									
Norm. $\mu V/(V/m)^2$		Sensor X		Sensor Y		Sensor Z		Unc. (k=2)	
0.45		0.45		0.45		0.45		± 10.1 %	
DCP (mV/P)		101.4		106.0		106.9			
<b>Calibration Results for Modulation Response</b>									
UID	Communication System Name	A dB	B dB-μV	C	D dB	VR mV	Max dBm	Max Unc <sup>(k=2)</sup>	
0	CW	X: 0.00	0.00	1.00	0.00	143.5	± 3.0 %	± 4.7 %	
		Y: 0.00	0.00	1.00		130.3			
		Z: 0.00	0.00	1.00		130.0			
10303-AAA	Pulse Waveform (200Hz, 10%)	X: 3.33	68.90	11.66	10.00	60.0	± 3.5 %	± 9.8 %	
		Y: 4.03	70.70	12.35		60.0			
		Z: 1.63	61.25	6.76		60.0			
10303-AAA	Pulse Waveform (200Hz, 20%)	X: 3.00	70.65	11.31	6.99	60.0	± 2.4 %	± 9.6 %	
		Y: 11.51	81.32	14.72		80.0			
		Z: 0.83	60.00	5.11		60.0			
10304-AAA	Pulse Waveform (200Hz, 40%)	X: 7.41	79.85	12.51	3.98	80.0	± 2.7 %	± 9.8 %	
		Y: 20.00	81.42	15.41		90.0			
		Z: 0.18	138.38	0.01		90.0			
10305-AAA	Pulse Waveform (200Hz, 60%)	X: 2.07	71.13	9.50	2.22	120.0	± 1.7 %	± 9.6 %	
		Y: 20.00	91.58	16.29		120.0			
		Z: 1.04	138.57	16.87		120.0			
10307-AAA	GRK Waveform, 1 MHz	X: 1.47	64.88	13.82	1.00	100.0	± 4.2 %	± 9.6 %	
		Y: 1.96	66.27	14.65		100.0			
		Z: 0.43	67.88	11.05		100.0			
10308-AAA	GRK Waveform, 10 MHz	X: 0.96	66.27	14.65	0.00	100.0	± 1.1 %	± 9.6 %	
		Y: 2.06	67.33	15.38		100.0			
		Z: 0.41	64.75	13.18		100.0			
10306-AAA	64-QAM Waveform, 100 kHz	X: 2.63	68.51	18.25	3.01	100.0	± 1.0 %	± 9.6 %	
		Y: 1.12	67.12	15.99		100.0			
		Z: 1.70	64.72	15.99		100.0			
10308-AAA	64-QAM Waveform, 40 MHz	X: 3.38	66.82	15.23	0.00	100.0	± 2.0 %	± 9.6 %	
		Y: 1.12	67.12	15.99		100.0			
		Z: 2.70	65.12	14.74		100.0			
10414-AAA	WLAN CCDF, 64-QAM, 60MHz	X: 4.71	65.35	13.77	0.00	100.0	± 3.6 %	± 9.6 %	
		Y: 4.70	65.54	15.41		100.0			
		Z: 3.83	66.16	15.28		100.0			
Note: For details on UID parameters see Appendix									
The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.									
* The uncertainties of Norm X, Y, Z do not affect the E-field uncertainty result. (See Pages 1 and 2)									
* Numerical limitation parameter, uncertainty not required									
* Uncertainty is determined using the max. deviation from linear response applying rectangular distribution and is expressed for the square of the field value.									
Certificate No: EX3-7346_Mar22 Page 3 of 34									

EX3DV4 - SN:7346									
March 30, 2022									
DASY/EASY - Parameters of Probe: EX3DV4 - SN:7346									
<b>Sensor Model Parameters</b>									
C1	C2	a	T1	T2	T3	T4	T5	T6	
IP	IP	V <sup>2</sup>	ms.V <sup>2</sup>	ms.V <sup>2</sup>	ms	V <sup>2</sup>	V <sup>2</sup>		
X	39.2	291.80	35.10	5.63	5.02	1.42	0.12	1.01	
Y	37.1	270.84	34.12	6.29	5.01	1.82	0.05	1.01	
Z	9.7	69.74	33.37	4.96	0.00	4.94	0.01	0.00	1.00
<b>Other Probe Parameters</b>									
Sensor Arrangement									
Triangular									
Connector Angle (°)									
-166.1									
Mechanical Surface Detection Mode									
enabled									
Optical Surface Detection Mode									
disabled									
Probe Overall Length									
337 mm									
Probe Body Diameter									
10 mm									
Tip Length									
9 mm									
Tip Diameter									
2.5 mm									
Probe Tip to Sensor X Calibration Point									
1 mm									
Probe Tip to Sensor Y Calibration Point									
1 mm									
Probe Tip to Sensor Z Calibration Point									
1 mm									
Recommended Measurement Distance from Surface									
1.4 mm									
Note: Measurement distance from surface can be increased to 3-4 mm for an Area Scan job.									
Certificate No: EX3-7346_Mar22 Page 4 of 34									

EX3DV4 - SN:7346									
March 30, 2022									
DASY/EASY - Parameters of Probe: EX3DV4 - SN:7346									
<b>Calibration Parameter Determined in Head Tissue Simulating Media</b>									
f (MHz)	Relative Permittivity <sup>1</sup>	Conductivity (S/m) <sup>2</sup>	ConvF X	ConvF Y	ConvF Z	Alpha <sup>3</sup>	Depth <sup>4</sup> (mm)	Unc. (k=2)	
6500	34.5	6.07	5.30	5.30	5.30	0.20	2.50	± 18.0 %	
<sup>1</sup> Frequency validity above 300 MHz is ± 100 MHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band.									
<sup>2</sup> At frequencies 5-10 GHz, the validity of tissue parameters (ρ and α) can be related to ± 10% if liquid compensation formula is applied to measured S-parameters. The uncertainty is the RSS of the ConvF uncertainty for indicated target tissue parameters.									
<sup>3</sup> Alpha/Depth are determined during calibration. SP-6A0 warrants that the remaining deviation due to the boundary effect after compensation is always less than ± 1% for frequencies below 3 GHz, below ± 2% for frequencies between 3-6 GHz, and below ± 4% for frequencies between 6-10 GHz at any distance larger than half the probe tip diameter from the boundary.									
Certificate No: EX3-7346_Mar22 Page 5 of 34									

EX3DV4 - SN:7346									
March 30, 2022									
DASY/EASY - Parameters of Probe: EX3DV4 - SN:7346									
<b>Calibration Parameter Determined in Head Tissue Simulating Media</b>									
f (MHz)	Relative Permittivity <sup>1</sup>	Conductivity (S/m) <sup>2</sup>	ConvF X	ConvF Y	ConvF Z	Alpha <sup>3</sup>	Depth <sup>4</sup> (mm)	Unc. (k=2)	
750	41.9	0.89	10.56	10.56	10.56	0.55	0.85	± 12.0 %	
835	41.5	0.90	10.12	10.12	10.12	0.42	0.96	± 12.0 %	
900	41.5	0.97	10.10	10.10	10.10	0.53	0.80	± 12.0 %	
1430	40.5	1.20	9.26	9.26	9.26	0.50	0.80	± 12.0 %	
1750	40.1	1.37	8.83	8.83	8.83	0.34	0.86	± 12.0 %	
1900	40.0	1.40	8.48	8.48	8.48	0.35	0.86	± 12.0 %	
2000	40.0	1.40	8.35	8.35	8.35	0.34	0.86	± 12.0 %	
2300	39.5	1.67	7.86	7.86	7.86	0.39	0.90	± 12.0 %	
2450	39.2	1.80	7.63	7.63	7.63	0.41	0.90	± 12.0 %	
2600	39.0	1.96	7.33	7.33	7.33	0.44	0.90	± 12.0 %	
3300	38.2	2.71	7.15	7.15	7.15	0.30	1.35	± 13.1 %	
3500	37.8	2.91	7.14	7.14	7.14	0.30	1.35	± 13.1 %	
3750	37.7	3.12	6.85	6.85	6.85	0.30	1.35	± 13.1 %	
3900	37.5	3.32	6.71	6.71	6.71	0.40	1.60	± 13.1 %	
4100	37.2	3.53	6.58	6.58	6.58	0.40	1.60	± 13.1 %	
4200	37.1	3.63	6.30	6.30	6.30	0.40	1.70	± 13.1 %	
4400	36.9	3.84	6.24	6.24	6.24	0.40	1.70	± 13.1 %	
4600	36.7	4.04	6.11	6.11	6.11	0.40	1.70	± 13.1 %	
4800	36.4	4.25	6.08	6.08	6.08	0.40	1.80	± 13.1 %	
4900	36.3	4.40	5.84	5.84	5.84	0.40	1.80	± 13.1 %	
5200	36.0	4.66	5.25	5.25	5.25	0.40	1.80	± 13.1 %	
5300	35.9	4.78	5.12	5.12	5.12	0.40	1.80	± 13.1 %	
5500	35.6	4.96	4.85	4.85	4.85	0.40	1.80	± 13.1 %	
5600	35.5	5.07	4.70	4.70	4.70	0.40	1.80	± 13.1 %	
5800	35.3	5.27	4.75	4.75	4.75	0.40	1.80	± 13.1 %	
<sup>1</sup> Frequency validity above 300 MHz is ± 100 MHz only applies for DASY v4.4 and higher (see Page 2), else it is restricted to ± 50 MHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. (Frequency validity below 300 MHz is ± 10, 25, 40, 50 and 70 MHz for ConvF assessments at 34, 64, 126, 150 and 200 MHz respectively. Validity of ConvF assessments at 8 MHz is ± 8 MHz, and ConvF assessed at 13 MHz is ± 10 MHz. Above 5 GHz frequency validity can be extended to ± 110 MHz.)									
<sup>2</sup> At frequencies below 3 GHz, the validity of tissue parameters (ρ and α) can be related to ± 10% if liquid compensation formula is applied to measured S-parameters. The uncertainty is the RSS of the ConvF uncertainty for indicated target tissue parameters.									
<sup>3</sup> Alpha/Depth are determined during calibration. SP-6A0 warrants that the remaining deviation due to the boundary effect after compensation is always less than ± 1% for frequencies below 3 GHz and below ± 2% for frequencies between 3-6 GHz at any distance larger than half the probe tip diameter from the boundary.									
Certificate No: EX3-7346_Mar22 Page 6 of 34									




Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing/inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300  
中国·江苏·昆山市留学院创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgsgroup.com.cn  
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com





Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: [CN.Doccheck@sgs.com](mailto:CN.Doccheck@sgs.com)

No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300  
中国·江苏·昆山市留学院创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 [www.sgsgroup.com.cn](http://www.sgsgroup.com.cn)  
t(86-512)57355888 f(86-512)57370818 [sgs.china@sgs.com](mailto:sgs.china@sgs.com)



EK32V4-SN-7346		March 30, 2022		EK32V4-SN-7346		March 30, 2022	
Appendix: Modulation Calibration Parameters							
UID	Rev	Communication System Name	Group	FARK (dB)	Unc2 (dB)		
10015	CAA	SAR Validation (Station, 100MHz, 10m)	CW	0.00	+4.7%		
10016	CAB	UITS-FDD (WCDMA)	WCDMA	2.81	+9.6%		
10017	CAB	IEEE 802.11a WIF 2.4 GHz (DSSS, 1Mbps)	WLAN	1.87	+9.6%		
10018	CAB	IEEE 802.11a WIF 2.4 GHz (DSSS-OFDM, 6 Mbps)	WLAN	9.46	+9.6%		
10019	DAC	GSM-FDD (TDMA, GSM)	GSM	9.57	+9.6%		
10020	DAC	GSM-FDD (TDMA, GSM, TN-0)	GSM	6.56	+9.6%		
10021	DAC	EDGE-FDD (TDMA, BPSK, TN-0-1)	GSM	12.62	+9.6%		
10022	DAC	EDGE-FDD (TDMA, BPSK, TN-0-2)	GSM	9.65	+9.6%		
10023	DAC	GPRS-FDD (TDMA, GSM, TN-0-2)	GSM	4.80	+9.6%		
10024	DAC	GPRS-FDD (TDMA, GSM, TN-0-2b)	GSM	3.50	+9.6%		
10025	DAC	EDGE-FDD (TDMA, BPSK, TN-0-3)	GSM	7.78	+9.6%		
10030	CAA	IEEE 802.15.1 Bluetooth (GFSK, Dn)	Bluetooth	5.20	+9.6%		
10031	CAA	IEEE 802.15.1 Bluetooth (GFSK, Dn)	Bluetooth	1.87	+9.6%		
10032	CAA	IEEE 802.15.1 Bluetooth (GFSK, Dn)	Bluetooth	1.18	+9.6%		
10033	CAA	IEEE 802.15.1 Bluetooth (P4-QPSK, Dn)	Bluetooth	7.74	+9.6%		
10034	CAA	IEEE 802.15.1 Bluetooth (P4-QPSK, Dn)	Bluetooth	4.53	+9.6%		
10035	CAA	IEEE 802.15.1 Bluetooth (P4-QPSK, Dn)	Bluetooth	3.83	+9.6%		
10036	CAA	IEEE 802.15.1 Bluetooth (8-QPSK, Dn)	Bluetooth	8.01	+9.6%		
10037	CAA	IEEE 802.15.1 Bluetooth (8-QPSK, Dn)	Bluetooth	4.77	+9.6%		
10038	CAA	IEEE 802.15.1 Bluetooth (8-QPSK, Dn)	Bluetooth	4.32	+9.6%		
10039	CAB	CDMA2000 (1XRTT, RC)	CDMA2000	4.57	+9.6%		
10040	CAB	IS-94 (1.35 FDD (TDMA/FDD, P4-QPSK, Full Rate))	AMPS	7.78	+9.6%		
10044	CAA	IS-94 (1.35 FDD (TDMA/FDD, P4-QPSK, Full Rate))	AMPS	0.00	+9.6%		
10048	CAA	DECT (D13, TDMA/FDD, GFSK, Full Rate, 20)	DECT	13.80	+9.6%		
10049	CAA	DECT (D13, TDMA/FDD, GFSK, Full Rate, 20)	DECT	15.79	+9.6%		
10054	CAA	UMTS-FDD (TD-SCDMA, 1.28 Mbps)	TD-SCDMA	11.01	+9.6%		
10058	CAB	EDGE-FDD (TDMA, BPSK, TN-0-3)	WCDMA	6.92	+9.6%		
10059	CAB	IEEE 802.11a WIF 2.4 GHz (DSSS, 2 Mbps)	WLAN	2.12	+9.6%		
10060	CAB	IEEE 802.11a WIF 2.4 GHz (DSSS, 5.5 Mbps)	WLAN	2.83	+9.6%		
10061	CAB	IEEE 802.11a WIF 2.4 GHz (DSSS, 11 Mbps)	WLAN	3.80	+9.6%		
10062	CAB	IEEE 802.11a WIF 2.4 GHz (OFDM, 6 Mbps)	WLAN	8.68	+9.6%		
10063	CAB	IEEE 802.11a WIF 2.4 GHz (OFDM, 9 Mbps)	WLAN	8.63	+9.6%		
10064	CAB	IEEE 802.11a WIF 2.4 GHz (OFDM, 12 Mbps)	WLAN	9.08	+9.6%		
10065	CAB	IEEE 802.11a WIF 2.4 GHz (OFDM, 18 Mbps)	WLAN	9.00	+9.6%		
10066	CAB	IEEE 802.11a WIF 2.4 GHz (OFDM, 24 Mbps)	WLAN	9.38	+9.6%		
10067	CAB	IEEE 802.11a WIF 2.4 GHz (OFDM, 36 Mbps)	WLAN	10.12	+9.6%		
10068	CAB	IEEE 802.11a WIF 2.4 GHz (OFDM, 48 Mbps)	WLAN	10.24	+9.6%		
10069	CAB	IEEE 802.11a WIF 2.4 GHz (OFDM, 54 Mbps)	WLAN	10.56	+9.6%		
10070	CAB	IEEE 802.11a WIF 2.4 GHz (DSSS-OFDM, 2 Mbps)	WLAN	9.60	+9.6%		
10072	CAB	IEEE 802.11a WIF 2.4 GHz (DSSS-OFDM, 12 Mbps)	WLAN	9.62	+9.6%		
10073	CAB	IEEE 802.11a WIF 2.4 GHz (DSSS-OFDM, 18 Mbps)	WLAN	9.94	+9.6%		
10074	CAB	IEEE 802.11a WIF 2.4 GHz (DSSS-OFDM, 24 Mbps)	WLAN	10.30	+9.6%		
10075	CAB	IEEE 802.11a WIF 2.4 GHz (DSSS-OFDM, 36 Mbps)	WLAN	10.77	+9.6%		
10076	CAB	IEEE 802.11a WIF 2.4 GHz (DSSS-OFDM, 48 Mbps)	WLAN	10.84	+9.6%		
10077	CAB	IEEE 802.11a WIF 2.4 GHz (DSSS-OFDM, 54 Mbps)	WLAN	11.00	+9.6%		
10081	CAA	CDMA2000 (1XRTT, RC)	CDMA2000	3.97	+9.6%		
10085	CAB	IS-94 (1.35 FDD (TDMA/FDD, P4-QPSK, Full Rate))	AMPS	4.77	+9.6%		
10090	DAC	GSM-FDD (TDMA, GSM, TN-0-4)	GSM	6.56	+9.6%		
10091	CAB	UMTS-FDD (HSPA, Subnet 2)	WCDMA	3.98	+9.6%		
10098	CAB	UMTS-FDD (HSPA, Subnet 2)	WCDMA	3.98	+9.6%		
10099	DAC	EDGE-FDD (TDMA, BPSK, TN-0-4)	GSM	9.55	+9.6%		
10100	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	5.67	+9.6%		
10101	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, 16-QAM)	LTE-FDD	6.42	+9.6%		
10102	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, 64-QAM)	LTE-FDD	8.60	+9.6%		
10103	CAE	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-TDD	9.29	+9.6%		
10104	CAE	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, 16-QAM)	LTE-TDD	9.97	+9.6%		
10105	CAE	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, 64-QAM)	LTE-TDD	10.01	+9.6%		
10106	CAG	LTE-FDD (SC-FDMA, 100% RB, 10 MHz, QPSK)	LTE-FDD	5.80	+9.6%		
10107	CAG	LTE-FDD (SC-FDMA, 100% RB, 10 MHz, 16-QAM)	LTE-FDD	6.43	+9.6%		
10108	CAG	LTE-FDD (SC-FDMA, 100% RB, 10 MHz, 64-QAM)	LTE-FDD	8.61	+9.6%		
10109	CAF	LTE-FDD (SC-FDMA, 100% RB, 5 MHz, QPSK)	LTE-FDD	5.75	+9.6%		
10110	CAF	LTE-FDD (SC-FDMA, 100% RB, 5 MHz, 16-QAM)	LTE-FDD	6.44	+9.6%		
10111	CAF	LTE-FDD (SC-FDMA, 100% RB, 5 MHz, 64-QAM)	LTE-FDD	6.59	+9.6%		
10112	CAF	LTE-FDD (SC-FDMA, 100% RB, 10 MHz, 64-QAM)	LTE-FDD	6.62	+9.6%		
10113	CAF	IEEE 802.11n (HT Greenfield, 13.5 Mbps, BPSK)	WLAN	8.30	+9.6%		
10114	CAF	IEEE 802.11n (HT Greenfield, 13.5 Mbps, BPSK)	WLAN	8.46	+9.6%		
10115	CAF	IEEE 802.11n (HT Greenfield, 13.5 Mbps, BPSK)	WLAN	8.15	+9.6%		
10116	CAF	IEEE 802.11n (HT MIMO, 13.5 Mbps, BPSK)	WLAN	8.07	+9.6%		
10117	CAF	IEEE 802.11n (HT MIMO, 13.5 Mbps, BPSK)	WLAN	8.59	+9.6%		
10118	CAF	IEEE 802.11n (HT MIMO, 13.5 Mbps, BPSK)	WLAN	8.13	+9.6%		
10119	CAF	LTE-FDD (SC-FDMA, 100% RB, 15 MHz, 16-QAM)	LTE-FDD	6.48	+9.6%		
10120	CAF	LTE-FDD (SC-FDMA, 100% RB, 15 MHz, 64-QAM)	LTE-FDD	6.53	+9.6%		
10121	CAF	LTE-FDD (SC-FDMA, 100% RB, 15 MHz, QPSK)	LTE-FDD	5.73	+9.6%		
10122	CAF	LTE-FDD (SC-FDMA, 100% RB, 15 MHz, 16-QAM)	LTE-FDD	6.38	+9.6%		
10123	CAF	LTE-FDD (SC-FDMA, 100% RB, 15 MHz, 64-QAM)	LTE-FDD	6.65	+9.6%		
10124	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.4 MHz, QPSK)	LTE-FDD	5.76	+9.6%		
10125	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.4 MHz, 16-QAM)	LTE-FDD	6.47	+9.6%		
10126	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.4 MHz, 64-QAM)	LTE-FDD	6.72	+9.6%		
10127	CAF	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, 16-QAM)	LTE-FDD	6.42	+9.6%		
10128	CAF	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	6.60	+9.6%		
10129	CAF	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-TDD	9.28	+9.6%		
10130	CAF	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, 16-QAM)	LTE-TDD	9.97	+9.6%		
10131	CAF	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, 64-QAM)	LTE-TDD	10.05	+9.6%		
10132	CAF	LTE-FDD (SC-FDMA, 100% RB, 10 MHz, QPSK)	LTE-FDD	5.73	+9.6%		
10133	CAF	LTE-FDD (SC-FDMA, 100% RB, 10 MHz, 16-QAM)	LTE-FDD	6.43	+9.6%		
10134	CAF	LTE-FDD (SC-FDMA, 100% RB, 10 MHz, 64-QAM)	LTE-FDD	5.79	+9.6%		
10135	CAF	LTE-FDD (SC-FDMA, 100% RB, 10 MHz, 16-QAM)	LTE-FDD	6.42	+9.6%		
10136	CAF	LTE-FDD (SC-FDMA, 100% RB, 10 MHz, 64-QAM)	LTE-FDD	6.62	+9.6%		
10137	CAF	LTE-FDD (SC-FDMA, 100% RB, 15 MHz, 16-QAM)	LTE-FDD	6.66	+9.6%		
10138	CAF	LTE-FDD (SC-FDMA, 100% RB, 15 MHz, QPSK)	LTE-FDD	5.82	+9.6%		
10139	CAF	LTE-FDD (SC-FDMA, 100% RB, 15 MHz, 16-QAM)	LTE-FDD	6.43	+9.6%		
10140	CAF	LTE-FDD (SC-FDMA, 100% RB, 15 MHz, 64-QAM)	LTE-FDD	6.53	+9.6%		
10141	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.4 MHz, QPSK)	LTE-FDD	5.46	+9.6%		
10142	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.4 MHz, 16-QAM)	LTE-FDD	6.19	+9.6%		
10143	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.4 MHz, 64-QAM)	LTE-FDD	6.79	+9.6%		
10144	CAF	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	5.73	+9.6%		
10145	CAF	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, 16-QAM)	LTE-FDD	6.32	+9.6%		
10146	CAF	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, 64-QAM)	LTE-FDD	6.49	+9.6%		
10147	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.8 MHz, QPSK)	LTE-FDD	5.73	+9.6%		
10148	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.8 MHz, 16-QAM)	LTE-FDD	6.49	+9.6%		
10149	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.8 MHz, 64-QAM)	LTE-FDD	6.53	+9.6%		
10150	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.8 MHz, QPSK)	LTE-FDD	5.76	+9.6%		
10151	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.8 MHz, 16-QAM)	LTE-FDD	6.50	+9.6%		
10152	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.8 MHz, 64-QAM)	LTE-FDD	6.53	+9.6%		
10153	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.8 MHz, QPSK)	LTE-FDD	5.73	+9.6%		
10154	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.8 MHz, 16-QAM)	LTE-FDD	6.50	+9.6%		
10155	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.8 MHz, 64-QAM)	LTE-FDD	6.53	+9.6%		
10156	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.8 MHz, QPSK)	LTE-FDD	5.73	+9.6%		
10157	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.8 MHz, 16-QAM)	LTE-FDD	6.50	+9.6%		
10158	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.8 MHz, 64-QAM)	LTE-FDD	6.53	+9.6%		
10159	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.8 MHz, QPSK)	LTE-FDD	5.73	+9.6%		
10160	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.8 MHz, 16-QAM)	LTE-FDD	6.50	+9.6%		
10161	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.8 MHz, 64-QAM)	LTE-FDD	6.53	+9.6%		
10162	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.8 MHz, QPSK)	LTE-FDD	5.73	+9.6%		
10163	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.8 MHz, 16-QAM)	LTE-FDD	6.50	+9.6%		
10164	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.8 MHz, 64-QAM)	LTE-FDD	6.53	+9.6%		
10165	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.8 MHz, QPSK)	LTE-FDD	5.73	+9.6%		
10166	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.8 MHz, 16-QAM)	LTE-FDD	6.50	+9.6%		
10167	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.8 MHz, 64-QAM)	LTE-FDD	6.53	+9.6%		
10168	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.8 MHz, QPSK)	LTE-FDD	5.73	+9.6%		
10169	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.8 MHz, 16-QAM)	LTE-FDD	6.50	+9.6%		
10170	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.8 MHz, 64-QAM)	LTE-FDD	6.53	+9.6%		
10171	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.8 MHz, QPSK)	LTE-FDD	5.73	+9.6%		
10172	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.8 MHz, 16-QAM)	LTE-FDD	6.50	+9.6%		
10173	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.8 MHz, 64-QAM)	LTE-FDD	6.53	+9.6%		
10174	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.8 MHz, QPSK)	LTE-FDD	5.73	+9.6%		
10175	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.8 MHz, 16-QAM)	LTE-FDD	6.50	+9.6%		
10176	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.8 MHz, 64-QAM)	LTE-FDD	6.53	+9.6%		
10177	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.8 MHz, QPSK)	LTE-FDD	5.73	+9.6%		
10178	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.8 MHz, 16-QAM)	LTE-FDD	6.50	+9.6%		
10179	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.8 MHz, 64-QAM)	LTE-FDD	6.53	+9.6%		
10180	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.8 MHz, QPSK)	LTE-FDD	5.73	+9.6%		
10181	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.8 MHz, 16-QAM)	LTE-FDD	6.50	+9.6%		
10182	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.8 MHz, 64-QAM)	LTE-FDD	6.53	+9.6%		
10183	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.8 MHz, QPSK)	LTE-FDD	5.73	+9.6%		
10184	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.8 MHz, 16-QAM)	LTE-FDD	6.50	+9.6%		
10185	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.8 MHz, 64-QAM)	LTE-FDD	6.53	+9.6%		
10186	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.8 MHz, QPSK)	LTE-FDD	5.73	+9.6%		
10187	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.8 MHz, 16-QAM)	LTE-FDD	6.50	+9.6%		
10188	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.8 MHz, 64-QAM)	LTE-FDD	6.53	+9.6%		
10189	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.8 MHz, QPSK)	LTE-FDD	5.73	+9.6%		
10190	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.8 MHz, 16-QAM)	LTE-FDD	6.50	+9.6%		
10191	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.8 MHz, 64-QAM)	LTE-FDD	6.53	+9.6%		
10192	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.8 MHz, QPSK)	LTE-FDD	5.73	+9.6%		
10193	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.8 MHz, 16-QAM)	LTE-FDD	6.50	+9.6%		
10194	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.8 MHz, 64-QAM)	LTE-FDD	6.53	+9.6%		
10195	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.8 MHz, QPSK)	LTE-FDD	5.73	+9.6%		
10196	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.8 MHz, 16-QAM)	LTE-FDD				





EX3044-SN-7346				March 30, 2022			
10414	AAA	WLAN-CDP: 64-QAM, 4096Q	Generic	8.54	± 0.6 %		
10415	AAA	IEEE 802.11b WFI 2.4 GHz (DSSS, 1 Mbps, R90c-0)	WLAN	1.54	± 0.6 %		
10416	AAA	IEEE 802.11g WFI 2.4 GHz (DSSS-OFDM, 5.5 Mbps, R90c-0)	WLAN	8.23	± 0.6 %		
10417	AAC	IEEE 802.11n WFI 5 GHz (OFDM, 4 Mbps, R90c-0)	WLAN	8.23	± 0.6 %		
10418	AAA	IEEE 802.11g WFI 2.4 GHz (DSSS-OFDM, 5.5 Mbps, R90c-Long)	WLAN	8.14	± 0.6 %		
10419	AAA	IEEE 802.11g WFI 2.4 GHz (DSSS-OFDM, 5.5 Mbps, R90c-Short)	WLAN	8.19	± 0.6 %		
10422	AAC	IEEE 802.11n HT Overhead: 7.2 Mbps, R90c-0	WLAN	8.32	± 0.6 %		
10421	AAC	IEEE 802.11n HT Overhead: 6.3 Mbps, 16-QAM, UL Sub-0	WLAN	8.47	± 0.6 %		
10424	AAC	IEEE 802.11n HT Overhead: 7.2 Mbps, 64-QAM, UL Sub-0	WLAN	8.40	± 0.6 %		
10425	AAC	IEEE 802.11n HT Overhead: 14 Mbps, R90c-0	WLAN	8.41	± 0.6 %		
10428	AAC	IEEE 802.11n HT Overhead: 30 Mbps, 16-QAM, UL Sub-0	WLAN	8.45	± 0.6 %		
10427	AAC	IEEE 802.11n HT Overhead: 18 Mbps, 64-QAM, UL Sub-0	WLAN	8.41	± 0.6 %		
10430	AAD	LTE-FDD (FDDMA, 5 MHz, E-UTRA 1.1, Clipping 44%)	LTE-FDD	8.28	± 0.6 %		
10431	AAD	LTE-FDD (FDDMA, 10 MHz, E-UTRA 1.1, Clipping 44%)	LTE-FDD	8.38	± 0.6 %		
10432	AAC	LTE-FDD (FDDMA, 15 MHz, E-UTRA 1.1, Clipping 44%)	LTE-FDD	8.34	± 0.6 %		
10433	AAC	LTE-FDD (FDDMA, 20 MHz, E-UTRA 1.1, Clipping 44%)	LTE-FDD	8.34	± 0.6 %		
10434	AAA	W-CDMA (BS Test Model 1.6a DPM3, Clipping 44%)	WCDMA	8.60	± 0.6 %		
10435	AAF	LTE-TDD (SC-FDMA, 1.8 MHz, 30 MHz, QPSK, UL Sub-0)	LTE-TDD	7.82	± 0.6 %		
10436	AAF	LTE-TDD (SC-FDMA, 1.8 MHz, 1.4 MHz, 16-QAM, UL Sub-0)	LTE-TDD	7.86	± 0.6 %		
10440	AAD	LTE-FDD (FDDMA, 10 MHz, E-UTRA 1.1, Clipping 44%)	LTE-FDD	7.93	± 0.6 %		
10449	AAC	LTE-FDD (FDDMA, 15 MHz, E-UTRA 1.1, Clipping 44%)	LTE-FDD	7.93	± 0.6 %		
10450	AAC	LTE-FDD (FDDMA, 20 MHz, E-UTRA 1.1, Clipping 44%)	LTE-FDD	7.98	± 0.6 %		
10451	AAA	W-CDMA (BS Test Model 1.6a DPM3, Clipping 44%)	WCDMA	7.99	± 0.6 %		
10452	AAD	Validation (Spurs: 10ms, 10ms)	Test	19.90	± 0.6 %		
10459	AAA	IEEE 802.11ac WFI (160MHz, 80 MHz, R90c-0)	WLAN	8.63	± 0.6 %		
10457	AAA	UMTS FDD (GSM-R, Rev. 6, 3.1 MHz)	WCDMA	6.62	± 0.6 %		
10458	AAA	CDMA2000 (1xEV-DO, Rev. 6, 3.1 MHz)	CDMA2000	8.18	± 0.6 %		
10459	AAA	CDMA2000 (1xEV-DO, Rev. 6, 3.1 MHz)	CDMA2000	8.25	± 0.6 %		
10460	AAA	UMTS FDD (GSM-R, Rev. 6, 3.1 MHz)	WCDMA	2.39	± 0.6 %		
10461	AAB	LTE-TDD (SC-FDMA, 1.8 MHz, 1.4 MHz, QPSK, UL Sub-0)	LTE-TDD	7.82	± 0.6 %		
10462	AAB	LTE-TDD (SC-FDMA, 1.8 MHz, 1.4 MHz, 16-QAM, UL Sub-0)	LTE-TDD	8.30	± 0.6 %		
10463	AAB	LTE-TDD (SC-FDMA, 1.8 MHz, 1.4 MHz, 64-QAM, UL Sub-0)	LTE-TDD	8.30	± 0.6 %		
10464	AAC	LTE-TDD (SC-FDMA, 1.8 MHz, 30 MHz, QPSK, UL Sub-0)	LTE-TDD	7.82	± 0.6 %		
10465	AAC	LTE-TDD (SC-FDMA, 1.8 MHz, 30 MHz, 16-QAM, UL Sub-0)	LTE-TDD	8.32	± 0.6 %		
10466	AAC	LTE-TDD (SC-FDMA, 1.8 MHz, 30 MHz, 64-QAM, UL Sub-0)	LTE-TDD	8.37	± 0.6 %		
10467	AAF	LTE-TDD (SC-FDMA, 1.8 MHz, 30 MHz, QPSK, UL Sub-0)	LTE-TDD	7.82	± 0.6 %		
10468	AAF	LTE-TDD (SC-FDMA, 1.8 MHz, 30 MHz, 16-QAM, UL Sub-0)	LTE-TDD	8.32	± 0.6 %		
10469	AAF	LTE-TDD (SC-FDMA, 1.8 MHz, 30 MHz, 64-QAM, UL Sub-0)	LTE-TDD	8.36	± 0.6 %		
10470	AAF	LTE-TDD (SC-FDMA, 1.8 MHz, 30 MHz, QPSK, UL Sub-0)	LTE-TDD	7.82	± 0.6 %		
10471	AAF	LTE-TDD (SC-FDMA, 1.8 MHz, 30 MHz, 16-QAM, UL Sub-0)	LTE-TDD	8.32	± 0.6 %		
10472	AAF	LTE-TDD (SC-FDMA, 1.8 MHz, 30 MHz, 64-QAM, UL Sub-0)	LTE-TDD	8.37	± 0.6 %		
10473	AAB	LTE-TDD (SC-FDMA, 1.8 MHz, 30 MHz, QPSK, UL Sub-0)	LTE-TDD	7.82	± 0.6 %		
10474	AAB	LTE-TDD (SC-FDMA, 1.8 MHz, 30 MHz, 16-QAM, UL Sub-0)	LTE-TDD	8.32	± 0.6 %		
10475	AAB	LTE-TDD (SC-FDMA, 1.8 MHz, 30 MHz, 64-QAM, UL Sub-0)	LTE-TDD	8.32	± 0.6 %		
10476	AAF	LTE-TDD (SC-FDMA, 1.8 MHz, 30 MHz, QPSK, UL Sub-0)	LTE-TDD	7.82	± 0.6 %		
10477	AAF	LTE-TDD (SC-FDMA, 1.8 MHz, 30 MHz, 16-QAM, UL Sub-0)	LTE-TDD	8.32	± 0.6 %		
10478	AAF	LTE-TDD (SC-FDMA, 1.8 MHz, 30 MHz, 64-QAM, UL Sub-0)	LTE-TDD	8.37	± 0.6 %		
10479	AAB	LTE-TDD (SC-FDMA, 1.8 MHz, 30 MHz, QPSK, UL Sub-0)	LTE-TDD	7.74	± 0.6 %		
10480	AAB	LTE-TDD (SC-FDMA, 1.8 MHz, 30 MHz, 16-QAM, UL Sub-0)	LTE-TDD	8.18	± 0.6 %		
10481	AAB	LTE-TDD (SC-FDMA, 1.8 MHz, 30 MHz, 64-QAM, UL Sub-0)	LTE-TDD	8.45	± 0.6 %		
10482	AAC	LTE-TDD (SC-FDMA, 1.8 MHz, 30 MHz, QPSK, UL Sub-0)	LTE-TDD	7.71	± 0.6 %		
10483	AAC	LTE-TDD (SC-FDMA, 1.8 MHz, 30 MHz, 16-QAM, UL Sub-0)	LTE-TDD	8.30	± 0.6 %		
10484	AAC	LTE-TDD (SC-FDMA, 1.8 MHz, 30 MHz, 64-QAM, UL Sub-0)	LTE-TDD	8.47	± 0.6 %		
10485	AAF	LTE-TDD (SC-FDMA, 1.8 MHz, 30 MHz, QPSK, UL Sub-0)	LTE-TDD	7.79	± 0.6 %		
10486	AAF	LTE-TDD (SC-FDMA, 1.8 MHz, 30 MHz, 16-QAM, UL Sub-0)	LTE-TDD	8.38	± 0.6 %		
10487	AAF	LTE-TDD (SC-FDMA, 1.8 MHz, 30 MHz, 64-QAM, UL Sub-0)	LTE-TDD	8.60	± 0.6 %		
10488	AAF	LTE-TDD (SC-FDMA, 1.8 MHz, 30 MHz, QPSK, UL Sub-0)	LTE-TDD	7.70	± 0.6 %		

Certificate No. EX3-7346\_Mar22 Page 16 of 24

EX3044-SN-7346				March 30, 2022			
10489	AAF	LTE-TDD (SC-FDMA, 1.8 MHz, 30 MHz, QPSK, UL Sub-0)	LTE-TDD	8.31	± 0.6 %		
10490	AAF	LTE-TDD (SC-FDMA, 1.8 MHz, 30 MHz, 16-QAM, UL Sub-0)	LTE-TDD	8.34	± 0.6 %		
10491	AAF	LTE-TDD (SC-FDMA, 1.8 MHz, 30 MHz, 64-QAM, UL Sub-0)	LTE-TDD	8.34	± 0.6 %		
10492	AAC	LTE-TDD (SC-FDMA, 1.8 MHz, 30 MHz, QPSK, UL Sub-0)	LTE-TDD	8.31	± 0.6 %		
10493	AAC	LTE-TDD (SC-FDMA, 1.8 MHz, 30 MHz, 16-QAM, UL Sub-0)	LTE-TDD	8.34	± 0.6 %		
10494	AAF	LTE-TDD (SC-FDMA, 1.8 MHz, 30 MHz, QPSK, UL Sub-0)	LTE-TDD	7.74	± 0.6 %		
10495	AAF	LTE-TDD (SC-FDMA, 1.8 MHz, 30 MHz, 16-QAM, UL Sub-0)	LTE-TDD	8.37	± 0.6 %		
10496	AAF	LTE-TDD (SC-FDMA, 1.8 MHz, 30 MHz, 64-QAM, UL Sub-0)	LTE-TDD	8.37	± 0.6 %		
10497	AAB	LTE-TDD (SC-FDMA, 1.8 MHz, 30 MHz, QPSK, UL Sub-0)	LTE-TDD	7.87	± 0.6 %		
10498	AAB	LTE-TDD (SC-FDMA, 1.8 MHz, 30 MHz, 16-QAM, UL Sub-0)	LTE-TDD	8.40	± 0.6 %		
10499	AAB	LTE-TDD (SC-FDMA, 1.8 MHz, 30 MHz, 64-QAM, UL Sub-0)	LTE-TDD	8.40	± 0.6 %		
10500	AAC	LTE-TDD (SC-FDMA, 1.8 MHz, 30 MHz, QPSK, UL Sub-0)	LTE-TDD	7.87	± 0.6 %		
10501	AAC	LTE-TDD (SC-FDMA, 1.8 MHz, 30 MHz, 16-QAM, UL Sub-0)	LTE-TDD	8.44	± 0.6 %		
10502	AAC	LTE-TDD (SC-FDMA, 1.8 MHz, 30 MHz, 64-QAM, UL Sub-0)	LTE-TDD	8.44	± 0.6 %		
10503	AAF	LTE-TDD (SC-FDMA, 1.8 MHz, 30 MHz, QPSK, UL Sub-0)	LTE-TDD	7.72	± 0.6 %		
10504	AAF	LTE-TDD (SC-FDMA, 1.8 MHz, 30 MHz, 16-QAM, UL Sub-0)	LTE-TDD	8.31	± 0.6 %		
10505	AAF	LTE-TDD (SC-FDMA, 1.8 MHz, 30 MHz, 64-QAM, UL Sub-0)	LTE-TDD	8.34	± 0.6 %		
10506	AAF	LTE-TDD (SC-FDMA, 1.8 MHz, 30 MHz, QPSK, UL Sub-0)	LTE-TDD	7.86	± 0.6 %		
10507	AAF	LTE-TDD (SC-FDMA, 1.8 MHz, 30 MHz, 16-QAM, UL Sub-0)	LTE-TDD	8.36	± 0.6 %		
10508	AAF	LTE-TDD (SC-FDMA, 1.8 MHz, 30 MHz, 64-QAM, UL Sub-0)	LTE-TDD	8.35	± 0.6 %		
10509	AAC	LTE-TDD (SC-FDMA, 1.8 MHz, 30 MHz, QPSK, UL Sub-0)	LTE-TDD	7.99	± 0.6 %		
10510	AAC	LTE-TDD (SC-FDMA, 1.8 MHz, 30 MHz, 16-QAM, UL Sub-0)	LTE-TDD	8.49	± 0.6 %		
10511	AAC	LTE-TDD (SC-FDMA, 1.8 MHz, 30 MHz, 64-QAM, UL Sub-0)	LTE-TDD	8.51	± 0.6 %		
10512	AAF	LTE-TDD (SC-FDMA, 1.8 MHz, 30 MHz, QPSK, UL Sub-0)	LTE-TDD	7.74	± 0.6 %		
10513	AAF	LTE-TDD (SC-FDMA, 1.8 MHz, 30 MHz, 16-QAM, UL Sub-0)	LTE-TDD	8.42	± 0.6 %		
10514	AAF	LTE-TDD (SC-FDMA, 1.8 MHz, 30 MHz, 64-QAM, UL Sub-0)	LTE-TDD	8.45	± 0.6 %		
10515	AAA	IEEE 802.11ac WFI 2.4 GHz (DSSS, 2 Mbps, R90c-0)	WLAN	1.58	± 0.6 %		
10516	AAA	IEEE 802.11ac WFI 2.4 GHz (DSSS, 5.5 Mbps, R90c-0)	WLAN	1.57	± 0.6 %		
10517	AAA	IEEE 802.11ac WFI 2.4 GHz (DSSS, 11 Mbps, R90c-0)	WLAN	1.58	± 0.6 %		
10518	AAC	IEEE 802.11ac WFI 5 GHz (OFDM, 5 Mbps, R90c-0)	WLAN	8.74	± 0.6 %		
10519	AAC	IEEE 802.11ac WFI 5 GHz (OFDM, 12 Mbps, R90c-0)	WLAN	8.39	± 0.6 %		
10520	AAC	IEEE 802.11ac WFI 5 GHz (OFDM, 18 Mbps, R90c-0)	WLAN	8.12	± 0.6 %		
10521	AAC	IEEE 802.11ac WFI 5 GHz (OFDM, 24 Mbps, R90c-0)	WLAN	7.97	± 0.6 %		
10522	AAC	IEEE 802.11ac WFI 5 GHz (OFDM, 36 Mbps, R90c-0)	WLAN	8.45	± 0.6 %		
10523	AAC	IEEE 802.11ac WFI 5 GHz (OFDM, 48 Mbps, R90c-0)	WLAN	8.48	± 0.6 %		
10524	AAC	IEEE 802.11ac WFI 5 GHz (OFDM, 54 Mbps, R90c-0)	WLAN	8.27	± 0.6 %		
10525	AAC	IEEE 802.11ac WFI (20MHz, MCS1, R90c-0)	WLAN	2.36	± 0.6 %		
10526	AAC	IEEE 802.11ac WFI (20MHz, MCS1, R90c-0)	WLAN	8.42	± 0.6 %		
10527	AAC	IEEE 802.11ac WFI (20MHz, MCS1, R90c-0)	WLAN	8.21	± 0.6 %		
10528	AAC	IEEE 802.11ac WFI (20MHz, MCS1, R90c-0)	WLAN	8.32	± 0.6 %		
10529	AAC	IEEE 802.11ac WFI (20MHz, MCS1, R90c-0)	WLAN	8.36	± 0.6 %		
10530	AAC	IEEE 802.11ac WFI (20MHz, MCS1, R90c-0)	WLAN	8.36	± 0.6 %		
10531	AAC	IEEE 802.11ac WFI (20MHz, MCS1, R90c-0)	WLAN	8.29	± 0.6 %		
10532	AAC	IEEE 802.11ac WFI (20MHz, MCS1, R90c-0)	WLAN	8.29	± 0.6 %		
10533	AAC	IEEE 802.11ac WFI (20MHz, MCS1, R90c-0)	WLAN	8.30	± 0.6 %		
10534	AAC	IEEE 802.11ac WFI (20MHz, MCS1, R90c-0)	WLAN	8.45	± 0.6 %		
10535	AAC	IEEE 802.11ac WFI (20MHz, MCS1, R90c-0)	WLAN	8.45	± 0.6 %		
10536	AAC	IEEE 802.11ac WFI (20MHz, MCS1, R90c-0)	WLAN	8.32	± 0.6 %		
10537	AAC	IEEE 802.11ac WFI (20MHz, MCS1, R90c-0)	WLAN	8.44	± 0.6 %		
10538	AAC	IEEE 802.11ac WFI (20MHz, MCS1, R90c-0)	WLAN	8.34	± 0.6 %		
10539	AAC	IEEE 802.11ac WFI (20MHz, MCS1, R90c-0)	WLAN	8.39	± 0.6 %		
10540	AAC	IEEE 802.11ac WFI (20MHz, MCS1, R90c-0)	WLAN	8.46	± 0.6 %		
10541	AAC	IEEE 802.11ac WFI (20MHz, MCS1, R90c-0)	WLAN	8.45	± 0.6 %		
10542	AAC	IEEE 802.11ac WFI (20MHz, MCS1, R90c-0)	WLAN	8.45	± 0.6 %		
10543	AAC	IEEE 802.11ac WFI (20MHz, MCS1, R90c-0)	WLAN	8.47	± 0.6 %		
10544	AAC	IEEE 802.11ac WFI (20MHz, MCS1, R90c-0)	WLAN	8.55	± 0.6 %		
10545	AAC	IEEE 802.11ac WFI (20MHz, MCS1, R90c-0)	WLAN	8.55	± 0.6 %		
10546	AAC	IEEE 802.11ac WFI (20MHz, MCS1, R90c-0)	WLAN	8.35	± 0.6 %		

Certificate No. EX3-7346\_Mar22 Page 16 of 24

EX3044-SN-7346				March 30, 2022			
10547	AAC	IEEE 802.11ac WFI (20MHz, MCS1, R90c-0)	WLAN	8.49	± 0.6 %		
10548	AAC	IEEE 802.11ac WFI (20MHz, MCS1, R90c-0)	WLAN	8.37	± 0.6 %		
10549	AAC	IEEE 802.11ac WFI (20MHz, MCS1, R90c-0)	WLAN	8.39	± 0.6 %		
10550	AAC	IEEE 802.11ac WFI (20MHz, MCS1, R90c-0)	WLAN	8.30	± 0.6 %		
10551	AAC	IEEE 802.11ac WFI (20MHz, MCS1, R90c-0)	WLAN	8.42	± 0.6 %		
10552	AAC	IEEE 802.11ac WFI (20MHz, MCS1,					



EX3D4- 9N7346

March 30, 2022

10673	AAC	IEEE 802.11ax (20MHz, MCS2, 90pdc)	WLAN	8.78	9.8%
10674	AAC	IEEE 802.11ax (20MHz, MCS3, 90pdc)	WLAN	8.74	9.6%
10675	AAC	IEEE 802.11ax (20MHz, MCS4, 90pdc)	WLAN	8.80	9.6%
10676	AAC	IEEE 802.11ax (20MHz, MCS5, 90pdc)	WLAN	8.77	9.6%
10677	AAC	IEEE 802.11ax (20MHz, MCS6, 90pdc)	WLAN	8.73	9.6%
10678	AAC	IEEE 802.11ax (20MHz, MCS7, 90pdc)	WLAN	8.78	9.6%
10679	AAC	IEEE 802.11ax (20MHz, MCS8, 90pdc)	WLAN	8.89	9.6%
10680	AAC	IEEE 802.11ax (20MHz, MCS9, 90pdc)	WLAN	8.80	9.6%
10681	AAC	IEEE 802.11ax (20MHz, MCS10, 90pdc)	WLAN	8.62	9.6%
10682	AAC	IEEE 802.11ax (20MHz, MCS11, 90pdc)	WLAN	8.83	9.6%
10683	AAC	IEEE 802.11ax (20MHz, MCS9, 90pdc)	WLAN	8.43	9.6%
10684	AAC	IEEE 802.11ax (20MHz, MCS11, 90pdc)	WLAN	8.26	9.6%
10685	AAC	IEEE 802.11ax (20MHz, MCS2, 90pdc)	WLAN	8.33	9.6%
10686	AAC	IEEE 802.11ax (20MHz, MCS3, 90pdc)	WLAN	8.28	9.6%
10687	AAC	IEEE 802.11ax (20MHz, MCS4, 90pdc)	WLAN	8.45	9.6%
10688	AAC	IEEE 802.11ax (20MHz, MCS5, 90pdc)	WLAN	8.29	9.6%
10689	AAC	IEEE 802.11ax (20MHz, MCS6, 90pdc)	WLAN	8.55	9.6%
10690	AAC	IEEE 802.11ax (20MHz, MCS7, 90pdc)	WLAN	8.29	9.6%
10691	AAC	IEEE 802.11ax (20MHz, MCS8, 90pdc)	WLAN	8.25	9.6%
10692	AAC	IEEE 802.11ax (20MHz, MCS9, 90pdc)	WLAN	8.29	9.6%
10693	AAC	IEEE 802.11ax (20MHz, MCS10, 90pdc)	WLAN	8.25	9.6%
10694	AAC	IEEE 802.11ax (20MHz, MCS11, 90pdc)	WLAN	8.57	9.6%
10695	AAC	IEEE 802.11ax (20MHz, MCS9, 90pdc)	WLAN	8.78	9.6%
10696	AAC	IEEE 802.11ax (20MHz, MCS11, 90pdc)	WLAN	8.91	9.6%
10697	AAC	IEEE 802.11ax (20MHz, MCS2, 90pdc)	WLAN	8.61	9.6%
10698	AAC	IEEE 802.11ax (20MHz, MCS3, 90pdc)	WLAN	8.89	9.6%
10699	AAC	IEEE 802.11ax (20MHz, MCS4, 90pdc)	WLAN	8.82	9.6%
10700	AAC	IEEE 802.11ax (20MHz, MCS5, 90pdc)	WLAN	8.73	9.6%
10701	AAC	IEEE 802.11ax (20MHz, MCS6, 90pdc)	WLAN	8.86	9.6%
10702	AAC	IEEE 802.11ax (20MHz, MCS7, 90pdc)	WLAN	8.70	9.6%
10703	AAC	IEEE 802.11ax (20MHz, MCS8, 90pdc)	WLAN	8.82	9.6%
10704	AAC	IEEE 802.11ax (20MHz, MCS9, 90pdc)	WLAN	8.56	9.6%
10705	AAC	IEEE 802.11ax (20MHz, MCS10, 90pdc)	WLAN	8.69	9.6%
10706	AAC	IEEE 802.11ax (20MHz, MCS11, 90pdc)	WLAN	8.68	9.6%
10707	AAC	IEEE 802.11ax (20MHz, MCS9, 90pdc)	WLAN	8.32	9.6%
10708	AAC	IEEE 802.11ax (20MHz, MCS11, 90pdc)	WLAN	8.56	9.6%
10709	AAC	IEEE 802.11ax (20MHz, MCS2, 90pdc)	WLAN	8.33	9.6%
10710	AAC	IEEE 802.11ax (20MHz, MCS3, 90pdc)	WLAN	8.29	9.6%
10711	AAC	IEEE 802.11ax (20MHz, MCS4, 90pdc)	WLAN	8.39	9.6%
10712	AAC	IEEE 802.11ax (20MHz, MCS5, 90pdc)	WLAN	8.67	9.6%
10713	AAC	IEEE 802.11ax (20MHz, MCS6, 90pdc)	WLAN	8.33	9.6%
10714	AAC	IEEE 802.11ax (20MHz, MCS7, 90pdc)	WLAN	8.26	9.6%
10715	AAC	IEEE 802.11ax (20MHz, MCS8, 90pdc)	WLAN	8.43	9.6%
10716	AAC	IEEE 802.11ax (20MHz, MCS9, 90pdc)	WLAN	8.30	9.6%
10717	AAC	IEEE 802.11ax (20MHz, MCS10, 90pdc)	WLAN	8.48	9.6%
10718	AAC	IEEE 802.11ax (20MHz, MCS11, 90pdc)	WLAN	8.24	9.6%
10719	AAC	IEEE 802.11ax (80MHz, MCS0, 90pdc)	WLAN	8.81	9.6%
10720	AAC	IEEE 802.11ax (80MHz, MCS1, 90pdc)	WLAN	8.87	9.6%
10721	AAC	IEEE 802.11ax (80MHz, MCS2, 90pdc)	WLAN	8.78	9.6%
10722	AAC	IEEE 802.11ax (80MHz, MCS3, 90pdc)	WLAN	8.55	9.6%
10723	AAC	IEEE 802.11ax (80MHz, MCS4, 90pdc)	WLAN	8.79	9.6%
10724	AAC	IEEE 802.11ax (80MHz, MCS5, 90pdc)	WLAN	8.90	9.6%
10725	AAC	IEEE 802.11ax (80MHz, MCS6, 90pdc)	WLAN	8.74	9.6%
10726	AAC	IEEE 802.11ax (80MHz, MCS7, 90pdc)	WLAN	8.72	9.6%
10727	AAC	IEEE 802.11ax (80MHz, MCS8, 90pdc)	WLAN	8.86	9.6%
10728	AAC	IEEE 802.11ax (80MHz, MCS9, 90pdc)	WLAN	8.65	9.6%

Certificate No: EX3-7346\_Mar22

Page 19 of 24

EX3D4- 9N7346

March 30, 2022

10729	AAC	IEEE 802.11ax (80MHz, MCS10, 90pdc)	WLAN	8.64	9.6%
10730	AAC	IEEE 802.11ax (80MHz, MCS11, 90pdc)	WLAN	8.67	9.6%
10731	AAC	IEEE 802.11ax (80MHz, MCS0, 90pdc)	WLAN	8.42	9.6%
10732	AAC	IEEE 802.11ax (80MHz, MCS1, 90pdc)	WLAN	8.48	9.6%
10733	AAC	IEEE 802.11ax (80MHz, MCS2, 90pdc)	WLAN	8.40	9.6%
10734	AAC	IEEE 802.11ax (80MHz, MCS3, 90pdc)	WLAN	8.29	9.6%
10735	AAC	IEEE 802.11ax (80MHz, MCS4, 90pdc)	WLAN	8.33	9.6%
10736	AAC	IEEE 802.11ax (80MHz, MCS5, 90pdc)	WLAN	8.27	9.6%
10737	AAC	IEEE 802.11ax (80MHz, MCS6, 90pdc)	WLAN	8.38	9.6%
10738	AAC	IEEE 802.11ax (80MHz, MCS7, 90pdc)	WLAN	8.42	9.6%
10739	AAC	IEEE 802.11ax (80MHz, MCS8, 90pdc)	WLAN	8.29	9.6%
10740	AAC	IEEE 802.11ax (80MHz, MCS9, 90pdc)	WLAN	8.48	9.6%
10741	AAC	IEEE 802.11ax (80MHz, MCS10, 90pdc)	WLAN	8.40	9.6%
10742	AAC	IEEE 802.11ax (80MHz, MCS11, 90pdc)	WLAN	8.43	9.6%
10743	AAC	IEEE 802.11ax (160MHz, MCS0, 90pdc)	WLAN	8.84	9.6%
10744	AAC	IEEE 802.11ax (160MHz, MCS1, 90pdc)	WLAN	9.16	9.6%
10745	AAC	IEEE 802.11ax (160MHz, MCS2, 90pdc)	WLAN	8.93	9.6%
10746	AAC	IEEE 802.11ax (160MHz, MCS3, 90pdc)	WLAN	9.11	9.6%
10747	AAC	IEEE 802.11ax (160MHz, MCS4, 90pdc)	WLAN	9.04	9.6%
10748	AAC	IEEE 802.11ax (160MHz, MCS5, 90pdc)	WLAN	8.93	9.6%
10749	AAC	IEEE 802.11ax (160MHz, MCS6, 90pdc)	WLAN	8.96	9.6%
10750	AAC	IEEE 802.11ax (160MHz, MCS7, 90pdc)	WLAN	8.79	9.6%
10751	AAC	IEEE 802.11ax (160MHz, MCS8, 90pdc)	WLAN	8.82	9.6%
10752	AAC	IEEE 802.11ax (160MHz, MCS9, 90pdc)	WLAN	8.81	9.6%
10753	AAC	IEEE 802.11ax (160MHz, MCS10, 90pdc)	WLAN	9.00	9.6%
10754	AAC	IEEE 802.11ax (160MHz, MCS11, 90pdc)	WLAN	8.96	9.6%
10755	AAC	IEEE 802.11ax (160MHz, MCS0, 90pdc)	WLAN	8.84	9.6%
10756	AAC	IEEE 802.11ax (160MHz, MCS1, 90pdc)	WLAN	8.77	9.6%
10757	AAC	IEEE 802.11ax (160MHz, MCS2, 90pdc)	WLAN	8.77	9.6%
10758	AAC	IEEE 802.11ax (160MHz, MCS3, 90pdc)	WLAN	8.69	9.6%
10759	AAC	IEEE 802.11ax (160MHz, MCS4, 90pdc)	WLAN	8.58	9.6%
10760	AAC	IEEE 802.11ax (160MHz, MCS5, 90pdc)	WLAN	8.49	9.6%
10761	AAC	IEEE 802.11ax (160MHz, MCS6, 90pdc)	WLAN	8.58	9.6%
10762	AAC	IEEE 802.11ax (160MHz, MCS7, 90pdc)	WLAN	8.48	9.6%
10763	AAC	IEEE 802.11ax (160MHz, MCS8, 90pdc)	WLAN	8.53	9.6%
10764	AAC	IEEE 802.11ax (160MHz, MCS9, 90pdc)	WLAN	8.56	9.6%
10765	AAC	IEEE 802.11ax (160MHz, MCS10, 90pdc)	WLAN	8.84	9.6%
10766	AAC	IEEE 802.11ax (160MHz, MCS11, 90pdc)	WLAN	8.81	9.6%
10767	AAC	SO NR CP-OFDM 1, 1RB, 5 MHz, QPSK, 15 MHz	SO NR FRI TTD	7.99	8.6%
10768	AAC	SO NR CP-OFDM 1, 1RB, 10 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.01	8.6%
10769	AAC	SO NR CP-OFDM 1, 1RB, 15 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.01	8.6%
10770	AAC	SO NR CP-OFDM 1, 1RB, 20 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.02	8.6%
10771	AAC	SO NR CP-OFDM 1, 1RB, 25 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.02	8.6%
10772	AAC	SO NR CP-OFDM 1, 1RB, 30 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.23	8.6%
10773	AAC	SO NR CP-OFDM 1, 1RB, 40 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.03	8.6%
10774	AAC	SO NR CP-OFDM 1, 1RB, 50 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.02	8.6%
10775	AAC	SO NR CP-OFDM 50% RB, 5 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.31	8.6%
10776	AAC	SO NR CP-OFDM 50% RB, 10 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.30	8.6%
10777	AAC	SO NR CP-OFDM 50% RB, 15 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.30	8.6%
10778	AAC	SO NR CP-OFDM 50% RB, 20 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.34	8.6%
10779	AAC	SO NR CP-OFDM 50% RB, 25 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.42	8.6%
10780	AAC	SO NR CP-OFDM 50% RB, 30 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.38	8.6%
10781	AAC	SO NR CP-OFDM 50% RB, 40 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.38	8.6%
10782	AAC	SO NR CP-OFDM 50% RB, 50 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.43	8.6%
10783	AAC	SO NR CP-OFDM 100% RB, 5 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.31	8.6%
10784	AAC	SO NR CP-OFDM 100% RB, 10 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.29	8.6%

Certificate No: EX3-7346\_Mar22

Page 20 of 24

EX3D4- 9N7346

Mar 30, 2022

10785	AAC	SO NR CP-OFDM 100% RB, 15 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.40	8.6%
10786	AAC	SO NR CP-OFDM 100% RB, 20 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.41	8.6%
10787	AAC	SO NR CP-OFDM 100% RB, 25 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.41	8.6%
10788	AAC	SO NR CP-OFDM 100% RB, 30 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.41	8.6%
10789	AAC	SO NR CP-OFDM 100% RB, 40 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.41	8.6%
10790	AAC	SO NR CP-OFDM 100% RB, 50 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.41	8.6%
10791	AAC	SO NR CP-OFDM 100% RB, 5 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.41	8.6%
10792	AAC	SO NR CP-OFDM 100% RB, 10 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.41	8.6%
10793	AAC	SO NR CP-OFDM 100% RB, 15 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.41	8.6%
10794	AAC	SO NR CP-OFDM 100% RB, 20 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.41	8.6%
10795	AAC	SO NR CP-OFDM 100% RB, 25 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.41	8.6%
10796	AAC	SO NR CP-OFDM 100% RB, 30 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.41	8.6%
10797	AAC	SO NR CP-OFDM 100% RB, 40 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.41	8.6%
10798	AAC	SO NR CP-OFDM 100% RB, 50 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.41	8.6%
10799	AAC	SO NR CP-OFDM 100% RB, 5 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.41	8.6%
10800	AAC	SO NR CP-OFDM 100% RB, 10 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.41	8.6%
10801	AAC	SO NR CP-OFDM 100% RB, 15 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.41	8.6%
10802	AAC	SO NR CP-OFDM 100% RB, 20 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.41	8.6%
10803	AAC	SO NR CP-OFDM 100% RB, 25 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.41	8.6%
10804	AAC	SO NR CP-OFDM 100% RB, 30 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.41	8.6%
10805	AAC	SO NR CP-OFDM 100% RB, 40 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.41	8.6%
10806	AAC	SO NR CP-OFDM 100% RB, 50 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.41	8.6%
10807	AAC	SO NR CP-OFDM 100% RB, 5 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.41	8.6%
10808	AAC	SO NR CP-OFDM 100% RB, 10 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.41	8.6%
10809	AAC	SO NR CP-OFDM 100% RB, 15 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.41	8.6%
10810	AAC	SO NR CP-OFDM 100% RB, 20 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.41	8.6%
10811	AAC	SO NR CP-OFDM 100% RB, 25 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.41	8.6%
10812	AAC	SO NR CP-OFDM 100% RB, 30 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.41	8.6%
10813	AAC	SO NR CP-OFDM 100% RB, 40 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.41	8.6%
10814	AAC	SO NR CP-OFDM 100% RB, 50 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.41	8.6%
10815	AAC	SO NR CP-OFDM 100% RB, 5 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.41	8.6%
10816	AAC	SO NR CP-OFDM 100% RB, 10 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.41	8.6%
10817	AAC	SO NR CP-OFDM 100% RB, 15 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.41	8.6%
10818	AAC	SO NR CP-OFDM 100% RB, 20 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.41	8.6%
10819	AAC	SO NR CP-OFDM 100% RB, 25 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.41	8.6%
10820	AAC	SO NR CP-OFDM 100% RB, 30 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.41	8.6%
10821	AAC	SO NR CP-OFDM 100% RB, 40 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.41	8.6%
10822	AAC	SO NR CP-OFDM 100% RB, 50 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.41	8.6%
10823	AAC	SO NR CP-OFDM 100% RB, 5 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.41	8.6%
10824	AAC	SO NR CP-OFDM 100% RB, 10 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.41	8.6%
10825	AAC	SO NR CP-OFDM 100% RB, 15 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.41	8.6%
10826	AAC	SO NR CP-OFDM 100% RB, 20 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.41	8.6%
10827	AAC	SO NR CP-OFDM 100% RB, 25 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.41	8.6%
10828	AAC	SO NR CP-OFDM 100% RB, 30 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.41	8.6%
10829	AAC	SO NR CP-OFDM 100% RB, 40 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.41	8.6%
10830	AAC	SO NR CP-OFDM 100% RB, 50 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.41	8.6%
10831	AAC	SO NR CP-OFDM 100% RB, 5 MHz, QPSK, 15 MHz	SO NR FRI TTD	8.41	8.6%
10832	AAC	SO NR CP-OFDM 100% RB, 10 MHz, QPSK, 15 MHz	SO NR FRI TTD</		



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx>, and for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is deemed to have read, understood and accepted the Terms and Conditions of the document. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or expiration date of this document will render it null and void. The original document and/or any other document otherwise stated in the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing /Inspection report or email: [CN.Doccheck@sgs.com](mailto:CN.Doccheck@sgs.com)  
No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300  
中国·江苏·昆山市留学院创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgsgroup.com.cn  
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

<p>EXCIV4-SN 7346</p> <p>March 30, 2022</p> <p>15985 AAA 50 NR DL (CP-CFOM, TM 3.1, 40 MHz, 64-QAM, 30 kHz) SG NR PR1 TDD 9.54 ± 9.8 %</p> <p>15986 AAA 50 NR DL (CP-CFOM, TM 3.1, 50 MHz, 64-QAM, 30 kHz) SG NR PR1 TDD 9.50 ± 9.8 %</p> <p>15987 AAA 50 NR DL (CP-CFOM, TM 3.1, 60 MHz, 64-QAM, 30 kHz) SG NR PR1 TDD 9.50 ± 9.8 %</p> <p>15988 AAA 50 NR DL (CP-CFOM, TM 3.1, 70 MHz, 64-QAM, 30 kHz) SG NR PR1 TDD 9.38 ± 9.8 %</p> <p>15989 AAA 50 NR DL (CP-CFOM, TM 3.1, 80 MHz, 64-QAM, 30 kHz) SG NR PR1 TDD 9.33 ± 9.8 %</p> <p>15990 AAA 50 NR DL (CP-CFOM, TM 3.1, 90 MHz, 64-QAM, 30 kHz) SG NR PR1 TDD 9.52 ± 9.8 %</p> <p><sup>1</sup> Uncertainty is determined using the max. deviation from linear response applying rectangular distribution and is expressed for the square of the field value.</p> <p>Certificate No. EXC-7346_Mar22</p> <p>Page 24 of 24</p>	<p>EXCIV4-SN 7346</p> <p>March 30, 2022</p> <p>15985 AAA 50 NR DL (CP-CFOM, TM 3.1, 40 MHz, 64-QAM, 30 kHz) SG NR PR1 TDD 9.54 ± 9.8 %</p> <p>15986 AAA 50 NR DL (CP-CFOM, TM 3.1, 50 MHz, 64-QAM, 30 kHz) SG NR PR1 TDD 9.50 ± 9.8 %</p> <p>15987 AAA 50 NR DL (CP-CFOM, TM 3.1, 60 MHz, 64-QAM, 30 kHz) SG NR PR1 TDD 9.50 ± 9.8 %</p> <p>15988 AAA 50 NR DL (CP-CFOM, TM 3.1, 70 MHz, 64-QAM, 30 kHz) SG NR PR1 TDD 9.38 ± 9.8 %</p> <p>15989 AAA 50 NR DL (CP-CFOM, TM 3.1, 80 MHz, 64-QAM, 30 kHz) SG NR PR1 TDD 9.33 ± 9.8 %</p> <p>15990 AAA 50 NR DL (CP-CFOM, TM 3.1, 90 MHz, 64-QAM, 30 kHz) SG NR PR1 TDD 9.52 ± 9.8 %</p> <p><sup>1</sup> Uncertainty is determined using the max. deviation from linear response applying rectangular distribution and is expressed for the square of the field value.</p> <p>Certificate No. EXC-7346_Mar22</p> <p>Page 24 of 24</p>
--	--

#### 4 Impedance and return loss

Dipole CLA150 SN 4025				
Head Liquid				
Date of Measurement	Return Loss(dB)	Δ %	Impedance (Ω)	ΔΩ
2021/4/26	-31.4	/	47.8	/
Dipole D450V3 SN 1103				
Head Liquid				
Date of Measurement	Return Loss(dB)	Δ %	Impedance (Ω)	ΔΩ
2021/4/21	-23	/	57.1	/



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: [CN.Doccheck@sgs.com](mailto:CN.Doccheck@sgs.com)

No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300  
中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 [www.sgsgroup.com.cn](http://www.sgsgroup.com.cn)  
t(86-512)57355888 f(86-512)57370818 [sgs.china@sgs.com](mailto:sgs.china@sgs.com)