



REPORT No.: SZ24100313S03

## Annex E DASY Calibration Certificate



Shenzhen Morlab Communications Technology Co., Ltd.  
FL1-1-3, Building A, FeiYang Science Park, No.8 LongChang Road,  
Block67, BaoAn District, ShenZhen , GuangDong Province, P. R. China

Tel: 86-755-36698555

[Http://www.morlab.cn](http://www.morlab.cn)

Fax: 86-755-36698525

E-mail: [service@morlab.cn](mailto:service@morlab.cn)



Accredited by the Swiss Accreditation Service (SAS)  
**The Swiss Accreditation Service is one of the signatories to the EA  
Multilateral Agreement for the recognition of calibration certificates**

Accreditation No.: **SCS 0108**

Client

**Morlab**  
Shenzhen City

Certificate No.

**ER-2434\_Oct24**

## CALIBRATION CERTIFICATE

Object

**ER3DV6 - SN:2434**

Calibration procedure(s)

**QA CAL-02.v9, QA CAL-25.v8**  
Calibration procedure for E-field probes optimized for close near field  
evaluations in air

Calibration date

**October 24, 2024**

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI).  
The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature  $(22 \pm 3)^\circ\text{C}$  and humidity  $< 70\%$ .

Calibration Equipment used (M&TE critical for calibration)

| Primary Standards          | ID               | Cal Date (Certificate No.)      | Scheduled Calibration |
|----------------------------|------------------|---------------------------------|-----------------------|
| Power meter NRP2           | SN: 104778       | 26-Mar-24 (No. 217-04036/04037) | Mar-25                |
| Power sensor NRP-Z91       | SN: 103244       | 26-Mar-24 (No. 217-04036)       | Mar-25                |
| Power sensor NRP-Z91       | SN: 103245       | 26-Mar-24 (No. 217-04037)       | Mar-25                |
| Reference 20 dB Attenuator | SN: CC2552 (20x) | 26-Mar-24 (No. 217-04046)       | Mar-25                |
| DAE4                       | SN: 789          | 03-Oct-24 (No. DAE4-789_Oct24)  | Oct-25                |
| Reference Probe ER3DV6     | SN: 2328         | 01-Oct-24 (No. ER3-2328_Oct24)  | Oct-25                |

| Secondary Standards     | ID               | Check Date (in house)             | Scheduled Check        |
|-------------------------|------------------|-----------------------------------|------------------------|
| Power meter E4419B      | SN: GB41293874   | 06-Apr-16 (in house check Jun-24) | In house check: Jun-26 |
| Power sensor E4412A     | SN: MY41498087   | 06-Apr-16 (in house check Jun-24) | In house check: Jun-26 |
| Power sensor E4412A     | SN: 000110210    | 06-Apr-16 (in house check Jun-24) | In house check: Jun-26 |
| RF generator HP 8648C   | SN: US3642U01700 | 04-Aug-99 (in house check Jun-24) | In house check: Jun-26 |
| Network Analyzer E8358A | SN: US41080477   | 31-Mar-14 (in house check Sep-24) | In house check: Sep-26 |

|               |                |                       |           |
|---------------|----------------|-----------------------|-----------|
|               | Name           | Function              | Signature |
| Calibrated by | Jeton Kastrati | Laboratory Technician |           |
| Approved by   | Sven Kühn      | Technical Manager     |           |

Issued: October 24, 2024

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

## Calibration Laboratory of

Schmid & Partner  
Engineering AG

Zeughausstrasse 43, 8004 Zurich, Switzerland



**S** Schweizerischer Kalibrierdienst  
**C** Service suisse d'étalonnage  
**S** Servizio svizzero di taratura  
**S** Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)

**The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates**

Accreditation No.: **SCS 0108**

## Glossary

|                          |                                                                                                                                                      |
|--------------------------|------------------------------------------------------------------------------------------------------------------------------------------------------|
| NORM <sub>x,y,z</sub>    | sensitivity in free space                                                                                                                            |
| DCP                      | diode compression point                                                                                                                              |
| CF                       | crest factor (1/duty_cycle) of the RF signal                                                                                                         |
| A, B, C, D               | modulation dependent linearization parameters                                                                                                        |
| En                       | incident E-field orientation normal to probe axis                                                                                                    |
| Ep                       | incident E-field orientation parallel to probe axis                                                                                                  |
| Polarization $\varphi$   | $\varphi$ rotation around probe axis                                                                                                                 |
| Polarization $\vartheta$ | $\vartheta$ rotation around an axis that is in the plane normal to probe axis (at measurement center), i.e., $\vartheta = 0$ is normal to probe axis |
| Connector Angle          | information used in DASY system to align probe sensor X to the robot coordinate system                                                               |

## Calibration is Performed According to the Following Standards:

- IEEE Std 1309-2005, "IEEE Standard for calibration of electromagnetic field sensors and probes, excluding antennas, from 9 kHz to 40 GHz", December 2005
- CTIA Test Plan for Hearing Aid Compatibility, Rev 3.1.1, May 2017

## Methods Applied and Interpretation of Parameters:

- NORM<sub>x,y,z</sub>**: Assessed for E-field polarization  $\vartheta = 0$  for XY sensors and  $\vartheta = 90$  for Z sensor ( $f \leq 900$  MHz in TEM-cell;  $f > 1800$  MHz in R22 waveguide).
- NORM(f)<sub>x,y,z</sub> = NORM<sub>x,y,z</sub> \* frequency\_response** (see Frequency Response Chart).
- DCP<sub>x,y,z</sub>**: DCP are numerical linearization parameters assessed based on the data of power sweep with CW signal. DCP does not depend on frequency nor media.
- PAR**: PAR is the Peak to Average Ratio that is not calibrated but determined based on the signal characteristics
- A<sub>x,y,z</sub>; B<sub>x,y,z</sub>; C<sub>x,y,z</sub>; D<sub>x,y,z</sub>; VR<sub>x,y,z</sub>**: A, B, C, D are numerical linearization parameters assessed based on the data of power sweep for specific modulation signal. The parameters do not depend on frequency nor media. VR is the maximum calibration range expressed in RMS voltage across the diode.
- Spherical isotropy (3D deviation from isotropy)**: in a locally homogeneous field realized using an open waveguide setup
- Sensor Offset**: The sensor offset corresponds to the offset of virtual measurement center from the probe tip (on probe axis). No tolerance required.
- Connector Angle**: The angle is assessed using the information gained by determining the **NORM<sub>x</sub>** (no uncertainty required).

## Parameters of Probe: ER3DV6 - SN:2434

### Basic Calibration Parameters

|                          | Sensor X | Sensor Y | Sensor Z | Unc ( $k = 2$ ) |
|--------------------------|----------|----------|----------|-----------------|
| Norm ( $\mu V/(V/m)^2$ ) | 1.49     | 1.62     | 1.97     | $\pm 10.1\%$    |
| DCP (mV) <sup>B</sup>    | 97.3     | 99.0     | 99.3     | $\pm 4.7\%$     |

### Calibration Results for Frequency Response (30 MHz – 3 GHz)

| Frequency MHz | Target E-field (En) V/m | Measured E-field (En) V/m | Deviation E-field (En) | Target E-field (Ep) V/m | Measured E-field (Ep) V/m | Deviation E-field (Ep) | Unc ( $k = 2$ ) |
|---------------|-------------------------|---------------------------|------------------------|-------------------------|---------------------------|------------------------|-----------------|
| 30            | 77.1                    | 76.6                      | -0.6%                  | 77.1                    | 77.4                      | 0.4%                   | $\pm 5.1\%$     |
| 100           | 77.0                    | 78.2                      | 1.5%                   | 76.9                    | 78.0                      | 1.5%                   | $\pm 5.1\%$     |
| 450           | 77.2                    | 78.7                      | 1.9%                   | 77.2                    | 78.2                      | 1.3%                   | $\pm 5.1\%$     |
| 600           | 77.2                    | 78.3                      | 1.4%                   | 77.1                    | 77.8                      | 0.9%                   | $\pm 5.1\%$     |
| 750           | 77.2                    | 78.1                      | 1.2%                   | 77.1                    | 77.7                      | 0.8%                   | $\pm 5.1\%$     |
|               |                         |                           |                        |                         |                           |                        |                 |
| 1800          | 143.2                   | 142.9                     | -0.2%                  | 143.2                   | 142.6                     | -0.4%                  | $\pm 5.1\%$     |
| 2000          | 135.2                   | 132.9                     | -1.7%                  | 135.3                   | 132.5                     | -2.1%                  | $\pm 5.1\%$     |
| 2200          | 127.9                   | 127.8                     | -0.1%                  | 127.5                   | 129.4                     | 1.5%                   | $\pm 5.1\%$     |
| 2500          | 125.2                   | 123.8                     | -1.2%                  | 125.5                   | 125.7                     | 0.1%                   | $\pm 5.1\%$     |
| 3000          | 79.5                    | 79.1                      | -0.6%                  | 79.5                    | 82.1                      | 3.4%                   | $\pm 5.1\%$     |

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor  $k=2$ , which for a normal distribution corresponds to a coverage probability of approximately 95%.

<sup>B</sup> Linearization parameter uncertainty for maximum specified field strength.

<sup>E</sup> Uncertainty is determined using the max. deviation from linear response applying rectangular distribution and is expressed for the square of the field value.

## Parameters of Probe: ER3DV6 - SN:2434

### Calibration Results for Modulation Response

| UID   | Communication System Name   |   | A<br>dB | B<br>dB $\sqrt{\mu V}$ | C     | D<br>dB | VR<br>mV | Max<br>dev. | Max<br>Unc <sup>E</sup><br>k = 2 |
|-------|-----------------------------|---|---------|------------------------|-------|---------|----------|-------------|----------------------------------|
| 0     | CW                          | X | 0.00    | 0.00                   | 1.00  | 0.00    | 200.5    | ±3.5%       | ±4.7%                            |
|       |                             | Y | 0.00    | 0.00                   | 1.00  |         | 198.5    |             |                                  |
|       |                             | Z | 0.00    | 0.00                   | 1.00  |         | 213.1    |             |                                  |
| 10352 | Pulse Waveform (200Hz, 10%) | X | 10.79   | 83.87                  | 21.55 | 10.00   | 60.0     | ±1.9%       | ±9.6%                            |
|       |                             | Y | 11.59   | 85.42                  | 21.50 |         | 60.0     |             |                                  |
|       |                             | Z | 10.76   | 83.22                  | 22.43 |         | 60.0     |             |                                  |
| 10353 | Pulse Waveform (200Hz, 20%) | X | 11.06   | 85.16                  | 20.58 | 6.99    | 80.0     | ±2.4%       | ±9.6%                            |
|       |                             | Y | 13.74   | 88.34                  | 21.11 |         | 80.0     |             |                                  |
|       |                             | Z | 11.55   | 85.92                  | 21.92 |         | 80.0     |             |                                  |
| 10354 | Pulse Waveform (200Hz, 40%) | X | 20.00   | 93.79                  | 21.52 | 3.98    | 95.0     | ±3.2%       | ±9.6%                            |
|       |                             | Y | 20.00   | 93.75                  | 21.13 |         | 95.0     |             |                                  |
|       |                             | Z | 20.00   | 95.19                  | 23.02 |         | 95.0     |             |                                  |
| 10355 | Pulse Waveform (200Hz, 60%) | X | 20.00   | 93.34                  | 19.69 | 2.22    | 120.0    | ±3.3%       | ±9.6%                            |
|       |                             | Y | 20.00   | 93.68                  | 19.58 |         | 120.0    |             |                                  |
|       |                             | Z | 20.00   | 97.15                  | 22.26 |         | 120.0    |             |                                  |
| 10387 | QPSK Waveform, 1 MHz        | X | 1.99    | 66.94                  | 15.76 | 1.00    | 150.0    | ±1.5%       | ±9.6%                            |
|       |                             | Y | 1.99    | 67.01                  | 15.81 |         | 150.0    |             |                                  |
|       |                             | Z | 1.95    | 67.35                  | 15.81 |         | 150.0    |             |                                  |
| 10388 | QPSK Waveform, 10 MHz       | X | 2.49    | 68.75                  | 15.83 | 0.00    | 150.0    | ±1.2%       | ±9.6%                            |
|       |                             | Y | 2.66    | 69.86                  | 16.37 |         | 150.0    |             |                                  |
|       |                             | Z | 2.47    | 69.02                  | 16.04 |         | 150.0    |             |                                  |
| 10396 | 64-QAM Waveform, 100 kHz    | X | 5.39    | 80.43                  | 23.89 | 3.01    | 150.0    | ±0.7%       | ±9.6%                            |
|       |                             | Y | 4.67    | 77.45                  | 22.28 |         | 150.0    |             |                                  |
|       |                             | Z | 4.39    | 74.71                  | 20.73 |         | 150.0    |             |                                  |
| 10399 | 64-QAM Waveform, 40 MHz     | X | 3.58    | 66.96                  | 15.65 | 0.00    | 150.0    | ±1.6%       | ±9.6%                            |
|       |                             | Y | 3.73    | 67.74                  | 16.03 |         | 150.0    |             |                                  |
|       |                             | Z | 3.57    | 67.14                  | 15.74 |         | 150.0    |             |                                  |
| 10414 | WLAN CCDF, 64-QAM, 40 MHz   | X | 5.03    | 65.44                  | 15.44 | 0.00    | 150.0    | ±3.3%       | ±9.6%                            |
|       |                             | Y | 5.00    | 65.40                  | 15.40 |         | 150.0    |             |                                  |
|       |                             | Z | 4.99    | 65.63                  | 15.51 |         | 150.0    |             |                                  |

Note: For details on UID parameters see Appendix

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

<sup>B</sup> Linearization parameter uncertainty for maximum specified field strength.

<sup>E</sup> Uncertainty is determined using the max. deviation from linear response applying rectangular distribution and is expressed for the square of the field value.

Parameters of Probe: ER3DV6 - SN:2434

Sensor Frequency Model Parameters

|                      | Sensor X | Sensor Y | Sensor Z |
|----------------------|----------|----------|----------|
| Frequency Corr. (LF) | -1.26    | -1.36    | -0.06    |
| Frequency Corr. (HF) | 0.00     | 0.00     | 0.00     |


Sensor Model Parameters

|   | C1<br>fF | C2<br>fF | $\alpha$<br>V <sup>-1</sup> | T1<br>msV <sup>-2</sup> | T2<br>msV <sup>-1</sup> | T3<br>ms | T4<br>V <sup>-2</sup> | T5<br>V <sup>-1</sup> | T6   |
|---|----------|----------|-----------------------------|-------------------------|-------------------------|----------|-----------------------|-----------------------|------|
| x | 101.3    | 491.02   | 37.02                       | 28.09                   | 2.04                    | 5.10     | 0.00                  | 0.65                  | 1.02 |
| y | 99.2     | 475.49   | 36.40                       | 26.15                   | 1.43                    | 5.10     | 0.00                  | 0.64                  | 1.02 |
| z | 86.3     | 409.31   | 35.84                       | 29.76                   | 3.23                    | 5.10     | 0.00                  | 0.82                  | 1.01 |


Other Probe Parameters

|                                         |             |
|-----------------------------------------|-------------|
| Sensor Arrangement                      | Rectangular |
| Connector Angle                         | 84.6°       |
| Mechanical Surface Detection Mode       | enabled     |
| Optical Surface Detection Mode          | disabled    |
| Probe Overall Length                    | 337 mm      |
| Probe Body Diameter                     | 10 mm       |
| Tip Length                              | 10 mm       |
| Tip Diameter                            | 8 mm        |
| Probe Tip to Sensor X Calibration Point | 2.5 mm      |
| Probe Tip to Sensor Y Calibration Point | 2.5 mm      |
| Probe Tip to Sensor Z Calibration Point | 2.5 mm      |


Receiving Pattern ( $\phi$ ),  $\vartheta = 0^\circ$




Receiving Pattern ( $\phi$ ),  $\vartheta = 90^\circ$



Receiving Pattern ( $\phi$ ),  $\vartheta = 0^\circ$





Receiving Pattern ( $\phi$ ),  $\vartheta = 90^\circ$



Dynamic Range f(E-field)



(TEM cell,  $f_{eval} = 900\text{MHz}$ )



Uncertainty of Linearity Assessment:  $\pm 0.6\%$  ( $k=2$ )

## Deviation from Isotropy in Air

Error ( $\phi, \theta$ ),  $f = 900\text{MHz}$



Uncertainty of Spherical Isotropy Assessment:  $\pm 2.6\%$  ( $k=2$ )