



Date: 27.04.2015

Federal Communications Commission
Authorization and Evaluation Division
7435 Oakland Mills Road
Columbia, MD 21046

Attn: OET Dept.

Ref: FCC Class II Permissive change for FCC ID: 2ACT5-Z
(Original Grant date:22/01/2015)
Applicant: Cilag GmbH International

Dear Examiner,

This is to request a Class II permissive change for the Blood Glucose Meter Model: OneTouch Verio Flex, originally granted on 22/01/2015. FCC Certification No. 2ACT5-Z

The major change filed under this application is:

- **Meter housing colour changed from white to blue.**
- **Meter branding changed (details not yet finalized).**
- **SPC changed to switch position of the long and short contacts.**
(except for SPC – Strip Port Connector)
- **Keypad colour changed to match blue housing.**
- **RF part is identical for both Models**

If you have any questions regarding this application, please feel free to contact me.

Sincerely,

A handwritten signature in blue ink, appearing to be 'T. Ring'.

Mr. Thomas Ring
Manager Certification