

## Label - SN Air11 Bilevel AMER

### **ResMed Confidential**

Proprietary Information, not to be reproduced or made available to third parties without prior consent from ResMed and not to be used in any unauthorised way. © 2022 ResMed..

SPEC NO.	R390-432	
REV	1.0	
PAGE	1 of 1	

Note: Barcodes are representative examples only of scale and placement

1.Scope: This specification defines the Blank Label, the Artwork, the Ink Ribbon

and the Master Spec to be used for printing the

Label - SN Air11 Bilevel AMER.

and is for reference and (where specified) positioning purposes only.

2. Requirement Only the specified Ink Ribbon & Blank Label materials shall be

used on this label, in order to ensure compliance to the durability

requirements for mandatory product information

(text and symbols).

NOTE: Any changes to this specification MUST be reviewed by

the IT Department

3. Blank Label For all label requirements, refer to the latest revision of

**R390-404** Label - SN 80x25 Silver (Blank)

4. Artwork File: For the artwork, use the CodeSoft "R390-432.001.lab" file.

5. Ink Ribbon For all ink ribbon requirements, refer to the latest revision of


**Z9467** Thermal Print Ribbon

6. Master Specification: For all other printing information, refer to the latest revision of

R390-405 Master Spec - Label-SN Printing

7. Unique Device For UDI labelling information, refer to R001-341 Barcode Labels – UDI

Identifier: 2D.



# DRAFT - NOT AN APPROVED RELEASE

## NOTES:

**Artwork** 

This box = 40mm x 4mm

Rev	CN No.	Date Drafted	Details	Drawn By	Checked By
1.0	K010094-00	21 Jan 2022	Initial Release of Specification	Melanie Webster	Annick De Silva

**COLOURS:** CODES:



- 1. Scale 1:1 (A3)
- 2. Dimensions in mm
- 3. GENERAL TOLERANCES
- (unless otherwise stated)
- Linear Dimenions +/- 0.25mm
- Angular Dimensions +/- 0.5°
- Printing tolerance between colours +/- 0.25mm
- Graphics registration with part +/- 0.4mm