

D2450V2 Sn:738

In Collaboration with
TTL **SPEAG**
CALIBRATION LABORATORY

Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-42394633-2079 Fax: +86-10-42394633-2504
E-mail: cti@chinaatl.com http://www.chinaatl.cn

中国认可
国际互认
CALIBRATION
CNAS L5570

Client: **SRTC** Certificate No: **Z17-97140**

CALIBRATION CERTIFICATE

Object: **D2450V2 - SN: 738**

Calibration Procedure(s): **FF-Z11-003-01**
Calibration Procedures for dipole validation kits

Calibration date: **September 18, 2017**

This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements(SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature(22±3)°C and humidity<70%.

Calibration Equipment used (M&TE critical for calibration)

Primary Standards	ID #	Cal Date(Calibrated by Certificate No.)	Scheduled Calibration
Power Meter NRVD	102196	02-Mar-17 (CTTL No.J17X01254)	Mar-18
Power sensor NR-V5	100596	02-Mar-17 (CTTL No.J17X01254)	Mar-18
Reference Probe EX3DV4	SN 7433	26-Sep-16(SPEAG No EX3-7433_Sep16)	Sep-17
D4E4	SN 1331	19-Jan-17(CTTL-SPEAG No Z17-97015)	Jan-18

Secondary Standards	ID #	Cal Date(Calibrated by Certificate No.)	Scheduled Calibration
Signal Generator E4438C	MY48071430	13-Jan-17 (CTTL No.J17X00285)	Jan-18
Network Analyzer E5071C	MY46110673	13-Jan-17 (CTTL No.J17X00285)	Jan-18

Calibrated by: **Zhao Jing** SAR Test Engineer Signature: *[Signature]*

Reviewed by: **Yu Zongying** SAR Test Engineer Signature: *[Signature]*

Approved by: **Qi Dianyuan** SAR Project Leader Signature: *[Signature]*

Issued: September 21, 2017

This calibration certificate shall not be reproduced except in full without written approval of the laboratory

Certificate No: Z17-97140

Page 1 of 8

In Collaboration with
TTL **SPEAG**
CALIBRATION LABORATORY

Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-42394633-2079 Fax: +86-10-42394633-2504
E-mail: cti@chinaatl.com http://www.chinaatl.cn

Glossary:
TSL tissue simulating liquid
ConvF sensitivity in TSL / NORMx,y,z
N/A not applicable or not measured

Calibration is Performed According to the Following Standards:

- IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013
- IEC 62209-1, "Measurement procedure for assessment of specific absorption rate of human exposure to radio frequency fields from hand-held and body-mounted wireless communication devices- Part 1: Device used next to the ear (Frequency range of 300MHz to 6GHz)", July 2016
- IEC 62209-2, "Procedure to measure the Specific Absorption Rate (SAR) For wireless communication devices used in close proximity to the human body (frequency range of 30MHz to 6GHz)", March 2010
- KDB865864, SAR Measurement Requirements for 100 MHz to 6 GHz

Additional Documentation:

- DASY4/5 System Handbook

Methods Applied and Interpretation of Parameters:

- Measurement Conditions:** Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL:** The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss:** These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay:** One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured:** SAR measured at the stated antenna input power.
- SAR normalized:** SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters:** The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution Corresponds to a coverage probability of approximately 95%.

Certificate No: Z17-97140

Page 2 of 8

In Collaboration with
TTL **SPEAG**
CALIBRATION LABORATORY

Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-42394633-2079 Fax: +86-10-42394633-2504
E-mail: cti@chinaatl.com http://www.chinaatl.cn

Measurement Conditions

DASY system configuration, as far as not given on page 1.

DASY Version	DASY52	52.10.0.1448
Extrapolation	Advanced Extrapolation	
Phantom	Triple Flat Phantom 5.1C	
Distance Dipole Center - TSL	10 mm	with Spacer
Zoom Scan Resolution	dx, dy, dz = 5 mm	
Frequency	2450 MHz ± 1 MHz	

Head TSL parameters

The following parameters and calculations were applied:

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	39.2	1.80 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	38.7 ± 6 %	1.79 mho/m ± 6 %
Head TSL temperature change during test	<1.0 °C	---	---

SAR result with Head TSL

SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	
SAR measured	250 mW input power	13.1 mW / g
SAR for nominal Head TSL parameters	normalized to 1W	52.4 mW / g ± 18.8 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition	
SAR measured	250 mW input power	6.10 mW / g
SAR for nominal Head TSL parameters	normalized to 1W	24.4 mW / g ± 18.7 % (k=2)

Body TSL parameters

The following parameters and calculations were applied:

	Temperature	Permittivity	Conductivity
Nominal Body TSL parameters	22.0 °C	52.7	1.95 mho/m
Measured Body TSL parameters	(22.0 ± 0.2) °C	52.5 ± 6 %	1.98 mho/m ± 6 %
Body TSL temperature change during test	<1.0 °C	---	---

SAR result with Body TSL

SAR averaged over 1 cm ³ (1 g) of Body TSL	Condition	
SAR measured	250 mW input power	13.2 mW / g
SAR for nominal Body TSL parameters	normalized to 1W	52.3 mW / g ± 18.8 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Body TSL	Condition	
SAR measured	250 mW input power	6.10 mW / g
SAR for nominal Body TSL parameters	normalized to 1W	24.3 mW / g ± 18.7 % (k=2)

Certificate No: Z17-97140

Page 3 of 8

In Collaboration with
TTL **SPEAG**
CALIBRATION LABORATORY

Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-42394633-2079 Fax: +86-10-42394633-2504
E-mail: cti@chinaatl.com http://www.chinaatl.cn

Appendix (Additional assessments outside the scope of CNAS L5070)

Antenna Parameters with Head TSL

Impedance, transformed to feed point	51.30 ± 5.93Ω
Return Loss	- 24.5dB

Antenna Parameters with Body TSL

Impedance, transformed to feed point	47.60 ± 6.38Ω
Return Loss	- 23.1dB

General Antenna Parameters and Design

Electrical Delay (one direction)	1.268 ns
----------------------------------	----------

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.


Additional EUT Data

Manufactured by	SPEAG
-----------------	-------

Certificate No: Z17-97140


Page 4 of 8

D2450V2 Sn:738




Certificate No: Z17-97140

Page 5 of 8




Certificate No: Z17-97140

Page 6 of 8



Certificate No: Z17-97140

Page 7 of 8



Certificate No: Z17-97140

Page 8 of 8

D5GHzV2 Sn:1079 (1/4)

In Collaboration with
TTL **s p e e a g**
CALIBRATION LABORATORY

Add: No.31 Xuyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2304
E-mail: cti@china.ttl.com.cn http://www.china.ttl.com.cn

Client: **SRTC** Certificate No: **Z17-97133**

CALIBRATION CERTIFICATE

Object: **D5GHzV2 - SN: 1079**

Calibration Procedure(s): **FF-Z11-003-01**
Calibration Procedures for dipole validation kits

Calibration date: **September 25, 2017**

This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements(SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature(22±3)°C and humidity<70%.

Calibration Equipment used (MATE critical for calibration)

Primary Standards	ID #	Cal Date/Calibrated by, Certificate No.	Scheduled Calibration
Power Meter NRP2	102196	02-Mar-17 (CTTL, No.J17X01254)	Mar-18
Power sensor NRP-291	100599	02-Mar-17 (CTTL, No.J17X01254)	Mar-18
ReferenceProbe EXG3V4	SN 3846	13-Jan-17(CTTL-SPEAG, No.Z15-97251)	Jan-18
DAE4	SN 1331	19-Jan-17(CTTL-SPEAG, No.Z17-97015)	Jan-18

Secondary Standards	ID #	Cal Date/Calibrated by, Certificate No.	Scheduled Calibration
Signal Generator E4438C	MY49071430	13-Jan-17 (CTTL, No.J17X00286)	Jan-18
NetworkAnalyzer E5071C	MY46110673	13-Jan-17 (CTTL, No.J17X00286)	Jan-18

Calibrated by: **Zhao Jing** SAR Test Engineer Signature: *[Signature]*

Reviewed by: **Yu Zangying** SAR Test Engineer Signature: *[Signature]*

Approved by: **Qi Dianyan** SAR Project Leader Signature: *[Signature]*

Issued: September 28, 2017

This calibration certificate shall not be reproduced except in full without written approval of the laboratory

Certificate No: Z17-97133

Page 1 of 16

In Collaboration with
TTL **s p e e a g**
CALIBRATION LABORATORY

Add: No.31 Xuyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2304
E-mail: cti@china.ttl.com.cn http://www.china.ttl.com.cn

Glossary:
TSL: tissue simulating liquid
CorrF: sensitivity in TSL / NCRMs, y, z
N/A: not applicable or not measured

Calibration is Performed According to the Following Standards:
a) IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013
b) IEC 62209-1, "Measurement procedure for assessment of specific absorption rate of human exposure to radio frequency fields from hand-held and body-mounted wireless communication devices-Part 1: Device used next to the ear (Frequency range of 300MHz to 6GHz)", July 2016
c) IEC 62209-2, "Procedure to measure the Specific Absorption Rate (SAR) For wireless communication devices used in close proximity to the human body (frequency range of 30MHz to 6GHz)", March 2010
d) KDB655664, SAR Measurement Requirements for 100 MHz to 6 GHz.

Additional Documentation:
a) DASYS System Handbook

Methods Applied and Interpretation of Parameters:
• Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
• Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
• Feed Point Impedance and Return Loss: These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
• Electrical Delay: One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
• SAR measured: SAR measured at the stated antenna input power.
• SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector.
• SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of Measurement, multiplied by the coverage factor k=2, which for a normal distribution Corresponds to a coverage probability of approximately 95%.

Certificate No: Z17-97133

Page 2 of 16

In Collaboration with
TTL **s p e e a g**
CALIBRATION LABORATORY

Add: No.31 Xuyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2304
E-mail: cti@china.ttl.com.cn http://www.china.ttl.com.cn

Measurement Conditions
DASY system configuration, as far as not given on page 1.

DASY Version	DASY2	52.10.0.1446
Extrapolation	Advanced Extrapolation	
Phantom	Triple Flat Phantom 5.1C	
Distance Dipole Center - TSL	10 mm	with Spacer
Zoom Scan Resolution	6k, dy = 4 mm, dz = 1.4 mm	Graded Ratio = 1.4 (Z direction)
Frequency	5250 MHz ± 1 MHz 5300 MHz ± 1 MHz 5500 MHz ± 1 MHz 5600 MHz ± 1 MHz 5600 MHz ± 1 MHz	

Head TSL parameters at 5200 MHz
The following parameters and calculations were applied

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	35.0	4.66 nhm/s
Measured Head TSL parameters	(22.0 ± 0.2) °C	35.7 ± 5 %	4.62 nhm/s ± 6 %
Head TSL temperature change during test	<1.0 °C	---	---

SAR result with Head TSL at 5200 MHz

	Condition	
SAR averaged over 1 cm ³ (1 g) of Head TSL		
SAR measured	100 mW input power	7.77 mW / g
SAR for nominal Head TSL parameters	normalized to 1W	77.6 mW / g ± 24.4 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Head TSL		
SAR measured	100 mW input power	2.24 mW / g
SAR for nominal Head TSL parameters	normalized to 1W	22.3 mW / g ± 24.2 % (k=2)

Certificate No: Z17-97133

Page 3 of 16

In Collaboration with
TTL **s p e e a g**
CALIBRATION LABORATORY

Add: No.31 Xuyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2304
E-mail: cti@china.ttl.com.cn http://www.china.ttl.com.cn

Head TSL parameters at 5300 MHz
The following parameters and calculations were applied

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	35.9	4.76 nhm/s
Measured Head TSL parameters	(22.0 ± 0.2) °C	36.1 ± 6 %	4.67 nhm/s ± 6 %
Head TSL temperature change during test	<1.0 °C	---	---

SAR result with Head TSL at 5300 MHz

	Condition	
SAR averaged over 1 cm ³ (1 g) of Head TSL		
SAR measured	100 mW input power	8.13 mW / g
SAR for nominal Head TSL parameters	normalized to 1W	81.3 mW / g ± 24.4 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Head TSL		
SAR measured	100 mW input power	2.32 mW / g
SAR for nominal Head TSL parameters	normalized to 1W	23.2 mW / g ± 24.2 % (k=2)

Head TSL parameters at 5500 MHz
The following parameters and calculations were applied

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	35.6	4.96 nhm/s
Measured Head TSL parameters	(22.0 ± 0.2) °C	36.9 ± 6 %	4.93 nhm/s ± 6 %
Head TSL temperature change during test	<1.0 °C	---	---

SAR result with Head TSL at 5500 MHz

	Condition	
SAR averaged over 1 cm ³ (1 g) of Head TSL		
SAR measured	100 mW input power	8.24 mW / g
SAR for nominal Head TSL parameters	normalized to 1W	82.6 mW / g ± 24.4 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Head TSL		
SAR measured	100 mW input power	2.37 mW / g
SAR for nominal Head TSL parameters	normalized to 1W	23.8 mW / g ± 24.2 % (k=2)

Certificate No: Z17-97133

Page 4 of 16

D5GHzV2 Sn:1079 (2/4)

TTL In Collaboration with **speag**
CALIBRATION LABORATORY

Addr: No.31 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-42394633-2079 Fax: +86-10-42394633-2504
E-mail: cti@china.ttl.com http://www.chinatitl.cn

Head TSL parameters at 5600 MHz
The following parameters and calculations were applied:

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	35.5	5.07 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	35.7 ± 6 %	4.98 mho/m ± 6 %
Head TSL temperature change during test	<1.0 °C	-----	-----

SAR result with Head TSL at 5600 MHz

	Condition	
SAR averaged over 1 cm ³ (1 g) of Head TSL	100 mW input power	8.16 mW / g
SAR measured	normalized to 1W	81.6 mW / g ± 24.4 % (k=2)
SAR for nominal Head TSL parameters	Condition	
SAR averaged over 10 cm ³ (10 g) of Head TSL	100 mW input power	2.34 mW / g
SAR measured	normalized to 1W	23.4 mW / g ± 24.2 % (k=2)

Head TSL parameters at 5800 MHz
The following parameters and calculations were applied:

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	35.3	5.27 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	35.8 ± 6 %	5.18 mho/m ± 6 %
Head TSL temperature change during test	<1.0 °C	-----	-----

SAR result with Head TSL at 5800 MHz

	Condition	
SAR averaged over 1 cm ³ (1 g) of Head TSL	100 mW input power	7.85 mW / g
SAR measured	normalized to 1W	78.7 mW / g ± 24.4 % (k=2)
SAR for nominal Head TSL parameters	Condition	
SAR averaged over 10 cm ³ (10 g) of Head TSL	100 mW input power	2.25 mW / g
SAR measured	normalized to 1W	22.6 mW / g ± 24.2 % (k=2)

Certificate No: Z17-97133

Page 5 of 16

TTL In Collaboration with **speag**
CALIBRATION LABORATORY

Addr: No.31 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-42394633-2079 Fax: +86-10-42394633-2504
E-mail: cti@china.ttl.com http://www.chinatitl.cn

Body TSL parameters at 5200 MHz
The following parameters and calculations were applied:

	Temperature	Permittivity	Conductivity
Nominal Body TSL parameters	22.0 °C	49.0	5.30 mho/m
Measured Body TSL parameters	(22.0 ± 0.2) °C	49.5 ± 6 %	5.38 mho/m ± 5 %
Body TSL temperature change during test	<1.0 °C	-----	-----

SAR result with Body TSL at 5200 MHz

	Condition	
SAR averaged over 1 cm ³ (1 g) of Body TSL	100 mW input power	7.52 mW / g
SAR measured	normalized to 1W	75.4 mW / g ± 24.4 % (k=2)
SAR for nominal Body TSL parameters	Condition	
SAR averaged over 10 cm ³ (10 g) of Body TSL	100 mW input power	2.12 mW / g
SAR measured	normalized to 1W	21.3 mW / g ± 24.2 % (k=2)

Body TSL parameters at 5300 MHz
The following parameters and calculations were applied:

	Temperature	Permittivity	Conductivity
Nominal Body TSL parameters	22.0 °C	48.9	5.42 mho/m
Measured Body TSL parameters	(22.0 ± 0.2) °C	49.2 ± 6 %	5.50 mho/m ± 6 %
Body TSL temperature change during test	<1.0 °C	-----	-----

SAR result with Body TSL at 5300 MHz

	Condition	
SAR averaged over 1 cm ³ (1 g) of Body TSL	100 mW input power	7.68 mW / g
SAR measured	normalized to 1W	76.9 mW / g ± 24.4 % (k=2)
SAR for nominal Body TSL parameters	Condition	
SAR averaged over 10 cm ³ (10 g) of Body TSL	100 mW input power	2.18 mW / g
SAR measured	normalized to 1W	21.9 mW / g ± 24.2 % (k=2)

Certificate No: Z17-97133

Page 6 of 16

TTL In Collaboration with **speag**
CALIBRATION LABORATORY

Addr: No.31 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-42394633-2079 Fax: +86-10-42394633-2504
E-mail: cti@china.ttl.com http://www.chinatitl.cn

Body TSL parameters at 5500 MHz
The following parameters and calculations were applied:

	Temperature	Permittivity	Conductivity
Nominal Body TSL parameters	22.0 °C	48.6	5.65 mho/m
Measured Body TSL parameters	(22.0 ± 0.2) °C	49.0 ± 6 %	5.72 mho/m ± 6 %
Body TSL temperature change during test	<1.0 °C	-----	-----

SAR result with Body TSL at 5500 MHz

	Condition	
SAR averaged over 1 cm ³ (1 g) of Body TSL	100 mW input power	8.22 mW / g
SAR measured	normalized to 1W	82.4 mW / g ± 24.4 % (k=2)
SAR for nominal Body TSL parameters	Condition	
SAR averaged over 10 cm ³ (10 g) of Body TSL	100 mW input power	2.35 mW / g
SAR measured	normalized to 1W	23.6 mW / g ± 24.2 % (k=2)

Body TSL parameters at 5600 MHz
The following parameters and calculations were applied:

	Temperature	Permittivity	Conductivity
Nominal Body TSL parameters	22.0 °C	48.5	5.77 mho/m
Measured Body TSL parameters	(22.0 ± 0.2) °C	48.4 ± 6 %	5.73 mho/m ± 6 %
Body TSL temperature change during test	<1.0 °C	-----	-----

SAR result with Body TSL at 5600 MHz

	Condition	
SAR averaged over 1 cm ³ (1 g) of Body TSL	100 mW input power	8.08 mW / g
SAR measured	normalized to 1W	80.7 mW / g ± 24.4 % (k=2)
SAR for nominal Body TSL parameters	Condition	
SAR averaged over 10 cm ³ (10 g) of Body TSL	100 mW input power	2.30 mW / g
SAR measured	normalized to 1W	23.0 mW / g ± 24.2 % (k=2)

Certificate No: Z17-97133

Page 7 of 16

TTL In Collaboration with **speag**
CALIBRATION LABORATORY

Addr: No.31 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-42394633-2079 Fax: +86-10-42394633-2504
E-mail: cti@china.ttl.com http://www.chinatitl.cn

Body TSL parameters at 5800 MHz
The following parameters and calculations were applied:

	Temperature	Permittivity	Conductivity
Nominal Body TSL parameters	22.0 °C	48.2	6.00 mho/m
Measured Body TSL parameters	(22.0 ± 0.2) °C	49.0 ± 6 %	5.94 mho/m ± 6 %
Body TSL temperature change during test	<1.0 °C	-----	-----

SAR result with Body TSL at 5800 MHz

	Condition	
SAR averaged over 1 cm ³ (1 g) of Body TSL	100 mW input power	7.73 mW / g
SAR measured	normalized to 1W	77.5 mW / g ± 24.4 % (k=2)
SAR for nominal Body TSL parameters	Condition	
SAR averaged over 10 cm ³ (10 g) of Body TSL	100 mW input power	2.17 mW / g
SAR measured	normalized to 1W	21.8 mW / g ± 24.2 % (k=2)

Certificate No: Z17-97133

Page 8 of 16

D5GHzV2 Sn:1079 (3/4)

TTL In Collaboration with SPEAG CALIBRATION LABORATORY Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com http://www.china.ttl.cn	
Appendix (Additional assessments outside the scope of CNAS L0570)	
Antenna Parameters with Head TSL at 5200 MHz	
Impedance, transformed to feed point	47.6Ω - 8.77jΩ
Return Loss	-20.7dB
Antenna Parameters with Head TSL at 5300 MHz	
Impedance, transformed to feed point	45.5Ω - 6.82jΩ
Return Loss	-21.4dB
Antenna Parameters with Head TSL at 5500 MHz	
Impedance, transformed to feed point	50.7Ω - 7.14jΩ
Return Loss	-23.0dB
Antenna Parameters with Head TSL at 5600 MHz	
Impedance, transformed to feed point	55.2Ω - 4.09jΩ
Return Loss	-24.1dB
Antenna Parameters with Head TSL at 5800 MHz	
Impedance, transformed to feed point	52.2Ω - 5.29jΩ
Return Loss	-21.6dB
Antenna Parameters with Body TSL at 5200 MHz	
Impedance, transformed to feed point	50.8Ω - 10.1jΩ
Return Loss	-20.0dB
Antenna Parameters with Body TSL at 5300 MHz	
Impedance, transformed to feed point	48.5Ω - 8.56jΩ
Return Loss	-21.1dB

Certificate No: Z17-97133 Page 9 of 16


TTL In Collaboration with SPEAG CALIBRATION LABORATORY Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com http://www.china.ttl.cn	
Antenna Parameters with Body TSL at 5500 MHz	
Impedance, transformed to feed point	54.9Ω - 8.85jΩ
Return Loss	-21.9dB
Antenna Parameters with Body TSL at 5600 MHz	
Impedance, transformed to feed point	56.6Ω - 2.28jΩ
Return Loss	-23.7dB
Antenna Parameters with Body TSL at 5800 MHz	
Impedance, transformed to feed point	56.7Ω - 8.10jΩ
Return Loss	-20.2dB
General Antenna Parameters and Design	
Electrical Delay (one direction)	1.313 ns
After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured. The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.	
Additional EUT Data	
Manufactured by	SPEAG

Certificate No: Z17-97133 Page 10 of 16

TTL In Collaboration with SPEAG CALIBRATION LABORATORY Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com http://www.china.ttl.cn	
DASY5 Validation Report for Head TSL	
Test Laboratory: TTL, Beijing, China DUT: Dipole 5GHz; Type: D5GHzV2; Serial: D5GHzV2 - SN: 1079 Communication System: CW; Frequency: 5200 MHz; Frequency: 5300 MHz; Frequency: 5500 MHz; Frequency: 5600 MHz; Frequency: 5800 MHz; Medium parameters used: f = 5200 MHz; α = 4.616 mho/m; εr = 35.72; ρ = 1000 kg/m3; Medium parameters used: f = 5300 MHz; α = 4.668 mho/m; εr = 36.09; ρ = 1000 kg/m3; Medium parameters used: f = 5500 MHz; α = 4.934 mho/m; εr = 35.92; ρ = 1000 kg/m3; Medium parameters used: f = 5600 MHz; α = 4.984 mho/m; εr = 35.73; ρ = 1000 kg/m3; Medium parameters used: f = 5800 MHz; α = 5.159 mho/m; εr = 35.83; ρ = 1000 kg/m3; Phantom section: Left Section Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007) DASY5 Configuration: • Probe: EX3DV4 - SN3846; ConvF(5.37.5.37.5.37); Calibrated: 1/13/2017, ConvF(4.72.4.72.4.72); Calibrated: 1/13/2017, ConvF(4.72.4.72.4.72); Calibrated: 1/13/2017, ConvF(4.95.4.95.4.95); Calibrated: 1/13/2017; • Sensor-Surface: 1.4mm (Mechanical Surface Detection) • Electronics: DAE4 Sn1331; Calibrated: 2017/1/19 • Phantom: Triple Flat Phantom 5.1C; Type: QD 000 P51 CA; Serial: 1161/3 • Measurement SW: DASY52, Version 52.10 (0); SEMCAD X Version 14.6.10 (7417)	
Dipole Calibration /Pin=100mW, d=10mm, f=5200 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 58.81 V/m; Power Drift = -0.01 dB Peak SAR (extrapolated) = 30.8 W/kg SAR(1 g) = 7.77 W/kg; SAR(10 g) = 2.24 W/kg Maximum value of SAR (measured) = 18.2 W/kg	
Dipole Calibration /Pin=100mW, d=10mm, f=5300 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 65.19 V/m; Power Drift = 0.05 dB Peak SAR (extrapolated) = 33.7 W/kg SAR(1 g) = 8.13 W/kg; SAR(10 g) = 2.32 W/kg Maximum value of SAR (measured) = 19.3 W/kg	

Certificate No: Z17-97133 Page 11 of 16

TTL In Collaboration with SPEAG CALIBRATION LABORATORY Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com http://www.china.ttl.cn	
Dipole Calibration /Pin=100mW, d=10mm, f=5500 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 57.80 V/m; Power Drift = 0.02 dB Peak SAR (extrapolated) = 34.3 W/kg SAR(1 g) = 8.24 W/kg; SAR(10 g) = 2.37 W/kg Maximum value of SAR (measured) = 19.6 W/kg	
Dipole Calibration /Pin=100mW, d=10mm, f=5600 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 57.89 V/m; Power Drift = 0.04 dB Peak SAR (extrapolated) = 35.7 W/kg SAR(1 g) = 8.16 W/kg; SAR(10 g) = 2.34 W/kg Maximum value of SAR (measured) = 20.0 W/kg	
Dipole Calibration /Pin=100mW, d=10mm, f=5800 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 53.56 V/m; Power Drift = -0.06 dB Peak SAR (extrapolated) = 35.0 W/kg SAR(1 g) = 7.85 W/kg; SAR(10 g) = 2.25 W/kg Maximum value of SAR (measured) = 19.7 W/kg	




Certificate No: Z17-97133 Page 12 of 16

D5GHzV2 Sn:1079 (4/4)



Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62304633-2079 Fax: +86-10-62304633-2504
E-mail: cnf@chinanet.com <http://www.chinanet.cn>

Impedance Measurement Plot for Head TSI



Certificate No: Z17-97133

Page 13 of 14



Add: No 51 Xueyun Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62704833-2079 Fax: +86-10-62704633-2504
E-mail: cnli@chinanet.com <http://www.chinanet.cn>

DASY5 Validation Report for Body TSL

Date: 09.25.2017

Test Laboratory: CTTL, Beijing, China

DUT: Dipole 5GHz; Type: D5GHzV2; Serial: D5GHzV2 - SN: 1079

Communication System: CW, Frequency: 5200 MHz, $\sigma = 5.382$ mho/m, $\epsilon_r = 49.47$, $\rho = 1000$ kg/m³, Frequency: 5500 MHz, Frequency: 5600 MHz, Frequency: 5800 MHz, $\sigma = 5.498$ mho/m, $\epsilon_r = 49.21$, $\rho = 1000$ kg/m³, Medium parameters used: $f = 5200$ MHz, $\sigma = 5.382$ mho/m, $\epsilon_r = 49.47$, $\rho = 1000$ kg/m³, Medium parameters used: $f = 5300$ MHz, $\sigma = 5.498$ mho/m, $\epsilon_r = 49.21$, $\rho = 1000$ kg/m³, Medium parameters used: $f = 5500$ MHz, $\sigma = 5.722$ mho/m, $\epsilon_r = 49.03$, $\rho = 1000$ kg/m³, Medium parameters used: $f = 5600$ MHz, $\sigma = 5.733$ mho/m, $\epsilon_r = 48.37$, $\rho = 1000$ kg/m³, Medium parameters used: $f = 5800$ MHz, $\sigma = 5.935$ mho/m, $\epsilon_r = 48.99$, $\rho = 1000$ kg/m³.

Phantom section: Center Section

- Measurement Standard: DASYS (IEEE/IEC/ANSI C63.19-2007)
- DASY5 Configuration:
- Probe: EX3DV4 - SN3846; Calibrated/F4.95, 95.45; Calibrated: 1/13/2017, Conf/F4.95, 95.45, 95.45; Calibrated: 1/13/2017, Conf/F4.18, 18.4, 18.4; Calibrated: 1/13/2017, Conf/F4.18, 18.4, 18.4; Calibrated: 1/13/2017, Conf/F4.53, 53.4, 53.4; Calibrated: 1/13/2017,
 - Sensor-Surface: 14mm (Mechanical Surface Detection)
 - Electronics: DASYS2, Serial 16163, Calibrated: 2/17/2018
 - Phantom: Triple Flat Phantom 5, IC, Type 150, QD 000 P51 CA, Serial: 116163
 - Measurement SW: DASY52, Version 52.10 (0); SEMCAD X, Version 14.6.10 (7417)

Dipole Calibration /Pin=100mW, d=10mm, f=5200 MHz/Zoom Scan,
dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm
Reference Value = 55.18 V/m; Power Drift = 0.01 dB
Peak SAR (extrapolated) = 30.0 W/kg
SAR(1g) = 7.52 W/kg; SAR(10g) = 2.12 W/kg
Maximum value of SAR (measured) = 18.2 W/kg

Dipole Calibration /Pin=100mW, d=10mm, f=5300 MHz/Zoom Scan,
dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm
Reference Value = 53.94 V/m; Power Drift = 0.01 dB
Peak SAR (extrapolated) = 31.9 W/kg
SAR(1 g) = 7.68 W/kg; SAR(10 g) = 2.18 W/kg
Maximum value of SAR (measured) = 18.3 W/kg

Certificate No. Z17-97133

Page 12 of 16




Add: No 51 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62304633-2079 Fax: +86-10-62304633-2504
E-mail: cti@chinanet.com <http://www.chinanet.cn>

Dipole Calibration (Pin=100mW, d=10mm, f=5500 MHz/Zoom Scan, dia=1.4mm (8x8x7)Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm
Reference Value = 66.70 V/m; Power Drift = -0.03 dB
Peak SAR (extrapolated) = 33.7 W/kg
SAR(1 g) = 8.22 W/kg; SAR(10 g) = 2.35 W/kg
Maximum value of SAR (measured) = 19.8 W/kg

Dipole Calibration /Pin=100mW, d=10mm, f=5600 MHz/Zoom Scan,
dist=1.4mm (8x8x7)Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm
Reference Value = 67.75 V/m; Power Drift = 0.01 dB
Peak SAR (extrapolated) = 34.2 W/kg
SAR(1 g) = 8.08 W/kg; SAR(10 g) = 2.3 W/kg
Maximum value of SAR (measured) = 19.3 W/kg

Dipole Calibration (Pin=100mW, d=10mm, f=5800 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm
Reference Value = 68.20 V/m: Power Drift = 0.05 dB
Peak SAR (extrapolated) = 33.3 W/kg
SAR(1 g) = 7.73 W/kg; SAR(10 g) = 2.17 W/kg
Maximum value of SAR (measured) = 18.3 W/kg




Certificate No. Z17-97131

Page 14 of 14



Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62304633-2079 Fax: +86-10-62304633-2704
E-mail: ctf@chinaetf.com <http://www.chinaetf.cn>

Impedance Measurement Plot for Body TSI



Certificate No: Z17-97131

Page 16 of 16