

Date: 18 May 2022

DECLARATION OF EQUIVALENCE

This document declares that the following designated products are equivalent to the unit under test **39428**.

For USA:

Model Name / Product Code	Marketing Name
39491	AIRCURVE 11 ASV USA
39492	AIRCURVE 11 S USA
39493	AIRCURVE 11 ST USA
39494	AIRCURVE 11 VAUTO USA

For Canada:

Model Name / Product Code	Marketing Name
39495	AIRCURVE 11 VAUTO CAN
39496	AIRCURVE 11 S CAN
39497	AIRCURVE 11 ST CAN
39498	AIRCURVE 11 ASV CAN

For USA & Canada:

Model Name / Product Code	Marketing Name
51400	AIRCURVE 11 ST NORTH AMERICA

All the above stated products have the same hardware, cellular firmware and Bluetooth firmware.

Different models are derived for marketing purposes. The differences between S, ST, VAuto and ASV models are the therapy modes and customer/region centric features supported by the device. These settings are configured at manufacturing via a configuration file and is independent of the cellular, Bluetooth and FG software.

In summary, the differences between the device models mentioned herein are:

- Therapy modes
- Data settings
- UI language
- Name plaque
- Data connection parameters.



Applicant:

Company Name: ResMed Pty Ltd
Address: 1 Elizabeth Macarthur Drive,
Bella Vista NSW 2153
Australia

By,

Christopher Jenkins

Christopher Jenkins

Title: Associate Manager – Systems Engineering
Company: ResMed Pty Ltd
Telephone: +61 2 8884 1517
e-mail: Christopher.jenkins@resmed.com.au