

Maximum Permissible Exposure Report

Product : AUGi Pendant
Model Name : IN6P001
FCC ID : 2BNIV-PND-1
Test Regulation : 47 CFR FCC Part 2.1093
Received Date : 2024/12/30
Test Date : 2025/1/3 ~ 2025/1/10
Issued Date : 2025/4/8
Applicant : All Inspire Health
19 Morris Avenue, Building 128, Brooklyn, NY 11205
Issued By : Underwriters Laboratories Taiwan Co., Ltd.
Building A, B and E, No. 372-7, Sec. 4, Zhongxing Rd.,
Zhudong Township, Hsinchu County, Taiwan



The results reported herein have been performed in accordance with the laboratory's terms of accreditation. This report shall not be reproduced except in full without the written approval of the Laboratory. The results in this report are responsible of the test sample(s) provided by the client only and are not to be used to indicate applicability to other similar products.


Underwriters Laboratories Taiwan Co., Ltd.

Building A, B and E, No. 372-7, Sec. 4, Zhongxing Rd., Zhudong Township, Hsinchu County, Taiwan

Telephone : +886-2-7737-3000

Facsimile (FAX) : +886-3-583-7948

Doc No: Form-ULID-004725 (DCS:17-EM-F0864) / 6.0



Original Test Report No.: 4791618243B-US-R1-V0

Doc No: Form-ULID-004725 (DCS:17-EM-F0864) / 6.0

Table of Contents

1. Attestation of Test Results.....	4
2. Test Methodology and Reference Procedures	5
3. Facilities and Accreditation	5
4. Equipment Under Test	6
4.1. Description of EUT	6
4.2. Description of Available Antennas	7
5. Requirement	8
6. Radio Frequency Radiation Exposure Evaluation	9

1. Attestation of Test Results

APPLICANT: All Inspire Health
19 Morris Avenue, Building 128, Brooklyn, NY 11205

MANUFACTURER: InnoComm Mobile Technology Corporation
3F, No. 6, Hsin Ann Rd., Hsinchu Science Park, Hsinchu 300092,
Taiwan

EUT DESCRIPTION: AUGi Pendant

BRAND: All Inspire Health

MODEL: IN6P001


SAMPLE STAGE: Engineering Verification Test Sample

APPLICABLE STANDARDS	
STANDARD	Test Results
47 CFR FCC Part 2.1093	PASS

Underwriters Laboratories Taiwan Co., Ltd. tested the above equipment in accordance with the requirements set forth in the above standards. All indications of Pass/Fail in this report are opinions expressed by Underwriters Laboratories Taiwan Co., Ltd. based on interpretations and/or observations of test results. The test results show that the equipment tested is capable of demonstrating compliance with the requirements as documented in this report.


Note: The results documented in this report apply only to the tested sample, under the conditions and modes of operation as described herein. This document may not be altered or revised in any way unless done so by Underwriters Laboratories Taiwan Co., Ltd. and all revisions are duly noted in the revisions section. Any alteration of this document not carried out by Underwriters Laboratories Taiwan Co., Ltd. will constitute fraud and shall nullify the document. This report must not be used by the client to claim product certification, approval, or endorsement by NVLAP, NIST, any agency of the Federal Government, or any agency of any government.

Prepared By:


Cindy Hsin
Project Handler

Date : 2025/4/8

Approved and Authorized By:


Kent Liu
Senior Laboratory Engineer

Date : 2025/4/8

Underwriters Laboratories Taiwan Co., Ltd.

Building A, B and E, No. 372-7, Sec. 4, Zhongxing Rd., Zhudong Township, Hsinchu County, Taiwan

Telephone : +886-2-7737-3000

Facsimile (FAX) : +886-3-583-7948

Doc No: Form-ULID-004725 (DCS:17-EM-F0864) / 6.0

2. Test Methodology and Reference Procedures

The tests documented in this report were performed in accordance with KDB 447498 D01 General RF Exposure Guidance v06.

3. Facilities and Accreditation

Test Location	Underwriters Laboratories Taiwan Co., Ltd.
Address	Building A, B and E, No. 372-7, Sec. 4, Zhongxing Rd., Zhudong Township, Hsinchu County, Taiwan
Accreditation Certificate	Underwriters Laboratories Taiwan Co., Ltd. is accredited by TAF, Laboratory Code 3398.

Underwriters Laboratories Taiwan Co., Ltd.

Building A, B and E, No. 372-7, Sec. 4, Zhongxing Rd., Zhudong Township, Hsinchu County, Taiwan

Telephone : +886-2-7737-3000

Facsimile (FAX) : +886-3-583-7948

Doc No: Form-ULID-004725 (DCS:17-EM-F0864) / 6.0

4. Equipment Under Test

4.1. Description of EUT

Product	AUGi Pendant
Brand Name	All Inspire Health
Model Name	IN6P001
Normal Voltage	3Vdc from Battery

Operating Frequency	BT LE: 2402MHz ~ 2480MHz
Sample ID	Conducted Test:7980208
	Radiated Test:7980209

Note:

1. For this report measurement uncertainty, statement of conformity, determining compliance, it is necessary to refer to the original measurement report of EUT.
2. The above EUT information is declared by manufacturer and for more detailed features description, please refer the manufacturer's or user's manual, the laboratory shall not be held responsible.

4.2. Description of Available Antennas

Ant. No.	Transmitter Circuit	Frequency Range	Brand Name	Model Name	Maximum Gain (dBi)	Ant. Type	Connector Type
BT	Chain0	2402MHz ~ 2480MHz	Innocomm	TAG	2.01	IFA	None

Note: The above antenna information was provided from customer and for more detailed features description, please refer the manufacturer's specification or user's manual, the laboratory shall not be held responsible.

5. Requirement

Following FCC KDB 447498 D01 “General SAR test exclusion guidance”

The corresponding SAR Exclusion Threshold condition, listed below:

- 1) The 1-g and 10-g SAR test exclusion thresholds for 100 MHz to 6 GHz at test separation distances ≤ 50 mm are determined by:

$[(\text{max. power of channel, including tune-up tolerance, mW})/(\text{min. test separation distance, mm})]$

$\cdot [\sqrt{f(\text{GHz})}] \leq 3.0$ for 1-g SAR and ≤ 7.5 for 10-g extremity SAR, where

- $f(\text{GHz})$ is the RF channel transmit frequency in GHz.
- Power and distance are rounded to the nearest mW and mm before calculation.
- The result is rounded to one decimal place for comparison. The test exclusions are applicable only when the minimum test separation distance is ≤ 50 mm and for transmission frequencies between 100 MHz and 6 GHz. When the minimum test separation distance is < 5 mm, a distance of 5 mm is applied to determine SAR test exclusion.

- 2) At 100 MHz to 6 GHz and for test separation distances > 50 mm, the SAR test exclusion threshold is determined according to the following:

- a) $[\text{Threshold at 50 mm in step 1}) + (\text{test separation distance} - 50\text{mm}) \cdot (f(\text{MHz})/150)]$ mW, at 100MHz to 1500 MHz
- b) $[\text{Threshold at 50 mm in step 1}) + (\text{test separation distance} - 50 \text{ mm}) \cdot 10]$ mW at > 1500 MHz and ≤ 6 GHz

- 3) At frequencies below 100 MHz, the following may be considered for SAR test exclusion.

- a) The threshold at the corresponding test separation distance at 100 MHz in step 2) is multiplied by $[1 + \log(100/f(\text{MHz}))]$ for test separation distances > 50 mm and < 200 mm.
- b) The threshold determined by the equation in a) for 50 mm and 100 MHz is multiplied by $\frac{1}{2}$ for test separation distances ≤ 50 mm.
- c) SAR measurement procedures are not established below 100 MHz. When SAR test exclusion cannot be applied, a KDB inquiry is required to determine SAR evaluation requirements for any test results to be acceptable.

6. Radio Frequency Radiation Exposure Evaluation

Operating Mode	Evaluation Frequency	Max. Average power EIRP	Antenna Gain	Min. test separation distance	SAR test exclusion calculation value	1-g SAR test exclusion thresholds	Result
	(MHz)	(mW)	(dBi)	(mm)			
BT LE	2402 ~ 2480	8.453	2.01	5	2.662	3	PASS

Note:

1. Calculate SAR test exclusion thresholds from section 5.1 formulas.

Conclusion:

Since maximum time-averaged power or effective radiated power (ERP) is below SAR exempt power thresholds, the SAR evaluation is not required.

END OF REPORT

Underwriters Laboratories Taiwan Co., Ltd.

Building A, B and E, No. 372-7, Sec. 4, Zhongxing Rd., Zhudong Township, Hsinchu County, Taiwan

Telephone : +886-2-7737-3000

Facsimile (FAX) : +886-3-583-7948

Doc No: Form-ULID-004725 (DCS:17-EM-F0864) / 6.0