

Report No. : TESA2207000226ES Rev: 01 Page: 1 of 13

Appendix C

Phantom Description

C/NII

d & Partner Engineering 40	S	p	е	a	g	_

Zeiighautstratze 43, 8004 Zurich, Switzerland Ptione +41 44 245 9700, Fan +41 44 245 9779 info@spead.com. Mitp //www.spead.com

Certificate of Conformity / First Article Inspection

Item	Oval Flat Phantom ELI 5.0
Type No	QD OVA 002 A
Series No	1108 and higher
Manufacturer	Untersee Composites Knebelstrasse 8, CH-6268 Mannenbach, Switzerland

Tests

Complete tests were made on the prototype units QD OVA 001 A, pre-series units QD OVA 001 B as well as on some series units QD OVA 001 B. Some tests are made on all series units QD OVA 002 A.

Test	Requirement	Details	Units tested
Shape	Internal dimensions, depth and sagging are compatible with standards.	Bottom elliptical 600 x 400 mm, Depth 190 mm, dimension compliant with [1] for t > 375 MHz	Prototypes
Material thickness	Botton: 2 Gnim +/- 0 2mm	dimension compliant with [3] for f > 800 MHz	al
Material parameters	rel. permittivity $2 = 5$, loss tangent ≤ 0.05 , at $f \leq 6$. GHz	rel. permittivity 3.5 +/- 0.5 loss tangent < 0.05	Material samples
Material resistivity	Compatibility with tissue simulating liquids .	Compatible with SPEAG liquids, **	Phantoms, Material sample
Sagging	Sagging of the flat section in folerance when filled with tissue simulating liquid.	within tolerance for filling height up to 155 mm	Prototypes, samples

Note: Compatibility restrictions apply certain liquid components mentioned in the standard, containing e.g. DGBE, DGMHE or Triton X-100. Observe technical note on material compatibility.

Standards


- OET Bulletin 65, Supplement C, "Evaluating Compliance with FCC Guidelines for Human Exposure to Radiofrequency Electromagnetic Fields", Edition 01-01
 IEEE 1528-2003, "Recommended Practice for Determining the Peak Spatial-Average Specific: Absorption Rate (SAR) in the Human Head from Wireless Communications Devices. Measurement
- Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques, December 2003
 [3] IEC 62209-1 ed1.0, "Human exposure to radio frequency fields from hand-heid and body-mounted wireless communication devices Human models, instrumentation, and procedures Part 1: Procedure to determine the specific absorption rate (SAR) for hand-heid devices used in close proximity to the ear (frequency range of 300 MHz to 3 GHz)", 2005-02-18
 [4] IEC 62209-2 ed1.0, "Human exposure to radio frequency fields from hand-heid and body-mounted wireless communication devices Human models, instrumentation, and procedures Part 2: Procedure to determine the specific devices of the ear (frequency fields from hand-heid and body-mounted wireless communication devices Human models, instrumentation, and procedures Part 2: Procedure to the ear (frequency fields from hand-heid and body-mounted wireless communication devices Human models, instrumentation, and procedures Part 2: Procedure to the ear (frequency fields from hand-heid excedure to the ear (frequency fields from h
- Procedure to determine the specific absorption rate (SAR) for wireless communication devices used in close proximity to the human body (frequency range of 30 MHz to 6 GHz)", 2010-03-30

Conformity

Based on the sample tests above, we certify that this item is in compliance with the uncertainty requirements of body-worn SAR measurements and system performance checks as specified in [1 - 4] and further standards

Date 25.7.2011

Signature / Stamp



Doc No SS1-QD OVA 002 A - A

1102 Page

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 90 days only.

除非另有說明,此報告結果僅對測試之樣品負責,同時此樣品僅保留90天。本報告未經本公司書面許可,不可部份複製。 This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at http://www.sgs.com.tw/Terms-and-Conditions and for electronic format documents, subject to Terms and Conditions for Electronic Documents at http://www.sgs.com.tw/Terms-and-Conditions. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law

No.134,Wu Kung Road, New Taipei Industrial Park, Wuku District, New Taipei City, Taiwan/新北市五股區新北產業園區五工路 134 號

www.sqs.com.tw



System Validation from Original Equipment Supplier

Client SGS		Certificate No: Z2	1-60368
CALIBRATION CI	ERTIFICAT	Е	
Object	D835V	2 - SN: 4d063	
Calibration Procedure(s)	FF-711	-003-01	
		tion Procedures for dipole validation kits	
Calibration date:	Octobe	r 18, 2021	
pages and are part of the ce		he placed laboratory facility and	
All calibrations have been humidity<70%.	conducted in t	the closed laboratory facility: environment b or calibration)	emperature (22±3)°C an
All calibrations have been humidity<70%. Calibration Equipment used	conducted in t		emperature (22±3)°C an Scheduled Calibration
All calibrations have been numidity<70%. Calibration Equipment used Primary Standards Power Meter NRP2	conducted in t	or calibration) Cal Date (Calibrated by, Certificate No.) 24-Sep-21 (CTTL, No.J21X08326)	
All calibrations have been humidity<70%. Calibration Equipment used Primary Standards Power Meter NRP2 Power sensor NRP8S	Conducted in t (M&TE critical fr ID # 106277 104291	or calibration) Cal Date (Calibrated by, Certificate No.) 24-Sep-21 (CTTL, No.J21X08326) 24-Sep-21 (CTTL, No.J21X08326)	Scheduled Calibration Sep-22 Sep-22
All calibrations have been numidity<70%. Calibration Equipment used Primary Standards Power Meter NRP2 Power sensor NRP8S Reference Probe EX3DV4	conducted in t (M&TE critical fo ID # 106277 104291 SN 7517	or calibration) Cal Date (Calibrated by, Certificate No.) 24-Sep-21 (CTTL, No.J21X08326) 24-Sep-21 (CTTL, No.J21X08326) 03-Feb-21 (CTTL, SPEAG, No.Z21-60001)	Scheduled Calibration Sep-22 Sep-22 Feb-22 Feb-22
All calibrations have been numidity<70%. Calibration Equipment used Primary Standards Power Meter NRP2 Power sensor NRP8S	Conducted in t (M&TE critical fr ID # 106277 104291	or calibration) Cal Date (Calibrated by, Certificate No.) 24-Sep-21 (CTTL, No.J21X08326) 24-Sep-21 (CTTL, No.J21X08326)	Scheduled Calibration Sep-22 Sep-22
All calibrations have been humidity<70%. Calibration Equipment used Primary Standards Power Meter NRP2 Power sensor NRP8S Reference Probe EX3DV4	conducted in t (M&TE critical fo ID # 106277 104291 SN 7517	or calibration) Cal Date (Calibrated by, Certificate No.) 24-Sep-21 (CTTL, No.J21X08326) 24-Sep-21 (CTTL, No.J21X08326) 03-Feb-21 (CTTL, SPEAG, No.Z21-60001)	Scheduled Calibration Sep-22 Sep-22 Feb-22 Feb-22
All calibrations have been humidity<70%. Calibration Equipment used Primary Standards Power Meter NRP2 Power sensor NRP8S Reference Probe EX3DV4 DAE4	conducted in t (M&TE critical fo ID # 106277 104291 SN 7517 SN 1556	Cal Date (Calibrated by, Certificate No.) 24-Sep-21 (CTTL, No.J21X08326) 24-Sep-21 (CTTL, No.J21X08326) 03-Feb-21 (CTTL-SPEAG,No.Z21-60001) 15-Jan-21 (SPEAG,No.DAE4-1556_Jan21)	Scheduled Calibration Sep-22 Sep-22 Feb-22 Jan-22
All calibrations have been humidity<70%. Calibration Equipment used Primary Standards Power Meter NRP2 Power sensor NRP8S Reference Probe EX3DV4 DAE4 Secondary Standards	conducted in t (M&TE critical fr 10# 106277 104291 SN 7517 SN 1556 ID #	Cal Date (Calibrated by, Certificate No.) 24-Sep-21 (CTTL, No.J21X08326) 24-Sep-21 (CTTL, No.J21X08326) 03-Feb-21 (CTTL-SPEAG,No.Z21-60001) 15-Jan-21 (SPEAG,No.DAE4-1556_Jan21) Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration Sep-22 Sep-22 Feb-22 Jan-22 Scheduled Calibration
All calibrations have been humidity<70%. Calibration Equipment used Primary Standards Power Meter NRP2 Power sensor NRP8S Reference Probe EX3DV4 DAE4 Secondary Standards Signal Generator E4438C	conducted in t (M&TE critical fo 106277 104291 SN 7517 SN 1556 ID # MY49071430	Cal Date (Calibrated by, Certificate No.) 24-Sep-21 (CTTL, No.J21X08326) 24-Sep-21 (CTTL, No.J21X08326) 03-Feb-21 (CTTL, SPEAG,No.Z21-60001) 15-Jan-21 (SPEAG,No.DAE4-1556_Jan21) Cal Date (Calibrated by, Certificate No.) 01-Feb-21 (CTTL, No.J21X00593)	Scheduled Calibration Sep-22 Sep-22 Feb-22 Jan-22 Scheduled Calibration Jan-22
All calibrations have been humidity<70%. Calibration Equipment used Primary Standards Power Meter NRP2 Power sensor NRP8S Reference Probe EX3DV4 DAE4 Secondary Standards Signal Generator E4438C NetworkAnalyzer E5071C	conducted in t (M&TE critical fd 106277 104291 SN 7517 SN 1556 ID # MY49071430 MY46110673	or calibration) Cal Date (Calibrated by, Certificate No.) 24-Sep-21 (CTTL, No.J21X08326) 24-Sep-21 (CTTL, No.J21X08326) 03-Feb-21 (CTTL, SPEAG,No.Z21-60001) 15-Jan-21 (SPEAG,No.DAE4-1556_Jan21) Cal Date (Calibrated by, Certificate No.) 01-Feb-21 (CTTL, No.J21X00593) 14-Jan-21 (CTTL, No.J21X00232)	Scheduled Calibration Sep-22 Sep-22 Feb-22 Jan-22 Scheduled Calibration Jan-22 Jan-22
All calibrations have been humidity<70%. Calibration Equipment used Primary Standards Power Meter NRP2 Power sensor NRP8S Reference Probe EX3DV4 DAE4 Secondary Standards Signal Generator E4438C	conducted in t (M&TE critical fd 106277 104291 SN 7517 SN 1556 ID # MY49071430 MY46110673 Name	Calibration) Cal Date (Calibrated by, Certificate No.) 24-Sep-21 (CTTL, No.J21X08326) 24-Sep-21 (CTTL, No.J21X08326) 03-Feb-21 (CTTL, SPEAG,No.Z21-60001) 15-Jan-21 (SPEAG,No.DAE4-1556_Jan21) Cal Date (Calibrated by, Certificate No.) 01-Feb-21 (CTTL, No.J21X00593) 14-Jan-21 (CTTL, No.J21X00232) Function	Scheduled Calibration Sep-22 Sep-22 Feb-22 Jan-22 Scheduled Calibration Jan-22 Jan-22

Certificate No: Z21-60368 Page 1 of 6

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 90 days only.

t (886-2) 2299-3279

除非另有說明,此報告結果僅對測試之樣品負責,同時此樣品僅保留90天。本報告未經本公司書面許可,不可部份複製。 This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at http://www.sgs.com.tw/Terms-and-Conditions and for electronic format documents subject to Terms and Conditions for Electronic Documents at <u>http://www.sgs.com.tw/Terms-and-Conditions</u>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

No.134,Wu Kung Road, New Taipei Industrial Park, Wuku District, New Taipei City, Taiwan/新北市五股區新北產業園區五工路 134 號

f (886-2) 2298-0488

```
www.sgs.com.tw
```





No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, Chins 86-10-62304633-2079 Fax: +86-10-62304633-2504 http://www.chinatil.com

Glossary: TSL

ConvF N/A

tissue simulating liquid sensitivity in TSL / NORMx,y,z not applicable or not measured

- Calibration is Performed According to the Following Standards: a) IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013 b) IEC 62209-1, "Measurement procedure for assessment of specific absorption rate of human exposure to radio frequency fields from hand-held and body-mounted wireless
- communication devices Part 1: Device used next to the ear (Frequency range of 300MHz to 6GHz)", July 2016 c) IEC 62209-2, "Procedure to measure the Specific Absorption Rate (SAR) For wireless
- communication devices used in close proximity to the human body (frequency range of 30MHz to 6GHz)", March 2010 d) KDB865664, SAR Measurement Requirements for 100 MHz to 6 GHz

Additional Documentation: e) DASY4/5 System Handbook

- Methods Applied and Interpretation of Parameters: Measurement Conditions: Further details are available from the Validation Report at the end
 - of the certificate. All figures stated in the certificate are valid at the frequency indicated. Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented
- parallel to the body axis. Feed Point Impedance and Return Loss: These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low
- reflected power. No uncertainty required. Electrical Delay: One-way delay between the SMA connector and the antenna feed point. No uncertainty required. SAR measured: SAR measured at the stated antenna input power. SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna
- connector.
- SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result. .

The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution Corresponds to a coverage probability of approximately 95%.

Certificate No: Z21-60368

Page 2 of 6

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 90 days only.

t (886-2) 2299-3279

除非另有說明,此報告結果僅對測試之樣品負責,同時此樣品僅保留90天。本報告未經本公司書面許可,不可部份複製。 This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at http://www.sgs.com.tw/Terms-and-Conditions and for electronic format documents, subject to Terms and Conditions for Electronic Documents at http://www.sgs.com.tw/Terms-and-Conditions. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law

No.134,Wu Kung Road, New Taipei Industrial Park, Wuku District, New Taipei City, Taiwan/新北市五股區新北產業園區五工路 134 號

f (886-2) 2298-0488





YuanBei Road, Haidian District, Beijing, 100191, China 04633-2079 Fax: +86-10-62304633-2504 inattl.com http://www.chinattl.cn

Measurement Conditions

DASY Version	DASY52	V52.10.4
Extrapolation	Advanced Extrapolation	
Phantom	Triple Flat Phantom 5.1C	
Distance Dipole Center - TSL	15 mm	with Spacer
Zoom Scan Resolution	dx, dy, dz = 5 mm	
Frequency	835 MHz ± 1 MHz	

Head TSL parameters

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	41.5	0.90 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	40.9 ± 6 %	0.89 mho/m ± 6 %
Head TSL temperature change during test	<1.0 °C		

SAR result with Head TSL

SAR averaged over 1 cm3 (1 g) of Head TSL	Condition	
SAR measured	250 mW input power	2.39 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	9.64 W/kg ± 18.8 % (k=2)
SAR averaged over 10 cm^3 (10 g) of Head TSL	Condition	
SAR measured	250 mW input power	1.56 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	6.29 W/kg ± 18.7 % (k=2)

Certificate No: Z21-60368

Page 3 of 6

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 90 days only.

t (886-2) 2299-3279

Unless other was baled une solute of the state and the state report refer only to the sample(s) lessed and state and provide the sample(s) lessed and state prosecuted to the fullest extent of the law.

No.134,Wu Kung Road, New Taipei Industrial Park, Wuku District, New Taipei City, Taiwan/新北市五股區新北產業園區五工路 134 號

f (886-2) 2298-0488

www.sgs.com.tw





laidian District, Beijing, 100191, China Fax: +86-10-62304633-2504 http://www.chinattl.en

Appendix (Additional assessments outside the scope of CNAS L0570)

Antenna Parameters with Head TSL

Impedance, transformed to feed point	51.1Ω- 3.03jΩ
Return Loss	- 29.9dB

Electrical Delay (one direction) 1.299 ns

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

Additional EUT Data

Manufactured by	SPEAG
-----------------	-------

Certificate No: Z21-60368

Page 4 of 6

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 90 days only.

除非另有說明,此報告結果僅對測試之樣品負責,同時此樣品僅保留90天。本報告未經本公司書面許可,不可部份複製。 This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at http://www.sgs.com.tw/Terms-and-Conditions and for electronic format documents, subject to Terms and Conditions for Electronic Documents at http://www.sgs.com.tw/Terms-and-Conditions. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law

No.134,Wu Kung Road, New Taipei Industrial Park, Wuku District, New Taipei City, Taiwan/新北市五股區新北產業園區五工路 134 號

f (886-2) 2298-0488

www.sgs.com.tw





d: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China :+86-10-62304633-2079 Fax: +86-10-62304633-2504 nail: ettl@chinatil.com http://www.chinatil.cn

Date: 10.18.2021

DASY5 Validation Report for Head TSL Test Laboratory: CTTL, Beijing, China DUT: Dipole 835 MHz; Type: D835V2; Serial: D835V2 - SN: 4d063


 $\begin{array}{l} \mbox{Communication System: UID 0, CW; Frequency: 835 MHz; Duty Cycle: 1:1 \\ \mbox{Medium parameters used: } f = 835 MHz; \sigma = 0.886 \ S/m; \epsilon_r = 40.9; \rho = 1000 \ kg/m^3 \end{array}$

Phantom section: Right Section DASY5 Configuration:

- Probe: EX3DV4 SN7517; ConvF(9.81, 9.81, 9.81) @ 835 MHz; Calibrated:
- 2021-02-03 .
- Sensor-Surface: 1.4mm (Mechanical Surface Detection) Electronics: DAE4 Sn1556; Calibrated: 2021-01-15
- . .
- Phantom: MFP_V5.1C (20deg probe tilt); Type: QD 000 P51 Cx; Serial: 1062 Measurement SW: DASY52, Version 52.10 (4); SEMCAD X Version 14.6.14 (7501)

Dipole Calibration/Zoom Scan (7x7x7) (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm Reference Value = 59.22 V/m; Power Drift = -0.01 dB

Peak SAR (extrapolated) = 3.68 W/kg SAR(1 g) = 2.39 W/kg; SAR(10 g) = 1.56 W/kg Smallest distance from peaks to all points 3 dB below = 20.6 mm Ratio of SAR at M2 to SAR at M1 = 64.8% Maximum value of SAR (measured) = 3.24 W/kg



0 dB = 3.24 W/kg = 5.11 dBW/kg

Certificate No: Z21-60368

Page 5 of 6

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 90 days only.

除非另有說明,此報告結果僅對測試之樣品負責,同時此樣品僅保留90天。本報告未經本公司書面許可,不可部份複製。 This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at http://www.sgs.com.tw/Terms-and-Conditions and for electronic format documents, subject to Terms and Conditions for Electronic Documents at http://www.sgs.com.tw/Terms-and-Conditions. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law

No.134,Wu Kung Road, New Taipei Industrial Park, Wuku District, New Taipei City, Taiwan/新北市五股區新北產業園區五工路 134 號


```
www.sgs.com.tw
```





laidian District, Beijing, 100191, China Fax: +86-10-62304633-2504 http://www.chinattl.cn : No.52 HuaYuanBet Ko +86-10-62304633-2079 ail: cttl@chinattl.com

Impedance Measurement Plot for Head TSL



Certificate No: Z21-60368

Page 6 of 6

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 90 days only.

除非另有說明,此報告結果僅對測試之樣品負責,同時此樣品僅保留90天。本報告未經本公司書面許可,不可部份複製。 This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at http://www.sgs.com.tw/Terms-and-Conditions and for electronic format documents, subject to Terms and Conditions for Electronic Documents at http://www.sgs.com.tw/Terms-and-Conditions. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law

No.134,Wu Kung Road, New Taipei Industrial Park, Wuku District, New Taipei City, Taiwan/新北市五股區新北產業園區五工路 134 號

f (886-2) 2298-0488

www.sgs.com.tw



Accredited by the Swiss Accreditati			ccreditation No.: SCS 0108
he Swiss Accreditation Service Iultilateral Agreement for the rea			
			D045010 707 4 00
Client SGS-TW (Auder	1)	Certificate N	o: D2450V2-727_Apr22
CALIBRATION C	ERTIFICATE		
Object	D2450V2 - SN:7	27	
Calibration procedure(s)	QA CAL-05.v11		
	Calibration Proce	edure for SAR Validation Sources	s between 0.7-3 GHz
Calibration date:	April 25, 2022		
All calibrations have been conduct	ed in the closed laborato	robability are given on the following pages a ry facility: environment temperature $(22 \pm 3)^{\circ}$	
All calibrations have been conduct Calibration Equipment used (M&TE	ed in the closed laborato		
All calibrations have been conduct Calibration Equipment used (M&T) Primary Standards	ed in the closed laborato E critical for calibration)	ry facility: environment temperature (22 ± 3)* Cal Date (Certificate No.)	C and humidity < 70%. Scheduled Calibration
All calibrations have been conduct Calibration Equipment used (M&T/ Primary Standards Power meter NBP	ed in the closed laborato E critical for calibration) ID # SN: 104778	ry facility: environment temperature (22 ± 3)* Cal Date (Certificate No.) 04-Apr-22 (No. 217-03525/03524)	C and humidity < 70%. Scheduled Calibration Apr:23
All calibrations have been conduct Calibration Equipment used (M&Tt Primary Standards Power meter NRP Power sensor NRP-Z91	ed in the closed laborato critical for calibration) ID # SN: 104778 SN: 103244	ry facility: environment temperature (22 ± 3)* Cal Date (Certificate No.) 04-Apr-22 (No. 217-03524) 04-Apr-22 (No. 217-03524)	C and humidity < 70%. Scheduled Calibration Apr-23 Apr-23
All calibrations have been conduct Calibration Equipment used (M&TI Primary Standards Power rensor NRP-291 Power sensor NRP-291 Power sensor NRP-291	ed in the closed laborato E critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245	ry facility: environment temperature (22 ± 3)* Cal Date (Certificate No.) 04-Apr-22 (No. 217-03524) 04-Apr-22 (No. 217-03524) 04-Apr-22 (No. 217-03525)	C and humidity < 70%. Scheduled Calibration Apr-23 Apr-23 Apr-23
All calibrations have been conduct Calibration Equipment used (M&TI Primary Standards Power raters NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator	ed in the closed laborato critical for calibration) ID # SN: 104778 SN: 103245 SN: 103245 SN: BH9394 (20k)	y facility: environment temperature (22 ± 3)" <u>Cal Date (Certificata No.)</u> 04-Apr-22 (No. 217-03524) 04-Apr-22 (No. 217-03524) 04-Apr-22 (No. 217-03525) 04-Apr-27 (No. 217-03527)	C and humidity < 70%. Scheduled Calibration Apr-23 Apr-23 Apr-23 Apr-23
All calibrations have been conduct Calibration Equipment used (M&TI Primary Standards Power sensor NRP-291 Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Type-N mismatch combination	ed in the closed laborato E critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245	ry facility: environment temperature (22 ± 3)* Cal Date (Certificate No.) 04-Apr22 (No. 217-03525/03524) 04-Apr-22 (No. 217-03525) 04-Apr-22 (No. 217-03525) 04-Apr-22 (No. 217-03527) 04-Apr-22 (No. 217-03528)	C and humidity < 70%. Scheduled Calibration Apr-23 Apr-23 Apr-23 Apr-23 Apr-23
All calibrations have been conduct Calibration Equipment used (M&TI Primary Standards Power mater KNP Power sansor NRP-291 Power sansor NRP-291 Reference 2:0 dB Attenuator Type-N mismatch combination Reference Probe EXSIDV4	ad in the closed laborato critical for calibration) ID # SN: 104778 SN: 103245 SN: 103245 SN: 819394 (20k) SN: 310827 06327	y facility: environment temperature (22 ± 3)" <u>Cal Date (Certificata No.)</u> 04-Apr-22 (No. 217-03524) 04-Apr-22 (No. 217-03524) 04-Apr-22 (No. 217-03525) 04-Apr-27 (No. 217-03527)	C and humidity < 70%. Scheduled Calibration Apr-23 Apr-23 Apr-23 Apr-23
All calibrations have been conduct Calibration Equipment used (M&TI Primary Standards Power meler NRP Power sansor NRP-Z91 Power sansor NRP-Z91 Reference 20 dB Attenuator Type-1 mismatch combination Reference Probe EX3DV4 DAE4	ad in the closed laborato E critical for calibration) ID # SN: 104778 SN: 104778 SN: 103244 SN: 103245 SN: 310982 / 06327 SN: 7549 SN: 601	ry facility: environment temperature (22 ± 3)* Cal Date (Certificata No.) 04-Apr-22 (No. 217-03525/03524) 04-Apr-22 (No. 217-03525) 04-Apr-22 (No. 217-03525) 04-Apr-22 (No. 217-03527) 04-Apr-22 (No. 217-03528) 31-Dec-21 (No. 217-03528) 31-Dec-21 (No. 217-03528)	C and humidity < 70%. Scheduled Calibration Apr-23 Apr-23 Apr-23 Apr-23 Apr-23 Dec-22 Nov-22
All calibrations have been conduct Calibration Equipment used (M&TI Primary Standards Power ansor NRP-291 Power sansor NRP-291 Power sansor NRP-291 Power sansor NRP-294 Reference 20 dB Attenuator Reference 20 dB Attenuator Reference Probe EX3DV4 DAE4 Secondary Standards	ed in the closed laborato critical for calibration) ID # SN: 103245 SN: 103245 SN: 103245 SN: 19394 (20k) SN: 310982 / 06327 SN: 310982 / 06327 SN: 310982 / 06327 SN: 3098 / 061	ry facility: environment temperature (22 ± 3)* Cal Date (Certificata No.) 04-Apr-22 (No. 217-0352503524) 04-Apr-22 (No. 217-035253) 04-Apr-22 (No. 217-03527) 04-Apr-22 (No. 217-03527) 04-Apr-22 (No. 217-03528) 31-Dec 21 (No. EX3-7340_Dec ² 1) 01-Nov-21 (No. DAE4-601_Nov21) Check Date (in house)	C and humidity < 70%. Scheduled Calibration Apr-23 Apr-23 Apr-23 Apr-23 Apr-23 Dec-22 Nov-22 Scheduled Check
All calibrations have been conduct Calibration Equipment used (M&TI Primary Standards Power smersor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Type-1 mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter E4419B	ad in the closed laborato E critical for calibration) ID # SN: 104778 SN: 104778 SN: 103244 SN: 103245 SN: 310982 / 06327 SN: 7549 SN: 601	Cal Date (Certificate No.) 04-Apr22 (No. 217-03525/03524) 04-Apr22 (No. 217-03525) 04-Apr22 (No. 217-03526) 04-Apr22 (No. 217-03527) 04-Apr22 (No. 217-03527) 04-Apr22 (No. 217-03527) 04-Apr22 (No. 217-03527) 04-Apr22 (No. 217-03528) 31-0ec>1 (No. EX3-7349_Dec>1) 01-Nov21 (No. DAF4-60_Nov21) Check Date (in house) 30-Qc1-4 (in souse check Qc1-20)	C and humidity < 70%. Scheduled Calibration Apr-23 Apr-23 Apr-23 Apr-23 Dec-22 Nov-22 Scheduled Check In house check: Oct:22
All calibrations have been conduct Calibration Equipment used (M&TI Primary Standards Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter E4198 Power sensor 1P 8481A	ed in the closed laborato E critical for calibration) ID # SN: 103244 SN: 103245 SN: 303245 SN: 3103245 SN: 3103245 SN: 3103245 SN: 3103245 SN: 303245 SN: 30325 SN: 30325	y facility: environment temperature (22 ± 3)* Cal Date (Certificata No.) 04-Apr-22 (No. 217-0352503524) 04-Apr-22 (No. 217-03525) 04-Apr-22 (No. 217-03527) 04-Apr-22 (No. 217-03527) 04-Apr-22 (No. 217-03528) 31-Doc 21 (No. EX3-7349_Dec21) 01-Nov-21 (No. DAE4-601_Nov21) Check Date (in house) 30-Qc1+14 (in house check Qc1-20) 07-Qc1+15 (in house check Qc1-20)	C and humidity < 70%. Scheduled Calibration Apr/23 Apr/23 Apr/23 Apr/23 Apr/23 Do-22 Nov-22 Scheduled Check In house check: Oct/22 In house check: Oct/22
All calibrations have been conduct Calibration Equipment used (M&TI Primary Standards Power sensor NNP-291 Power sensor NNP-291 Power sensor NNP-291 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power motor E44198 Power sensor HP 8481A Power sensor HP 8481A	ed in the closed laborato critical for calibration) ID # SN: 103245 SN: 103245 SN: 103245 SN: 103245 SN: 310982 / 06327 SN: 310982 / 06327 SN: 601 ID # SN: 0830512475 SN: US37292783	Cal Date (Certificate No.) 04-Apr22 (No. 217-03525/03524) 04-Apr22 (No. 217-03525) 04-Apr22 (No. 217-03526) 04-Apr22 (No. 217-03527) 04-Apr22 (No. 217-03527) 04-Apr22 (No. 217-03527) 04-Apr22 (No. 217-03527) 04-Apr22 (No. 217-03528) 31-0ec>1 (No. EX3-7349_Dec>1) 01-Nov21 (No. DAF4-60_Nov21) Check Date (in house) 30-Qc1-4 (in souse check Qc1-20)	C and humidity < 70%. Scheduled Calibration Apr/23 Apr/23 Apr/23 Apr/23 Doc-22 Nov-22 Scheduled Check In house check: Oct-22 In house check: Oct-22
All calibrations have been conduct Calibration Equipment used (M&TI Primary Standards Power meter NBP Power sensor NBP-291 Reference 2:0 dB Attenuator Type-N mismatch combination Reference PATURE Reference PATURE Reference PATURE Reference PATURE Reference PATURE Power sensor HP 8481A Power sensor HP 8481A Power sensor HP 8481A Power sensor HP 8481A	ed in the closed laborato critical for calibration) ID # SN: 103245 SN: 103245 SN: 103245 SN: 103245 SN: 103245 SN: 103245 SN: 103245 SN: 31082 / 06327 SN: 31082 / 06327 SN: 0537292783 SN: WA1093315	y facility: environment temperature (22 ± 3)* Cal Date (Certificate No.) 04-Apr-22 (No. 217-03525) 04-Apr-22 (No. 217-03524) 04-Apr-22 (No. 217-03525) 04-Apr-22 (No. 217-03527) 04-Apr-22 (No. 217-03527)	C and humidity < 70%. Scheduled Calibration Apr/23 Apr/23 Apr/23 Apr/23 Apr/23 Do-22 Nov-22 Scheduled Check In house check: Oct/22 In house check: Oct/22
All calibrations have been conduct Calibration Equipment used (M&TI Primary Standards Power mater KNP Power ansero NRP-291 Power sensor NRP-291 Reference 2:00 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary EA4198 Power sensor HP 8481A Power sensor HP 8481A Power sensor HP 8481A Power sensor HP 8481A	ad in the closed laborato E critical for calibration) ID # SN: 103244 SN: 103245 SN: 303245 SN: 310982 / 06327 SN: 310982 / 06327 SN: 310982 / 06327 SN: 310982 / 06327 SN: US37292783 SN: W74103315 SN: US41080477	Cal Date (Certificate No.) 04-Apr.22 (No. 217-03525/03524) 04-Apr.22 (No. 217-03525) 04-Apr.22 (No. 217-03523) 04-Apr.22 (No. 217-03523) 04-Apr.22 (No. 217-03523) 04-Apr.22 (No. 217-03523) 04-Apr.22 (No. 217-03528) 01-Apr.22 (No. 217-03528) 01-Apr.22 (No. 217-03528) 01-Apr.22 (No. 217-03528) 01-Nov-21 (No. DAE4-601_Nov21) Check Date (in house check Oct-20) 07-Oct-15 (in house check Oct-20) 07-Oct-15 (in house check Oct-20) 15-Jun-15 (in house check Oct-20) 31-Mar-14 (in house check Oct-20)	C and humidity < 70%. Scheduled Calibration Apr-23 Apr-23 Apr-23 Apr-23 Apr-23 Dec-22 Nov-22 Scheduled Check In house check: Oct-22 In house check: Oct-22
All calibrations have been conduct Calibration Equipment used (M&TI Primary Standards Power sensor NRP-291 Power sensor NRP-291 Reference 20 d8 Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power sensor HP 8481A Power sensor HP 8481A	ed in the closed laborato critical for calibration) ID # SN: 103244 SN: 103244 SN: 103245 SN: 103245 SN: 103245 SN: 310882 / 06327 SN: 310882 / 06327 SN: 3039512475 SN: US37282783 SN: W74109315 SN: US37282783 SN: W74109315 SN: US41080477 Name	ry facility: environment temperature (22 ± 3)* Cal Date (Certificata No.) 04-Apr-22 (No. 217-035203524) 04-Apr-22 (No. 217-03525) 04-Apr-22 (No. 217-03527) 04-Apr-22 (No. 217-03527) 04-Apr-22 (No. 217-03527) 04-Apr-22 (No. 217-03528) 31-Doc-21 (No. EX3-7340 Dec?1) 01-Nov-21 (No. EX3-7340 Dec?1) 01-Nov-21 (No. EX3-7340 Dec?1) 07-021-5 (in house check Oct-20) 07-021-5 (in house check Oct-20) 07-021-5 (in house check Oct-20) 07-021-5 (in house check Oct-20) 31-Mar-14 (in house check Oct-20) 31-Mar-14 (in house check Oct-20) Function	C and humidity < 70%. Scheduled Calibration Apr/23 Apr/23 Apr/23 Apr/23 Apr/23 Apr/23 De-22 Nov-22 Scheduled Check In house check: Oct-22 In house check: Oct-22 In house check: Oct-22 In house check: Oct-22 In house check: Oct-22
All calibrations have been conduct Calibration Equipment used (M&TI Primary Standards Power sensor NRP-291 Power sensor NRP-291 Reference 20 d8 Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power sensor HP 8481A Power sensor HP 8481A	ad in the closed laborato E critical for calibration) ID # SN: 103244 SN: 103245 SN: 303245 SN: 310982 / 06327 SN: 310982 / 06327 SN: 310982 / 06327 SN: 310982 / 06327 SN: US37292783 SN: W74103315 SN: US41080477	Cal Date (Certificate No.) 04-Apr.22 (No. 217-03525/03524) 04-Apr.22 (No. 217-03525) 04-Apr.22 (No. 217-03523) 04-Apr.22 (No. 217-03523) 04-Apr.22 (No. 217-03523) 04-Apr.22 (No. 217-03523) 04-Apr.22 (No. 217-03528) 01-Apr.22 (No. 217-03528) 01-Apr.22 (No. 217-03528) 01-Apr.22 (No. 217-03528) 01-Nov-21 (No. DAE4-601_Nov21) Check Date (in house check Oct-20) 07-Oct-15 (in house check Oct-20) 07-Oct-15 (in house check Oct-20) 15-Jun-15 (in house check Oct-20) 31-Mar-14 (in house check Oct-20)	C and humidity < 70%. Scheduled Calibration Apr-23 Apr-23 Apr-23 Apr-23 Apr-23 Dec-22 Nov-22 Scheduled Check In house check: Oct-22 In house check: Oct-22
All calibrations have been conduct Calibration Equipment used (M&TI Primary Standards Power meter NBP Power sensor NRP-291 Power sensor NRP-291 Power sensor NRP-291 Reference 20 d8 Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power sensor HP 8491A Power sensor HP 8491A Power sensor HP 8491A RF generator R&S SMI-06 Network Analyzer Agilent EB358A Calibrated by:	ed in the closed laborato critical for calibration) ID # SN: 103245 SN: 103245 SN: 103245 SN: 103245 SN: 103245 SN: 103247 SN: 310982 / 06327 SN: 310982 / 06327 SN: 0839512475 SN: 0839512475 SN: 1037228783 SN: WY41093315 SN: 103972 SN: US41080477 Name Michael Weber	ry facility: environment temperature (22 ± 3)* Cal Date (Certificata No.) 04-Apr-22 (No. 217-0352/3) 04-Apr-22	C and humidity < 70%. Scheduled Calibration Apr-23 Apr-23 Apr-23 Apr-23 Apr-23 Dec-22 Nov-22 Scheduled Check In house check: Oct-22 In house check: Oct-22
All calibrations have been conduct Calibration Equipment used (M&TI Primary Standards Power smoor NRP-291 Power sensor NRP-291 Power sensor NRP-293 Reference 20 d8 Attenuator Type-P misimatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter E44198 Power sensor IP 8481A Power sensor IP 8481A RF generator R&S SMI-06 Network Analyzer Agilent E8358A Calibrated by:	ed in the closed laborato critical for calibration) ID # SN: 103244 SN: 103244 SN: 103245 SN: 103245 SN: 103245 SN: 310882 / 06327 SN: 310882 / 06327 SN: 3039512475 SN: US37282783 SN: W74109315 SN: US37282783 SN: W74109315 SN: US41080477 Name	ry facility: environment temperature (22 ± 3)* Cal Date (Certificata No.) 04-Apr-22 (No. 217-035203524) 04-Apr-22 (No. 217-03525) 04-Apr-22 (No. 217-03527) 04-Apr-22 (No. 217-03527) 04-Apr-22 (No. 217-03527) 04-Apr-22 (No. 217-03528) 31-Doc-21 (No. EX3-7340 Dec?1) 01-Nov-21 (No. EX3-7340 Dec?1) 01-Nov-21 (No. EX3-7340 Dec?1) 07-021-5 (in house check Oct-20) 07-021-5 (in house check Oct-20) 07-021-5 (in house check Oct-20) 07-021-5 (in house check Oct-20) 31-Mar-14 (in house check Oct-20) 31-Mar-14 (in house check Oct-20) Function	C and humidity < 70%. Scheduled Calibration Apr-23 Apr-23 Apr-23 Apr-23 Apr-23 Dec-22 Nov-22 Scheduled Check In house check: Oct-22 In house check: Oct-22
All calibrations have been conduct Calibration Equipment used (M&TI Primary Standards Power meter NBP Power sensor NRP-291 Power sensor NRP-291 Power sensor NRP-291 Reference 20 d8 Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power sensor HP 8491A Power sensor HP 8491A Power sensor HP 8491A RF generator R&S SMI-06 Network Analyzer Agilent EB358A Calibrated by:	ed in the closed laborato critical for calibration) ID # SN: 103245 SN: 103245 SN: 103245 SN: 103245 SN: 103245 SN: 103247 SN: 310982 / 06327 SN: 310982 / 06327 SN: 0839512475 SN: 0839512475 SN: 1037228783 SN: WY41093315 SN: 103972 SN: US41080477 Name Michael Weber	ry facility: environment temperature (22 ± 3)* Cal Date (Certificata No.) 04-Apr-22 (No. 217-0352/3) 04-Apr-22	C and humidity < 70%. Scheduled Calibration Apr-23 Apr-23 Apr-23 Apr-23 Apr-23 Dec-22 Nov-22 Scheduled Check In house check: Oct-22 In house check: Oct-22
	ed in the closed laborato critical for calibration) ID # SN: 103245 SN: 103245 SN: 103245 SN: 103245 SN: 103245 SN: 103247 SN: 310982 / 06327 SN: 310982 / 06327 SN: 0839512475 SN: 0839512475 SN: 1037228783 SN: WY41093315 SN: 103972 SN: US41080477 Name Michael Weber	ry facility: environment temperature (22 ± 3)* Cal Date (Certificata No.) 04-Apr-22 (No. 217-0352/3) 04-Apr-22	C and humidity < 70%. Scheduled Calibration Apr-23 Apr-23 Apr-23 Apr-23 Dec-22 Nov-22 Scheduled Check In house check: Oct-22 In house check: Oct-22

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 90 days only.


t (886-2) 2299-3279

Unless other was baled une solute of the state and the state report refer only to the sample(s) lessed and state and provide the sample(s) lessed and state prosecuted to the fullest extent of the law.

No.134,Wu Kung Road, New Taipei Industrial Park, Wuku District, New Taipei City, Taiwan/新北市五股區新北產業園區五工路 134 號

f (886-2) 2298-0488





- SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna
- connector.
- SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result. .

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

Certificate No: D2450V2-727_Apr22

Page 2 of 6

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 90 days only.

t (886-2) 2299-3279

除非另有說明,此報告結果僅對測試之樣品負責,同時此樣品僅保留90天。本報告未經本公司書面許可,不可部份複製。 This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at http://www.sgs.com.tw/Terms-and-Conditions and for electronic format documents, subject to Terms and Conditions for Electronic Documents at http://www.sgs.com.tw/Terms-and-Conditions. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law

No.134,Wu Kung Road, New Taipei Industrial Park, Wuku District, New Taipei City, Taiwan/新北市五股區新北產業園區五工路 134 號

f (886-2) 2298-0488

www.sqs.com.tw



Report No. : TESA2207000226ES Rev: 01 Page: 10 of 13

Measurement Conditions

DASY Version	DASY52	V52.10.4
Extrapolation	Advanced Extrapolation	
Phantom	Modular Flat Phantom	
Distance Dipole Center - TSL	10 mm	with Spacer
Zoom Scan Resolution	dx, dy, dz = 5 mm	
Frequency	2450 MHz ± 1 MHz	

Head TSL parameters The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	39.2	1.80 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	37.8 ± 6 %	1.87 mho/m ± 6 %
Head TSL temperature change during test	< 0.5 °C		

SAR result with Head TSL

SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	
SAR measured	250 mW input power	13.6 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	52.8 W/kg ± 17.0 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Head TSL	condition	
SAR averaged over 10 cm ³ (10 g) of Head TSL. SAR measured	condition 250 mW input power	6.34 W/kg

Certificate No: D2450V2-727 Apr22

Page 3 of 6

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 90 days only.

Unless otherwise stated the results shown in this test report reter only to the sample(s) lested and such sample(s) are retained to 90 days only. Mir#JSfabity i, Lt&Bateger and the company subject to its General Conditions of Service printed overleaf, available on request or accessible at http://www.sgs.com.tw/Terms-and-Conditions and for electronic format documents, subject to Terms and Conditions for Electronic Documents at http://www.sgs.com.tw/Terms-and-Conditions and for electronic format defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fulleest extent of the law. prosecuted to the fullest extent of the law.

No.134,Wu Kung Road, New Taipei Industrial Park, Wuku District, New Taipei City, Taiwan/新北市五股區新北產業園區五工路 134 號

f (886-2) 2298-0488

www.sgs.com.tw



Report No. : TESA2207000226ES Rev: 01 Page: 11 of 13

Appendix (Additional assessments outside the scope of SCS 0108)

Antenna Parameters with Head TSL

Impedance, transformed to feed point	55.7 Ω + 3.0 jΩ	
Return Loss	- 24.4 dB	

General Antenna Parameters and Design

Electrical Delay (one direction) 1.149 ns

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. according to the Standard. No excessive force must b sive force must be applied to the dipole arms, because they might bend or the soldered connections near the

feedpoint may be damaged.

Additional EUT Data

Manufactured by	SPEAG

Certificate No: D2450V2-727 Apr22

Page 4 of 6

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 90 days only.

除非另有說明,此報告結果僅對測試之樣品負責,同時此樣品僅保留90天。本報告未經本公司書面許可,不可部份複製。 This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at http://www.sgs.com.tw/Terms-and-Conditions and for electronic format documents, subject to Terms and Conditions for Electronic Documents at http://www.sgs.com.tw/Terms-and-Conditions. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law

No.134,Wu Kung Road, New Taipei Industrial Park, Wuku District, New Taipei City, Taiwan/新北市五股區新北產業園區五工路 134 號

f (886-2) 2298-0488

www.sgs.com.tw



DASY5 Validation Report for Head TSL

Test Laboratory: SPEAG, Zurich, Switzerland

Date: 25.04.2022

DUT: Dipole 2450 MHz; Type: D2450V2; Serial: D2450V2 - SN: 727 Communication System: UID 0 - CW; Frequency: 2450 MHz Medium parameters used: f = 2450 MHz; σ = 1.87 S/m; ϵ_r = 37.8; ρ = 1000 kg/m³ Phantom section: Flat Section


Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2011)

DASY52 Configuration:

- Probe: EX3DV4 SN7349; ConvF(7.96, 7.96, 7.96) @ 2450 MHz; Calibrated: 31.12.2021
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 01.11.2021
- Phantom: Flat Phantom 5.0 (front); Type: QD 000 P50 AA; Serial: 1001
- DASY52 52.10.4(1535); SEMCAD X 14.6.14(7501)

Dipole Calibration for Head Tissue/Pin=250 mW, d=10mm/Zoom Scan (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm Reference Value = 115.6 V/m; Power Drift = 0.09 dB

Reference value = 11.5.0 v/m, rower D/m = 0.09 uB Peak SAR (extrapolated) = 26.5 W/kg SAR(1 g) = 13.6 W/kg; SAR(10 g) = 6.34 W/kg Smallest distance from peaks to all points 3 dB below = 9 mm Ratio of SAR at M2 to SAR at M1 = 51% Maximum value of SAR (measured) = 22.1 W/kg



Certificate No: D2450V2-727 Apr22

Page 5 of 6

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 90 days only.

t (886-2) 2299-3279


除非另有說明,此報告結果僅對測試之樣品負責,同時此樣品僅保留90天。本報告未經本公司書面許可,不可部份複製。 This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at http://www.sgs.com.tw/Terms-and-Conditions and for electronic format documents, subject to Terms and Conditions for Electronic Documents at http://www.sgs.com.tw/Terms-and-Conditions. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law

No.134,Wu Kung Road, New Taipei Industrial Park, Wuku District, New Taipei City, Taiwan/新北市五股區新北產業園區五工路 134 號

f (886-2) 2298-0488



Impedance Measurement Plot for Head TSI



Certificate No: D2450V2-727_Apr22 Page 6 of 6

- End of report -

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 90 days only.

除非另有說明,此報告結果僅對測試之樣品負責,同時此樣品僅保留90天。本報告未經本公司書面許可,不可部份複製。 This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at http://www.sgs.com.tw/Terms-and-Conditions and for electronic format documents, subject to Terms and Conditions for Electronic Documents at http://www.sgs.com.tw/Terms-and-Conditions. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

No.134,Wu Kung Road, New Taipei Industrial Park, Wuku District, New Taipei City, Taiwan/新北市五股區新北產業園區五工路 134 號

f (886-2) 2298-0488

www.sgs.com.tw