

## Appendix (Additional assessments outside the scope of SCS 0108)

#### Antenna Parameters with Head TSL

Impedance, transformed to feed point	54.1 Ω + 3.9 jΩ		
Return Loss	- 25.2 dB		

#### Antenna Parameters with Body TSL

Impedance, transformed to feed point	48.5 Ω + 6.0 jΩ	
Return Loss	- 24.1 dB	

### General Antenna Parameters and Design

Electrical Delay (one direction)	1.161 ns
----------------------------------	----------

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard.

No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

#### **Additional EUT Data**

Manufactured by	SPEAG
Manufactured on	November 10, 2009

Certificate No: D2450V2-853\_Jul18



### **DASY5 Validation Report for Head TSL**

Date: 24.07.2018

Test Laboratory: SPEAG, Zurich, Switzerland

### DUT: Dipole 2450 MHz; Type: D2450V2; Serial: D2450V2 - SN:853

Communication System: UID 0 - CW; Frequency: 2450 MHz Medium parameters used: f = 2450 MHz;  $\sigma$  = 1.85 S/m;  $\epsilon_r$  = 37.8;  $\rho$  = 1000 kg/m<sup>3</sup> Phantom section: Flat Section Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2011)

DASY52 Configuration:

- Probe: EX3DV4 SN7349; ConvF(7.88, 7.88, 7.88) @ 2450 MHz; Calibrated: 30.12.2017
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 26.10.2017
- Phantom: Flat Phantom 5.0 (front); Type: QD 000 P50 AA; Serial: 1001
- DASY52 52.10.1(1476); SEMCAD X 14.6.11(7439)

Dipole Calibration for Head Tissue/Pin=250 mW, d=10mm/Zoom Scan (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm Reference Value = 115.3 V/m; Power Drift = -0.00 dB Peak SAR (extrapolated) = 26.1 W/kg SAR(1 g) = 13.2 W/kg; SAR(10 g) = 6.13 W/kg Maximum value of SAR (measured) = 21.6 W/kg



0 dB = 21.6 W/kg = 13.34 dBW/kg

Certificate No: D2450V2-853\_Jul18

Page 5 of 8



## Impedance Measurement Plot for Head TSL



Certificate No: D2450V2-853\_Jul18

Page 6 of 8



### **DASY5 Validation Report for Body TSL**

Date: 16.07.2018

Test Laboratory: SPEAG, Zurich, Switzerland

## DUT: Dipole 2450 MHz; Type: D2450V2; Serial: D2450V2 - SN:853

Communication System: UID 0 - CW; Frequency: 2450 MHz Medium parameters used: f = 2450 MHz;  $\sigma$  = 2.02 S/m;  $\epsilon_r$  = 51.9;  $\rho$  = 1000 kg/m<sup>3</sup> Phantom section: Flat Section Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2011)

DASY52 Configuration:

- Probe: EX3DV4 SN7349; ConvF(8.01, 8.01, 8.01) @ 2450 MHz; Calibrated: 30.12.2017
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 26.10.2017
- Phantom: Flat Phantom 5.0 (back); Type: QD 000 P50 AA; Serial: 1002
- DASY52 52.10.1(1476); SEMCAD X 14.6.11(7439)

Dipole Calibration for Body Tissue/Pin=250 mW, d=10mm/Zoom Scan (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm Reference Value = 108.0 V/m; Power Drift = -0.03 dB Peak SAR (extrapolated) = 25.6 W/kg SAR(1 g) = 13.1 W/kg; SAR(10 g) = 6.1 W/kg Maximum value of SAR (measured) = 21.0 W/kg



0 dB = 21.0 W/kg = 13.22 dBW/kg

Certificate No: D2450V2-853\_Jul18



### Impedance Measurement Plot for Body TSL



Certificate No: D2450V2-853\_Jul18

Page 8 of 8



### 2600 MHz Dipole Calibration Certificate

Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland



Schweizerischer Kalibrierdienst Service suisse d'étalonnage Servizio svizzero di taratura Swiss Calibration Service

Accreditation No.: SCS 0108

S

С

S

Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

		Certificate No:	: D2600V2-1012_Jul18
CALIBRATION C	ERTIFICATE		
Dbject	D2600V2 - SN:10	012	
Calibration procedure(s)	QA CAL-05.v10		
	Calibration proce	dure for dipole validation kits abo	ve 700 MHz
Calibration date:	July 26, 2018		
Ill calibrations have been conduct Calibration Equipment used (M&T)	ed in the closed laborator E critical for calibration)	ry facility: environment temperature (22 $\pm$ 3)°C	C and humidity < 70%.
Primary Standards	ID #	Cal Date (Certificate No.)	Scheduled Calibration
Power meter NRP	SN: 104778	04-Apr-18 (No. 217-02672/02673)	Apr-19
ower sensor NRP-Z91	SN: 103244	04-Apr-18 (No. 217-02672)	Apr-19
ower sensor NRP-Z91	SN: 103245	04-Apr-18 (No. 217-02673)	Apr-19
Reference 20 dB Attenuator	SN: 5058 (20k)	04-Apr-18 (No. 217-02682)	Apr-19
ype-N mismatch combination	SN: 5047.2 / 06327	04-Apr-18 (No. 217-02683)	Apr-19
Reference Probe EX3DV4	SN: 7349	30-Dec-17 (No. EX3-7349_Dec17)	Dec-18
DAE4	SN: 601	26-Oct-17 (No. DAE4-601_Oct17)	Oct-18
	ID #	Check Date (in house)	
Secondary Standards			Scheduled Check
Secondary Standards Power meter EPM-442A	SN: GB37480704	07-Oct-15 (in house check Oct-16)	Scheduled Check In house check: Oct-18
Secondary Standards Power meter EPM-442A Power sensor HP 8481A	SN: GB37480704 SN: US37292783	07-Oct-15 (in house check Oct-16) 07-Oct-15 (in house check Oct-16)	Scheduled Check In house check: Oct-18 In house check: Oct-18
Secondary Standards Power meter EPM-442A Power sensor HP 8481A Power sensor HP 8481A	SN: GB37480704 SN: US37292783 SN: MY41092317	07-Oct-15 (in house check Oct-16) 07-Oct-15 (in house check Oct-16) 07-Oct-15 (in house check Oct-16)	Scheduled Check In house check: Oct-18 In house check: Oct-18 In house check: Oct-18
Secondary Standards Power meter EPM-442A Power sensor HP 8481A Power sensor HP 8481A RF generator R&S SMT-06	SN: GB37480704 SN: US37292783 SN: MY41092317 SN: 100972	07-Oct-15 (in house check Oct-16) 07-Oct-15 (in house check Oct-16) 07-Oct-15 (in house check Oct-16) 15-Jun-15 (in house check Oct-16)	Scheduled Check In house check: Oct-18 In house check: Oct-18 In house check: Oct-18 In house check: Oct-18
Secondary Standards Power meter EPM-442A Power sensor HP 8481A Power sensor HP 8481A RF generator R&S SMT-06 Network Analyzer Agilent E8358A	SN: GB37480704 SN: US37292783 SN: MY41092317 SN: 100972 SN: US41080477	07-Oct-15 (in house check Oct-16) 07-Oct-15 (in house check Oct-16) 07-Oct-15 (in house check Oct-16) 15-Jun-15 (in house check Oct-16) 31-Mar-14 (in house check Oct-17)	Scheduled Check In house check: Oct-18 In house check: Oct-18 In house check: Oct-18 In house check: Oct-18 In house check: Oct-18
Secondary Standards Power meter EPM-442A Power sensor HP 8481A Power sensor HP 8481A RF generator R&S SMT-06 Network Analyzer Agilent E8358A	SN: GB37480704 SN: US37292783 SN: MY41092317 SN: 100972 SN: US41080477 Name	07-Oct-15 (in house check Oct-16) 07-Oct-15 (in house check Oct-16) 07-Oct-15 (in house check Oct-16) 15-Jun-15 (in house check Oct-16) 31-Mar-14 (in house check Oct-17) Function	Scheduled Check In house check: Oct-18 In house check: Oct-18 In house check: Oct-18 In house check: Oct-18 In house check: Oct-18 Signature
Secondary Standards Power meter EPM-442A Power sensor HP 8481A Power sensor HP 8481A RF generator R&S SMT-06 Network Analyzer Agilent E8358A Calibrated by:	SN: GB37480704 SN: US37292783 SN: MY41092317 SN: 100972 SN: US41080477 Name Michael Weber	07-Oct-15 (in house check Oct-16) 07-Oct-15 (in house check Oct-16) 07-Oct-15 (in house check Oct-16) 15-Jun-15 (in house check Oct-16) 31-Mar-14 (in house check Oct-17) Function Laboratory Technician	Scheduled Check In house check: Oct-18 In house check: Oct-18 In house check: Oct-18 In house check: Oct-18 In house check: Oct-18 Signature
Secondary Standards Power meter EPM-442A Power sensor HP 8481A Power sensor HP 8481A RF generator R&S SMT-06 Network Analyzer Agilent E8358A Calibrated by:	SN: GB37480704 SN: US37292783 SN: MY41092317 SN: 100972 SN: US41080477 Name Michael Weber Katja Pokovic	07-Oct-15 (in house check Oct-16) 07-Oct-15 (in house check Oct-16) 07-Oct-15 (in house check Oct-16) 15-Jun-15 (in house check Oct-16) 31-Mar-14 (in house check Oct-17) Function Laboratory Technician Technical Manager	Scheduled Check In house check: Oct-18 Signature

Certificate No: D2600V2-1012\_Jul18

Page 1 of 8



## No.I19Z61624-SEM03 Page 170 of 217

#### Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland



Schweizerischer Kalibrierdienst Service suisse d'étalonnage С Servizio svizzero di taratura Swiss Calibration Service

S

S

Accreditation No.: SCS 0108

Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

#### Glossary:

TSL	tissue simulating liquid
ConvF	sensitivity in TSL / NORM x,y,z
N/A	not applicable or not measured

### Calibration is Performed According to the Following Standards:

- a) IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013
- b) IEC 62209-1, "Measurement procedure for the assessment of Specific Absorption Rate (SAR) from hand-held and body-mounted devices used next to the ear (frequency range of 300 MHz to 6 GHz)", July 2016
- c) IEC 62209-2, "Procedure to determine the Specific Absorption Rate (SAR) for wireless communication devices used in close proximity to the human body (frequency range of 30 MHz to 6 GHz)", March 2010
- d) KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

#### **Additional Documentation:**

e) DASY4/5 System Handbook

### Methods Applied and Interpretation of Parameters:

- Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss: These parameters are measured with the dipole • positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay: One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured: SAR measured at the stated antenna input power.
- SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

Certificate No: D2600V2-1012\_Jul18

Page 2 of 8



#### **Measurement Conditions**

DASY system configuration, as far as not given on page 1.

DASY Version	DASY5	V52.10.1
Extrapolation	Advanced Extrapolation	
Phantom	Modular Flat Phantom	54 N
Distance Dipole Center - TSL	10 mm	with Spacer
Zoom Scan Resolution	dx, dy, dz = 5 mm	
Frequency	2600 MHz ± 1 MHz	

### **Head TSL parameters**

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	39.0	1.96 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	37.2 ± 6 %	2.02 mho/m ± 6 %
Head TSL temperature change during test	< 0.5 °C		

### SAR result with Head TSL

SAR averaged over 1 cm <sup>3</sup> (1 g) of Head TSL	Condition	
SAR measured	250 mW input power	14.2 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	55.4 W/kg ± 17.0 % (k=2)
SAR averaged over 10 cm <sup>3</sup> (10 g) of Head TSL	condition	
SAR averaged over 10 cm <sup>3</sup> (10 g) of Head TSL SAR measured	condition 250 mW input power	6.33 W/kg

### **Body TSL parameters**

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Body TSL parameters	22.0 °C	52.5	2.16 mho/m
Measured Body TSL parameters	(22.0 ± 0.2) °C	51.5 ± 6 %	2.20 mho/m ± 6 %
Body TSL temperature change during test	< 0.5 °C		

### SAR result with Body TSL

SAR averaged over 1 cm <sup>3</sup> (1 g) of Body TSL	Condition	
SAR measured	250 mW input power	13.7 W/kg
SAR for nominal Body TSL parameters	normalized to 1W	54.1 W/kg ± 17.0 % (k=2)
a second s	and a stand of the	the same intervent a needed and the term
SAR averaged over 10 cm <sup>3</sup> (10 g) of Body TSL	condition	
SAR averaged over 10 cm <sup>3</sup> (10 g) of Body TSL SAR measured	condition 250 mW input power	6.17 W/kg

Certificate No: D2600V2-1012\_Jul18



### Appendix (Additional assessments outside the scope of SCS 0108)

#### Antenna Parameters with Head TSL

Impedance, transformed to feed point	47.4 Ω - 7.4 jΩ
Return Loss	- 21.9 dB

#### Antenna Parameters with Body TSL

Impedance, transformed to feed point	44.1 Ω - 4.9 jΩ
Return Loss	- 21.8 dB

#### General Antenna Parameters and Design

Electrical Delay (one direction)	1.154 ns

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard.

No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

#### **Additional EUT Data**

Manufactured by	SPEAG
Manufactured on	October 30, 2007

Page 4 of 8



### **DASY5** Validation Report for Head TSL

Date: 26.07.2018

Test Laboratory: SPEAG, Zurich, Switzerland

#### DUT: Dipole 2600 MHz; Type: D2600V2; Serial: D2600V2 - SN:1012

Communication System: UID 0 - CW; Frequency: 2600 MHz Medium parameters used: f = 2600 MHz;  $\sigma$  = 2.02 S/m;  $\epsilon_r$  = 37.2;  $\rho$  = 1000 kg/m<sup>3</sup> Phantom section: Flat Section Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2011)

DASY52 Configuration:

- Probe: EX3DV4 SN7349; ConvF(7.7, 7.7, 7.7) @ 2600 MHz; Calibrated: 30.12.2017
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 26.10.2017
- Phantom: Flat Phantom 5.0 (front); Type: QD 000 P50 AA; Serial: 1001
- DASY52 52.10.1(1476); SEMCAD X 14.6.11(7439)

Dipole Calibration for Head Tissue/Pin=250 mW, d=10mm/Zoom Scan (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm Reference Value = 118.3 V/m; Power Drift = -0.04 dB Peak SAR (extrapolated) = 28.3 W/kg SAR(1 g) = 14.2 W/kg; SAR(10 g) = 6.33 W/kg Maximum value of SAR (measured) = 23.7 W/kg



0 dB = 23.7 W/kg = 13.75 dBW/kg

Certificate No: D2600V2-1012\_Jul18

Page 5 of 8



### Impedance Measurement Plot for Head TSL



Certificate No: D2600V2-1012\_Jul18

Page 6 of 8



### **DASY5 Validation Report for Body TSL**

Date: 26.07.2018

Test Laboratory: SPEAG, Zurich, Switzerland

#### DUT: Dipole 2600 MHz; Type: D2600V2; Serial: D2600V2 - SN:1012

Communication System: UID 0 - CW; Frequency: 2600 MHz Medium parameters used: f = 2600 MHz;  $\sigma$  = 2.2 S/m;  $\epsilon_r$  = 51.5;  $\rho$  = 1000 kg/m<sup>3</sup> Phantom section: Flat Section Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2011)

DASY52 Configuration:

- Probe: EX3DV4 SN7349; ConvF(7.81, 7.81, 7.81) @ 2600 MHz; Calibrated: 30.12.2017
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 26.10.2017
- Phantom: Flat Phantom 5.0 (back); Type: QD 000 P50 AA; Serial: 1002
- DASY52 52.10.1(1476); SEMCAD X 14.6.11(7439)

#### **Dipole Calibration for Body Tissue/Pin=250 mW, d=10mm/Zoom Scan (7x7x7)/Cube 0:** Measurement grid: dx=5mm, dy=5mm, dz=5mm Reference Value = 107.5 V/m; Power Drift = -0.07 dB Peak SAR (extrapolated) = 27.7 W/kg

Peak SAR (extrapolated) = 27.7 W/kgSAR(1 g) = 13.7 W/kg; SAR(10 g) = 6.17 W/kgMaximum value of SAR (measured) = 22.6 W/kg



0 dB = 22.6 W/kg = 13.54 dBW/kg



### Impedance Measurement Plot for Body TSL



Certificate No: D2600V2-1012\_Jul18

Page 8 of 8





# ANNEX I Sensor Triggering Data Summary

Antenna	Trigger Position	Trigger Distance(mm)
	Rear	16
l# Main Antenna	Bottom	16
Main Antenna	Front	11

Note: According to customer requirements, Front edge is tested with normal power of 10mm.



## No.I19Z61624-SEM03 Page 178 of 217

According to the above description, this device was tested by the manufacturer to determine the SAR sensor triggering distances for the rear and bottom edge of the device. The measured power state within  $\pm$ 5mm of the triggering points (or until touching the phantom) is included for rear and each applicable edge.

To ensure all production units are compliant it is necessary to test SAR at a distance 1mm less than the smallest distance from the device and SAR phantom with the device at maximum output power without power reduction.

We tested the power and got the different proximity sensor triggering distances for rear and bottom edge. But the manufacturer has declared 16mm is the most conservative triggering distance for main antenna. So base on the most conservative triggering distance of 16mm, additional SAR measurements were required at 15mm from the highest SAR position between rear and bottom edge of main antenna.

### Rear

Moving device toward the phantom:

The power state											
Distance [mm]	21	20	19	18	17	16	15	14	13	12	11
Main antenna	Normal	Normal	Normal	Normal	Normal	Low	Low	Low	Low	Low	Low

Moving device away from the phantom:

The power state											
Distance [mm]	11	12	13	14	15	16	17	18	19	20	21
Main antenna	Low	Low	Low	Low	Low	Low	Normal	Normal	Normal	Normal	Normal

## **Bottom Edge**

Moving device toward the phantom:

The power state											
Distance [mm]	21	20	19	18	17	16	15	14	13	12	11
Main antenna	Low										

Moving device away from the phantom:

The power state											
Distance [mm]	11	12	13	14	15	16	17	18	19	20	21
Main antenna	Low										

## No.I19Z61624-SEM03 Page 179 of 217



The influence of table tilt angles to proximity sensor triggering is determined by positioning each edge that contains a transmitting antenna, perpendicular to the flat phantom, at the smallest sensor triggering test distance by rotating the device around the edge next to the phantom in  $\leq 10^{\circ}$  increments until the tablet is  $\pm 45^{\circ}$  or more from the vertical position at 0°.



## The bottom edge evaluation for main antenna

Based on the above evaluation, we come to the conclusion that the sensor triggering is not released and normal maximum output power is not restored within the  $\pm 45^{\circ}$  range at the smallest sensor triggering test distance declared by manufacturer.



# ANNEX J Extended Calibration SAR Dipole

Referring to KDB865664 D01, if dipoles are verified in return loss ( <-20dBm, within 20% of prior calibration), and in impedance ( within 5 ohm of prior calibration), the annual calibration is not necessary and the calibration interval can be extended.

Justification of Extended Calibration SAR Dipole D835V2- serial no.4d069

Head										
Date of Measurement	Return-Loss (dB)	Delta (%)	Real Impedance (ohm)	Delta (ohm)	Imaginary Impedance (johm)	Delta (johm)				
2018-7-23	-33.0		50.8		-2.1					
2019-7-21	-30.3	8.1	53.0	2.2	-1.0	1.1				

Body										
Date of Measurement	Return-Loss (dB)	Delta (%)	Real Impedance (ohm)	Delta (ohm)	Imaginary Impedance (johm)	Delta (johm)				
2018-7-23	-23.4		46.1		-5.2					
2019-7-21	-25.5	9.0	48.5	2.4	-5.0	0.2				

Justification of Extended Calibration SAR Dipole D1900V2– serial no.5d101

Head											
Date of Measurement	Return-Loss (dB)	Delta (%)	Real Impedance (ohm)	Delta (ohm)	Imaginary Impedance (johm)	Delta (johm)					
2018-7-24	-25.6		50.7		5.3						
2019-7-22	-22.9	10.5	50.6	-0.1	7.2	1.9					

Body						
Date of Measurement	Return-Loss (dB)	Delta (%)	Real Impedance (ohm)	Delta (ohm)	Imaginary Impedance (johm)	Delta (johm)
2017-7-26	-21.2		44.9		6.5	
2018-7-24	-21.4	0.9	46.4	1.5	7.4	0.9



### Justification of Extended Calibration SAR Dipole D2450V2– serial no.853

Head						
Date of Measurement	Return-Loss (dB)	Delta (%)	Real Impedance (ohm)	Delta (ohm)	Imaginary Impedance (johm)	Delta (johm)
2018-7-24	-25.2		54.1		3.9	
2019-7-22	-23.1	9.5	53.6	-0.5	6.3	2.4

			( - )		V - /	
2018-7-24	-25.2		54.1		3.9	
2019-7-22	-23.1	9.5	53.6	-0.5	6.3	2.4
Body						
			Real		Imaginary	

Boay						
Date of Measurement	Return-Loss (dB)	Delta (%)	Real Impedance (ohm)	Delta (ohm)	Imaginary Impedance (johm)	Delta (johm)
2018-7-24	-24.1		48.5		6.0	
2019-7-22	-22.0	8.7	50.4	1.9	8.0	2

### Justification of Extended Calibration SAR Dipole D2600V2- serial no.1012

Head						
Date of Measurement	Return-Loss (dB)	Delta (%)	Real Impedance (ohm)	Delta (ohm)	Imaginary Impedance (johm)	Delta (johm)
2018-7-26	-21.9		47.4		-7.4	
2019-7-24	-23.7	8.2	47.7	0.3	-5.9	1.5

Body						
Date of Measurement	Return-Loss (dB)	Delta (%)	Real Impedance (ohm)	Delta (ohm)	Imaginary Impedance (johm)	Delta (johm)
2018-7-26	-21.8		44.1		-4.9	
2019-7-24	-22.4	2.7	44.4	0.3	-4.5	0.4



# ANNEX K SPOT CHECK

## K.1 Dielectric Performance and System Validation

### Table K.1-1: Dielectric Performance of Head Tissue Simulating Liquid

Measurement Date (yyyy-mm-dd)	Туре	Frequency	Permittivity ٤	Drift (%)	Conductivity σ (S/m)	Drift (%)
2019/9/23	Head	2450 MHz	38.54	-1.68	1.777	-1.28
	Body	2450 MHz	52.61	-0.17	1.957	0.36

### Table K.1-2: System Validation of Head

Measurement		Target value (W/kg)		Measured	value(W/kg)	Deviation		
Date	Frequency	10 g 1 g 10 g 1 g 10 g		1 g				
(yyyy-mm-dd)		Average	Average	Average	Average	Average	Average	
2019/9/23	2450 MHz	24.2	51.6	24.6	51.72	1.65%	0.23%	

### Table K.1-3: System Validation of Body

Measurement		Target value (W/kg)		Measured	value(W/kg)	Deviation		
Date	Frequency	10 g	1 g	10 g	1 g	10 g	1 g	
(yyyy-mm-dd)		Average	Average	Average	Average	Average	Average	
2019/9/23	2450 MHz	24.5	52.3	24.84	52.72	1.39%	0.80%	

## K.2 Conducted power of selected case

### Table K.2-1: The conducted Power for WLAN-Normal Power

Mode / data rate	Channel	Measured Power (dBm)
802.11b – 1Mbps	1	17.91

### Table K.2-2: The conducted Power for WLAN-Low Power

Mode / data rate	Channel	Measured Power (dBm)
802.11b – 1Mbps	1	14.82

## K.3 SAR results for Main antenna

Frequency Band	Channel Number	Frequency (MHz)	Tune up (dBm)	EUT Measured Power (dBm)	Test setup	Measured SAR 10g (W/kg)	Measured SAR 1g (W/kg)	Calculated SAR 10g (W/kg)	Calculated SAR 1g (W/kg)	Power Drift	
WLAN	1	2412	16.00	14.82	Left Tilt	0.160	0.407	0.21	0.53	0.03	
WLAN	1	2412	19.00	17.91	Top 10mm	0.113	0.263	0.15	0.34	0.16	



## K.4 Reported SAR Comparison

		Highest	Highest	
Exposure	Technology	Reported SAR	Reported SAR	Equipment
Configuration	Band	1g(W/kg)	1g(W/kg)	Class
		original	spot check	
	GSM 850	0.26	\	
	PCS 1900	0.14	\	
	UMTS FDD 2	0.14	١	
	UMTS FDD 5	0.29	\	
Hood	CDMA BC0	0.31	١	PCE
пеац	LTE Band 2	0.28	\	
	LTE Band 5	0.21	١	
	LTE Band 7	0.24	١	
	LTE Band 41	0.07	١	
	WLAN 2.4	1.20	0.53	DTS
	GSM 850	0.54	١	
	PCS 1900	0.46	١	
	UMTS FDD 2	0.39	١	
	UMTS FDD 5	0.32	١	
Hotepot	CDMA BC0	0.40	١	PCE
riotspot	LTE Band 2	0.40	١	
	LTE Band 5	0.24	١	
	LTE Band 7	0.20	١	
	LTE Band 41	0.18	\	
	WLAN 2.4	0.74	0.34	DTS

Note: The spot check results marked blue are larger than the original result.



## No.I19Z61624-SEM03 Page 184 of 217

## **K.5 MAIN TEST INSTRUMENTS**

	-				-
No.	Name	Туре	Serial Number	Calibration Date	Valid Period
01	Network analyzer	E5071C	MY46110673	January 24, 2019	One year
02	Power meter	NRVD	102083	Ostak az 04. 0040	0
03	Power sensor	NRV-Z5	100542	October 24, 2018	One year
04	Power sensor	NRP6A	101369	April 11, 2019	One Year
05	Signal Generator	E4438C	MY49070393	January 4, 2019	One Year
06	Amplifier	60S1G4	0331848	No Calibration R	equested
07	Directional Coupler	778D	MY48220584	No Calibration R	equested
08	Directional Coupler	772D	MY46151265	No Calibration R	equested
09	E-field Probe	SPEAG EX3DV4	3617	January 31, 2019	One year
10	DAE	SPEAG DAE4	771	January 11,2019	One year
11	Dipole Validation Kit	SPEAG D2450V2	853	July 17, 2019	One year

### Table K.5-1: List of Main Instruments



## **K.6 GRAPH RESULTS**

## WLAN2450\_CH1 Left Tilt

Date: 9/23/2019 Electronics: DAE4 Sn771 Medium: head 2450 MHz Medium parameters used: f = 2412;  $\sigma = 1.741$  mho/m;  $\epsilon r = 38.59$ ;  $\rho = 1000$  kg/m<sup>3</sup> Ambient Temperature: 22.5°C, Liquid Temperature: 22.3°C Communication System: WLAN2450 2412 MHz Duty Cycle: 1:1 Probe: EX3DV4 – SN3617 ConvF(7.62,7.62,7.62)

**Area Scan (71x121x1):** Interpolated grid: dx=1.000 mm, dy=1.000 mm Maximum value of SAR (interpolated) = 0.908 W/kg

Zoom Scan (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm Reference Value = 4.624 V/m; Power Drift = 0.03 dB Peak SAR (extrapolated) = 1.08 W/kg SAR(1 g) = 0.407 W/kg; SAR(10 g) = 0.16 W/kg Maximum value of SAR (measured) = 0.603 W/kg







## WLAN2450\_CH1 Top

Date: 9/23/2019Electronics: DAE4 Sn771 Medium: body 2450 MHz Medium parameters used: f = 2412;  $\sigma = 1.921$  mho/m;  $\epsilon r = 52.66$ ;  $\rho = 1000$  kg/m<sup>3</sup> Ambient Temperature: 22.5°C, Liquid Temperature: 22.3°C Communication System: WLAN2450 2412 MHz Duty Cycle: 1:1 Probe: EX3DV4 – SN3617 ConvF(7.79,7.79,7.79)

**Area Scan (71x121x1):** Interpolated grid: dx=1.000 mm, dy=1.000 mm Maximum value of SAR (interpolated) = 0.466 W/kg

Zoom Scan (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mmReference Value = 4.518 V/m; Power Drift = 0.16 dB Peak SAR (extrapolated) = 0.586 W/kg SAR(1 g) = 0.263 W/kg; SAR(10 g) = 0.113 W/kg Maximum value of SAR (measured) = 0.465 W/kg



Fig K.6-2



## K.7 ANNEX SYSTEM VALIDATION RESULTS

## 2450 MHz

Date: 9/23/2019 Electronics: DAE4 Sn771 Medium: Head 2450 MHz Medium parameters used: f = 2450 MHz;  $\sigma = 1.777$  mho/m;  $\epsilon_r = 38.54$ ;  $\rho = 1000$  kg/m<sup>3</sup> Ambient Temperature: 22.5°C Liquid Temperature: 22.3°C Communication System: CW Frequency: 2450 MHz Duty Cycle: 1:1 Probe: EX3DV4 – SN3617 ConvF(7.62,7.62,7.62)

System Validation /Area Scan (81x191x1): Interpolated grid: dx=1.000 mm, dy=1.000
mm
Reference Value = 117.48 V/m; Power Drift = 0.06
Fast SAR: SAR(1 g) = 12.97 W/kg; SAR(10 g) = 6.03 W/kg
Maximum value of SAR (interpolated) = 22.26 W/kg

**System Validation /Zoom Scan (7x7x7)/Cube 0:** Measurement grid: dx=5mm, dy=5mm, dz=5mm Reference Value =117.48 V/m; Power Drift = 0.06 dB Peak SAR (extrapolated) = 25.48 W/kg

SAR(1 g) = 12.93 W/kg; SAR(10 g) = 6.15 W/kg

Maximum value of SAR (measured) = 21.09 W/kg



0 dB = 21.09 W/kg = 13.24 dB W/kg





## 2450 MHz

Date: 9/23/2019 Electronics: DAE4 Sn771 Medium: Body 2450 MHz Medium parameters used: f = 2450 MHz;  $\sigma = 1.957$  mho/m;  $\epsilon_r = 52.61$ ;  $\rho = 1000$  kg/m<sup>3</sup> Ambient Temperature: 22.5°C Liquid Temperature: 22.3°C Communication System: CW Frequency: 2450 MHz Duty Cycle: 1:1 Probe: EX3DV4 – SN3617 ConvF(7.79,7.79,7.79)

System Validation /Area Scan (81x191x1): Interpolated grid: dx=1.000 mm, dy=1.000 mm Reference Value = 109.15 V/m; Power Drift = -0.02

Fast SAR: SAR(1 g) = 13.14 W/kg; SAR(10 g) = 6.03 W/kg Maximum value of SAR (interpolated) = 21.04 W/kg

System Validation /Zoom Scan (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm Reference Value =109.15 V/m; Power Drift = -0.02 dB Peak SAR (extrapolated) = 26.74 W/kg SAR(1 g) = 13.18 W/kg; SAR(10 g) = 6.21 W/kg

Maximum value of SAR (measured) = 21.84 W/kg









## K.8 System Validation

The SAR system must be validated against its performance specifications before it is deployed. When SAR probes, system components or software are changed, upgraded or recalibrated, these must be validated with the SAR system(s) that operates with such components.

Probe SN.	Liquid name	Validation date	Frequency point	Status (OK or Not)
3617	Head 750MHz	Feb.14,2019	750 MHz	OK
3617	Head 850MHz	Feb.14,2019	835 MHz	OK
3617	Head 900MHz	Feb.14,2019	900 MHz	OK
3617	Head 1750MHz	Feb.14,2019	1750 MHz	OK
3617	Head 1810MHz	Feb.14,2019	1810 MHz	OK
3617	Head 1900MHz	Feb.15,2019	1900 MHz	OK
3617	Head 2000MHz	Feb.15,2019	2000 MHz	OK
3617	Head 2100MHz	Feb.15,2019	2100 MHz	OK
3617	Head 2300MHz	Feb.15,2019	2300 MHz	OK
3617	Head 2450MHz	Feb.15,2019	2450 MHz	OK
3617	Head 2600MHz	Feb.16,2019	2600 MHz	OK
3617	Head 3500MHz	Feb.16,2019	3500 MHz	OK
3617	Head 3700MHz	Feb.16,2019	3700 MHz	OK
3617	Head 5200MHz	Feb.16,2019	5250 MHz	OK
3617	Head 5500MHz	Feb.16,2019	5600 MHz	OK
3617	Head 5800MHz	Feb.16,2019	5800 MHz	OK
3617	Body 750MHz	Feb.16,2019	750 MHz	OK
3617	Body 850MHz	Feb.13,2019	835 MHz	OK
3617	Body 900MHz	Feb.13,2019	900 MHz	OK
3617	Body 1750MHz	Feb.13,2019	1750 MHz	OK
3617	Body 1810MHz	Feb.13,2019	1810 MHz	OK
3617	Body 1900MHz	Feb.13,2019	1900 MHz	OK
3617	Body 2000MHz	Feb.17,2019	2000 MHz	OK
3617	Body 2100MHz	Feb.17,2019	2100 MHz	OK
3617	Body 2300MHz	Feb.17,2019	2300 MHz	OK
3617	Body 2450MHz	Feb.17,2019	2450 MHz	OK
3617	Body 2600MHz	Feb.17,2019	2600 MHz	OK
3617	Body 3500MHz	Feb.12,2019	3500 MHz	OK
3617	Body 3700MHz	Feb.12,2019	3700 MHz	OK
3617	Body 5200MHz	Feb.12,2019	5250 MHz	OK
3617	Body 5500MHz	Feb.12,2019	5600 MHz	OK
3617	Body 5800MHz	Feb.12,2019	5800 MHz	OK

### Table F.1: System Validation for 3617



## K.9 Probe Calibration Certificate

### Probe 3617 Calibration Certificate

Engineering AG Zeughausstrasse 43, 8004 Zui	rich, Switzerland	C S	Service suisse d'étalonnage Servizio svizzero di taratura Swiss Calibration Service
Accredited by the Swiss Accred	itation Service (SAS) ice is one of the signatories	to the EA	reditation No.: SCS 0108
Autiliateral Agreement for the	recognition of calibration c	ertificates	EV0.0047 1 40
Client CTTL (Auden	)	Certificate No:	EX3-3617_Jan19
CALIBRATION	CERTIFICATE		
Object	EX3DV4 - SN:361	7	
Calibration procedure(s)	QA CAL-01.v9, QA QA CAL-25.v7 Calibration procee	A CAL-12.v9, QA CAL-14.v5, QA lure for dosimetric E-field probes	CAL-23.v5,
Calibration date:	January 31, 2019		
The measurements and the uncompared to the control of the control	certainties with confidence pro lucted in the closed laboratory &TE critical for calibration)	facility: environment temperature (22 $\pm$ 3)°C a	or measurements (S)). are part of the certificate. and humidity < 70%.
All calibrations have been cond Calibrations have been cond Calibration Equipment used (M	Certainties with confidence pro- lucted in the closed laboratory &TE critical for calibration)	facility: environment temperature (22 ± 3)°C a	or measurements (S)). are part of the certificate. and humidity < 70%.
The measurements and the unit All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP	Certainties with confidence pro- lucted in the closed laboratory &TE critical for calibration)	Cal Date (Certificate No.)	are part of the certificate. and humidity < 70%.
The measurements and the unit All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-Z91	Certainties with confidence pro- lucted in the closed laboratory &TE critical for calibration)	Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672)	or measurements (SI), are part of the certificate, and humidity < 70%, Scheduled Calibration Apr-19 Apr-19
The measurements and the unit All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91	Certainties with confidence pro- lucted in the closed laboratory &TE critical for calibration) ID SN: 104778 SN: 103244 SN: 103245	Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02672)	or measurements (SI). are part of the certificate. and humidity < 70%. Scheduled Calibration Apr-19 Apr-19 Apr-19
The measurements and the una All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator	Certainties with confidence pro- lucted in the closed laboratory &TE critical for calibration) ID SN: 104778 SN: 104244 SN: 103244 SN: 103245 SN: S5277 (20x)	Cal Date (Certificate No.)           04-Apr-18 (No. 217-02672)           04-Apr-18 (No. 217-02673)           04-Apr-18 (No. 217-02672)	Scheduled Calibration         Apr-19         Apr-19         Apr-19
All calibration Setundate doca All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator DAE4	ID SN: 104778 SN: 103245 SN: 103245 SN: 55277 (20x) SN: 660	Cal Date (Certificate No.) Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02672) 104-Apr-18 (No. 217-02682) 19-Dec-18 (No. DAE4-660_Dec18)	Scheduled Calibration         Apr-19         Apr-19         Dec-19
The measurements and the under the measurement of the measur	ATE critical for calibration) LID SN: 104778 SN: 103244 SN: 103245 SN: 55277 (20x) SN: 660 SN: 3013	Cal Date (Certificate No.)           04-Apr-18 (No. 217-02672/02673)           04-Apr-18 (No. 217-02672/02673)           04-Apr-18 (No. 217-02672)           04-Apr-18 (No. 217-02673)           04-Apr-18 (No. 217-02682)           19-Dec-18 (No. DAE4-660_Dec18)           31-Dec-18 (No. ES3-3013_Dec18)	Scheduled Calibration         Apr-19         Apr-19         Apr-19         Apr-19         Dec-19         Dec-19
All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator DAE4 Reference Probe ES3DV2 Secondary Standards	ID SN: 104778 SN: 104778 SN: 103244 SN: 103245 SN: 55277 (20x) SN: 3013 ID	Cal Date (Certificate No.)           04-Apr-18 (No. 217-02672/02673)           04-Apr-18 (No. 217-02672/02673)           04-Apr-18 (No. 217-02672)           04-Apr-18 (No. 217-02672)           04-Apr-18 (No. 217-02673)           04-Apr-18 (No. 217-02672)           04-Apr-18 (No. 217-02673)           04-Apr-18 (No. 217-02672)           19-Dec-18 (No. DAE4-660_Dec18)           31-Dec-18 (No. ES3-3013_Dec18)           Check Date (in house)	Scheduled Calibration         Apr-19         Apr-19         Apr-19         Apr-19         Scheduled Calibration         Scheduled Calibration         Apr-19         Apr-19         Scheduled Calibration         Scheduled Calibration         Scheduled Calibration         Scheduled Calibration         Scheduled Check
The measurements and the under the measurement of the measur	ATE critical for calibration) LID SN: 104778 SN: 104778 SN: 103244 SN: 103245 SN: 55277 (20x) SN: 660 SN: 3013 ID SN: GB41293874	Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02682) 19-Dec-18 (No. DAE4-60_Dec18) 31-Dec-18 (No. ES3-3013_Dec18) Check Date (in house) 06-Apr-16 (in house check Jun-18)	Scheduled Calibration         Apr-19         Apr-19         Apr-19         Apr-19         Scheduled Calibration         Scheduled Calibration         Apr-19         Apr-19         Scheduled Calibration         Apr-19         Scheduled Calibration         Image: Scheduled Calibration         Apr-19         Apr-19         Dec-19         Dec-19         Image: Scheduled Check         In house check: Jun-20
The measurements and the under docard the measurements and the under docard the docard t	ATE critical for calibration) LID SN: 104778 SN: 104778 SN: 103244 SN: 103245 SN: 55277 (20x) SN: 660 SN: 3013 ID SN: GB41293874 SN: MY41498087	Cal Date (Certificate No.)           04-Apr-18 (No. 217-02672/02673)           04-Apr-18 (No. 217-02672/02673)           04-Apr-18 (No. 217-02672)           04-Apr-18 (No. 217-02672)           04-Apr-18 (No. 217-02672)           04-Apr-18 (No. 217-02673)           04-Apr-18 (No. 217-02673)           04-Apr-18 (No. 217-02682)           19-Dec-18 (No. DAE4-660_Dec18)           31-Dec-18 (No. ES3-3013_Dec18)           Check Date (in house)           06-Apr-16 (in house check Jun-18)           06-Apr-16 (in house check Jun-18)	Scheduled Calibration         Apr-19         Apr-19         Apr-19         Apr-19         Scheduled Calibration         Scheduled Calibration         Apr-19         Apr-19         Scheduled Calibration         Image: Scheduled Calibration         Apr-19         Scheduled Calibration         In house check: Jun-20         In house check: Jun-20
All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator DAE4 Reference Probe ES3DV2 Secondary Standards Power meter E4419B Power sensor E4412A Power sensor E4412A	ID       ID         ID       SN: 104778         SN: 103244       SN: 103244         SN: 103245       SN: 55277 (20x)         SN: 660       SN: 3013         ID       SN: 3013         SN: GB41293874       SN: MY41498087         SN: 000110210       SN: 000110210	Cal Date (Certificate No.)           04-Apr-18 (No. 217-02672/02673)           04-Apr-18 (No. 217-02672/02673)           04-Apr-18 (No. 217-02672)           04-Apr-18 (No. 217-02672)           04-Apr-18 (No. 217-02673)           04-Apr-18 (No. 217-02672)           04-Apr-18 (No. 217-02673)           04-Apr-18 (No. 217-02673)           04-Apr-18 (No. 217-02673)           04-Apr-18 (No. 217-02682)           19-Dec-18 (No. DAE4-660_Dec18)           31-Dec-18 (No. ES3-3013_Dec18)           Check Date (in house)           06-Apr-16 (in house check Jun-18)           06-Apr-16 (in house check Jun-18)           06-Apr-16 (in house check Jun-18)	Scheduled Calibration         Apr-19         Apr-19         Apr-19         Apr-19         Scheduled Calibration         Scheduled Calibration         Apr-19         Apr-19         Scheduled Calibration         Image: Scheduled Calibration         Scheduled Calibration         Apr-19         Apr-19         Dec-19         Scheduled Check         In house check: Jun-20         In house check: Jun-20         In house check: Jun-20
The measurements and the under the measurement of the measurements and the under the under the measurements and the under the un	ID         ID           ID         SN: 104778           SN: 103244         SN: 103244           SN: 103245         SN: 55277 (20x)           SN: 660         SN: 3013           ID         SN: 601           SN: 00110210         SN: WY41498087           SN: 000110210         SN: US3642U01700	Cal Date (Certificate No.)           04-Apr-18 (No. 217-02672/02673)           04-Apr-18 (No. 217-02672/02673)           04-Apr-18 (No. 217-02672)           04-Apr-18 (No. 217-02672)           04-Apr-18 (No. 217-02672)           04-Apr-18 (No. 217-02673)           04-Apr-18 (No. 217-02673)           04-Apr-18 (No. 217-02673)           04-Apr-18 (No. 217-02682)           19-Dec-18 (No. 217-02682)           19-Dec-18 (No. 237-02673)           04-Apr-16 (in bouse)           06-Apr-16 (in house check Jun-18)           06-Apr-16 (in house check Jun-18)           06-Apr-16 (in house check Jun-18)           04-Aug-99 (in house check Jun-18)	Scheduled Calibration         Apr-19         Apr-19         Apr-19         Dec-19         Dec-19         Scheduled Check         In house check: Jun-20
The measurements and the und All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator DAE4 Reference Probe ES3DV2 Secondary Standards Power meter E44198 Power sensor E4412A RF generator HP 8648C Network Analyzer E8358A	ID         ID           ID         SN: 104778           SN: 103244         SN: 103245           SN: 103245         SN: 55277 (20x)           SN: 660         SN: 3013           ID         SN: 6841293874           SN: 000110210         SN: WY41498087           SN: US3642U01700         SN: US41080477	Cal Date (Certificate No.)           04-Apr-18 (No. 217-02672/02673)           04-Apr-18 (No. 217-02672/02673)           04-Apr-18 (No. 217-02672)           04-Apr-18 (No. 217-02672)           04-Apr-18 (No. 217-02672)           04-Apr-18 (No. 217-02672)           04-Apr-18 (No. 217-02673)           04-Apr-18 (No. 217-02682)           19-Dec-18 (No. DAE4-660_Dec18)           31-Dec-18 (No. ES3-3013_Dec18)           0           06-Apr-16 (in house)           06-Apr-16 (in house check Jun-18)           06-Apr-16 (in house check Jun-18)           04-Apr-18 (in house check Jun-18)           04-Apr-14 (in house check Jun-18)           04-Apr-14 (in house check Jun-18)	Scheduled Calibration         Apr-19         Apr-19         Apr-19         Apr-19         Scheduled Calibration         Apr-19         Apr-19         Apr-19         Dec-19         Dec-19         In house check: Jun-20         In house check: Jun-20
All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator DAE4 Reference Probe ES3DV2 Secondary Standards Power meter E44198 Power sensor E4412A RF generator HP 8648C Network Analyzer E8358A	ATE critical for calibration) LID SN: 104778 SN: 104778 SN: 103244 SN: 103245 SN: 55277 (20x) SN: 660 SN: 3013 ID SN: GB41293874 SN: WY41498087 SN: 000110210 SN: US3642U01700 SN: US3642U01700 SN: US41080477 Name	Cal Date (Certificate No.)           04-Apr-18 (No. 217-02672/02673)           04-Apr-18 (No. 217-02672/02673)           04-Apr-18 (No. 217-02672)           04-Apr-18 (No. 217-02672)           04-Apr-18 (No. 217-02673)           04-Apr-18 (No. 217-02673)           04-Apr-18 (No. 217-02673)           04-Apr-18 (No. 217-02682)           19-Dec-18 (No. DAE4-660_Dec18)           31-Dec-18 (No. ES3-3013_Dec18)           0           06-Apr-16 (in house)           06-Apr-16 (in house check Jun-18)           06-Apr-16 (in house check Jun-18)           04-Apr-18 (in house check Jun-18)           04-Apr-14 (in house check Jun-18)           04-Apr-14 (in house check Jun-18)           05-Apr-16 (in house check Jun-18)           05-Apr-16 (in house check Jun-18)           06-Apr-16 (in house check Jun-18)           07-Aug-99 (in house check Jun-18)           07-Aug-99 (in house check Oct-18)	Scheduled Calibration         Apr-19         Apr-19         Apr-19         Dec-19         Dec-19         Scheduled Check         In house check: Jun-20         Signature
The measurements and the und All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator DAE4 Reference Probe ES3DV2 Secondary Standards Power meter E4419B Power sensor E4412A RF generator HP 8648C Network Analyzer E8358A Calibrated by:	Arrent and the constrainty of neurons of constrainty of neurons of constrainty of neurons of the constrainty of the constrainty of the constraints	Cal Date (Certificate No.) (Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02682) 19-Dec-18 (No. DAE4-660_Dec18) 31-Dec-18 (No. ES3-3013_Dec18) 06-Apr-16 (in house) 06-Apr-16 (in house check Jun-18) 06-Apr-16 (in house check Jun-18) 06-Apr-16 (in house check Jun-18) 06-Apr-16 (in house check Jun-18) 04-Aug-99 (in house check Jun-18) 31-Mar-14 (in house check Jun-18) 31-Mar-14 (in house check Jun-18) Function Laboratory Technician	Scheduled Calibration         Apr-19         Apr-19         Apr-19         Apr-19         Apr-19         Dec-19         Dec-19         In house check: Jun-20         In house check: Jun-20         In house check: Jun-20         In house check: Jun-20         Signature
All calibration of the decomposition of the decompo	ID         ID         SN: 104778         SN: 103244         SN: 103245         SN: 103245         SN: 3013         ID         SN: 660         SN: 00110210         SN: US4109087         SN: US41080477	Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02682) 19-Dec-18 (No. DAE4-660_Dec18) 31-Dec-18 (No. ES3-3013_Dec18) 06-Apr-16 (in house) 06-Apr-16 (in house check Jun-18) 06-Apr-16 (in house check Jun-18) 06-Apr-16 (in house check Jun-18) 06-Apr-16 (in house check Jun-18) 04-Aug-99 (in house check Jun-18) 04-Aug-99 (in house check Jun-18) 31-Mar-14 (in house check Jun-18) Tunction Laboratory Technician Technical Manager	Scheduled Calibration         Apr-19         Apr-19         Apr-19         Apr-19         Dec-19         Dec-19         Scheduled Check         In house check: Jun-20         In house check: Jun-20         In house check: Jun-20         In house check: Jun-20         Signature

Certificate No: EX3-3617\_Jan19

Page 1 of 19