



Hear better. Live better.

6700 Washington Avenue S
Eden Prairie, Minnesota 55344
T: 1.800.328.8602

Starkey.com

August 31, 2022

Federal Communications Commission
45 L Street, NE
Washington, DC 20554
Attention: Application Examiner

Reference:	Certification:	Part 15.247 Intentional Radiator
	Applicant:	Starkey Laboratories, Inc.
	Equipment:	Low Power Device (2400-2483.5 MHz)
	Model #:	Genesis AI ITE R custom hearing aid (left ear)
		FCC ID: EOA-24GENITELCH

Dear Sir/Madam:

Enclosed please find an FCC EOA-24GENITELCH certification application. The device is a wireless hearing aid intended for use by the hearing impaired. The wireless features are used during the initial fitting of a hearing aid to program the hearing aid and used during normal usage to communicate control information and digital audio to and from any properly configured Bluetooth Low Energy compliant accessory and between a binaural pair of Genesis hearing aids. Note that this is the left hearing aid. The corresponding right hearing aid is certified with FCC ID: EOA-24GENITERCH. Further details and supporting documents are attached.

Should you have any questions, please feel free to contact the undersigned.

Sincerely,

A handwritten signature in black ink, appearing to read "William J. Mitchell".

William J. Mitchell, PE
Principal Wireless and Safety Regulatory Engineer
952-918-5683
bill_mitchell@starkey.com