

TEST RESULT SUMMARY

EN 300 330-2 V1.1.1: 2001 Subclauses 7.2, 7.3, 7.4

MANUFACTURER'S NAME Medtronic

NAME OF EQUIPMENT Patient Assistant

MODEL NUMBER 2696

TYPE OF EQUIPMENT Pocket-sized remote communications device

MANUFACTURER'S ADDRESS 7000 Central Avenue NE, MS CW300

Minneapolis MN 55432-3576

TEST REPORT NUMBER NC300445.2

TEST DATE 03 February 2003

According to testing performed at TÜV Product Service Inc, the above-mentioned unit is in compliance with the emission requirements defined in European Telecommunication Standard EN 300 330.

It is the manufacturer's responsibility to assure that additional production units of this model are manufactured with identical electrical and mechanical characteristics. Any modifications necessary for compliance made during testing on the above mentioned date(s) must be implemented in all production units for compliance to be maintained.

TÜV Product Service Inc, as an independent testing laboratory, declares that the equipment tested as specified above conforms to the emission requirements of European Telecommunication Standard EN 300 330: "Electromagnetic Compatibility and Radio Spectrum Matters (ERM); Short Range Devices (SRD); Technical Characteristics and Test Methods for Radio Equipment in the Frequency Range 9 kHz to 25 MHz and Inductive Loop Systems in the Frequency Range 9 kHz to 30 MHz."

Date: 14 March 2003

Location: Taylors Falls MN

USA

Tested By: R. M. Johnson

Technical Writer

& C. Sausan Thomas K. Swamon

Not Transferable

T. K. Swanson



DIRECTORY/SUB-CLAUSE PARAMETER TO BE MEASURED PAGE

A)	A) Documentation				
	Test Result Summary	-	1		
	Directory/Parameters to be measured	-	2 - 8		
The	Test Setup Photos complete list of measurements called for in EN 300	330 is given below	A1 – A3		
Trar	smitter parameters - Ambient temperature20	°C Relative humidity	<i>/</i> 22%		
7.2.1	Transmitter Carrier Output Levels		3		
7.2.2	RF Carrier Current	Class 3 Only	N/A		
7.2.3	Radiated E-Field	Class 4 Only	N/A		
7.3.1	Permitted Frequency Range of Modulation bandwidth	_	4		
7.4.2	Conducted Spurious Emissions (Operating)	Class 3 Only	N/A		
7.4.2	Conducted Spurious Emissions (Standby)	Class 3 Only	N/A		
7.4.3	Radiated Field Strength (Transmit < 30 MHz)		5		
7.4.3	Radiated Field Strength (Standby < 30 MHz)	_	5		
7.4.4	Radiated Field Strength (Transmit > 30 MHz)	<u>-</u>	6		
7.4.4	Radiated Field Strength (Standby > 30 MHz)	6			
Rec	eiver parameters -	-			
8.1.2	8.1.2 Receiver Spurious Radiation (Frequencies < 30 MHz)				
8.1.2	Receiver Spurious Radiation (Frequencies > 30 MHz)		N/A		

File No. NC300445.2, Page 2 of 8



H-FIELD FIELD STRENGTH - SUB-CLAUSE 7.2.1 (Class 1)

Rated field strength (maximum) -15.5 dBμA/m at 10 metres

Test conditions	Nominal System Operating Frequency: 200.0 kHz Maximum Transmitter Field Strength (dBμA/m)			
T _{nom} (20)°C	Final 10 Metre 10 Metre Margin of compl		Margin of compliance	
		-15.5	35.99	51.49
Maximum deviation from rated output under normal test conditions (dB)				
Measurement uncertainty (dBμA/m) ±1				
The noise floor measurement at 10 metres is –21.5dBuA/m.				

LIMIT SUB-CLAUSE 7.2.1.3

Frequency range - (MHz)	H-field field strength limit (Hf) - dBμA/m at 10 m
0,009 ≤ f < 0,03	72 or per note on loop coil antenna area
0,03 ≤ f < 0,07	72 at 0,03 MHz descending 3 dB/oct
0,119≤ f < 0,135	or per note on loop coil antenna area
0,05975 ≤ f <0,06025	42
0,07≤ f < 0,119	
0,135 ≤ f < 1,0	37,7 at 0,135 MHz descending 3 dB/oct
1,0 ≤ f < 4,642	29 at 1,0 MHz descending 9 dB/oct
4,642 ≤ f < 30	9
6,765 ≤ f ≤ 6,795 (ISM)	42
13,553 ≤ f ≤ 13,567 (ISM)	
$26,957 \le f \le 27,283 \text{ (ISM)}$	

REFERENCE NUMBER(S) OF TEST EQUIPMENT USED (for reference see test equipment listing)

1,2,3

Ambient temperature20.°C Relative humidity22%

File No. NC300445.2, Page 3 of 8



RF CARRIER CURRENT - SUB-CLAUSE 7.2.2 (Class 2)

N/A

RADIATED E-FIELD, FIELD STRENGTH (measured as H-field) - SUB-CLAUSE 7.2.3 (Class 4) N/A

PERMITTED FREQUENCY RANGE OF MODULATION BANDWIDTH - SUB-CLAUSE 7.3.1

Test not applicable since peak fundamental level is below the spurious emission limit.

CONDUCTED SPURIOUS EMISSIONS (Operating)- SUB-CLAUSE 7.4.2 (Class 3)

N/A

CONDUCTED SPURIOUS EMISSIONS (Standby) - SUB-CLAUSE 7.4.2 (Class 3)

N/A

File No. NC300445.2, Page 4 of 8



TRANSMITTER RADIATED SPURIOUS EMISSIONS (Transmit < 30 MHz) - SUB-CLAUSE 7.4.3

					Transmit	
					spec limit	
Freq.	1 Meter	3 Meter	10 M	10 M	dBuA/m	Limit Margin
MHz	dBuV/m	dBuV/m	dBuV/m	dBuA/m	300 330	dBuA/m
0.15	72	60	34	-17.5	14.76452881	32.2645288
0.6	57	40	16	-35.5	8.735560259	44.2355603
1	51	30	6	-45.5	6.513989065	52.0139891

Transmitter operating with normal internal modulation.

LIMIT SUB-CLAUSE 7.4.3.2

State	Frequency 9 kHz ≤ f < 10 MHz	Frequency 10 MHz ≤ f < 30 MHz
Transmit	27 dBμA/m descending 3 dB/oct	-3,5 dBμA/m

REFERENCE NUMBER(S) OF TEST EQUIPMENT USED (for reference see test equipment listing) 1,2,3

Ambient temperature20.°C Relative humidity22%

TRANSMITTER RADIATED SPURIOUS EMISSIONS (standby < 30 MHz) - SUB-CLAUSE 7.4.3

No spurious emissions detected in standby mode below 30 MHz.

LIMIT SUB-CLAUSE 7.4.3.2

State	Frequency 9 kHz ≤ f < 10 MHz	Frequency 10 MHz ≤ f < 30 MHz
Standby	6 dBμA/m descending 3 dB/oct	-24,5 dBμA/m

REFERENCE NUMBER(S) OF TEST EQUIPMENT USED (for reference see test equipment listing)

1,2,3

Ambient temperature20.°C Relative humidity22%

File No. NC300445.2, Page 5 of 8



TRANSMITTER RADIATED SPURIOUS EMISSIONS (Transmit > 30 MHz) - SUB-CLAUSE 7.4.4

No spurious emissions detected in transmit mode above 30 MHz.

LIMIT SUB-CLAUSE 7.4.4.2

State	47 MHz to 74 MHz	Other Frequencies between 30 to 1000 MHz
	87,5 MHz to 118 MHz	
	174 MHz to 230 MHz	
	470 MHz to 862 MHz	
Operating	4 nW	250 nW

REFERENCE NUMBER(S) OF TEST EQUIPMENT USED (for reference see test equipment listing) 4,5,6,7,8

Ambient temperature20.°C

Relative humidity22%

TRANSMITTER RADIATED SPURIOUS EMISSIONS (standby > 30 MHz) - SUB-CLAUSE 7.4.4

No spurious emissions detected in standby mode above 30 MHz.

LIMIT SUB-CLAUSE 7.4.4.2

State	47 MHz to 74 MHz	Other Frequencies between 30 to 1000 MHz	
	87,5 MHz to 118 MHz		
	174 MHz to 230 MHz		
	470 MHz to 862 MHz		
Standby	2 nW	2 nW	

REFERENCE NUMBER(S) OF TEST EQUIPMENT USED (for reference see test equipment listing) 4,5,6,7,8

Ambient temperature20.°C Relative humidity22%

File No. NC300445.2, Page 6 of 8



RECEIVER SPURIOUS RADIATION (< 30 MHz) - SUB-CLAUSE 8.1

N/A

RECEIVER SPURIOUS RADIATION (>30 MHz) SUB-CLAUSE 8.1

N/A



File No. NC300445.2, Page 7 of 8



TEST EQUIPMENT

To simplify the identification on each page of the test equipment used, on each page of the test report, each item of test equipment and ancillaries such as cables are identified (numbered) by the Test Laboratory, below.

Ref. No.	Instrument/Ancillary	Туре	Manufacturer	Serial No.
01	Loop Antenna	HFH2-Z2	Polarad	879285/036
02	EMI Receiver	ESHS-20	Rohde-Schwarz	837055/003
03	Coaxial cable		Polarad	
04	Spectrum Analyzer	8566B	HP	2430A00930
05	Analyzer Display	85662A	HP	2403A08134
06	Quasi-Peak Adapter	85650A	HP	2521A01006
07	Preamplifier	ZHL-1042J	Mini-Circuits	32296
08	Biconicalog Antenna	EM-6917B	Electro-Metrics	102

ADDITIONAL INFORMATION SUPPLEMENTARY TO THE TEST REPORT

Photographs other than of the test-setup are sent independently from report.



TEST SETUP PHOTOS





Test-setup photo(s): Radiated Emission





Test-setup photo(s): Radiated Emission

