



10 June 2021

Federal Communications Commission
Authorization and Evaluation Division
7435 Oakland Mills Road
Columbia, Maryland 21046

Re: FCC Application
FCC ID: 2AGO22580

To whom it may concern,

We, Inspire Medical Systems Inc., hereby apply for original certification of the radios included in the Model 2580 Patient Remote. The telemetry intentional radiator operates at 175kHz and is compliant to 47 CFR §15.209. The Bluetooth radio operates at 2.4GHz and is compliant to 47 CFR §15.247.

Sincerely,

A handwritten signature in blue ink that reads "Joel Aaberg". The signature is written in a cursive, flowing style.

Sr. Director Regulatory: Joel Aaberg

Telephone Number: 763-205-7972

Email: joelaaberg@inspiresleep.com