

Page:

Report No.: SZEM170500410002

No. 1 Workshop, M-10, Middle section, Science & Technology Park,

Shenzhen, Guangdong, China 518057 Telephone: +86 (0) 755 2601 2053 Fax: +86 (0) 755 2671 0594

Email: ee.shenzhen@sgs.com

SAR Evaluation Report

Application No.: SZEM1705004100CR

Applicant: Shenzhen Medica Technology Development Co., Ltd

Address of Applicant: 2F Building A, Tongfang Information Harbor, No. 11, East Langshan Road,

Nanshan District, Shenzhen, China

Manufacturer: Shenzhen Medica Technology Development Co., Ltd

Address of Manufacturer: 2F Building A, Tongfang Information Harbor, No. 11, East Langshan Road,

Nanshan District, Shenzhen, China

Factory: Apexto electronics Co., Ltd

Address of Factory: 2F, A block, Nanchang road 59, Gushu Xixiang, Baoan District, Shenzhen,

P.R. China

Equipment Under Test (EUT):

EUT Name: Smart Pillow

Model No.: P200

FCC ID: 2ADIOP200

Standards: 47 CFR Part 1.1307

47 CFR Part 2.1093

KDB447498D01 General RF Exposure Guidance v06

Date of Receipt: 2017-05-10

Date of Test: 2017-05-15 to 2017-06-02

Date of Issue: 2017-06-08

Test Result : PASS*

Authorized Signature:



Jack Zhang

EMC Laboratory Manager

The manufacturer should ensure that all products in series production are in conformity with the product sample detailed in this report. If the product in this report is used in any configuration other than that detailed in the report, the manufacturer must ensure the new system complies with all relevant standards. Any mention of SGS International Electrical Approvals or testing done by SGS International Electrical Approvals in connection with, distribution or use of the product described in this report must be approved by SGS International Electrical Approvals in writing.

The report must not be used by the client to claim product certification, approval, or endorsement by NVLAP, NIST, or any agency of the federal government. All test results in this report can be traceable to National or International Standards.

This document is issued by the Company subject to its General Conditions of Service printed overleaf,-available on request or accessible at http://www.sgs.com/en/Terms-and-Conditions.aspx and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at http://www.sgs.com/en/Terms-and-Conditions/Terms

^{*} In the configuration tested, the EUT complied with the standards specified above.



Report No.: SZEM170500410002

Page: 2 of 6

2 Version

Revision Record							
Version	Chapter	Date	Modifier	Remark			
01		2017-06-08		Original			

Authorized for issue by:		
	leo ti	
	Leo Li /Project Engineer	
	Eric Fu	
	Eric Fu /Reviewer	



Report No.: SZEM170500410002

Page: 3 of 6

3 Contents

		Page
1	COVER PAGE	1
2	VERSION	2
3	CONTENTS	3
4	GENERAL INFORMATION	4
	4.1 GENERAL DESCRIPTION OF EUT	4
	4.2 TEST LOCATION	5
	4.4 DEVIATION FROM STANDARDS	5
	4.5 ABNORMALITIES FROM STANDARD CONDITIONS	5
5	SAR EVALUATION	
	5.1 RF Exposure Compliance Requirement	6
	5.1.1 Standard Requirement	6
	5.1.2 Limits	6
	•	



Report No.: SZEM170500410002

Page: 4 of 6

4 General Information

4.1 General Description of EUT

Product Name:	Smart Pillow
Model No.:	P200
Operation Frequency:	2402MHz~2480MHz
Bluetooth Version:	V4.0 single mode
Modulation Type:	GFSK
Number of Channel:	40
Antenna Type:	Integral
Antenna Gain:	2.09dBi
Power Supply	Lithium Ion Battery: 3.7V 900mAh rechargeable battery which charged by USB port



Report No.: SZEM170500410002

Page: 5 of 6

4.2 Test Location

All tests were performed at:

SGS-CSTC Standards Technical Services Co., Ltd. Shenzhen Branch

No. 1 Workshop, M-10, Middle section, Science & Technology Park, Shenzhen, Guangdong, China 518057

Telephone: +86 (0) 755 2601 2053 Fax: +86 (0) 755 2671 0594

No tests were sub-contracted.

4.3 Test Facility

The test facility is recognized, certified, or accredited by the following organizations:

• CNAS (No. CNAS L2929)

CNAS has accredited SGS-CSTC Standards Technical Services Co., Ltd. Shenzhen Branch EMC Lab to ISO/IEC 17025:2005 General Requirements for the Competence of Testing and Calibration Laboratories (CNAS-CL01 Accreditation Criteria for the Competence of Testing and Calibration Laboratories) for the competence in the field of testing.

· A2LA (Certificate No. 3816.01)

SGS-CSTC Standards Technical Services Co., Ltd., Shenzhen EMC Laboratory is accredited by the American Association for Laboratory Accreditation(A2LA). Certificate No. 3816.01.

VCCI

The 10m Semi-anechoic chamber and Shielded Room of SGS-CSTC Standards Technical Services Co., Ltd. have been registered in accordance with the Regulations for Voluntary Control Measures with Registration No.: G-823, R-4188, T-1153 and C-2383 respectively.

• FCC - Registration No.: 556682

SGS-CSTC Standards Technical Services Co., Ltd., Shenzhen EMC Laboratory has been registered and fully described in a report filed with the (FCC) Federal Communications Commission. The acceptance letter from the FCC is maintained in our files. Registration No.: 556682.

Industry Canada (IC)

Two 3m Semi-anechoic chambers and the 10m Semi-anechoic chamber of SGS-CSTC Standards Technical Services Co., Ltd. Shenzhen Branch EMC Lab have been registered by Certification and Engineering Bureau of Industry Canada for radio equipment testing with Registration No.: 4620C-1, 4620C-2, 4620C-3.

4.4 Deviation from Standards

None.

4.5 Abnormalities from Standard Conditions

None.

4.6 Other Information Requested by the Customer

None.



Report No.: SZEM170500410002

Page: 6 of 6

5 SAR Evaluation

5.1 RF Exposure Compliance Requirement

5.1.1 Standard Requirement

According to KDB447498D01 General RF Exposure Guidance v06

4.3.1. Standalone SAR test exclusion considerations

Unless specifically required by the published RF exposure KDB procedures, standalone 1-g head or body and 10-g extremity SAR evaluation for general population exposure conditions, by measurement or numerical simulation, is not required when the corresponding SAR Exclusion Threshold condition, listed below, is satisfied.

5.1.2 Limits

The 1-g and 10-g SAR test exclusion thresholds for 100 MHz to 6 GHz at test separation distances ≤ 50 mm are determined by:

[(max. power of channel, including tune-up tolerance, mW)/(min. test separation distance, mm)] $\cdot [\sqrt{f(GHz)}] \le 3.0$ for 1-g SAR and ≤ 7.5 for 10-g extremity SAR, where

f(GHz) is the RF channel transmit frequency in GHz

Power and distance are rounded to the nearest mW and mm before calculation ¹⁷

The result is rounded to one decimal place for comparison

The test exclusions are applicable only when the minimum test separation distance is \leq 50 mm and for transmission frequencies between 100 MHz and 6 GHz. When the minimum test separation distance is < 5 mm, a distance of 5 mm is applied to determine SAR test exclusion

5.1.3 EUT RF Exposure

The Max Conducted Peak Output Power is	-1.53	dBm in highest channel	2.402	GHz
-1.53 dBm logarithmic terms convert to numeric				
According to the formula. calculate the test excl				
[(max. power of channel, including tune-up tolerance, mW)/				
(min. test separation distance, mm)] $\cdot [\sqrt{f(GHz)}]$				
General RF Exposure = $(0.70 \text{ mW} / 5 \text{ mm}) \times \sqrt{2.402 \text{ GHz}} = 0.22$			(1)	
SAR requirement:				
S = 3.0				
(1) < (2)			(2)	
So the SAR report is not required.				