

BIOTRONIK SE & Co. KG · Postfach 470255 · 12311 Berlin

Federal Communication Commission
Authorization and Evaluation Division
Equipment Authorization Branch
7435 Oakland Mills Road
Columbia, MD 21046

Tel +49 (0) 30 68905-1291
 Fax +49 (0) 30 68905-1921
 dirk.koenig@biotronik.com
 www.biotronik.com

Berlin, 25 March 2020

Statement of re-use of original test results of approved radio modules
according to KDB 996369 D02 Module Q and A v01 Q1(a)

Reference of new FCC ID application:

Applicant: BIOTRONIK SE & Co. KG	
FCC ID:	QRI-RENAMICNEO

We, the undersigned, hereby declare that our radio device hosts the following approved radio module(s)

Grantee: Telit Communications S.p.A.			
Type of Module: BlueMod+S50			
FCC ID:	RFR-S50	Report Reference	1-5687/17-01-07-A

The module(s) is/are installed according to the installation instructions of the module manufacturer(s) without modification. The original module test report(s) accurately represent all conducted measured test results under the new conditions. All radiated measurements according to the applicable radio standard(s), the RF-exposure and simultaneous transmission evaluation are performed and reported for the entire host.

Sincerely,



Dirk König
 Regulatory Affairs Senior Manager
 BIOTRONIK SE & Co. KG