

Lumicare Medtech Co., Ltd

3/20/2025

Federal Communications Commission
Equipment Authorization Branch
7435 Oakland Mills Road
Columbia, MD 21046

Modular Approval Request

FCC ID: 2BM9Y0001


The following attestation addresses the requirements to support modular approval:

Modular approval requirement	Yes (provide brief statement)	No *
(a) The radio elements must have the radio frequency circuitry shielded. Physical components and tuning capacitor(s) may be located external to the shield, but must be on the module assembly	The product is produced with permanently attached RF shield during production.	/
(b) The module must have buffered modulation/data inputs to ensure that the device will comply with Part 15 requirements with any type of input signal	The product buffers the modulation and data inputs ensuring compliance with regulations.	/
(c) The module must contain power supply regulation on the module	The product incorporates regulators to ensure operation in specified range.	/
(d) The module must contain a permanently attached antenna, or contain a unique antenna connector, and be marketed and operated only with specific antenna(s), per Sections 15.203, 15.204(b), 15.204(c), 15.212(a), 2.929(b)	This product contain a permanently attached antenna.	/
(e) The module must demonstrate compliance in a stand-alone configuration	The product was tested in a stand-alone configuration (placed on development/interface fixture) without enclosure.	/
(f) The module must be labelled with its permanently affixed FCC ID label, or use an electronic display (See KDB Publication 784748 about labelling requirements)	The final product is affixed with a permanent FCC ID label.	/
(g) The module must comply with all specific rules applicable to the transmitter. The grantee must provide comprehensive instructions to explain compliance requirements	The product fully complies with all specific applicable rules and regulatory precautions are addressed in manual.	/

Modular approval requirement	Yes (provide brief statement)	No *
(h) The module must comply with RF exposure requirements	The product complies with applicable RF exposure requirements. Installation and RF exposure precautions are addressed in manual.	/

* Please provide a detailed explanation if the answer is "No."

Sincerely,

Signed: 

Printed name: Tong Liu

Title: R&D Manager