

Schmid & Partner Engineering AG

s p e a g

Zeughausstrasse 43, 8004 Zurich, Switzerland
Phone +41 1 245 9700, Fax +41 1 245 9779
info@speag.com, <http://www.speag.com>

Dr. Chris Zombolas
EMC Technologies Pty. Ltd.
57 Assembly Drive
Tullamarine, Vic
Australia 3043

Zurich, September 6, 2004

Re: Additional Conversion Factor for Dosimetric E-Field Probe (Your P.O. No. 1138)

Dear Chris,

Attached please find additional conversion factor for dosimetric E-field probe ET3DV6 - SN:1380.

Should you have any additional questions, please do not hesitate to contact us. We are always honored to offer our products and services to EMC.

Best regards,


Katja Pokovic

Encl.

Additional Conversion Factors

for Dosimetric E-Field Probe

Type:

ET3DV6

Serial Number:

1380

Place of Assessment:

Zurich

Date of Assessment:


September 6, 2004

Probe Calibration Date:

July 14, 2004

Schmid & Partner Engineering AG hereby certifies that conversion factor(s) of this probe have been evaluated on the date indicated above. The assessment was performed using the FDTD numerical code SEMCAD of Schmid & Partner Engineering AG. The uncertainty of the numerical assessment is based on the extrapolation from measured values.

Assessed by:



Dosimetric E-Field Probe ET3DV6 SN:1380

Conversion factor ($\pm$ standard deviation)

1610 MHz ConvF $5.3 \pm 10\%$

$\epsilon_r = 40.3 \pm 5\%$
 $\sigma = 1.29 \pm 5\% \text{ mho/m}$
(head tissue)

Important Note:

For numerically assessed probe conversion factors, parameters Alpha and Delta in the DASY software must have the following entries: Alpha = 0 and Delta = 1.

Please see also Section 4.7 of the DASY4 Manual.