

# APPENDIX C: SAR TISSUE SPECIFICATIONS

## **Measurement Procedure for Tissue verification:**

- 1) The network analyzer and probe system was configured and calibrated.
- 2) The probe was immersed in the tissue. The tissue was placed in a nonmetallic container. Trapped air bubbles beneath the flange were minimized by placing the probe at a slight angle.
- 3) The complex admittance with respect to the probe aperture was measured
- 4) The complex relative permittivity  $\epsilon$  can be calculated from the below equation (Pournaropoulos and Misra):

$$Y = \frac{j2\omega\varepsilon_{r}\varepsilon_{0}}{\left[\ln(b/a)\right]^{2}} \int_{a}^{b} \int_{a}^{b} \int_{0}^{\pi} \cos\phi' \frac{\exp\left[-j\omega r(\mu_{0}\varepsilon_{r}'\varepsilon_{0})^{1/2}\right]}{r} d\phi' d\rho' d\rho$$

where Y is the admittance of the probe in contact with the sample, the primed and unprimed coordinates refer to source and observation points, respectively,  $r^2 = \rho^2 + \rho'^2 - 2\rho\rho'\cos\phi'$ ,  $\omega$  is the angular frequency, and  $j = \sqrt{-1}$ .

2 Mixtures escription: Aqueous solution with	surfactants and inhibitors	
clarable, or hazardous compon		
AS: 107-21-1	Ethanediol	>1.0-4.9%
INECS: 203-473-3	STOT RE 2, H373;	
leg.nr.: 01-2119456816-28-0000	Acute Tox. 4, H302	
AS: 68608-26-4	Sodium petroleum sulfonate	< 2.9%
INECS: 271-781-5	Eye Irrit. 2, H319	
leg.nr.: 01-2119527859-22-0000		
AS: 107-41-5	Hexylene Glycol / 2-Methyl-pentane-2,4-diol	< 2.9%
INECS: 203-489-0	Skin Irrit. 2, H315; Eye Irrit. 2, H319	
leg.nr.: 01-2119539582-35-0000	-	
AS: 68920-66-1	Alkoxylated alcohol, > C <sub>16</sub>	< 2.0%
ILP: 500-236-9	Aquatic Chronic 2, H411;	
leg.nr.: 01-2119489407-26-0000	Skin Irrit. 2, H315; Eye Irrit. 2, H319	
Iditional information:	-	
r the wording of the listed risk phra	ases refer to section 16.	
t mentioned CAS-, EINECS- or re	gistration numbers are to be regarded as Proprietary	Confidential.

## Figure C-1

Note: Liquid recipes are proprietary SPEAG. Since the composition is approximate to the actual liquids utilized, the manufacturer tissue-equivalent liquid data sheets are provided below.

FCC ID A3LSMF721U	SAR EVALUATION REPORT	Approved by: Technical Manager
<b>DUT Type:</b> Portable Handset		APPENDIX C: Page 1 of 4



Schmid & Partner Engineering AG

S p e a g

5500

5500

1.9

-1.0 -6.2

-3.2 3.6

-3.3 4.2

4500

3.55

5.30 -2.6 1.7

5.42

5.77

6.00 -3.5 4.6 -3.7 5.5

6.23

3500 ncy MHz 51.3 3.31 -0.9 -6.3

49.0 5.36 -2.7 -2.8 2.2

48.6 5.65 -3.0 3.2

48.3 5.88

Zeughausstrasse 43, 8004 Zurich, Switzerland Phone +41 44 245 9700, Fax +41 44 245 9779 www.speag.swiss, info@speag.swiss

## Measurement Certificate / Material Test

Body Tissue Simulating Liquid (MBBL600-6000V6)

Item Name SL AAM U16 BC (Batch: 210621-3) Product No.

SPEAG Manufacturer

## Measurement Method

TSL dielectric parameters measured using calibrated DAK probe.

## Target Parameters

Target parameters as defined in the KDB 865664 compliance standard.

## **Test Condition**

Ambient Condition 22°C; 30% humidity

TSL Temperature 22°C Test Date 23-Jun-21 Operator WM

Additional Information
TSL Density

TSL Heat-capacity

TILL	Measu	red		Targe	t	Diff.to Targ	get [%]	15.0
[MHz]	e'	е"	sigma	eps	sigma	∆-eps	∆-sigma	10.0
600	55.7	26.7	0.89	56.1	0.95	-0.7	-6.3	% 5.0 -
750	55.3	22.5	0.94	55.5	0.96	-0.4	-2.1	0.0 -5.0 -5.0
800	55.1	21.5	0.96	55.3	0.97	-0.4	-1.0	
825	55.1	21.1	0.97	55.2	0.98	-0.3	-1.0	
835	55.1	20.8	0.97	55.1	0.99	0.0	-1.5	ð-10.0
850	55.0	20.6	0.97	55.2	0.99	-0.3	-2.0	-15.0 1500 2500 35
900	54.9	19.9	0.99	55.0	1.05	-0.2	-5.7	500 1500 2500 35 Frequency M
1400	54.1	15.9	1.24	54.1	1.28	0.0	-3.1	15.0
1450	54.0	15.7	1.27	54.0	1.30	0.0	-2.3	10.0
1600	53.8	15.3	1.36	53.8	1.39	0.0	-2.2	% A
1625	53.8	15.2	1.38	53.8	1.41	0.1	-2.1	Souductivity On Proceedings
1640	53.8	15.2	1.39	53.7	1.42	0.1	-2.1	B 0.0 A
1650	53.7	15.1	1.39	53.7	1.43	0.0	-2.8	
1700	53.7	15.0	1.42	53.6	1.46	0.3	-2.7	÷10.0
1750	53.6	14.9	1.45	53.4	1.49	0.3	-2.7	-15.0
1800	53.5	14.9	1.49	53.3	1.52	0.4	-2.0	500 1500 2500 33 Frequency M
1810	53.5	14.9	1.50	53.3	1.52	0.4	-1.3	3500 50.9 15.9 3.10 51.
1825	53.5	14.8	1.51	53.3	1.52	0.4	-0.7	3700 50.6 16.2 3.33 51.
1850	53.5	14.8	1.52	53.3	1.52	0.4	0.0	5200 47.7 18.6 5.39 49.
1900	53.4	14.8	1.56	53.3	1.52	0.2	2.6	5250 47.6 18.7 5.46 49.
1950	53.4	14.7	1.60	53.3	1.52	0.2	5.3	5300 47.5 18.8 5.54 48.
2000	53.3	14.7	1.63	53.3	1.52	0.0	7.2	5500 47.1 19.1 5.83 48.
2050	53.3	14.7	1.67	53.2	1.57	0.1	6.4	5600 46.9 19.2 5.98 48.
2100	53.2	14.7	1.71	53.2	1.62	0.1	5.6	5700 46.7 19.3 6.13 48.
2150	53.1	14.7	1.75	53.1	1.66	0.0	5.4	5800 46.5 19.4 6.27 48.
2200	53.1	14.7	1.80	53.0	1.71	0.1	5.3	6000 46.1 19.7 6.57 47.
2250	53.0	14.7	1.84	53.0	1.76	0.1	4.5	6500
2300	52.9	14.7	1.88	52.9	1.81	0.0	3.9	7000
2350	52.9	14.8	1.93	52.8	1.85	0.1	4.3	7500
2400	52.8	14.8	1.98	52.8	1.90	0.1	4.2	8000
2450	52.7	14.8	2.02	52.7	1.95	0.0	3.6	8500
2500	52.6	14.9	2.07	52.6	2.02	-0.1	2.5	9000
2550	52.5	14.9	2.12	52.6	2.09	-0.1	1.4	9500
2600	52.5	15.0	2.16	52.5	2.16	0.0	0.0	10000

Figure C-2 600 - 6000 MHz Body Tissue Equivalent Matter

FCC ID A3LSMF721U	SAR EVALUATION REPORT	Approved by: Technical Manager
<b>DUT Type:</b> Portable Handset		APPENDIX C: Page 2 of 4



Schmid & Partner Engineering AG

0 e a g S

Zeughausstrasse 43, 8004 Zurich, Switzerland Phone +41 44 245 9700, Fax +41 44 245 9779 www.speag.swiss, info@speag.swiss

## Measurement Certificate / Material Test

Head Tissue Simulating Liquid (HBBL600-10000V6)

Product No. SL AAH U16 BC (Batch: 210629-3)

Manufacturer SPEAG

## Measurement Method

TSL dielectric parameters measured using calibrated DAK probe.

### **Target Parameters**

Target parameters as defined in the IEEE 1528 and IEC 62209 compliance standards.

## Test Condition

Ambient Condition 22°C; 30% humidity

TSL Temperature 22°C Test Date Operator WM Additional Information

TSL Density

TSL Heat-capacity

	Measu	ıred		Targe	t	Diff.to Target [%]	
f [MHz]	e'	e"	sigma	eps	sigma	∆-eps	∆-sigma
600	44.7	25.5	0.85	42.7	0.88	4.6	-3.6
750	44.1	21.6	0.90	41.9	0.89	5.1	0.7
800	44.0	20.6	0.92	41.7	0.90	5.6	2.5
825	44.0	20.2	0.93	41.6	0.91	5.8	2.6
835	44.0	20.0	0.93	41.5	0.91	5.9	2.0
850	43.9	19.8	0.93	41.5	0.92	5.8	1.5
900	43.8	19.0	0.95	41.5	0.97	5.5	-2.1
1400	42.8	15.1	1.18	40.6	1.18	5.4	0.0
1450	42.7	14.9	1.20	40.5	1.20	5.4	0.0
1600	42.4	14.4	1.28	40.3	1.28	5.2	-0.3
1625	42.4	14.3	1.30	40.3	1.30	5.3	0.1
1640	42.4	14.3	1.31	40.3	1.31	5.3	0.3
1650	42.3	14.3	1.31	40.2	1.31	5.1	-0.2
1700	42.3	14.2	1.34	40.2	1.34	5.3	-0.2
1750	42.2	14.1	1.37	40.1	1.37	5.3	-0.1
1800	42.1	14.0	1.40	40.0	1.40	5.3	0.0
1810	42.1	13.9	1.41	40.0	1.40	5.3	0.7
1825	42.1	13.9	1.42	40.0	1.40	5.3	1.4
1850	42.0	13.9	1.43	40.0	1.40	5.0	2.1
1900	42.0	13.8	1.46	40.0	1.40	5.0	4.3
1950	41.9	13.8	1.49	40.0	1.40	4.7	6.4
2000	41.8	13.7	1.53	40.0	1.40	4.5	9.3
2050	41.8	13.7	1.56	39.9	1.44	4.7	8.0
2100	41.7	13.7	1.59	39.8	1.49	4.7	6.8
2150	41.6	13.6	1.63	39.7	1.53	4.7	6.3
2200	41.6	13.6	1.67	39.6	1.58	4.9	5.8
2250	41.5	13.6	1.70	39.6	1.62	4.9	4.8
2300	41.4	13.6	1.74	39.5	1.67	4.9	4.4
2350	41.3	13.6	1.78	39.4	1.71	4.9	4.0
2400	41.3	13.6	1.82	39.3	1.76	5.1	3.7
2450	41.2	13.6	1.86	39.2	1.80	5.1	3.3
2500	41.1	13.6	1.90	39.1	1.85	5.0	2.5
2550	41.0	13.7	1.94	39.1	1.91	4.9	1.6
2600	41.0	13.7	1.98	39.0	1.96	5.1	0.8




Figure C-3 600 - 10000 MHz Head Tissue Equivalent Matter

FCC ID A3LSMF721U	SAR EVALUATION REPORT	Approved by: Technical Manager
<b>DUT Type:</b> Portable Handset		APPENDIX C: Page 3 of 4



Schmid & Partner Engineering AG

Zeughausstrasse 43, 8004 Zurich, Switzerland Phone +41 44 245 9700, Fax +41 44 245 9779 www.speag.swiss, info@speag.swiss Measurement Certificate / Material Test Head Tissue Simulating Liquid (HBBL4-250V3) item Name Product No. SL AAH 005 AD (Batch: 210601-1) Manufacturer SPEAG Measurement Method
TSL dielectric parameters measured using calibrated DAK probe. Setup Validation Validation results were within ± 2.5% towards the target values of Methanol. Target Parameters

Target parameters as defined in the IEEE 1528 and IEC 62209 compliance standards. Test Condition
Ambient
TSL Temperature Environment temperatur (22 ± 3)°C and humidity < 70%. 3-Jun-21 Test Date WM Operator Additional Information TSL Density 1.042 g/cm3
TSL Heat-capacity 3.574 kJ/(kg\*K) 10.0 -7.5 -5.0 -2.5 -0.0 -Dev. Permithity % -2.5 -5.0 -7.5 5 25 45 65 7.5 5.0 2.5 -2.5 5.0 7.5 10.0 5 25 45 65 85 105 125 145 165 185 205 225 245 Proquency Mitz

e a g

S P

Figure C-4 5 – 250 MHz Head Tissue Equivalent Matter

Pana 1 of 1

SAR EVALUATION REPORT	Tarkairal Managan
	Technical Manager  APPENDIX C: Page 4 of 4