## Medtronic

September 07, 2016

Federal Communications Commission Authorization and Evaluation Division 7435 Oakland Mills Road Columbia, Maryland 21046

To whom it may concern.

Reference: Application for CareLink SmartSync Device Manager Base FCC ID: LF524970A

I, the undersigned, as the authorized signatory for the Medtronic, Inc. hereby apply to seek original grant for 24970A CareLink SmartSync Device Manager Base.

Authorization is sought under FCC rule parts 15.207 and 951.

The 24970A CareLink SmartSync Device Manager Base pairs Medtronic apps with the patient's mobile device to Interrogate and/or program Implantable Medtronic Devices.

The Base unit is AC powered, uses a MICS (Medical Implant Communication Service) radio Module operating in the 402-405 MHz band to communicate with the patients implant and Bluetooth (BT) classic preapproved radio module operating at 2.4 GHz band to communicate with the Medtronic apps running on the patient's mobile devices.

The following reports are being submitted to support this filing. The results are representative of the radios installed in the Patient Connector. The radios are electrically and mechanically identical to those in the reports.

Thank you in advance, should you have any questions please feel free to contact me.

Sincerely,

Daniel Johnson

Radio Regulatory Affairs Specialist Cardiac Rhythm Heart Failure Tel 763-526-0820 Daniel.L.Johnson@Medtronic.com

Medtronic Inc. 710 Medtronic Parkway Minneapolis MN 55432