



#### SAR REFERENCE DIPOLE CALIBRATION REPORT

Ref: ACR.329.17.21.BES.A


#### 7.2 SAR MEASUREMENT RESULT WITH HEAD LIQUID

At those frequencies, the target SAR value can not be generic. Hereunder is the target SAR value defined by MVG, within the uncertainty for the system validation. All SAR values are normalized to 1 W net power. In bracket, the measured SAR is given with the used input power.

Software	OPENSAR V5		
Phantom	SN 13/09 SAM68		
Probe	SN 41/18 EPGO333		
Liquid	Head Liquid Values 5200 MHz: eps': 34.44 sigma: 4.64 Head Liquid Values 5400 MHz: eps': 33.63 sigma: 4.88 Head Liquid Values 5600 MHz: eps': 32.80 sigma: 5.12 Head Liquid Values 5800 MHz: eps': 32.63 sigma: 5.31		
Distance between dipole and liquid	10 mm		
Area scan resolution	dx=8mm/dy=8mm		
Zoon Scan Resolution	dx=4mm/dy=4m/dz=2mm		
Frequency	5200 MHz 5400 MHz 5600 MHz 5800 MHz		
Input power	20 dBm		
Liquid Temperature	20 +/- 1 °C		
Lab Temperature	20 +/- 1 °C		
Lab Humidity	30-70 %		

Frequency (MHz)	1 g SA	R (W/kg)	10 g SAR (W/kg)	
	required	measured	required	measured
5200	76.50	76.41 (7.64)	21.60	21.86 (2.19)
5400	· <del>-</del> 3	80.52 (8.05)	) <del>-</del>	22.91 (2.29)
5600	\$ <del>7</del> .4	79.08 (7.91)	65-17	22.73 (2.27)
5800	78.00	76.49 (7.65)	21.90	22.03 (2.20)

## SAR MEASUREMENT PLOTS @ 5200 MHz



Page: 8/13







SAR REFERENCE DIPOLE CALIBRATION REPORT

Ref: ACR.329.17.21.BES.A






## SAR MEASUREMENT PLOTS @ 5600 MHz



## SAR MEASUREMENT PLOTS @ 5800 MHz



Page: 9/13





## SAR REFERENCE DIPOLE CALIBRATION REPORT

Ref: ACR.329.17.21.BES.A

#### 7.3 BODY LIQUID MEASUREMENT

Frequency MHz	Relative permittivity $(\mathbf{s}_{r}')$		Conductivity (σ) S/m	
	required	m easured	required	measured
5200	49.0 <b>±10</b> %	45.50	5.30 <b>±10</b> %	5.63
5300	48.9 <b>±10</b> %		5.42 <b>±10</b> %	
5400	48.7 <b>±10</b> %	44.78	5.53 <b>±10</b> %	5.95
5500	48.6 <b>±10</b> %		5.65 <b>±10</b> %	
5600	48.5 <b>±10</b> %	44.85	5.77 <b>±10</b> %	6.26
5800	48.2 <b>±10</b> %	44.45	6.00 <b>±10</b> %	6.58

#### 7.4 SAR MEASUREMENT RESULT WITH BODY LIQUID

Software	OPENSAR V5	
Phantom	SN 13/09 SAM68	
Probe	SN 41/18 EPGO333	
Liquid	Body Liquid Values 5200 MHz: eps': 45.50 sigma: 5. Body Liquid Values 5400 MHz: eps': 44.78 sigma: 5 Body Liquid Values 5600 MHz: eps': 44.85 sigma: 6 Body Liquid Values 5800 MHz: eps': 44.45 sigma: 6	
Distance between dipole and liquid	10 mm	
Area scan resolution	dx=8mm/dy=8mm	
Zoon Scan Resolution	dx=4mm/dy=4m/dz=2mm	
Frequency	5200 MHz 5400 MHz 5600 MHz 5800 MHz	
Input power	20 dBm	
Liquid Temperature	20 +/- 1 °C	
Lab Temperature	20 +/- 1 °C	
Lab Humidity	30-70 %	

Frequency (MHz)	1 g SAR (W/kg)	10 g SAR (W/kg)
3007 46 3000	measured	measured
5200	73.02 (7.30)	20.58 (2.06)
5400	77.86 (7.79)	21.85 (2.19)
5600	79.90 (7.99)	22.73 (2.27)
5800	71.90 (7.19)	20.50 (2.05)


Page: 10/13







SAR REFERENCE DIPOLE CALIBRATION REPORT

Ref: ACR.329.17.21.BES.A



BODY SAR MEASUREMENT PLOTS @ 5400 MHz






Page: 11/13





SAR REFERENCE DIPOLE CALIBRATION REPORT

Ref: ACR.329.17.21.BES.A



Page: 12/13





#### SAR REFERENCE DIPOLE CALIBRATION REPORT

Ref: ACR.329.7.21.BES.A

## 8 LIST OF EQUIPMENT

Equipment Summary Sheet				
Equipment Description	Manufacturer / Model	Identification No.	Current Calibration Date	Next Calibration Date
SAM Phantom	MVG	SN 13/09 SAM68	Validated. No cal required.	Validated. No cal required.
COMOSAR Test Bench	Version 3	NA	Validated. No cal required.	Validated. No cal required.
Network Analyzer	Rohde & Schwarz ZVM	100203	08/2021	08/2024
Network Analyzer	Agilent 8753ES	MY40003210	10/2022	10/2025
Network Analyzer – Calibration kit	Rohde & Schwarz ZV-Z235	101223	05/2012	05/2025
Network Analyzer – Calibration kit	HP 85033D	3423A08186	06/2021	06/2027
Calipers	Mitutoyo	SN 0009732	10/2022	10/2025
Reference Probe	MVG	SN 41/18 EPGO333	10/2022	10/2025
Multimeter	Keithley 2000	1160271	02/2023	02/2026
Signal Generator	Rohde & Schwarz SMB	106589	04/2022	04/2025
Amplifier	MVG	MODU-023-C-0002	Characterized prior to test. No cal required.	Characterized prior to test. No cal required.
Power Meter	NI-USB 5680	170100013	06/2021	06/2024
Power Meter	Rohde & Schwarz NRVD	832839-056	11/2022	11/2025
Directional Coupler	Krytar 158020	131467	Characterized prior to test. No cal required.	Characterized prior to test. No cal required.
Temperature / Humidity Sensor	Testo 184 H1	44225320	06/2021	06/2024

Page: 13/13

Template ACR.DDD.N.YY.MVGB.ISSUE SAR Reference Dipole v1

This document shall not be reproduced, except in full or in part, without the written approval of MVG. The information contained herein is to be used only for the purpose for which it is submitted and is not to be released in whole or part without written approval of MVG.

Page 95 of 100 No.: BCTC/RF-EMC-005 Edition: B.0



## 17. EUT Photographs

#### **EUT Front View**




#### **EUT Back View**




No.: BCTC/RF-EMC-005 Page 96 of 100 Edition/ B.0



# 18. Photographs Of The Liquid



Photograph of the depth in the Body Phantom (600-10000MHz, depth >15cm)




No.: BCTC/RF-EMC-005 Page 97 of 100




# 19. EUT Test Setup Photographs

## Front



Back



No.: BCTC/RF-EMC-005 Page 98 of 100 Edition / B.0



Left



No.: BCTC/RF-EMC-005 Page 99 of 100 Edition / B.0



### **STATEMENT**

- 1. The equipment lists are traceable to the national reference standards.
- 2. The test report can not be partially copied unless prior written approval is issued from our lab.
- 3. The test report is invalid without the "special seal for inspection and testing".
- 4. The test report is invalid without the signature of the approver.
- 5. The test process and test result is only related to the Unit Under Test.
- 6. Sample information is provided by the client and the laboratory is not responsible for its authenticity.
- 7. The quality system of our laboratory is in accordance with ISO/IEC17025.
- 8. If there is any objection to this test report, the client should inform issuing laboratory within 15 days from the date of receiving test report.

#### Address:

1-2/F., Building B, Pengzhou Industrial Park, No.158, Fuyuan 1st Road, Zhancheng, Fuhai Subdistrict, Bao'an District, Shenzhen, Guangdong, China

TEL: 400-788-9558

P.C.: 518103

FAX: 0755-33229357

Website: http://www.chnbctc.com

E-Mail: bctc@bctc-lab.com.cn

\*\*\*\* END \*\*\*\*

No.: BCTC/RF-EMC-005 Page 100 of 100

Edition: B.0