

## Radio Similarity Declaration

**Manufacturer:**

**Name:** Medtronic Inc.  
**Address:** 710 Medtronic Parkway N.E.  
Minneapolis, MN, USA  
55432

**Equipment:**

**Model(s):** 8667-20 and 8667-40

We, Medtronic Inc., declare that the radio equipment models listed above are of identical electronic construction (RF hybrid layout, antenna), sharing the same electronic hardware design (BOM, Electrical Schematic), differences being in the drug delivery volume features that allow a volume of 20ml for 8667-20 and 40 ml for 8667-40. The volume increase changes the titanium enclosure but doesn't change any radio parameters.

The Radio active and passive components of these devices are entirely identical. These implants share the same RF specifications.

**Signed for and on behalf of:** Medtronic Inc.

**Place:** Minneapolis MN, USA. **Date:** November 11, 2021

**Name:**  
Guillaume Girard

**Function:**  
Senior Compliance Manager

**Signature:**

