

Measurement Report for Device
Device Under Test Properties

Model, Manufacturer	Dimensions [mm]	Software Version	DUT Type
Device,	70.0 x 50.0 x 8.0	3.2.0.1840	5G Verification Source

Exposure Conditions

Phantom Section	Position, Test Distance [mm]	Frequency [MHz]	Conversion Factor
5G	FRONT, 10.00	10000.0	1.0

Hardware Setup

Phantom	Medium	Probe, Calibration Date	DAE, Calibration Date
mmWave – 1044	Air –	EUmmWV4 – SN9441_F1–55GHz, 2023–11–17	DAE4 Sn1424, 2023–12–07

Scans Setup

Scan Type	5G Scan
Grid Extents [mm]	120.0 x 120.0
Grid Steps [lambda]	0.25 x 0.25
Sensor Surface [mm]	10.0

Measurement Results

Date	2024–05–13
Avg. Area [cm²]	4.00
psPDn+ [W/m²]	61.9
psPDtot+ [W/m²]	62.0
H _{max} [A/m]	0.432
E _{max} [V/m]	163
max _(Stot) [W/m²]	70.0
Power Drift [dB]	0.01
iPDn	117

