

Date: March 31, 2019

C2PC Declaration

1. Augmedics will be using the WiFi module certified on 03/11/2019 under FCC ID: 2AR2O-VOB-P3310 in its own device known as the “xvision-spine (XVS)” system.
2. The module will be inserted in the headset of the “xvision-spine (XVS)” and will now be body worn with a separation distance of 4cm from the user.
3. Only the DTS WiFi g transmitter of the module will be operational. The other 2 transmitters (DSS and NII transmitters) and BLE (DTS) transmitter are non-operational by software.
4. In support of our request for a C2PC we are submitting the following:
 - a. C2PC 731 form
 - b. C2PC declaration
 - c. Host label
 - d. Exterior photos of host and photo of module in host
 - e. SAR exemption calculation
 - f. Host user manual
 - g. Photos and/or technical drawings of the device to show the 4 cm separation distance.
5. The documents support our request for a C2PC.

Thank you,



Stuart Wolf, CTO