

WSAUD A/S  
Nymoellevej 6  
3540 Lynge  
Denmark

Federal Communication Commission  
Authorization and Evaluation Division  
7435 Oakland Mills Road  
Columbia, MD 21046  
USA

Wednesday, June 7, 2023

## Attestation Statement according to 2.911(d)(7)

**FCC ID: 2AXDT-RFM015**

TO WHOM IT MAY CONCERN:

We, WSAUD A/S with FCC Registration Number (FRN) 0029767068 designate WS Audiology USA Inc. with FCC Registration Number (FRN) 0033434770, to act as U.S. agent for service of process.

WS Audiology USA Inc. with FCC Registration Number (FRN) 0033434770 hereby acknowledges the obligation of the U.S. agent for service of process. Further WSAUD A/S with FCC Registration Number (FRN) 0029767068, accepts to maintain an agent for no less than one year after the grantee has terminated all marketing and importation or the conclusion of any Commission-related proceeding involving the equipment.

Sincerely

**WSAudiology**

WSAUD A/S  
Nymøllevej 6  
DK-3540 Lynge  
Denmark  
Tel. +45 44 35 56 00



(signature & company stamp)

Ulrik Bjørn  
Vice President AP (Advanced Projects)

WSAUD A/S  
Nymoellevej 6  
3540 Lynge  
Denmark  
+45 30 84 01 60

ulrik.bjoern@wsa.com



(signature)

La-Shaunda Joseph  
US Regulatory Affairs Manager

WS Audiology USA Inc.  
101 Wood Ave  
Iselin, NJ 08830, United States  
+1 561 785 9321

lashaunda.joseph@wsa.com

(d) The module shall comply with the provisions for external power amplifiers and antennas detailed in the applicable RSS and rule part. The equipment certification application shall contain: <ul style="list-style-type: none"> <li>i. a detailed description of the configuration of highest antenna gain for each type of transmitting antenna for licence-exempt modules;</li> <li>ii. the maximum transmitting antenna gain for licence modules; and</li> <li>iii. a detailed description of the configuration of lowest antenna gain for each type of receiving antenna for Dynamic Frequency Selection (DFS) modules with removable antenna(s).</li> </ul>	X	
(e) The module shall be tested for compliance with the applicable standard in a stand-alone configuration (i.e., the module shall not be inside another product during testing).		X
<i>The modular transmitter has been tested integrated on a typical host, a hearing aid instrument that best represents all the future hosts where the module will be used by WS Audiology. Due to the size limitations of the hearing aid instruments, the radio module is in the same PCB of the host making impossible to test it on a standalone configuration.</i>		
(f) The module complies or will comply with applicable RSS-102 exposure requirements and any applicable FCC RF exposure requirement (§§1.1307(b), 1.1310, 2.1091, and 2.1093) in its intended configuration/integration in a host.	X	
(g) The module must be labeled with its permanently affixed label (indicating ISCED certification number, HVIN and FCC identifier), or use an electronic display (see KDB Publication 784748 and RSP-100, section 5).		X
<i>The size of the module is too small for the regulatory FCC/ISED label in its short form to be placed any, which are therefore displayed in the user manual.</i>		
(h) The module must comply with all specific rules applicable to the transmitter, including all the conditions provided in the integration instructions by the grantee.	X	
<b>Only applicable for FCC certification:</b>		
(i) The module must contain a permanently attached antenna, or contain a unique antenna connector, and be marketed and operated only with specific antenna(s), per §§ 15.203, 15.204(b), 15.204(c), 15.212(a), 2.929(b). For further information concerning antenna connectors see: DA-00-1087.	X	

If you have any questions, please feel free to contact us at the address shown below.

Sincerely,




(signature & company stamp)

Ulrik Bjørn  
Vice President AP (Advanced Projects)

WSAUD A/S  
Nymoellevej 6  
3540 Lynge  
Denmark  
+45 30 84 01 60

ulrik.bjoern@wsa.com



(Date/signature)

Kristine Pedersen  
Senior Director Global Regulatory Affairs

WSAUD A/S  
Nymoellevej 6  
3540 Lynge  
Denmark  
+45 21 37 09 38

kristine.pedersen@wsa.com

**WSAudiology**  
WSAUD A/S  
Nymoellevej 6  
DK-3540 Lynge  
Denmark  
Tel. +45 44 35 56 00