

### testo Smart Probe user manual

## **Chapter 1 Project overview**

Testo Smart probe project involve development on APP and eight types of probe which APP can get measurement data from probe and also realize extend function on refrigerant/VAC/heating system.







## **Chapter 2 Product description**

### 2.1 Smart Probe App

## Main Features & Benefits


Control of the Smart Line instrument via smartphone / tablet	
Check measured data wireless on smartphone / tablet	
Do trendings as graph or table on smartphone / tablet	
Send measured values in protocol as .pdf or excel file	

## **Technical Product Data**

Wireless connection	Bluetooth Low Energy
Range	20 m (unobstructed field), 5m for handhold probe
Compatibility	Android 4.3 or higher / iOS 7.0 or higher







Page 2 last saved: 2021-12-29



## 2.1 Smart Probe testo 405i

## Main Features & Benefits

Measures air velocity, volume flow and temperature	
Easy set up for volume flow measurement via APP	
Easy determination of timed or multi-point average via APP	
Telescopic probe that can be extended up to 400 mm	
Determine BTUh in combination with testo 605i	
Check measured data wireless on mobile device	
Send measured values in protocol as .pdf or excel file	

## **Technical Product Data**

Measuring range	0 to 20 m/s / 0 to 3940 fpm -20 to 60 °C / -4 to 140 °F
Accuracy	±(0.05 m/s + 2.5% o. mv.) (0 +2 m/s) ±(0.15 m/s + 2.5% o. mv.) (2 +15 m/s) ±(10 fpm + 2.5% o. mv.) (0 +394 fpm) ±(29.5 fpm + 2.5% o. mv.) (394 +3.940 fpm) ±0.5 °C / ±0.9°F
Resolution	0.01 m/s / 1 fpm 0.1 °C / 0.1 °F
Dimensions	200mm x 30 mm x 41 mm Extendable telescope up to 400 mm
Measuring cycle	1/sec
Battery life	15 h
Battery type	3x AAA
Operating temperature	-20 °C + 50 °C / -4 to 122 °F
Storage temperature	-20°C to 60°C / -4 to 140 °F

Page 3 last saved: 2021-12-29



### 2.2 Smart Probe testo 410i

## Main Features & Benefits

Measures air velocity, volume flow and temperature	
Easy set up for volume flow measurement via APP	
Easy determination of timed or multi-point average via APP	
Comparison of volume flows of different registers on one screen for	
balancing	
Check measured data wireless on mobile device	
Send measured values in protocol as .pdf or excel file	

## Technical Product Data

I common I loudet Dute	
Measuring range	0,4 to 30 m/s / 78.74 to 5905.51 fpm -20 to 60 °C / -4 to 140 °F
Accuracy	±(0.2 m/s + 2% of mv) (0,4 to 20 m/s) '±(40 fpm + 2% of mv) (80 to 4.000 fpm)
	±0.5 °C / ±0.9°F
Resolution	0.1 °C / 0.1 °F 0.1 m/s / 1 fpm
Dimensions	154mm x 43mm x 21mm 40mm vane diameter
Measuring cycle	1/sec
Battery life	140 h
Battery type	3x AAA
Operating temperature	-20 °C + 50 °C / -4 to 122 °F
Storage temperature	-20°C to 60°C / -4 to 140 °F

Page 4 last saved: 2021-12-29



### 2.3 Smart Probe testo 510i

## Main Features & Benefits

Can be used to determine differential pressure, air velocity and volume flow		
Easy set up for volume flow measurement via APP		
Easy determination of timed or multi-point average via APP		
Menu to perfom tightness / leak test on gas pipes		
Check measured data wireless on mobile device		
Do trendings as graph or table on mobile device		
Send measured values in protocol as .pdf or excel file		

## Technical Product Data

Measuring range	0 150 hPa / 60 in wc
	±0.05 hPa (0 1.00 hPa) /
	±0.02 in wc (0 0.4 in wc)
Accuracy	
	±0.2 hPa + 1.5 % o. mv. (1.01 150 hPa)
	±0.08 in wc + 1.5 % o. mv. (0.41 60 in wc)
Resolution	0.01 hPa / 0.01 inch wc
Dimensions	148mm x 36mm x 23mm
Measuring cycle	2/sec
Battery life	130 h
Battery type	3x AAA
Operating temperature	-20 °C + 50 °C / -4 to 122 °F
Storage temperature	-20°C to 60°C / -4 to 140 °F

Page 5 last saved: 2021-12-29



### 2.4 Smart Probe testo 549i

## Main Features & Benefits

## Technical Product Data

The state of the s	
Measuring range	0 60 bar / 0 870 psi
Accuracy	0.5% full scale
Resolution	0.01 bar / 0.1 psi
Diagonalismo	125mm x 32 mm x 31mm
Dimensions	1x 7/16" UNF female connection
Measuring cycle	2/sec
Battery life	130 h
Battery type	3x AAA
Operating temperature	-20 °C + 50 °C / -4 to 122 °F
Storage temperature	-20°C to 60°C / -4 to 140 °F
Refrigerants	CFC, HFC, HCFC, N, H20, CO2 as pressure media

Page 6 last saved: 2021-12-29



### 2.5 Smart Probe testo 605i

## Main Features & Benefits

Automatic calculation of wet bulb and dew point in App	
Determine BTU/H in combination with testo 405i	
Check measured data wireless on mobile device	
Do trendings as graph or table on mobile device	
Send measured values in protocol as .pdf or excel file	

### **Technical Product Data**

recrimical Product Data	
Measuring range	-20 to 60 °C, -4 to 140°F,
	0 to 100 %rH
	±1.8 %rH + 3% of rdg. at +25 °C (5 to 80 %rH)
	±0.03 %rH / K (0 to 60 °C)
Accuracy	
	±0.8 °C (-20 0 °C) / ±1.44 °F (-4 32°F)
	±0.5 °C (0 +60 °C) / ±0.9 °F (32 140°F)
Resolution	0.1 °F / 0.1 °C
	0.1 %rH
Dimensions	243mm x 30mm x 24mm
	100mm probe tip
Measuring cycle	1/sec
Battery life	280 h
Battery type	3x AAA
Operating temperature	-20 °C + 50 °C / -4 to 122 °F
Storage temperature	-20°C to 60°C / -4 to 140 °F

Page 7 last saved: 2021-12-29



### 2.6 Smart Probe testo 115i

### Main Features & Benefits

Work on systems with measuring points for temperature that have big distance in between

Calculate all necessary refrigeration parameters in App in combination with smart temperature probes

Check measured data wireless on mobile device

Do trendings as graph or table on mobile device

Send measured values in protocol as .pdf or excel file

### Technical Product Data

I delillion I I deduct Dute		
Measuring range	-50 to 150°C / -58 to 302 °F	
Accuracy	± 1.3 °C (-2085 °C)	
	± 2.34 °F (4185 °F)	
Resolution	0.1 °C / 0.1 °F	
Dimensions	183mm x 90 mm x 30mm	
	max. 35mm pipe diameter	
Measuring cycle	1/sec	
Battery life	240 h	
Battery type	3x AAA	
Operating temperature	-20 °C + 50 °C / -4 to 122 °F	
Storage temperature	-20°C to 60°C / -4 to 140 °F	

Page 8 last saved: 2021-12-29



### 2.7 Smart Probe testo 905i

## Main Features & Benefits

Check measured data wireless on mobile device		
Do trendings as graph or table on mobile device		
Send measured values in protocol as .pdf or excel file		

Tec	hnical	P	rodu	ct	Data

I walling and I describe the same	
Measuring range	-50 to 150°C / -58 to 302 °F
Accuracy	±1 °C / ±.1.8 °F
Resolution	0.1 °C / 0.1 °F
Dimensions	222mm x 30mm x 24mm
	100mm probe tip / Diameter: 4mm
Measuring cycle	1/sec
Battery life	190 h
Battery type	3x AAA
Operating temperature	-20 °C + 50 °C / -4 to 122 °F
Storage temperature	-20°C to 60°C / -4 to 140 °F

Page 9 last saved: 2021-12-29



### 2.8 Smart Probe testo 805i

### Main Features & Benefits

Diffractive optic as laser marker (laser circle)

Picture with temperature marking can be generated in App for documentation

Check measured data wireless on mobile device

Send measured values in protocol as .pdf or excel file

### Technical Product Data

reclinical Product Data	
Measuring range	-30°C to 250 °C / -22 to 482 °F
	±1.5 °C or ±1.5 % of m.v. (rest)
	±2.0 °C (-20.0 to -0.1 °C )
	±2.5 °C (-30.0 to -20.1 °C)
Accuracy	
	±2.7 °F or ±1.5 % of m.v. (rest)
	±3.6 °F (4 to 32 °F )
	±4.5 °F (-22 to4 °F)
Resolution	0.1 °C / 0.1 °F
Optics	10:1
Laser Sighting	Diffractive optics (laser circle)
Measuring cycle	2/sec
Dimensions	140mm x 36mm x 25mm
Battery life	25 h
Battery type	3x AAA
Operating temperature	-10 °C + 50 °C / 14 to 122 °F
Storage temperature	-20°C to 60°C / -4 to 140 °F

Page 10 last saved: 2021-12-29



### **Chapter 3 Safety**

### 3.1 Safety with testo 510i and testo 605i

## MARNING

## Magnetic field!

May be harmful to those with pacemakers.

 Keep a minimum distance of 10 cm between pacemaker and instrument.

### 3.2 Safety with testo 549i

## MARNING

### Risk of injury caused by pressurized, hot, cold or toxic refrigerants/media!

- Only to be used by qualified staff.
- Wear protective goggles and safety gloves.
- Before applying pressure to the measuring instrument: always fix the instrument tightly onto the pressure connection.
- Comply with the permissible measuring range (0 to 60 bar). Pay particular attention to this in systems with R744 refrigerant, since these are frequently operated with higher pressures!
- Use with A2L refrigerants

Testo measuring instruments (as of July 2020) can be used in compliance with the prescribed laws, standards, directives and safety regulations for refrigeration systems and refrigerants as well as regulations of the manufacturers of refrigerants of safety group A2L as per ISO 817.

Regional standardization and interpretation must always be observed.

For example, DIN EN 378-Part 1-4 applies to the scope of the EN standards.

During maintenance work, the employer must ensure that a hazardous explosive atmosphere is prevented (see also

TRBS1112, TRBS2152 VDMA 24020-3).

A hazardous and potentially explosive atmosphere must be anticipated during maintenance and repair work on refrigeration systems with flammable refrigerants (e.g. those of category A2L and A3).

Maintenance, repairs, removal of refrigerants and commissioning of systems may only be carried out by qualified personnel.



## 3.3 Safety with testo 805i

# **A** CAUTION



Laser radiation! Class 2 laser.

- Do not look into the laser beam!

Page 12 last saved: 2021-12-29



### **Chapter 4 Operation principle**

### 4.1 App operation principle

App which install in smart phone shall be received and display measurement value from Smart probes when it's opening via Bluetooth. Customer also can configure units or other parameters by App.

### 4.2 testo 405i operation principle

testo 405i has a NTC sensor, which shall change resistant value related ambient temperature. Circuit is response for detect NTC's resistant and calculate ambient temperature through NTC's feature by MCU ..

Then MCU control the Bluetooth module to send data to smart phone via Bluetooth and also received parameter changed and action protocol from smart phone.

Key is use for start, stop and hold value.

Led is use for indicate status of device.

Red: warning, such as low battery level

Yellow: finding App in smart phone via Bluetooth

Green: communication success.

### 4.3 testo 410i operation principle

testo 410i has a vane, which can recognize flow velocity by vane's speed of revolution. Speed's analog signal from vane can transform to digital signal for MCU deal.

Then MCU control the Bluetooth module to send data to smart phone via Bluetooth and also received parameter changed and action protocol from smart phone.

Key is use for start ,stop and hold value.

Led is use for indicate status of device. Red: warning, such as low battery level

Yellow: finding App in smart phone via Bluetooth

Green: communication success.



### 4.4 testo 510i operation principle

testo 510i has a differential pressure sensor, which can recognize differential pressure signal from gas pipes. Sensor can transform Pressure signal to electronic signal automatically, and transform analog signal from sensor to digital signal for MCU by A/D.

Then MCU control the Bluetooth module to send data to smart phone via Bluetooth and also received parameter changed and action protocol from smart phone.

Key is use for start, stop and hold value.

Led is use for indicate status of device. Red: warning, such as low battery level

Yellow: finding App in smart phone via Bluetooth

Green: communication success.

### 4.5 testo 549i operation principle

testo 549i has a high pressure sensor, which can recognize high pressure from gas pipe. Sensor can transform Pressure signal to electronic signal automatically, and transform analog signal from sensor to digital signal for MCU by A/D.

Then MCU control the Bluetooth module to send data to smart phone via Bluetooth and also received parameter changed and action protocol from smart phone.

Key is use for start, stop and hold value.

Led is use for indicate status of device. Red: warning, such as low battery level

Yellow: finding App in smart phone via Bluetooth

Green: communication success.

### 4.6 testo 605i operation principle

testo 605i has a temperature senor and humidity sensor, which can recognize ambient temperature and humidity in air. Sensors can transform temperature and humidity signals to electronic signal automatically, and transform analog signal from sensor to digital signal for MCU by A/D.

Then MCU control the Bluetooth module to send data to smart phone via Bluetooth and also received parameter changed and action protocol from smart phone.



Key is use for start, stop and hold value.

Led is use for indicate status of device. Red: warning, such as low battery level

Yellow: finding App in smart phone via Bluetooth

Green: communication success.

### 4.7 testo 115i operation principle

testo 605i has a NTC sensor in clamp, which shall change resistant value related surface temperature. Circuit is response for detect NTC's resistant and calculate surface temperature of pipe, which clamp catch, through NTC's feature by MCU.

Then MCU control the Bluetooth module to send data to smart phone via Bluetooth and also received parameter changed and action protocol from smart phone.

Key is use for start, stop and hold value.

Led is use for indicate status of device. Red: warning, such as low battery level

Yellow: finding App in smart phone via Bluetooth

Green: communication success.

### 4.8 testo 905i operation principle

testo 905i has a TC sensor, which shall change electronic capability value related ambient temperature. Circuit is response for detect TC's capability and calculate ambient temperature in air through TC's feature by MCU.

Then MCU control the Bluetooth module to send data to smart phone via Bluetooth and also received parameter changed and action protocol from smart phone.

Key is use for start, stop and hold value.

Led is use for indicate status of device. Red: warning, such as low battery level

Yellow: finding App in smart phone via Bluetooth

Green: communication success.



### 4.9 testo 805i operation principle

testo 805i has a IR sensor, which can recognize surface temperature by IR reflect. Analog signal from sensor shall be transformed to digital signal by ADC for MCU.

Laser shall be turn on automatically when measuring.

Then MCU control the Bluetooth module to send data to smart phone via Bluetooth and also received parameter changed and action protocol from smart phone.

Key is use for start, stop and hold value.

Led is use for indicate status of device. Red: warning, such as low battery level

Yellow: finding App in smart phone via Bluetooth

Green: communication success.

Testo SE & Co. KGaA

Celsiusstr. 2, 79822 Titisee-Neustadt, Germany

Tel.: +49 7653 681-0 E-Mail: info@testo.de

www.testo.com

# **Approval and Certification**

Product 产品名称	testo 405i testo 410i testo 510i
MatNo. 型号	0560 1405 0560 1410 0560 1510
Date 日期	04.15.2022

The use of the wireless module is subject to the regulations and stipulations of the respective country of use, and the module may only be used in countries for which a country certification has been granted. The user and every owner has the obligation to adhere to these regulations and prerequisites for use, and acknowledges that the re-sale, export, import etc. in particular in countries without wireless permits, is his responsibility.

countries without wheless permits, is his responsibility.			
Country	Comments		
Australia	◬	E 1561	
Canada	testo 405i: IC: 6127B-05601405 testo 410i: IC: 6127B-05601410 testo 510i: IC: 6127B-05601510 see IC Warnings		
Europa + EFTA (Länderliste einfügen)	The EU Declaration of Conformity can be found on the testo homepage www.testo.com under the product specific downloads.  EU countries: Belgium (BE), Bulgaria (BG), Denmark (DK), Germany (DE), Estonia (EE), Finland (FI), France (FR), Greece (GR), Ireland (IE), Italy (IT), Latvia (LV), Lithuania (LT), Luxembourg (LU), Malta (MT), Netherlands (NL), Austria (AT), Poland (PL), Portugal (PT), Romania (RO), Sweden (SE), Slovakia (SK), Slovenia (SI), Spain (ES), Czech Republic (CZ), Hungary (HU), United Kingdom (GB), Republic of Cyprus (CY).  EFTA countries: Iceland, Liechtenstein, Norway,		
United Kingdom (UK)	can be found		
USA	homepage www.testo.com under the product specific downloads.  testo 405i: WAF-05601405 testo 410i: WAF-05601410 testo 510i: WAF-05601510 See FCC Warnings		

Korea	see KCC Warning testo 405i: R-R-te2-05601405 testo 410i: R-R-te2-05601410 testo 510i: R-R-te2-05601510		
Japan	Bluetooth module:  R 201-200983  See Japan information		
Bluetooth® Information 蓝牙信息	Feature 特征与参数 Bluetooth® range 范围 radio type 型号 Bluetooth® company 制造商 RF Band 射频频段 power output [E.I.R.P] 输出功率	Values 数值 typical 80 m 80 米无障碍场地 Bluetooth® Low Energy (BLE) 4.2 Lierda Science & Technology Group Co., Ltd 2402 – 2480MHz	
Bluetooth® SIG Listing	Feature Declaration ID member company	Values D043363 Testo SE & Co. KGaA	

### IC Warnings

### CAN ICES-003(B)/NMB-003(B):

This Class B digital apparatus complies with Canadian ICES-003. Cet appareil numérique de la classe B est conforme à la norme NMB-003 du Canada.

#### RSS-Gen & RSS-247 statement:

This device contains licence-exempt transmitter(s)/receiver(s) that comply with Innovation, Science and Economic Development Canada's licence-exempt RSS(s).

Operation is subject to the following two conditions:

- (1) this device may not cause interference, and
- (2) this device must accept any interference, including interference that may cause undesired operation of the device.

L'émetteur/récepteur exempt de licence contenu dans le présent appareil est conforme aux CNR d'Innovation, Sciences et Développement économique Canada applicables aux appareils radio exempts de licence. L'exploitation est autorisée aux deux conditions suivantes:

- (1) L'appareil ne doit pas produire de brouillage
- (2) l'appareil doit accepter tout brouillage radioélectrique subi, même si le brouillage est susceptible d'en compromettre le fonctionnement.

Caution: Radio Frequency Radiation Exposure

The product comply with the Canada portable RF exposure limit set forth for an uncontrolled environment and are safe for intended operation as described in this manual. The further RF exposure reduction can be achieved if the product can be kept as far as possible from the user body or set the device to lower output power if such fuction is available.

#### Co-Location:

This device must not be co-located or operated in conjunction with any other antenna or transmitter.

Attention : exposition au rayonnement de radiofréquences Le produit est conforme aux limites d'exposition pour les appareils portables RF pour les Etats-Unis et le Canada établies pour un environnement non contrôlé.

Le produit est sûr pour un fonctionnement tel que décrit dans ce manuel. La réduction aux expositions RF peut être augmentée si l'appareil peut être conservé aussi loin que possible du corps de l'utilisateur ou que le dispositif est réglé sur la puissance de sortie la plus faible si une telle fonction est disponible.

### Co-location

Ce dispositif ne doit pas être utilisé à proximité d'une autre antenne ou d'un autre émetteur.

### FCC Warnings

Information from the FCC (Federal Communications Commission)

### For your own safety

Shielded cables should be used for a composite interface. This is to ensure continued protection against radio frequency interference.

### FCC warning statement

This equipment has been tested and found to comply with the limits for a Class B digital device, pursuant to Part 15 of the FCC Rules. These limits are designed to provide reasonable protection against harmful interference in a residential installation. This equipment generates, uses and can radiate radio frequency energy and, if not installed and used in accordance with the instructions, may cause harmful interference to radio communications. However, there is no guarantee that interference will not occur in a particular installation. If this equipment does cause harmful interference to radio or television reception, which can be determined by turning the equipment off and on, the user is encouraged

to try to correct the interference by one or more of the following measures:

- Reorient or relocate the receiving antenna.
- Increase the separation between the equipment and receiver.
- Connect the equipment into an outlet on a circuit different from that to which the receiver is connected.
- Consult the dealer or an experienced radio/TV technician for help.

#### Caution

Changes or modifications not expressly approved by the party responsible for compliance could void the user's authority to operate the equipment. Shielded interface cable must be used in order to comply with the emission limits.

#### Warning

This device complies with Part 15 of the FCC Rules. Operation is subject to the following two conditions:

- (1) this device may not cause harmful interference, and
- (2) this device must accept any interference received, including interference that may cause undesired operation.

### **Caution: Radio Frequency Radiation Exposure**

The device has been evaluated to meet general RF exposure requirement, The device can be used in portable exposure condition without restriction. Federal Communication Commission (FCC) Radiation Exposure Statement Power is so low that no RF exposure calculation is needed.

#### Co-Location:

This device must not be co-located or operated in conjunction with any other antenna or transmitter.

#### Japan Information

当該機器には電波法に基づく、技術基準適合証明等を受けた特定無線設備 を装着している。

### **KCC Warning**

해당 무선 설비는 운용 중 전파혼신 가능성이 있음